

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088948.D Time Analyzed: 13:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.070	8.977	9.177	47.560	50.000	-4.9
Endrin	6.572	6.477	6.677	51.120	50.000	2.2
gamma-BHC (Lindane)	4.328	4.232	4.432	56.060	50.000	12.1
Heptachlor	4.928	4.832	5.032	55.280	50.000	10.6
Heptachlor epoxide	5.689	5.593	5.793	54.050	50.000	8.1
Methoxychlor	7.491	7.395	7.595	48.520	50.000	-3.0
Tetrachloro-m-xylene	3.548	3.451	3.651	59.790	50.000	19.6

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088948.D Time Analyzed: 13:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.070	7.977	8.177	47.080	50.000	-5.8
Endrin	5.788	5.692	5.892	47.140	50.000	-5.7
gamma-BHC (Lindane)	3.728	3.631	3.831	50.040	50.000	0.1
Heptachlor	4.081	3.985	4.185	48.490	50.000	-3.0
Heptachlor epoxide	4.871	4.775	4.975	49.320	50.000	-1.4
Methoxychlor	6.752	6.658	6.858	46.380	50.000	-7.2
Tetrachloro-m-xylene	2.879	2.782	2.982	50.110	50.000	0.2