

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088958.D Time Analyzed: 19:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.070	8.977	9.177	49.060	50.000	-1.9
Endrin	6.573	6.477	6.677	52.090	50.000	4.2
gamma-BHC (Lindane)	4.328	4.232	4.432	56.160	50.000	12.3
Heptachlor	4.928	4.832	5.032	54.930	50.000	9.9
Heptachlor epoxide	5.690	5.593	5.793	54.570	50.000	9.1
Methoxychlor	7.491	7.395	7.595	49.000	50.000	-2.0
Tetrachloro-m-xylene	3.548	3.451	3.651	59.590	50.000	19.2

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088958.D Time Analyzed: 19:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.071	7.977	8.177	49.040	50.000	-1.9
Endrin	5.788	5.692	5.892	48.050	50.000	-3.9
gamma-BHC (Lindane)	3.729	3.631	3.831	50.820	50.000	1.6
Heptachlor	4.082	3.985	4.185	49.110	50.000	-1.8
Heptachlor epoxide	4.872	4.775	4.975	49.910	50.000	-0.2
Methoxychlor	6.753	6.658	6.858	47.410	50.000	-5.2
Tetrachloro-m-xylene	2.880	2.782	2.982	50.870	50.000	1.7