

Report of Analysis

Client:	ENTACT		Date Collected:	06/10/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/10/25	
Client Sample ID:	HAM-CONCRETEMS		SDG No.:	Q2301	
Lab Sample ID:	Q2285-05MS		Matrix:	TCLP	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088952.D	1	06/13/25 11:45	06/13/25 17:48	PB168474

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0060		0.000037	0.00050	mg/L
76-44-8	Heptachlor	0.0057		0.000027	0.00050	mg/L
1024-57-3	Heptachlor epoxide	0.0058		0.000096	0.00050	mg/L
72-20-8	Endrin	0.0057		0.000032	0.00050	mg/L
72-43-5	Methoxychlor	0.0054		0.00011	0.00050	mg/L
8001-35-2	Toxaphene	0.0017	U	0.0017	0.010	mg/L
57-74-9	Chlordane	0.00088	U	0.00088	0.0050	mg/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.0		30 (57) - 150 (171)	110%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.6		30 (61) - 150 (148)	108%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit