

Surrogate Summary

SDG No.: Q2301

Client: ENTACT

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088583.D	PIBLK-PD088583.D	Decachlorobiphenyl	1	20	18.4	92		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	16.5	83		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	18.0	90		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	16.9	84		30 (61)	150 (148)
I.BLK-PD088947.D	PIBLK-PD088947.D	Decachlorobiphenyl	1	20	20.1	100		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	21.5	107		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	19.9	99		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	20.0	100		30 (61)	150 (148)
PB168474BL	PB168474BL	Decachlorobiphenyl	1	20	20.9	105		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	22.4	112		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	20.9	105		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	20.6	103		30 (61)	150 (148)
PB168474BS	PB168474BS	Decachlorobiphenyl	1	20	24.3	121		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	25.5	128		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	24.5	122		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	24.2	121		30 (61)	150 (148)
Q2285-05MS	HAM-CONCRETEMS	Decachlorobiphenyl	1	20	22.0	110		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	21.6	108		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	22.0	110		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (61)	150 (148)
Q2285-05MSD	HAM-CONCRETEMSD	Decachlorobiphenyl	1	20	21.8	109		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	21.6	108		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	22.0	110		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	19.1	96		30 (61)	150 (148)
Q2301-03	WC-URBAN-FILL-C	Decachlorobiphenyl	1	20	22.6	113		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	22.8	114		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	23.0	115		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	21.1	106		30 (61)	150 (148)
PB168428TB	PB168428TB	Decachlorobiphenyl	1	20	20.9	105		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	21.7	108		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	21.3	107		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	20.4	102		30 (61)	150 (148)
I.BLK-PD088957.D	PIBLK-PD088957.D	Decachlorobiphenyl	1	20	20.6	103		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	21.4	107		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	20.9	104		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	20.1	100		30 (61)	150 (148)