

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2301</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ENTACT</u>	DataFile :	<u>PD088952.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2285-05MS (Column 1)	HAM-CONCRETEMS gamma-BHC (Lindane)	5	0	6.00	ug/L	120				30 (60)	150 (152)	
	Heptachlor	5	0	5.70	ug/L	114				30 (56)	150 (147)	
	Heptachlor epoxide	5	0	5.80	ug/L	116				30 (77)	150 (143)	
	Endrin	5	0	5.70	ug/L	114				30 (76)	150 (144)	
	Methoxychlor	5	0	5.40	ug/L	108				30 (70)	150 (142)	
Client Sample ID: Q2285-05MS (Column 2)	HAM-CONCRETEMS gamma-BHC (Lindane)	5	0	5.30	ug/L	106				30 (60)	150 (152)	
	Heptachlor	5	0	5.10	ug/L	102				30 (56)	150 (147)	
	Heptachlor epoxide	5	0	5.20	ug/L	104				30 (77)	150 (143)	
	Endrin	5	0	5.20	ug/L	104				30 (76)	150 (144)	
	Methoxychlor	5	0	5.10	ug/L	102				30 (70)	150 (142)	