

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2301

Client: ENTACT

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2287-02MSD		Client Sample ID: CONCRETE-SLABMSD		DataFile: BM050304.D							
Pyridine	500	0	330	ug/L	66	3			20 (10)	160 (109)	20 (20)
1,4-Dichlorobenzene	500	0	460	ug/L	92	9			70 (55)	130 (125)	20 (20)
2-Methylphenol	500	0	480	ug/L	96	6			70 (60)	130 (131)	20 (20)
3+4-Methylphenols	500	0	470	ug/L	94	4			20 (54)	160 (136)	20 (20)
Hexachloroethane	500	0	470	ug/L	94	7			20 (19)	160 (146)	20 (20)
Nitrobenzene	500	0	540	ug/L	108	8			70 (62)	130 (112)	20 (20)
Hexachlorobutadiene	500	0	510	ug/L	102	13			70 (52)	130 (125)	20 (20)
2,4,6-Trichlorophenol	500	0	550	ug/L	110	10			70 (78)	130 (112)	20 (20)
2,4,5-Trichlorophenol	500	0	540	ug/L	108	8			70 (71)	130 (111)	20 (20)
2,4-Dinitrotoluene	500	0	570	ug/L	114	4			70 (74)	130 (137)	20 (20)
Hexachlorobenzene	500	0	550	ug/L	110	10			70 (72)	130 (115)	20 (20)
Pentachlorophenol	1000	0	1100	ug/L	110	0			20 (52)	160 (162)	20 (20)