

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2301

Client: ENTACT

Analytical Method: 8270E DataFile: BM050316.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168473BS	Pyridine	50	42.5	ug/L	85				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	45.5	ug/L	91				70 (76)	130 (103)	
	2-Methylphenol	50	46.1	ug/L	92				70 (69)	130 (109)	
	3+4-Methylphenols	50	45.3	ug/L	91				20 (67)	160 (106)	
	Hexachloroethane	50	48.5	ug/L	97				20 (76)	160 (118)	
	Nitrobenzene	50	50.6	ug/L	101				70 (58)	130 (106)	
	Hexachlorobutadiene	50	47.3	ug/L	95				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	49.3	ug/L	99				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	48.6	ug/L	97				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	52.0	ug/L	104				70 (60)	130 (115)	
	Hexachlorobenzene	50	50.8	ug/L	102				70 (73)	130 (106)	
	Pentachlorophenol	100	110	ug/L	110				20 (47)	160 (114)	