

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/13/2025

Lab File ID: BM050300.D Time Analyzed: 13:56

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	431593	7.763	1716350	10.55	988714	14.39
UPPER LIMIT	863186	8.263	3432700	11.051	1977430	14.892
LOWER LIMIT	215797	7.263	858175	10.051	494357	13.892
EPA SAMPLE NO.						
01 CONCRETE-SLABMS	429386	7.76	1666090	10.55	945863	14.39
02 CONCRETE-SLABMSD	421105	7.76	1599280	10.55	894516	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.