

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/14/2025

Lab File ID: BM050314.D Time Analyzed: 01:01

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	461796	7.763	1832640	10.55	1033250	14.39
UPPER LIMIT	923592	8.263	3665280	11.051	2066500	14.892
LOWER LIMIT	230898	7.263	916320	10.051	516625	13.892
EPA SAMPLE NO.						
01 WC-URBAN-FILL-C	365030	7.76	1367330	10.55	730437	14.39
02 PB168428TB	427909	7.76	1575830	10.55	855467	14.39
03 PB168473BL	409459	7.76	1522100	10.55	824447	14.39
04 PB168473BS	445397	7.76	1717060	10.55	940759	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.