

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG No.: Q2301

Instrument ID: MSVOA_X Calibration Date/Time: 06/13/2025 12:30

Lab File ID: VX046682.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:42 12:57

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.630	0.534		-15.24	20
1,1-Dichloroethene	0.594	0.583		-1.85	20
2-Butanone	0.495	0.434		-12.32	20
Carbon Tetrachloride	0.588	0.572		-2.72	20
Chloroform	1.342	1.274		-5.07	20
Benzene	1.479	1.558		5.34	20
1,2-Dichloroethane	0.615	0.601		-2.28	20
Trichloroethene	0.376	0.405		7.71	20
Tetrachloroethene	0.334	0.365		9.28	20
Chlorobenzene	1.176	1.297	0.3	10.29	20
1,2-Dichloroethane-d4	0.890	0.809		-9.1	20
Dibromofluoromethane	0.368	0.364		-1.09	20
Toluene-d8	1.212	1.271		4.87	20
4-Bromofluorobenzene	0.502	0.541		7.77	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.