

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060625

Review By	Mahesh Dadoda	Review On	6/9/2025 1:15:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/9/2025 1:19:58 PM
SubDirectory	VX060625	HP Acquire Method	HP Processing Method 82X060625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134153
Initial Calibration Stds	VP134235,VP134236,VP134237,VP134238,VP134239,VP134240
CCC	VP134154
Internal Standard/PEM	
ICV/I.BLK	VP134241
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046516.D	06 Jun 2025 08:47		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046517.D	06 Jun 2025 09:13		JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046518.D	06 Jun 2025 09:42	for TCLP	JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX046519.D	06 Jun 2025 10:18	Comp #05 fail	JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046520.D	06 Jun 2025 10:40	LR-13,16,17	JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046521.D	06 Jun 2025 11:02		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046522.D	06 Jun 2025 11:25		JC/MD	Ok,M
8	IBLK	IBLK	VX046523.D	06 Jun 2025 11:47		JC/MD	Ok
9	VSTDIC001	VSTDIC001	VX046524.D	06 Jun 2025 12:57		JC/MD	Ok,M
10	VSTDICV050	ICV VX060625	VX046525.D	06 Jun 2025 14:11		JC/MD	Ok,M
11	VX0606MBL01	VX0606MBL01	VX046526.D	06 Jun 2025 14:39		JC/MD	Ok
12	VX0606WBL01	VX0606WBL01	VX046527.D	06 Jun 2025 15:02		JC/MD	Ok
13	VX0606WBS01	VX0606WBS01	VX046528.D	06 Jun 2025 15:25		JC/MD	Ok,M
14	VX0606MBS01	VX0606MBS01	VX046529.D	06 Jun 2025 15:51		JC/MD	Ok,M
15	Q2168-11MEDL	C2MEDL	VX046530.D	06 Jun 2025 16:13		JC/MD	Ok
16	VX0606WBSD01	VX0606WBSD01	VX046531.D	06 Jun 2025 16:36		JC/MD	Ok,M
17	Q2194-02	COMP-12	VX046532.D	06 Jun 2025 16:58	vial A pH#5.0	JC/MD	Ok
18	Q2194-04	COMP-13	VX046533.D	06 Jun 2025 17:21	vial A pH#5.0	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060625

Review By	Maresh Dadoda	Review On	6/9/2025 1:15:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/9/2025 1:19:58 PM
SubDirectory	VX060625	HP Acquire Method	HP Processing Method 82X060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134153 VP134235,VP134236,VP134237,VP134238,VP134239,VP134240		
CCC Internal Standard/PEM	VP134154		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134241		

19	Q2207-09	BU-703-COMP-01	VX046534.D	06 Jun 2025 17:43	vial A pH#5.0	JC/MD	Ok,M
20	Q2207-18	BU-703-COMP-02	VX046535.D	06 Jun 2025 18:06	vial A pH#5.0	JC/MD	Ok,M
21	Q2207-27	BU-703-COMP-03	VX046536.D	06 Jun 2025 18:28	vial A pH#5.0	JC/MD	Ok,M
22	Q2207-36	BU-703-COMP-04	VX046537.D	06 Jun 2025 18:51	vial A pH#5.0	JC/MD	Ok,M
23	Q2207-45	BU-703-COMP-05	VX046538.D	06 Jun 2025 19:13	vial A pH#5.0	JC/MD	Ok,M
24	Q2208-09	BU-703-COMP-06	VX046539.D	06 Jun 2025 19:35	vial A pH#5.0	JC/MD	Ok,M
25	Q2208-18	BU-703-COMP-07	VX046540.D	06 Jun 2025 19:58	vial A pH#5.0	JC/MD	Ok,M
26	Q2208-27	BU-703-COMP-08	VX046541.D	06 Jun 2025 20:20	vial A pH#5.0	JC/MD	Ok,M
27	Q2208-36	BU-703-COMP-09	VX046542.D	06 Jun 2025 20:42	vial A pH#5.0	JC/MD	Ok,M
28	Q2236-01	WC-A4-05A-G	VX046543.D	06 Jun 2025 21:04	vial A pH#5.0 Out of tune	JC/MD	Not Ok
29	VSTDCCC050	VSTDCCC050EC	VX046544.D	06 Jun 2025 21:26	Out of tune	JC/MD	Not Ok
30	BFB	BFB	VX046545.D	06 Jun 2025 23:59		JC/MD	Ok
31	VSTDCCC050	VSTDCCC050	VX046546.D	07 Jun 2025 00:34		JC/MD	Ok,M
32	VX0606WBL02	VX0606WBL02	VX046547.D	07 Jun 2025 01:17		JC/MD	Ok
33	VX0606WBS02	VX0606WBS02	VX046548.D	07 Jun 2025 02:00		JC/MD	Ok,M
34	VX0606WBSD02	VX0606WBSD02	VX046549.D	07 Jun 2025 02:22		JC/MD	Ok,M
35	PB168312TB	PB168312TB	VX046550.D	07 Jun 2025 02:43		JC/MD	Ok,M
36	PB168272TB	PB168272TB	VX046551.D	07 Jun 2025 03:05		JC/MD	Ok,M
37	Q2236-05	WC-A2-04-G	VX046552.D	07 Jun 2025 03:26	Surrogate Fail	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060625

Review By	Maresh Dadoda	Review On	6/9/2025 1:15:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/9/2025 1:19:58 PM
SubDirectory	VX060625	HP Acquire Method	HP Processing Method 82X060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134153 VP134235,VP134236,VP134237,VP134238,VP134239,VP134240		
CCC Internal Standard/PEM	VP134154		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134241		

38	Q2236-09	WC-A2-05-G	VX046553.D	07 Jun 2025 03:48	Internal Standard Fail	JC/MD	ReRun
39	Q2236-13	WC-A2-06-G	VX046554.D	07 Jun 2025 04:09	vial A pH#5.0	JC/MD	Ok
40	Q2236-17	WC-A2-07-G	VX046555.D	07 Jun 2025 04:31	Internal Standard Fail; Surrogate fail	JC/MD	ReRun
41	Q2227-04	TP07-MHH-WC	VX046556.D	07 Jun 2025 04:52		JC/MD	Ok,M
42	Q2228-04	TP08-MHI-WC	VX046557.D	07 Jun 2025 05:14	Internal Standard Fail	JC/MD	ReRun
43	Q2235-01	WC-A2-08-G	VX046558.D	07 Jun 2025 05:36	Internal Standard Fail	JC/MD	ReRun
44	Q2240-04	TP-3	VX046559.D	07 Jun 2025 05:57	Internal Standard Fail	JC/MD	ReRun
45	Q2240-08	TP-2	VX046560.D	07 Jun 2025 06:18	Internal Standard Fail	JC/MD	ReRun
46	Q2240-12	TP-1	VX046561.D	07 Jun 2025 06:40	Internal Standard Fail	JC/MD	ReRun
47	Q2241-04	TP-N	VX046562.D	07 Jun 2025 07:01	Internal Standard Fail	JC/MD	ReRun
48	Q2241-08	TP-S	VX046563.D	07 Jun 2025 07:23	Internal Standard Fail	JC/MD	ReRun
49	Q2226-04	TP06-MHI-WC	VX046564.D	07 Jun 2025 07:44	Internal Standard Fail	JC/MD	ReRun
50	VSTDCCC050	VSTDCCC050EC	VX046565.D	07 Jun 2025 08:06		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061325

Review By	John Carlone	Review On	6/16/2025 8:54:25 AM
Supervise By	Maresh Dadoda	Supervise On	6/16/2025 5:04:15 PM
SubDirectory	VX061325	HP Acquire Method	HP Processing Method 624X061025W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134332,VP134335
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134333,VP134334,VP134336,VP134337

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046674.D	13 Jun 2025 08:50		JC/MD	Ok
2	VSTDCCC020	VSTDCCC020	VX046675.D	13 Jun 2025 09:26	pH#Lot#V12668	JC/MD	Ok,M
3	VX0613WBS01	VX0613WBS01	VX046676.D	13 Jun 2025 09:58		JC/MD	Ok,M
4	VX0613WBSD01	VX0613WBSD01	VX046677.D	13 Jun 2025 10:25		JC/MD	Ok,M
5	VX0613WBL01	VX0613WBL01	VX046678.D	13 Jun 2025 10:47		JC/MD	Ok
6	Q2302-02	250611056-09-TRIP-BL	VX046679.D	13 Jun 2025 11:15	vial B pH#6.0 TB	JC/MD	Ok
7	Q2302-01	250611078-02-VOA	VX046680.D	13 Jun 2025 11:37	vial B pH#6.0	JC/MD	Ok
8	BFB	BFB	VX046681.D	13 Jun 2025 11:58		JC/MD	Ok
9	VSTDCCC050	VSTDCCC050	VX046682.D	13 Jun 2025 12:30		JC/MD	Ok,M
10	VX0613MBL01	VX0613MBL01	VX046683.D	13 Jun 2025 12:58		JC/MD	Ok
11	VX0613WBL02	VX0613WBL02	VX046684.D	13 Jun 2025 13:19		JC/MD	Ok
12	VX0613WBS02	VX0613WBS02	VX046685.D	13 Jun 2025 13:45		JC/MD	Ok,M
13	VX0613WBSD02	VX0613WBSD02	VX046686.D	13 Jun 2025 14:09		JC/MD	Ok,M
14	PB168429TB	PB168429TB	VX046687.D	13 Jun 2025 14:31		JC/MD	Ok
15	Q2287-02	CONCRETE-SLAB	VX046688.D	13 Jun 2025 14:53	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
16	Q2295-04	TP01-MH1-WC	VX046689.D	13 Jun 2025 15:14	vial A pH#5.0	JC/MD	Ok
17	Q2295-08	TP02-MH5-WC	VX046690.D	13 Jun 2025 15:36	vial A pH#5.0	JC/MD	Ok,M
18	Q2296-04	WC-1	VX046691.D	13 Jun 2025 15:57	vial A pH#5.0	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061325

Review By	John Carlone	Review On	6/16/2025 8:54:25 AM
Supervise By	Maresh Dadoda	Supervise On	6/16/2025 5:04:15 PM
SubDirectory	VX061325	HP Acquire Method	HP Processing Method 624X061025W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134332,VP134335		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134333,VP134334,VP134336,VP134337		

19	Q2296-08	WC-2	VX046692.D	13 Jun 2025 16:19	vial A pH#5.0	JC/MD	Ok
20	Q2296-12	WC-3	VX046693.D	13 Jun 2025 16:41	vial A pH#5.0	JC/MD	Ok
21	Q2296-16	WC-4	VX046694.D	13 Jun 2025 17:02	vial A pH#5.0	JC/MD	Ok
22	Q2296-20	WC-5	VX046695.D	13 Jun 2025 17:24	vial A pH#5.0	JC/MD	Ok
23	Q2296-24	WC-6	VX046696.D	13 Jun 2025 17:46	vial A pH#5.0	JC/MD	Ok
24	Q2297-04	TP-3	VX046697.D	13 Jun 2025 18:07	vial A pH#5.0	JC/MD	Ok
25	Q2301-01	WC-URBAN-FILL-G	VX046698.D	13 Jun 2025 18:29	vial A pH#5.0	JC/MD	Ok
26	Q2310-04	TP-7	VX046699.D	13 Jun 2025 18:51	vial A pH#5.0	JC/MD	Ok
27	Q2288-01	ETGI-329	VX046700.D	13 Jun 2025 19:12	vial A pH#5.0	JC/MD	Ok
28	Q2287-02RE	CONCRETE-SLABRE	VX046701.D	13 Jun 2025 19:34	vial B pH#5.0 Surrogate fail	JC/MD	Confirms
29	VSTDCCC050	VSTDCCC050EC	VX046702.D	13 Jun 2025 19:55		JC/MD	Ok,M

M : Manual Integration