

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2301

Client: ENTACT

Analytical Method: SW8260D

Datafile : VX046686.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0613WBSD02	Vinyl chloride	20	14.5	ug/L	73	9		70 (65)	130 (117)	20 (19)
	1,1-Dichloroethene	20	16.3	ug/L	81	16		70 (74)	130 (110)	20 (20)
	2-Butanone	100	110	ug/L	110	16		40 (65)	160 (122)	20 (26)
	Carbon Tetrachloride	20	20.1	ug/L	101	5		70 (77)	130 (113)	20 (15)
	Chloroform	20	21.6	ug/L	108	5		70 (79)	130 (113)	20 (20)
	Benzene	20	20.7	ug/L	104	1		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	21.3	ug/L	106	6		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.9	ug/L	100	1		70 (77)	130 (113)	20 (15)
	Tetrachloroethene	20	20.1	ug/L	101	4		70 (67)	130 (123)	20 (15)
	Chlorobenzene	20	21.3	ug/L	106	1		70 (82)	130 (109)	20 (15)