

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301
 Lab File ID: VX046682.D Date Analyzed: 06/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 12:30
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	65729	5.56	116296	6.76	106162	10.06
UPPER LIMIT	131458	6.056	232592	7.263	212324	10.555
LOWER LIMIT	32864.5	5.056	58148	6.263	53081	9.555
EPA SAMPLE NO.						
WC-URBAN-FILL-G	81744	5.57	146734	6.78	135324	10.06
VX0613WBL02	93820	5.57	167354	6.77	149855	10.06
VX0613WBS02	70456	5.56	119817	6.77	108883	10.06
VX0613WBSD02	65989	5.56	114290	6.77	103875	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.