

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2301 SAS No.: Q2301 SDG NO.: Q2301
 Lab File ID: VX046682.D Date Analyzed: 06/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 12:30
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	55407	12.018				
UPPER LIMIT	110814	12.518				
LOWER LIMIT	27703.5	11.518				
EPA SAMPLE NO.						
WC-URBAN-FILL-G	66328	12.02				
VX0613WBL02	74343	12.02				
VX0613WBS02	57914	12.02				
VX0613WBSD02	55253	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.