

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GARD04

Lab Code: CHEM Case No.: Q2302 SAS No.: Q2302 SDG No.: Q2302

Instrument ID: MSVOA_X Calibration Date/Time: 06/13/2025 09:26

Lab File ID: VX046675.D Init. Calib. Date(s): 06/10/2025 06/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 20:19 21:44

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Acrolein	0.221	0.231		4.53	
Acrylonitrile	0.928	0.872		-6.03	
1,2-Dichloroethane-d4	2.428	2.437	0.01	0.37	
Toluene-d8	1.340	1.367	0.01	2.02	
4-Bromofluorobenzene	0.506	0.513	0.2	1.38	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.