

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GARD04
 Lab Code: CHEM Case No.: Q2302 SAS No.: Q2302 SDG NO.: Q2302
 Lab File ID: VN086967.D Date Analyzed: 06/12/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	185417	8.23	334881	9.11	291913	11.87
UPPER LIMIT	370834	8.729	669762	9.606	583826	12.365
LOWER LIMIT	92708.5	7.729	167441	8.606	145957	11.365
EPA SAMPLE NO.						
250611078-02-VOA	214885	8.24	415965	9.11	381430	11.87
250611056-09-TRIP-BLANK	289495	8.24	557887	9.11	494340	11.87
VN0612WBL01	293212	8.23	559485	9.11	496764	11.87
VN0612WBS01	173273	8.24	314156	9.11	274906	11.87
VN0612WBSD01	164097	8.23	292991	9.11	261207	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.