

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GARD04
 Lab Code: CHEM Case No.: Q2302 SAS No.: Q2302 SDG NO.: Q2302
 Lab File ID: VN086996.D Date Analyzed: 06/13/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:52
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	190412	8.23	349185	9.11	311622	11.87
UPPER LIMIT	380824	8.73	698370	9.606	623244	12.365
LOWER LIMIT	95206	7.73	174593	8.606	155811	11.365
EPA SAMPLE NO.						
250611078-02-VOARE	194559	8.23	391285	9.11	359346	11.87
VN0613WBL01	317214	8.24	606954	9.11	536496	11.87
VN0613WBS01	193820	8.23	342767	9.11	297937	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.