

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GARD04
 Lab Code: CHEM Case No.: Q2302 SAS No.: Q2302 SDG NO.: Q2302
 Lab File ID: VN086967.D Date Analyzed: 06/12/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	144672	13.794				
UPPER LIMIT	289344	14.294				
LOWER LIMIT	72336	13.294				
EPA SAMPLE NO.						
250611078-02-VOA	185553	13.79				
250611056-09-TRIP-BLANK	234791	13.79				
VN0612WBL01	239856	13.79				
VN0612WBS01	136822	13.79				
VN0612WBSD01	129235	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.