



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q2303

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB136147

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.503	0.5	101	90-110	06/13/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.488	0.5	98	90-110	06/13/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	06/13/2025