

LB136159


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

56030



A= 0.0000e+000

B= 1.8232e-004

C= -4.8492e-003

Rho= 0.9983008

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.001	0.001	34	0.000	34					10
0.2	0.200	0.220	0.020	1233	0.0 %	1233					2
2.5	2.500	2.543	0.043	13976	0.0 %	13976					3
5.0	5.000	5.173	0.173	28398	0.0 %	28398					6
7.5	7.500	7.052	-0.448	38704	0.0 %	38704					2
10.0	10.000	10.211	0.211	56030	0.0 %	56030					

LB136159 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Jun 2025 12:16:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7471B	PPB	16 Jun 2025 12:25:04	S	Std
0.2 - 1	50.2	1233	-	0.2000	7471B	PPB	16 Jun 2025 12:27:22	S	Std
2.5 - 1	52.5	13976	-	2.5000	7471B	PPB	16 Jun 2025 12:29:39	S	Std
5.0 - 1	55	28398	-	5.0000	7471B	PPB	16 Jun 2025 12:31:55	S	Std
7.5 - 1	57.5	38704	-	7.5000	7471B	PPB	16 Jun 2025 12:34:14	S	Std
10.0 - 1	510	56030	-	10.0000	7471B	PPB	16 Jun 2025 12:44:42	S	Std
ICV39 - 1	ICV39	21726	3.9563	-7471B	PPB	16 Jun 2025 12:52:01	U	SMPL	
ICB39 - 1	ICB39	-183	-0.0382	-7471B	PPB	16 Jun 2025 12:54:16	U	SMPL	
CCV45 - 1	CCV45	28523	5.1956	-7471B	PPB	16 Jun 2025 12:56:33	U	SMPL	
CCB45 - 1	CCB45	-303	-0.0601	-7471B	PPB	16 Jun 2025 12:58:49	U	SMPL	
CRA - 1	CRA	1182	0.2107	-7471B	PPB	16 Jun 2025 13:01:08	U	SMPL	
HighStd - 1	HighStd	55950	10.1962	-7471B	PPB	16 Jun 2025 13:03:23	U	SMPL	
ChkStd - 1	ChkStd	36499	6.6498	-7471B	PPB	16 Jun 2025 13:05:39	U	SMPL	
PB168482BL - 1	PBS	-243	-0.0492	-7471B	PPB	16 Jun 2025 13:08:05	U	SMPL	
PB168482BS - 1	LCSS	22235	4.0491	-7471B	PPB	16 Jun 2025 13:13:14	U	SMPL	
Q2303-01 - 1	B-165-SB01	16301	2.9672	-7471B	PPB	16 Jun 2025 13:15:30	U	SMPL	
Q2303-02 - 1	B-170-SB03	72054	13.1323	-7471B	PPB	16 Jun 2025 13:17:52	U	SMPL	
Q2303-02DUP - 1	B-170-SB03DUP	81827	14.9142	-7471B	PPB	16 Jun 2025 13:20:12	U	SMPL	
Q2303-02MS - 1	B-170-SB03MS	96969	17.6749	-7471B	PPB	16 Jun 2025 13:22:47	U	SMPL	
Q2303-02MSD - 1	B-170-SB03MSD	91271	16.6361	-7471B	PPB	16 Jun 2025 13:28:41	U	SMPL	
CCV46 - 1	CCV46	27236	4.9609	-7471B	PPB	16 Jun 2025 13:30:59	U	SMPL	
CCB46 - 1	CCB46	-392	-0.0763	-7471B	PPB	16 Jun 2025 13:33:32	U	SMPL	
Q2311-01 - 1	TP03-MH2MH3-WC	20620	3.7547	-7471B	PPB	16 Jun 2025 13:35:51	U	SMPL	
Q2311-05 - 1	TP04-MH2MH3-WC	3113	0.5627	-7471B	PPB	16 Jun 2025 13:38:07	U	SMPL	
Q2319-01 - 1	MH-B	433	0.0741	-7471B	PPB	16 Jun 2025 13:40:25	U	SMPL	
Q2328-01 - 1	CHRT25653	6681	1.2133	-7471B	PPB	16 Jun 2025 13:42:42	U	SMPL	
Q2303-02LX5 - 1		11956	2.1750	-7471B	PPB	16 Jun 2025 13:44:58	U	SMPL	
Q2303-02A - 1		94529	17.2301	-7471B	PPB	16 Jun 2025 13:47:17	U	SMPL	
CCV47 - 1	CCV47	27451	5.0001	-7471B	PPB	16 Jun 2025 13:49:36	U	SMPL	
CCB47 - 1	CCB47	-402	-0.0781	-7471B	PPB	16 Jun 2025 13:52:14	U	SMPL	
Q2303-02DLX2 - 1		37443	6.8219	-7471B	PPB	16 Jun 2025 13:54:35	U	SMPL	
Q2303-02DUPDLX2 - 1		41034	7.4766	-7471B	PPB	16 Jun 2025 13:56:51	U	SMPL	
Q2303-02MSDLX2 - 1		44672	8.1399	-7471B	PPB	16 Jun 2025 13:59:20	U	SMPL	
Q2303-02MSDDLX2 - 1		45852	8.3551	-7471B	PPB	16 Jun 2025 14:04:57	U	SMPL	
Q2303-02DLX2 - 1	52303-02DLX10	6606	1.1996	-7471B	PPB	16 Jun 2025 14:07:14	U	SMPL	
Q2303-02DLX2 - 1	52303-02DLX7	45330	8.2599	-7471B	PPB	16 Jun 2025 14:09:39	U	SMPL	
CCV48 - 1	CCV48	27664	5.0390	-7471B	PPB	16 Jun 2025 14:11:57	U	SMPL	
CCB48 - 1	CCB48	-347	-0.0681	-7471B	PPB	16 Jun 2025 14:14:20	U	SMPL	

NR
6/16/25