

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136159

Review By	MOHAN	Review On	6/16/2025 4:38:10 PM
Supervise By	jaswal	Supervise On	6/16/2025 5:53:17 PM

STD. NAME	STD REF.#
ICAL Standard	MP85971,MP85972,MP85973,MP85974,MP85975,MP85976
ICV Standard	MP85977
CCV Standard	MP85979
ICSA Standard	
CRI Standard	MP85981
LCS Standard	
Chk Standard	MP85978,MP85980,MP85982,MP85986

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/16/25 12:25		MOHAN	OK
2	S0.2	S0.2	CAL2	06/16/25 12:27		MOHAN	OK
3	S2.5	S2.5	CAL3	06/16/25 12:29		MOHAN	OK
4	S5	S5	CAL4	06/16/25 12:31		MOHAN	OK
5	S7.5	S7.5	CAL5	06/16/25 12:34		MOHAN	OK
6	S10	S10	CAL6	06/16/25 12:44		MOHAN	OK
7	ICV39	ICV39	ICV	06/16/25 12:52		MOHAN	OK
8	ICB39	ICB39	ICB	06/16/25 12:54		MOHAN	OK
9	CCV45	CCV45	CCV	06/16/25 12:56		MOHAN	OK
10	CCB45	CCB45	CCB	06/16/25 12:58		MOHAN	OK
11	CRA	CRA	CRDL	06/16/25 13:01		MOHAN	OK
12	HighStd	HighStd	HIGH STD	06/16/25 13:03		MOHAN	OK
13	ChkStd	ChkStd	SAM	06/16/25 13:05		MOHAN	OK
14	PB168482BL	PB168482BL	MB	06/16/25 13:08		MOHAN	OK
15	PB168482BS	PB168482BS	LCS	06/16/25 13:13		MOHAN	OK
16	Q2303-01	B-165-SB01	SAM	06/16/25 13:15		MOHAN	OK
17	Q2303-02	B-170-SB03	SAM	06/16/25 13:17	Hg High	MOHAN	Dilution
18	Q2303-02DUP	B-170-SB03DUP	DUP	06/16/25 13:20	Hg High	MOHAN	Dilution

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136159

Review By	MOHAN	Review On	6/16/2025 4:38:10 PM
Supervise By	jaswal	Supervise On	6/16/2025 5:53:17 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85971,MP85972,MP85973,MP85974,MP85975,MP85976		
ICV Standard	MP85977		
CCV Standard	MP85979		
ICSA Standard			
CRI Standard	MP85981		
LCS Standard			
Chk Standard	MP85978,MP85980,MP85982,MP85986		

19	Q2303-02MS	B-170-SB03MS	MS	06/16/25 13:22	Hg High	MOHAN	Dilution
20	Q2303-02MSD	B-170-SB03MSD	MSD	06/16/25 13:28	Hg High	MOHAN	Dilution
21	CCV46	CCV46	CCV	06/16/25 13:30		MOHAN	OK
22	CCB46	CCB46	CCB	06/16/25 13:33		MOHAN	OK
23	Q2311-01	TP03-MH2MH3-WC	SAM	06/16/25 13:35		MOHAN	OK
24	Q2311-05	TP04-MH4-WC	SAM	06/16/25 13:38		MOHAN	OK
25	Q2319-01	MH-B	SAM	06/16/25 13:40		MOHAN	OK
26	Q2328-01	CHRT25653	SAM	06/16/25 13:42		MOHAN	OK
27	Q2303-02L	B-170-SB03L	SD	06/16/25 13:44		MOHAN	OK
28	Q2303-02A	B-170-SB03A	PS	06/16/25 13:47	Hg High	MOHAN	Dilution
29	CCV47	CCV47	CCV	06/16/25 13:49		MOHAN	OK
30	CCB47	CCB47	CCB	06/16/25 13:52		MOHAN	OK
31	Q2303-02DL	B-170-SB03DL	SAM	06/16/25 13:54	2X For Hg	MOHAN	Confirms
32	Q2303-02DUPDL	B-170-SB03DUPDL	DUP	06/16/25 13:56	2X For Hg	MOHAN	Confirms
33	Q2303-02MSDL	B-170-SB03MSDL	MS	06/16/25 13:59	2X For Hg	MOHAN	Confirms
34	Q2303-02MSDDL	B-170-SB03MSDDL	MSD	06/16/25 14:04	2X For Hg	MOHAN	Confirms
35	Q2303-02LDL	B-170-SB03LDL	SD	06/16/25 14:07	Not Required	MOHAN	Not Ok
36	Q2303-02ADL	B-170-SB03ADL	PS	06/16/25 14:09	2X For Hg	MOHAN	Confirms
37	CCV48	CCV48	CCV	06/16/25 14:11		MOHAN	OK
38	CCB48	CCB48	CCB	06/16/25 14:14		MOHAN	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136164

Review By	Janvi	Review On	6/18/2025 2:48:50 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:49:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/16/25 12:52		Jaswal	OK
2	S1	S1	CAL2	06/16/25 12:56		Jaswal	OK
3	S2	S2	CAL3	06/16/25 13:01		Jaswal	OK
4	S3	S3	CAL4	06/16/25 13:05		Jaswal	OK
5	S4	S4	CAL5	06/16/25 13:09		Jaswal	OK
6	S5	S5	CAL6	06/16/25 13:13		Jaswal	OK
7	ICV01	ICV01	ICV	06/16/25 14:03		Jaswal	OK
8	LLICV01	LLICV01	LLICV	06/16/25 14:42		Jaswal	OK
9	ICB01	ICB01	ICB	06/16/25 14:49		Jaswal	OK
10	CRI01	CRI01	CRDL	06/16/25 14:53		Jaswal	OK
11	ICSA01	ICSA01	ICSA	06/16/25 14:59		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	06/16/25 15:06		Jaswal	OK
13	ICSADL	ICSADL	ICSA	06/16/25 15:12		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	06/16/25 15:20		Jaswal	OK
15	CCV01	CCV01	CCV	06/16/25 15:26		Jaswal	OK
16	CCB01	CCB01	CCB	06/16/25 15:30		Jaswal	OK
17	Q2287-02	CONCRETE-SLAB	SAM	06/16/25 15:35		Jaswal	OK
18	Q2295-04	TP01-MH1-WC	SAM	06/16/25 15:40		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136164

Review By	Janvi	Review On	6/18/2025 2:48:50 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:49:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

19	Q2295-08	TP02-MH5-WC	SAM	06/16/25 15:44		Jaswal	OK
20	Q2303-01	B-165-SB01	SAM	06/16/25 15:49		Jaswal	OK
21	Q2303-02	B-170-SB03	SAM	06/16/25 15:53		Jaswal	OK
22	Q2308-01	EO-02-06122025	SAM	06/16/25 15:58		Jaswal	OK
23	Q2328-01	CHRT25653	SAM	06/16/25 16:02		Jaswal	OK
24	PB168428TB	PB168428TB	MB	06/16/25 16:06		Jaswal	OK
25	PB168477BL	PB168477BL	MB	06/16/25 16:15		Jaswal	OK
26	PB168477BS	PB168477BS	LCS	06/16/25 16:19		Jaswal	OK
27	CCV02	CCV02	CCV	06/16/25 16:29		Jaswal	OK
28	CCB02	CCB02	CCB	06/16/25 16:34		Jaswal	OK
29	PB168472BL	PB168472BL	MB	06/16/25 16:44		Jaswal	OK
30	PB168472BS	PB168472BS	LCS	06/16/25 16:48		Jaswal	OK
31	Q2296-04	WC-1	SAM	06/16/25 16:54		Jaswal	OK
32	Q2296-08	WC-2	SAM	06/16/25 16:59		Jaswal	OK
33	Q2296-12	WC-3	SAM	06/16/25 17:03		Jaswal	OK
34	Q2296-16	WC-4	SAM	06/16/25 17:08		Jaswal	OK
35	Q2296-20	WC-5	SAM	06/16/25 17:12		Jaswal	OK
36	Q2296-24	WC-6	SAM	06/16/25 17:17		Jaswal	OK
37	Q2297-04	TP-3	SAM	06/16/25 17:21		Jaswal	OK
38	Q2301-03	WC-URBAN-FILL-C	SAM	06/16/25 17:26		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136164

Review By	Janvi	Review On	6/18/2025 2:48:50 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:49:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

39	CCV03	CCV03	CCV	06/16/25 17:31		Jaswal	OK
40	CCB03	CCB03	CCB	06/16/25 17:49		Jaswal	OK
41	Q2310-04	TP-7	SAM	06/16/25 17:54		Jaswal	OK
42	Q2310-04DUP	TP-7DUP	DUP	06/16/25 17:58		Jaswal	OK
43	Q2310-04L	TP-7L	SD	06/16/25 18:03		Jaswal	OK
44	Q2310-04MS	TP-7MS	MS	06/16/25 18:07		Jaswal	OK
45	Q2310-04MSD	TP-7MSD	MSD	06/16/25 18:11		Jaswal	OK
46	Q2310-04A	TP-7A	PS	06/16/25 18:16		Jaswal	OK
47	Q2298-01	AU-05-061125	SAM	06/16/25 18:20		Jaswal	OK
48	Q2305-01	TR-04-06122025	SAM	06/16/25 18:25		Jaswal	OK
49	CCV04	CCV04	CCV	06/16/25 18:29		Jaswal	OK
50	CCB04	CCB04	CCB	06/16/25 18:33		Jaswal	OK
51	Q2307-01	LINDEN-SAA	SAM	06/16/25 18:38	CCB5 fail for Cd	Jaswal	Not Ok
52	Q2310-01	TP-7	SAM	06/16/25 18:42	CCB5 fail for Cd	Jaswal	Not Ok
53	Q2310-01DUP	TP-7DUP	DUP	06/16/25 18:46	CCB5 fail for Cd	Jaswal	Not Ok
54	Q2310-01L	TP-7L	SD	06/16/25 18:51	CCB5 fail for Cd	Jaswal	Not Ok
55	Q2310-01MS	TP-7MS	MS	06/16/25 18:55	CCB5 fail for Cd	Jaswal	Not Ok
56	Q2310-01MSD	TP-7MSD	MSD	06/16/25 18:59	CCB5 fail for Cd	Jaswal	Not Ok
57	Q2310-01A	TP-7A	PS	06/16/25 19:04	CCB5 fail for Cd	Jaswal	Not Ok
58	Q2311-01	TP03-MH2MH3-WC	SAM	06/16/25 19:08	CCB5 fail for Cd	Jaswal	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136164

Review By	Janvi	Review On	6/18/2025 2:48:50 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:49:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

59	Q2311-05	TP04-MH4-WC	SAM	06/16/25 19:12	CCB5 fail for Cd	Jaswal	Not Ok
60	CCV05	CCV05	CCV	06/16/25 19:17		Jaswal	OK
61	CCB05	CCB05	CCB	06/16/25 19:21	Fail for Cd	Jaswal	OK
62	Q2296-01	WC-1	SAM	06/16/25 19:25	CCB5 fail for Cd	Jaswal	Not Ok
63	Q2296-01DUP	WC-1DUP	DUP	06/16/25 19:30	CCB5 fail for Cd	Jaswal	Not Ok
64	Q2296-01L	WC-1L	SD	06/16/25 19:34	CCB5 fail for Cd	Jaswal	Not Ok
65	Q2296-01MS	WC-1MS	MS	06/16/25 19:38	CCB5 fail for Cd	Jaswal	Not Ok
66	Q2296-01MSD	WC-1MSD	MSD	06/16/25 19:42	CCB5 fail for Cd	Jaswal	Not Ok
67	Q2296-01A	WC-1A	PS	06/16/25 19:47	CCB5 fail for Cd	Jaswal	Not Ok
68	Q2296-09	WC-3	SAM	06/16/25 19:51	CCB5 fail for Cd	Jaswal	Not Ok
69	Q2296-13	WC-4	SAM	06/16/25 19:55	CCB5 fail for Cd	Jaswal	Not Ok
70	Q2296-17	WC-5	SAM	06/16/25 20:00	CCB5 fail for Cd	Jaswal	Not Ok
71	CCV06	CCV06	CCV	06/16/25 20:04		Jaswal	OK
72	CCB06	CCB06	CCB	06/16/25 20:08		Jaswal	OK
73	Q2296-21	WC-6	SAM	06/16/25 20:13		Jaswal	Not Ok
74	Q2297-01	TP-3	SAM	06/16/25 20:17		Jaswal	OK
75	Q2307-03DL	LINDEN-SAA-WATER	SAM	06/16/25 20:22	Report straight 5X	Jaswal	Not Ok
76	Q2316-01	RW8-SP100-2025061	SAM	06/16/25 20:26		Jaswal	OK
77	Q2316-02	RW8-SP303-2025061	SAM	06/16/25 20:30		Jaswal	OK
78	Q2316-02DUP	RW8-SP303-2025061	DUP	06/16/25 20:35		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136164

Review By	Janvi	Review On	6/18/2025 2:48:50 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:49:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

79	Q2316-02L	RW8-SP303-2025061	SD	06/16/25 20:39		Jaswal	OK
80	Q2316-02MS	RW8-SP303-2025061	MS	06/16/25 20:43		Jaswal	OK
81	Q2316-02MSD	RW8-SP303-2025061	MSD	06/16/25 20:48		Jaswal	OK
82	CCV07	CCV07	CCV	06/16/25 20:52		Jaswal	OK
83	CCB07	CCB07	CCB	06/16/25 20:56		Jaswal	OK
84	Q2316-02A	RW8-SP303-2025061	PS	06/16/25 21:01		Jaswal	OK
85	CCV08	CCV08	CCV	06/16/25 21:14		Jaswal	OK
86	CCB08	CCB08	CCB	06/16/25 21:18		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136187

Review By	jaswal	Review On	6/19/2025 12:06:04 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:50:23 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872,MP85897
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/17/25 11:58		Jaswal	OK
2	S1	S1	CAL2	06/17/25 12:03		Jaswal	OK
3	S2	S2	CAL3	06/17/25 12:07		Jaswal	OK
4	S3	S3	CAL4	06/17/25 12:11		Jaswal	OK
5	S4	S4	CAL5	06/17/25 12:16		Jaswal	OK
6	S5	S5	CAL6	06/17/25 12:20		Jaswal	OK
7	ICV01	ICV01	ICV	06/17/25 12:25		Jaswal	OK
8	LLICV01	LLICV01	LLICV	06/17/25 12:41		Jaswal	OK
9	ICB01	ICB01	ICB	06/17/25 12:46		Jaswal	OK
10	CRI01	CRI01	CRDL	06/17/25 12:50		Jaswal	OK
11	ICSA01	ICSA01	ICSA	06/17/25 13:11		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	06/17/25 13:22		Jaswal	OK
13	ICSADL	ICSADL	ICSA	06/17/25 13:28		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	06/17/25 13:35		Jaswal	OK
15	CCV01	CCV01	CCV	06/17/25 13:40		Jaswal	OK
16	CCB01	CCB01	CCB	06/17/25 13:48		Jaswal	OK
17	Q2264-04	EF-WW	SAM	06/17/25 13:53		Jaswal	OK
18	Q2264-04L	EF-WWL	SD	06/17/25 14:02		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136187

Review By	jaswal	Review On	6/19/2025 12:06:04 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:50:23 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872,MP85897
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

19	Q2264-04MS	EF-WWMS	MS	06/17/25 14:06		Jaswal	OK
20	Q2264-04MSD	EF-WWMSD	MSD	06/17/25 14:11		Jaswal	OK
21	Q2264-04A	EF-WWA	PS	06/17/25 14:15	0.1ML M6004,M6013-10ML SAMPLE	Jaswal	OK
22	Q2264-04DUP	EF-WWDUP	DUP	06/17/25 14:21		Jaswal	OK
23	Q2290-01	Outfall 1	SAM	06/17/25 14:25		Jaswal	OK
24	Q2290-02	Outfall 2	SAM	06/17/25 14:30		Jaswal	OK
25	Q2291-01	Outfall 001	SAM	06/17/25 14:34		Jaswal	OK
26	Q2292-01	SW-1	SAM	06/17/25 14:39		Jaswal	OK
27	CCV02	CCV02	CCV	06/17/25 14:43		Jaswal	OK
28	CCB02	CCB02	CCB	06/17/25 14:50		Jaswal	OK
29	Q2292-02	SW-2	SAM	06/17/25 14:54		Jaswal	OK
30	Q2293-01	SW-3	SAM	06/17/25 14:59		Jaswal	OK
31	Q2312-01	TP-1	SAM	06/17/25 15:03		Jaswal	OK
32	Q2319-01	MH-B	SAM	06/17/25 15:07		Jaswal	OK
33	Q2322-01	CL-01-061325	SAM	06/17/25 15:12		Jaswal	OK
34	Q2323-01	PL-01-06132025	SAM	06/17/25 15:16		Jaswal	OK
35	Q2324-01	HD-01-6132025	SAM	06/17/25 15:20		Jaswal	OK
36	Q2325-01	TP-8	SAM	06/17/25 15:25		Jaswal	OK
37	Q2325-01DUP	TP-8DUP	DUP	06/17/25 15:29		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136187

Review By	jaswal	Review On	6/19/2025 12:06:04 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:50:23 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872,MP85897
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

38	Q2325-01L	TP-8L	SD	06/17/25 15:34		Jaswal	OK
39	CCV03	CCV03	CCV	06/17/25 15:38		Jaswal	OK
40	CCB03	CCB03	CCB	06/17/25 15:42		Jaswal	OK
41	Q2325-01MS	TP-8MS	MS	06/17/25 15:46		Jaswal	OK
42	Q2325-01MSD	TP-8MSD	MSD	06/17/25 15:51		Jaswal	OK
43	Q2325-01A	TP-8A	PS	06/17/25 15:55	0.1ML M6004,M6013-10ML SAMPLE	Jaswal	OK
44	Q2307-01	LINDEN-SAA	SAM	06/17/25 15:59		Jaswal	OK
45	Q2310-01	TP-7	SAM	06/17/25 16:04		Jaswal	OK
46	Q2310-01DUP	TP-7DUP	DUP	06/17/25 16:08		Jaswal	OK
47	Q2310-01L	TP-7L	SD	06/17/25 16:12		Jaswal	OK
48	Q2310-01MS	TP-7MS	MS	06/17/25 16:17		Jaswal	OK
49	Q2310-01MSD	TP-7MSD	MSD	06/17/25 16:21		Jaswal	OK
50	Q2310-01A	TP-7A	PS	06/17/25 16:25	0.1ML M6004,M6013-10ML SAMPLE	Jaswal	OK
51	CCV04	CCV04	CCV	06/17/25 16:30		Jaswal	OK
52	CCB04	CCB04	CCB	06/17/25 16:34		Jaswal	OK
53	PB168444BL	PB168444BL	MB	06/17/25 16:40		Jaswal	OK
54	PB168444BS	PB168444BS	LCS	06/17/25 16:49		Jaswal	OK
55	PB168489BL	PB168489BL	MB	06/17/25 17:00		Jaswal	OK
56	PB168489BS	PB168489BS	LCS	06/17/25 17:05		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136187

Review By	jaswal	Review On	6/19/2025 12:06:04 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:50:23 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872,MP85897
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

57	PB168475BL	PB168475BL	MB	06/17/25 17:10		Jaswal	OK
58	PB168475BS	PB168475BS	LCS	06/17/25 17:14		Jaswal	OK
59	PB168490BL	PB168490BL	MB	06/17/25 17:19		Jaswal	OK
60	PB168490BS	PB168490BS	LCS	06/17/25 17:23		Jaswal	OK
61	Q2311-08	TP04-MH4-WC	SAM	06/17/25 17:27		Jaswal	OK
62	Q2311-04	TP03-MH2MH3-WC	SAM	06/17/25 17:32		Jaswal	OK
63	CCV05	CCV05	CCV	06/17/25 17:37		Jaswal	OK
64	CCB05	CCB05	CCB	06/17/25 17:41		Jaswal	OK
65	Q2334-01	MW3	SAM	06/17/25 17:50		Jaswal	OK
66	Q2334-01DUP	MW3DUP	DUP	06/17/25 17:54		Jaswal	OK
67	Q2334-01L	MW3L	SD	06/17/25 17:59		Jaswal	OK
68	Q2334-01MS	MW3MS	MS	06/17/25 18:03		Jaswal	OK
69	Q2334-01MSD	MW3MSD	MSD	06/17/25 18:07		Jaswal	OK
70	Q2334-02	MW4	SAM	06/17/25 18:16		Jaswal	OK
71	Q2334-01A	MW3A	PS	06/17/25 18:20	0.1ML M6004,M6013-10ML SAMPLE	Jaswal	OK
72	Q2334-03DL	GBTW1DL	SAM	06/17/25 18:25		Jaswal	OK
73	Q2296-13	WC-4	SAM	06/17/25 18:29		Jaswal	OK
74	Q2296-17	WC-5	SAM	06/17/25 18:33		Jaswal	OK
75	CCV06	CCV06	CCV	06/17/25 18:38		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136187

Review By	jaswal	Review On	6/19/2025 12:06:04 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:50:23 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872,MP85897		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

76	CCB06	CCB06	CCB	06/17/25 18:42		Jaswal	OK
77	Q2311-01	TP03-MH2MH3-WC	SAM	06/17/25 18:47		Jaswal	OK
78	Q2311-05	TP04-MH4-WC	SAM	06/17/25 18:51		Jaswal	OK
79	Q2296-01	WC-1	SAM	06/17/25 18:55		Jaswal	OK
80	Q2296-01DUP	WC-1DUP	DUP	06/17/25 19:00		Jaswal	OK
81	Q2296-01L	WC-1L	SD	06/17/25 19:04		Jaswal	OK
82	Q2296-01MS	WC-1MS	MS	06/17/25 19:08		Jaswal	OK
83	Q2296-01MSD	WC-1MSD	MSD	06/17/25 19:12		Jaswal	OK
84	Q2296-01A	WC-1A	PS	06/17/25 19:17	0.1ML M6004,M6013-10ML SAMPLE	Jaswal	OK
85	Q2296-09	WC-3	SAM	06/17/25 19:21		Jaswal	OK
86	CCV07	CCV07	CCV	06/17/25 19:25		Jaswal	OK
87	CCB07	CCB07	CCB	06/17/25 19:30		Jaswal	OK
88	Q2296-21	WC-6	SAM	06/17/25 19:34		Jaswal	OK
89	Q2297-01	TP-3	SAM	06/17/25 19:38		Jaswal	OK
90	Q2307-03DL	LINDEN-SAA-WATER	SAM	06/17/25 19:43		Jaswal	OK
91	Q2316-01	RW8-SP100-2025061	SAM	06/17/25 19:47		Jaswal	OK
92	Q2316-02	RW8-SP303-2025061	SAM	06/17/25 19:51		Jaswal	OK
93	Q2316-02DUP	RW8-SP303-2025061	DUP	06/17/25 19:56		Jaswal	OK
94	Q2316-02L	RW8-SP303-2025061	SD	06/17/25 20:00		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136187

Review By	jaswal	Review On	6/19/2025 12:06:04 PM
Supervise By	MOHAN	Supervise On	6/19/2025 12:50:23 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872,MP85897		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

95	Q2316-02MS	RW8-SP303-2025061	MS	06/17/25 20:04		Jaswal	OK
96	Q2316-02MSD	RW8-SP303-2025061	MSD	06/17/25 20:09		Jaswal	OK
97	Q2316-02A	RW8-SP303-2025061	PS	06/17/25 20:13	0.1ML M6004,M6013-10ML SAMPLE	Jaswal	OK
98	CCV08	CCV08	CCV	06/17/25 20:17		Jaswal	OK
99	CCB08	CCB08	CCB	06/17/25 20:22		Jaswal	OK