

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV39	Mercury	3.96	4.0	99	90 - 110	CV	06/16/2025	12:52	LB136159

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV45	Mercury	5.20	5.0	104	90 - 110	CV	06/16/2025	12:56	LB136159
CCV46	Mercury	4.96	5.0	99	90 - 110	CV	06/16/2025	13:30	LB136159
CCV47	Mercury	5.00	5.0	100	90 - 110	CV	06/16/2025	13:49	LB136159
CCV48	Mercury	5.04	5.0	101	90 - 110	CV	06/16/2025	14:11	LB136159

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7940	8000	99	90 - 110	P	06/16/2025	14:03	lb136164
	Antimony	4090	4000	102	90 - 110	P	06/16/2025	14:03	lb136164
	Arsenic	3820	4000	96	90 - 110	P	06/16/2025	14:03	lb136164
	Barium	7830	8000	98	90 - 110	P	06/16/2025	14:03	lb136164
	Beryllium	203	200	102	90 - 110	P	06/16/2025	14:03	lb136164
	Cadmium	1950	2000	98	90 - 110	P	06/16/2025	14:03	lb136164
	Calcium	19700	20000	98	90 - 110	P	06/16/2025	14:03	lb136164
	Chromium	812	800	102	90 - 110	P	06/16/2025	14:03	lb136164
	Cobalt	1970	2000	99	90 - 110	P	06/16/2025	14:03	lb136164
	Copper	1030	1000	103	90 - 110	P	06/16/2025	14:03	lb136164
	Iron	3880	4000	97	90 - 110	P	06/16/2025	14:03	lb136164
	Lead	3820	4000	96	90 - 110	P	06/16/2025	14:03	lb136164
	Magnesium	19600	20000	98	90 - 110	P	06/16/2025	14:03	lb136164
	Manganese	1950	2000	98	90 - 110	P	06/16/2025	14:03	lb136164
	Nickel	1980	2000	99	90 - 110	P	06/16/2025	14:03	lb136164
	Potassium	20300	20000	102	90 - 110	P	06/16/2025	14:03	lb136164
	Selenium	3940	4000	99	90 - 110	P	06/16/2025	14:03	lb136164
	Silver	963	1000	96	90 - 110	P	06/16/2025	14:03	lb136164
	Sodium	20200	20000	101	90 - 110	P	06/16/2025	14:03	lb136164
	Thallium	4040	4000	101	90 - 110	P	06/16/2025	14:03	lb136164
	Vanadium	1940	2000	97	90 - 110	P	06/16/2025	14:03	lb136164
	Zinc	1980	2000	99	90 - 110	P	06/16/2025	14:03	lb136164

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	116	100	116	80 - 120	P	06/16/2025	14:42	lb136164
	Antimony	53.4	50.0	107	80 - 120	P	06/16/2025	14:42	lb136164
	Arsenic	16.6	20.0	83	80 - 120	P	06/16/2025	14:42	lb136164
	Barium	101	100	101	80 - 120	P	06/16/2025	14:42	lb136164
	Beryllium	6.39	6.0	106	80 - 120	P	06/16/2025	14:42	lb136164
	Cadmium	5.68	6.0	95	80 - 120	P	06/16/2025	14:42	lb136164
	Calcium	2050	2000	103	80 - 120	P	06/16/2025	14:42	lb136164
	Chromium	9.87	10.0	99	80 - 120	P	06/16/2025	14:42	lb136164
	Cobalt	30.1	30.0	100	80 - 120	P	06/16/2025	14:42	lb136164
	Copper	22.6	20.0	113	80 - 120	P	06/16/2025	14:42	lb136164
	Iron	101	100	101	80 - 120	P	06/16/2025	14:42	lb136164
	Lead	12.5	12.0	104	80 - 120	P	06/16/2025	14:42	lb136164
	Magnesium	2090	2000	105	80 - 120	P	06/16/2025	14:42	lb136164
	Manganese	20.8	20.0	104	80 - 120	P	06/16/2025	14:42	lb136164
	Nickel	39.0	40.0	98	80 - 120	P	06/16/2025	14:42	lb136164
	Potassium	1620	2000	81	80 - 120	P	06/16/2025	14:42	lb136164
	Selenium	18.6	20.0	93	80 - 120	P	06/16/2025	14:42	lb136164
	Silver	10.8	10.0	108	80 - 120	P	06/16/2025	14:42	lb136164
	Sodium	1840	2000	92	80 - 120	P	06/16/2025	14:42	lb136164
	Thallium	37.2	40.0	93	80 - 120	P	06/16/2025	14:42	lb136164
	Vanadium	41.0	40.0	103	80 - 120	P	06/16/2025	14:42	lb136164
	Zinc	41.4	40.0	104	80 - 120	P	06/16/2025	14:42	lb136164

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10400	10000	104	90 - 110	P	06/16/2025	15:26	lb136164
	Antimony	5180	5000	104	90 - 110	P	06/16/2025	15:26	lb136164
	Arsenic	5140	5000	103	90 - 110	P	06/16/2025	15:26	lb136164
	Barium	10300	10000	103	90 - 110	P	06/16/2025	15:26	lb136164
	Beryllium	262	250	105	90 - 110	P	06/16/2025	15:26	lb136164
	Cadmium	2580	2500	103	90 - 110	P	06/16/2025	15:26	lb136164
	Calcium	26100	25000	104	90 - 110	P	06/16/2025	15:26	lb136164
	Chromium	1050	1000	105	90 - 110	P	06/16/2025	15:26	lb136164
	Cobalt	2560	2500	102	90 - 110	P	06/16/2025	15:26	lb136164
	Copper	1300	1250	104	90 - 110	P	06/16/2025	15:26	lb136164
	Iron	5140	5000	103	90 - 110	P	06/16/2025	15:26	lb136164
	Lead	5100	5000	102	90 - 110	P	06/16/2025	15:26	lb136164
	Magnesium	25400	25000	102	90 - 110	P	06/16/2025	15:26	lb136164
	Manganese	2570	2500	103	90 - 110	P	06/16/2025	15:26	lb136164
	Nickel	2580	2500	103	90 - 110	P	06/16/2025	15:26	lb136164
	Potassium	25300	25000	101	90 - 110	P	06/16/2025	15:26	lb136164
	Selenium	5180	5000	104	90 - 110	P	06/16/2025	15:26	lb136164
	Silver	1320	1250	106	90 - 110	P	06/16/2025	15:26	lb136164
	Sodium	25600	25000	102	90 - 110	P	06/16/2025	15:26	lb136164
	Thallium	5260	5000	105	90 - 110	P	06/16/2025	15:26	lb136164
	Vanadium	2560	2500	102	90 - 110	P	06/16/2025	15:26	lb136164
	Zinc	2580	2500	103	90 - 110	P	06/16/2025	15:26	lb136164
CCV02	Aluminum	10100	10000	101	90 - 110	P	06/16/2025	16:29	lb136164
	Antimony	5070	5000	101	90 - 110	P	06/16/2025	16:29	lb136164
	Arsenic	5040	5000	101	90 - 110	P	06/16/2025	16:29	lb136164
	Barium	10100	10000	101	90 - 110	P	06/16/2025	16:29	lb136164
	Beryllium	260	250	104	90 - 110	P	06/16/2025	16:29	lb136164
	Cadmium	2520	2500	101	90 - 110	P	06/16/2025	16:29	lb136164
	Calcium	25500	25000	102	90 - 110	P	06/16/2025	16:29	lb136164
	Chromium	1020	1000	102	90 - 110	P	06/16/2025	16:29	lb136164
	Cobalt	2500	2500	100	90 - 110	P	06/16/2025	16:29	lb136164
	Copper	1290	1250	103	90 - 110	P	06/16/2025	16:29	lb136164
	Iron	4960	5000	99	90 - 110	P	06/16/2025	16:29	lb136164
	Lead	4990	5000	100	90 - 110	P	06/16/2025	16:29	lb136164

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24800	25000	99	90 - 110	P	06/16/2025	16:29	lb136164
	Manganese	2490	2500	100	90 - 110	P	06/16/2025	16:29	lb136164
	Nickel	2510	2500	100	90 - 110	P	06/16/2025	16:29	lb136164
	Potassium	25100	25000	100	90 - 110	P	06/16/2025	16:29	lb136164
	Selenium	5090	5000	102	90 - 110	P	06/16/2025	16:29	lb136164
	Silver	1280	1250	102	90 - 110	P	06/16/2025	16:29	lb136164
	Sodium	25500	25000	102	90 - 110	P	06/16/2025	16:29	lb136164
	Thallium	5110	5000	102	90 - 110	P	06/16/2025	16:29	lb136164
	Vanadium	2490	2500	100	90 - 110	P	06/16/2025	16:29	lb136164
	Zinc	2510	2500	100	90 - 110	P	06/16/2025	16:29	lb136164
CCV03	Aluminum	10100	10000	101	90 - 110	P	06/16/2025	17:31	lb136164
	Antimony	5060	5000	101	90 - 110	P	06/16/2025	17:31	lb136164
	Arsenic	5060	5000	101	90 - 110	P	06/16/2025	17:31	lb136164
	Barium	10200	10000	102	90 - 110	P	06/16/2025	17:31	lb136164
	Beryllium	258	250	103	90 - 110	P	06/16/2025	17:31	lb136164
	Cadmium	2530	2500	101	90 - 110	P	06/16/2025	17:31	lb136164
	Calcium	25300	25000	101	90 - 110	P	06/16/2025	17:31	lb136164
	Chromium	1010	1000	101	90 - 110	P	06/16/2025	17:31	lb136164
	Cobalt	2510	2500	100	90 - 110	P	06/16/2025	17:31	lb136164
	Copper	1280	1250	102	90 - 110	P	06/16/2025	17:31	lb136164
	Iron	4950	5000	99	90 - 110	P	06/16/2025	17:31	lb136164
	Lead	4990	5000	100	90 - 110	P	06/16/2025	17:31	lb136164
	Magnesium	25000	25000	100	90 - 110	P	06/16/2025	17:31	lb136164
	Manganese	2480	2500	99	90 - 110	P	06/16/2025	17:31	lb136164
	Nickel	2510	2500	100	90 - 110	P	06/16/2025	17:31	lb136164
	Potassium	24900	25000	99	90 - 110	P	06/16/2025	17:31	lb136164
	Selenium	5070	5000	102	90 - 110	P	06/16/2025	17:31	lb136164
	Silver	1260	1250	101	90 - 110	P	06/16/2025	17:31	lb136164
	Sodium	25600	25000	102	90 - 110	P	06/16/2025	17:31	lb136164
	Thallium	5150	5000	103	90 - 110	P	06/16/2025	17:31	lb136164
	Vanadium	2490	2500	100	90 - 110	P	06/16/2025	17:31	lb136164
	Zinc	2500	2500	100	90 - 110	P	06/16/2025	17:31	lb136164
CCV04	Aluminum	10000	10000	100	90 - 110	P	06/16/2025	18:29	lb136164
	Antimony	5030	5000	101	90 - 110	P	06/16/2025	18:29	lb136164

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4970	5000	99	90 - 110	P	06/16/2025	18:29	lb136164
	Barium	10200	10000	102	90 - 110	P	06/16/2025	18:29	lb136164
	Beryllium	258	250	103	90 - 110	P	06/16/2025	18:29	lb136164
	Cadmium	2500	2500	100	90 - 110	P	06/16/2025	18:29	lb136164
	Calcium	25000	25000	100	90 - 110	P	06/16/2025	18:29	lb136164
	Chromium	988	1000	99	90 - 110	P	06/16/2025	18:29	lb136164
	Cobalt	2490	2500	99	90 - 110	P	06/16/2025	18:29	lb136164
	Copper	1290	1250	103	90 - 110	P	06/16/2025	18:29	lb136164
	Iron	4870	5000	97	90 - 110	P	06/16/2025	18:29	lb136164
	Lead	4960	5000	99	90 - 110	P	06/16/2025	18:29	lb136164
	Magnesium	24700	25000	99	90 - 110	P	06/16/2025	18:29	lb136164
	Manganese	2450	2500	98	90 - 110	P	06/16/2025	18:29	lb136164
	Nickel	2490	2500	100	90 - 110	P	06/16/2025	18:29	lb136164
	Potassium	25700	25000	103	90 - 110	P	06/16/2025	18:29	lb136164
	Selenium	5010	5000	100	90 - 110	P	06/16/2025	18:29	lb136164
	Silver	1230	1250	99	90 - 110	P	06/16/2025	18:29	lb136164
	Sodium	25900	25000	104	90 - 110	P	06/16/2025	18:29	lb136164
	Thallium	5090	5000	102	90 - 110	P	06/16/2025	18:29	lb136164
	Vanadium	2480	2500	99	90 - 110	P	06/16/2025	18:29	lb136164
	Zinc	2520	2500	101	90 - 110	P	06/16/2025	18:29	lb136164
CCV05	Aluminum	10100	10000	101	90 - 110	P	06/16/2025	19:17	lb136164
	Antimony	5050	5000	101	90 - 110	P	06/16/2025	19:17	lb136164
	Arsenic	4990	5000	100	90 - 110	P	06/16/2025	19:17	lb136164
	Barium	10400	10000	104	90 - 110	P	06/16/2025	19:17	lb136164
	Beryllium	267	250	107	90 - 110	P	06/16/2025	19:17	lb136164
	Cadmium	2510	2500	100	90 - 110	P	06/16/2025	19:17	lb136164
	Calcium	25300	25000	101	90 - 110	P	06/16/2025	19:17	lb136164
	Chromium	989	1000	99	90 - 110	P	06/16/2025	19:17	lb136164
	Cobalt	2480	2500	99	90 - 110	P	06/16/2025	19:17	lb136164
	Copper	1320	1250	106	90 - 110	P	06/16/2025	19:17	lb136164
	Iron	4930	5000	99	90 - 110	P	06/16/2025	19:17	lb136164
	Lead	4950	5000	99	90 - 110	P	06/16/2025	19:17	lb136164
	Magnesium	24800	25000	99	90 - 110	P	06/16/2025	19:17	lb136164
	Manganese	2470	2500	99	90 - 110	P	06/16/2025	19:17	lb136164

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2500	2500	100	90 - 110	P	06/16/2025	19:17	lb136164
	Potassium	26600	25000	106	90 - 110	P	06/16/2025	19:17	lb136164
	Selenium	5040	5000	101	90 - 110	P	06/16/2025	19:17	lb136164
	Silver	1240	1250	99	90 - 110	P	06/16/2025	19:17	lb136164
	Sodium	26400	25000	106	90 - 110	P	06/16/2025	19:17	lb136164
	Thallium	5130	5000	103	90 - 110	P	06/16/2025	19:17	lb136164
	Vanadium	2500	2500	100	90 - 110	P	06/16/2025	19:17	lb136164
	Zinc	2530	2500	101	90 - 110	P	06/16/2025	19:17	lb136164
CCV06	Aluminum	10200	10000	102	90 - 110	P	06/16/2025	20:04	lb136164
	Antimony	5030	5000	100	90 - 110	P	06/16/2025	20:04	lb136164
	Arsenic	4950	5000	99	90 - 110	P	06/16/2025	20:04	lb136164
	Barium	10100	10000	101	90 - 110	P	06/16/2025	20:04	lb136164
	Beryllium	266	250	106	90 - 110	P	06/16/2025	20:04	lb136164
	Cadmium	2500	2500	100	90 - 110	P	06/16/2025	20:04	lb136164
	Calcium	25200	25000	101	90 - 110	P	06/16/2025	20:04	lb136164
	Chromium	1020	1000	102	90 - 110	P	06/16/2025	20:04	lb136164
	Cobalt	2490	2500	100	90 - 110	P	06/16/2025	20:04	lb136164
	Copper	1320	1250	106	90 - 110	P	06/16/2025	20:04	lb136164
	Iron	4870	5000	97	90 - 110	P	06/16/2025	20:04	lb136164
	Lead	4940	5000	99	90 - 110	P	06/16/2025	20:04	lb136164
	Magnesium	24100	25000	96	90 - 110	P	06/16/2025	20:04	lb136164
	Manganese	2410	2500	96	90 - 110	P	06/16/2025	20:04	lb136164
	Nickel	2500	2500	100	90 - 110	P	06/16/2025	20:04	lb136164
	Potassium	26600	25000	106	90 - 110	P	06/16/2025	20:04	lb136164
	Selenium	5010	5000	100	90 - 110	P	06/16/2025	20:04	lb136164
	Silver	1280	1250	102	90 - 110	P	06/16/2025	20:04	lb136164
	Sodium	26700	25000	107	90 - 110	P	06/16/2025	20:04	lb136164
	Thallium	5060	5000	101	90 - 110	P	06/16/2025	20:04	lb136164
	Vanadium	2420	2500	97	90 - 110	P	06/16/2025	20:04	lb136164
	Zinc	2480	2500	99	90 - 110	P	06/16/2025	20:04	lb136164
CCV07	Aluminum	9940	10000	99	90 - 110	P	06/16/2025	20:52	lb136164
	Antimony	4960	5000	99	90 - 110	P	06/16/2025	20:52	lb136164
	Arsenic	4910	5000	98	90 - 110	P	06/16/2025	20:52	lb136164
	Barium	10100	10000	101	90 - 110	P	06/16/2025	20:52	lb136164

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	263	250	105	90 - 110	P	06/16/2025	20:52	lb136164
	Cadmium	2460	2500	98	90 - 110	P	06/16/2025	20:52	lb136164
	Calcium	24600	25000	98	90 - 110	P	06/16/2025	20:52	lb136164
	Chromium	982	1000	98	90 - 110	P	06/16/2025	20:52	lb136164
	Cobalt	2460	2500	98	90 - 110	P	06/16/2025	20:52	lb136164
	Copper	1310	1250	105	90 - 110	P	06/16/2025	20:52	lb136164
	Iron	4710	5000	94	90 - 110	P	06/16/2025	20:52	lb136164
	Lead	4880	5000	98	90 - 110	P	06/16/2025	20:52	lb136164
	Magnesium	23800	25000	95	90 - 110	P	06/16/2025	20:52	lb136164
	Manganese	2380	2500	95	90 - 110	P	06/16/2025	20:52	lb136164
	Nickel	2460	2500	98	90 - 110	P	06/16/2025	20:52	lb136164
	Potassium	26400	25000	106	90 - 110	P	06/16/2025	20:52	lb136164
	Selenium	4960	5000	99	90 - 110	P	06/16/2025	20:52	lb136164
	Silver	1230	1250	99	90 - 110	P	06/16/2025	20:52	lb136164
	Sodium	26700	25000	107	90 - 110	P	06/16/2025	20:52	lb136164
	Thallium	4990	5000	100	90 - 110	P	06/16/2025	20:52	lb136164
	Vanadium	2410	2500	96	90 - 110	P	06/16/2025	20:52	lb136164
	Zinc	2440	2500	98	90 - 110	P	06/16/2025	20:52	lb136164
CCV08	Aluminum	9810	10000	98	90 - 110	P	06/16/2025	21:14	lb136164
	Antimony	4940	5000	99	90 - 110	P	06/16/2025	21:14	lb136164
	Arsenic	4870	5000	97	90 - 110	P	06/16/2025	21:14	lb136164
	Barium	9880	10000	99	90 - 110	P	06/16/2025	21:14	lb136164
	Beryllium	259	250	103	90 - 110	P	06/16/2025	21:14	lb136164
	Cadmium	2450	2500	98	90 - 110	P	06/16/2025	21:14	lb136164
	Calcium	24400	25000	98	90 - 110	P	06/16/2025	21:14	lb136164
	Chromium	984	1000	98	90 - 110	P	06/16/2025	21:14	lb136164
	Cobalt	2440	2500	98	90 - 110	P	06/16/2025	21:14	lb136164
	Copper	1290	1250	103	90 - 110	P	06/16/2025	21:14	lb136164
	Iron	4660	5000	93	90 - 110	P	06/16/2025	21:14	lb136164
	Lead	4850	5000	97	90 - 110	P	06/16/2025	21:14	lb136164
	Magnesium	23500	25000	94	90 - 110	P	06/16/2025	21:14	lb136164
	Manganese	2350	2500	94	90 - 110	P	06/16/2025	21:14	lb136164
	Nickel	2440	2500	98	90 - 110	P	06/16/2025	21:14	lb136164
	Potassium	26100	25000	104	90 - 110	P	06/16/2025	21:14	lb136164

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4930	5000	98	90 - 110	P	06/16/2025	21:14	1b136164
	Silver	1230	1250	99	90 - 110	P	06/16/2025	21:14	1b136164
	Sodium	26300	25000	105	90 - 110	P	06/16/2025	21:14	1b136164
	Thallium	4970	5000	100	90 - 110	P	06/16/2025	21:14	1b136164
	Vanadium	2380	2500	95	90 - 110	P	06/16/2025	21:14	1b136164
	Zinc	2440	2500	97	90 - 110	P	06/16/2025	21:14	1b136164

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7940	8000	99	90 - 110	P	06/17/2025	12:25	LB136187
	Antimony	4130	4000	103	90 - 110	P	06/17/2025	12:25	LB136187
	Arsenic	3880	4000	97	90 - 110	P	06/17/2025	12:25	LB136187
	Barium	8170	8000	102	90 - 110	P	06/17/2025	12:25	LB136187
	Beryllium	208	200	104	90 - 110	P	06/17/2025	12:25	LB136187
	Cadmium	1970	2000	99	90 - 110	P	06/17/2025	12:25	LB136187
	Calcium	19800	20000	99	90 - 110	P	06/17/2025	12:25	LB136187
	Chromium	816	800	102	90 - 110	P	06/17/2025	12:25	LB136187
	Cobalt	1990	2000	100	90 - 110	P	06/17/2025	12:25	LB136187
	Copper	1050	1000	104	90 - 110	P	06/17/2025	12:25	LB136187
	Iron	3960	4000	99	90 - 110	P	06/17/2025	12:25	LB136187
	Lead	3890	4000	97	90 - 110	P	06/17/2025	12:25	LB136187
	Magnesium	20200	20000	101	90 - 110	P	06/17/2025	12:25	LB136187
	Manganese	2010	2000	100	90 - 110	P	06/17/2025	12:25	LB136187
	Nickel	2000	2000	100	90 - 110	P	06/17/2025	12:25	LB136187
	Potassium	20500	20000	102	90 - 110	P	06/17/2025	12:25	LB136187
	Selenium	3980	4000	99	90 - 110	P	06/17/2025	12:25	LB136187
	Silver	956	1000	96	90 - 110	P	06/17/2025	12:25	LB136187
	Sodium	20700	20000	104	90 - 110	P	06/17/2025	12:25	LB136187
	Thallium	4050	4000	101	90 - 110	P	06/17/2025	12:25	LB136187
	Vanadium	2010	2000	101	90 - 110	P	06/17/2025	12:25	LB136187
	Zinc	2020	2000	101	90 - 110	P	06/17/2025	12:25	LB136187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	89.5	100	90	80 - 120	P	06/17/2025	12:41	LB136187
	Antimony	54.3	50.0	109	80 - 120	P	06/17/2025	12:41	LB136187
	Arsenic	19.7	20.0	99	80 - 120	P	06/17/2025	12:41	LB136187
	Barium	101	100	101	80 - 120	P	06/17/2025	12:41	LB136187
	Beryllium	6.43	6.0	107	80 - 120	P	06/17/2025	12:41	LB136187
	Cadmium	5.76	6.0	96	80 - 120	P	06/17/2025	12:41	LB136187
	Calcium	2060	2000	103	80 - 120	P	06/17/2025	12:41	LB136187
	Chromium	9.74	10.0	97	80 - 120	P	06/17/2025	12:41	LB136187
	Cobalt	29.8	30.0	100	80 - 120	P	06/17/2025	12:41	LB136187
	Copper	23.3	20.0	117	80 - 120	P	06/17/2025	12:41	LB136187
	Iron	110	100	110	80 - 120	P	06/17/2025	12:41	LB136187
	Lead	12.5	12.0	104	80 - 120	P	06/17/2025	12:41	LB136187
	Magnesium	2130	2000	106	80 - 120	P	06/17/2025	12:41	LB136187
	Manganese	21.1	20.0	105	80 - 120	P	06/17/2025	12:41	LB136187
	Nickel	39.7	40.0	99	80 - 120	P	06/17/2025	12:41	LB136187
	Potassium	1650	2000	82	80 - 120	P	06/17/2025	12:41	LB136187
	Selenium	20.9	20.0	105	80 - 120	P	06/17/2025	12:41	LB136187
	Silver	10.7	10.0	107	80 - 120	P	06/17/2025	12:41	LB136187
	Sodium	1840	2000	92	80 - 120	P	06/17/2025	12:41	LB136187
	Thallium	40.2	40.0	101	80 - 120	P	06/17/2025	12:41	LB136187
	Vanadium	44.7	40.0	112	80 - 120	P	06/17/2025	12:41	LB136187
	Zinc	41.5	40.0	104	80 - 120	P	06/17/2025	12:41	LB136187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9700	10000	97	90 - 110	P	06/17/2025	13:40	LB136187
	Antimony	4870	5000	97	90 - 110	P	06/17/2025	13:40	LB136187
	Arsenic	4780	5000	96	90 - 110	P	06/17/2025	13:40	LB136187
	Barium	9770	10000	98	90 - 110	P	06/17/2025	13:40	LB136187
	Beryllium	250	250	100	90 - 110	P	06/17/2025	13:40	LB136187
	Cadmium	2420	2500	97	90 - 110	P	06/17/2025	13:40	LB136187
	Calcium	24300	25000	97	90 - 110	P	06/17/2025	13:40	LB136187
	Chromium	983	1000	98	90 - 110	P	06/17/2025	13:40	LB136187
	Cobalt	2400	2500	96	90 - 110	P	06/17/2025	13:40	LB136187
	Copper	1250	1250	100	90 - 110	P	06/17/2025	13:40	LB136187
	Iron	4850	5000	97	90 - 110	P	06/17/2025	13:40	LB136187
	Lead	4790	5000	96	90 - 110	P	06/17/2025	13:40	LB136187
	Magnesium	24100	25000	96	90 - 110	P	06/17/2025	13:40	LB136187
	Manganese	2430	2500	97	90 - 110	P	06/17/2025	13:40	LB136187
	Nickel	2410	2500	96	90 - 110	P	06/17/2025	13:40	LB136187
	Potassium	25100	25000	100	90 - 110	P	06/17/2025	13:40	LB136187
	Selenium	4860	5000	97	90 - 110	P	06/17/2025	13:40	LB136187
	Silver	1230	1250	98	90 - 110	P	06/17/2025	13:40	LB136187
	Sodium	24800	25000	99	90 - 110	P	06/17/2025	13:40	LB136187
	Thallium	4910	5000	98	90 - 110	P	06/17/2025	13:40	LB136187
	Vanadium	2420	2500	97	90 - 110	P	06/17/2025	13:40	LB136187
	Zinc	2430	2500	97	90 - 110	P	06/17/2025	13:40	LB136187
CCV02	Aluminum	9910	10000	99	90 - 110	P	06/17/2025	14:43	LB136187
	Antimony	5000	5000	100	90 - 110	P	06/17/2025	14:43	LB136187
	Arsenic	4880	5000	98	90 - 110	P	06/17/2025	14:43	LB136187
	Barium	10100	10000	100	90 - 110	P	06/17/2025	14:43	LB136187
	Beryllium	253	250	101	90 - 110	P	06/17/2025	14:43	LB136187
	Cadmium	2460	2500	99	90 - 110	P	06/17/2025	14:43	LB136187
	Calcium	25200	25000	101	90 - 110	P	06/17/2025	14:43	LB136187
	Chromium	1000	1000	100	90 - 110	P	06/17/2025	14:43	LB136187
	Cobalt	2450	2500	98	90 - 110	P	06/17/2025	14:43	LB136187
	Copper	1260	1250	101	90 - 110	P	06/17/2025	14:43	LB136187
	Iron	4990	5000	100	90 - 110	P	06/17/2025	14:43	LB136187
	Lead	4900	5000	98	90 - 110	P	06/17/2025	14:43	LB136187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24800	25000	99	90 - 110	P	06/17/2025	14:43	LB136187
	Manganese	2490	2500	99	90 - 110	P	06/17/2025	14:43	LB136187
	Nickel	2460	2500	98	90 - 110	P	06/17/2025	14:43	LB136187
	Potassium	26000	25000	104	90 - 110	P	06/17/2025	14:43	LB136187
	Selenium	4930	5000	98	90 - 110	P	06/17/2025	14:43	LB136187
	Silver	1270	1250	102	90 - 110	P	06/17/2025	14:43	LB136187
	Sodium	25100	25000	100	90 - 110	P	06/17/2025	14:43	LB136187
	Thallium	5030	5000	100	90 - 110	P	06/17/2025	14:43	LB136187
	Vanadium	2500	2500	100	90 - 110	P	06/17/2025	14:43	LB136187
	Zinc	2510	2500	100	90 - 110	P	06/17/2025	14:43	LB136187
CCV03	Aluminum	9900	10000	99	90 - 110	P	06/17/2025	15:38	LB136187
	Antimony	4860	5000	97	90 - 110	P	06/17/2025	15:38	LB136187
	Arsenic	4740	5000	95	90 - 110	P	06/17/2025	15:38	LB136187
	Barium	9710	10000	97	90 - 110	P	06/17/2025	15:38	LB136187
	Beryllium	247	250	99	90 - 110	P	06/17/2025	15:38	LB136187
	Cadmium	2400	2500	96	90 - 110	P	06/17/2025	15:38	LB136187
	Calcium	24400	25000	98	90 - 110	P	06/17/2025	15:38	LB136187
	Chromium	981	1000	98	90 - 110	P	06/17/2025	15:38	LB136187
	Cobalt	2370	2500	95	90 - 110	P	06/17/2025	15:38	LB136187
	Copper	1230	1250	99	90 - 110	P	06/17/2025	15:38	LB136187
	Iron	4870	5000	97	90 - 110	P	06/17/2025	15:38	LB136187
	Lead	4740	5000	95	90 - 110	P	06/17/2025	15:38	LB136187
	Magnesium	23800	25000	95	90 - 110	P	06/17/2025	15:38	LB136187
	Manganese	2400	2500	96	90 - 110	P	06/17/2025	15:38	LB136187
	Nickel	2390	2500	96	90 - 110	P	06/17/2025	15:38	LB136187
	Potassium	25100	25000	100	90 - 110	P	06/17/2025	15:38	LB136187
	Selenium	4770	5000	96	90 - 110	P	06/17/2025	15:38	LB136187
	Silver	1230	1250	99	90 - 110	P	06/17/2025	15:38	LB136187
	Sodium	24400	25000	98	90 - 110	P	06/17/2025	15:38	LB136187
	Thallium	4890	5000	98	90 - 110	P	06/17/2025	15:38	LB136187
	Vanadium	2390	2500	96	90 - 110	P	06/17/2025	15:38	LB136187
	Zinc	2420	2500	97	90 - 110	P	06/17/2025	15:38	LB136187
CCV04	Aluminum	10000	10000	100	90 - 110	P	06/17/2025	16:30	LB136187
	Antimony	4980	5000	100	90 - 110	P	06/17/2025	16:30	LB136187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture SDG No.: Q2303
Contract: PORT06 Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4840	5000	97	90 - 110	P	06/17/2025	16:30	LB136187
	Barium	10100	10000	101	90 - 110	P	06/17/2025	16:30	LB136187
	Beryllium	260	250	104	90 - 110	P	06/17/2025	16:30	LB136187
	Cadmium	2460	2500	98	90 - 110	P	06/17/2025	16:30	LB136187
	Calcium	25400	25000	101	90 - 110	P	06/17/2025	16:30	LB136187
	Chromium	1010	1000	101	90 - 110	P	06/17/2025	16:30	LB136187
	Cobalt	2430	2500	97	90 - 110	P	06/17/2025	16:30	LB136187
	Copper	1290	1250	103	90 - 110	P	06/17/2025	16:30	LB136187
	Iron	4980	5000	100	90 - 110	P	06/17/2025	16:30	LB136187
	Lead	4850	5000	97	90 - 110	P	06/17/2025	16:30	LB136187
	Magnesium	24500	25000	98	90 - 110	P	06/17/2025	16:30	LB136187
	Manganese	2470	2500	99	90 - 110	P	06/17/2025	16:30	LB136187
	Nickel	2460	2500	98	90 - 110	P	06/17/2025	16:30	LB136187
	Potassium	26400	25000	105	90 - 110	P	06/17/2025	16:30	LB136187
	Selenium	4880	5000	98	90 - 110	P	06/17/2025	16:30	LB136187
	Silver	1280	1250	102	90 - 110	P	06/17/2025	16:30	LB136187
	Sodium	25400	25000	102	90 - 110	P	06/17/2025	16:30	LB136187
	Thallium	5020	5000	100	90 - 110	P	06/17/2025	16:30	LB136187
	Vanadium	2460	2500	98	90 - 110	P	06/17/2025	16:30	LB136187
	Zinc	2490	2500	100	90 - 110	P	06/17/2025	16:30	LB136187
CCV05	Aluminum	10000	10000	100	90 - 110	P	06/17/2025	17:37	LB136187
	Antimony	4950	5000	99	90 - 110	P	06/17/2025	17:37	LB136187
	Arsenic	4820	5000	96	90 - 110	P	06/17/2025	17:37	LB136187
	Barium	10000	10000	100	90 - 110	P	06/17/2025	17:37	LB136187
	Beryllium	257	250	103	90 - 110	P	06/17/2025	17:37	LB136187
	Cadmium	2460	2500	98	90 - 110	P	06/17/2025	17:37	LB136187
	Calcium	25000	25000	100	90 - 110	P	06/17/2025	17:37	LB136187
	Chromium	1010	1000	101	90 - 110	P	06/17/2025	17:37	LB136187
	Cobalt	2420	2500	97	90 - 110	P	06/17/2025	17:37	LB136187
	Copper	1280	1250	102	90 - 110	P	06/17/2025	17:37	LB136187
	Iron	4900	5000	98	90 - 110	P	06/17/2025	17:37	LB136187
	Lead	4850	5000	97	90 - 110	P	06/17/2025	17:37	LB136187
	Magnesium	24300	25000	97	90 - 110	P	06/17/2025	17:37	LB136187
	Manganese	2440	2500	98	90 - 110	P	06/17/2025	17:37	LB136187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2450	2500	98	90 - 110	P	06/17/2025	17:37	LB136187
	Potassium	25900	25000	104	90 - 110	P	06/17/2025	17:37	LB136187
	Selenium	4880	5000	98	90 - 110	P	06/17/2025	17:37	LB136187
	Silver	1270	1250	102	90 - 110	P	06/17/2025	17:37	LB136187
	Sodium	25200	25000	101	90 - 110	P	06/17/2025	17:37	LB136187
	Thallium	5010	5000	100	90 - 110	P	06/17/2025	17:37	LB136187
	Vanadium	2430	2500	97	90 - 110	P	06/17/2025	17:37	LB136187
	Zinc	2480	2500	99	90 - 110	P	06/17/2025	17:37	LB136187
CCV06	Aluminum	9900	10000	99	90 - 110	P	06/17/2025	18:38	LB136187
	Antimony	4920	5000	98	90 - 110	P	06/17/2025	18:38	LB136187
	Arsenic	4800	5000	96	90 - 110	P	06/17/2025	18:38	LB136187
	Barium	9840	10000	98	90 - 110	P	06/17/2025	18:38	LB136187
	Beryllium	256	250	102	90 - 110	P	06/17/2025	18:38	LB136187
	Cadmium	2440	2500	98	90 - 110	P	06/17/2025	18:38	LB136187
	Calcium	24800	25000	99	90 - 110	P	06/17/2025	18:38	LB136187
	Chromium	987	1000	99	90 - 110	P	06/17/2025	18:38	LB136187
	Cobalt	2420	2500	97	90 - 110	P	06/17/2025	18:38	LB136187
	Copper	1270	1250	102	90 - 110	P	06/17/2025	18:38	LB136187
	Iron	4900	5000	98	90 - 110	P	06/17/2025	18:38	LB136187
	Lead	4820	5000	96	90 - 110	P	06/17/2025	18:38	LB136187
	Magnesium	23900	25000	96	90 - 110	P	06/17/2025	18:38	LB136187
	Manganese	2400	2500	96	90 - 110	P	06/17/2025	18:38	LB136187
	Nickel	2430	2500	97	90 - 110	P	06/17/2025	18:38	LB136187
	Potassium	25800	25000	103	90 - 110	P	06/17/2025	18:38	LB136187
	Selenium	4890	5000	98	90 - 110	P	06/17/2025	18:38	LB136187
	Silver	1250	1250	100	90 - 110	P	06/17/2025	18:38	LB136187
	Sodium	25100	25000	100	90 - 110	P	06/17/2025	18:38	LB136187
	Thallium	4980	5000	100	90 - 110	P	06/17/2025	18:38	LB136187
	Vanadium	2400	2500	96	90 - 110	P	06/17/2025	18:38	LB136187
	Zinc	2420	2500	97	90 - 110	P	06/17/2025	18:38	LB136187
CCV07	Aluminum	9890	10000	99	90 - 110	P	06/17/2025	19:25	LB136187
	Antimony	4960	5000	99	90 - 110	P	06/17/2025	19:25	LB136187
	Arsenic	4860	5000	97	90 - 110	P	06/17/2025	19:25	LB136187
	Barium	10100	10000	100	90 - 110	P	06/17/2025	19:25	LB136187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	259	250	103	90 - 110	P	06/17/2025	19:25	LB136187
	Cadmium	2450	2500	98	90 - 110	P	06/17/2025	19:25	LB136187
	Calcium	24800	25000	99	90 - 110	P	06/17/2025	19:25	LB136187
	Chromium	986	1000	99	90 - 110	P	06/17/2025	19:25	LB136187
	Cobalt	2440	2500	98	90 - 110	P	06/17/2025	19:25	LB136187
	Copper	1300	1250	104	90 - 110	P	06/17/2025	19:25	LB136187
	Iron	4900	5000	98	90 - 110	P	06/17/2025	19:25	LB136187
	Lead	4870	5000	97	90 - 110	P	06/17/2025	19:25	LB136187
	Magnesium	24200	25000	97	90 - 110	P	06/17/2025	19:25	LB136187
	Manganese	2430	2500	97	90 - 110	P	06/17/2025	19:25	LB136187
	Nickel	2440	2500	98	90 - 110	P	06/17/2025	19:25	LB136187
	Potassium	26200	25000	105	90 - 110	P	06/17/2025	19:25	LB136187
	Selenium	4940	5000	99	90 - 110	P	06/17/2025	19:25	LB136187
	Silver	1240	1250	99	90 - 110	P	06/17/2025	19:25	LB136187
	Sodium	26000	25000	104	90 - 110	P	06/17/2025	19:25	LB136187
	Thallium	4990	5000	100	90 - 110	P	06/17/2025	19:25	LB136187
	Vanadium	2450	2500	98	90 - 110	P	06/17/2025	19:25	LB136187
	Zinc	2470	2500	99	90 - 110	P	06/17/2025	19:25	LB136187
CCV08	Aluminum	9820	10000	98	90 - 110	P	06/17/2025	20:17	LB136187
	Antimony	4930	5000	99	90 - 110	P	06/17/2025	20:17	LB136187
	Arsenic	4850	5000	97	90 - 110	P	06/17/2025	20:17	LB136187
	Barium	9910	10000	99	90 - 110	P	06/17/2025	20:17	LB136187
	Beryllium	251	250	100	90 - 110	P	06/17/2025	20:17	LB136187
	Cadmium	2440	2500	98	90 - 110	P	06/17/2025	20:17	LB136187
	Calcium	24500	25000	98	90 - 110	P	06/17/2025	20:17	LB136187
	Chromium	968	1000	97	90 - 110	P	06/17/2025	20:17	LB136187
	Cobalt	2440	2500	98	90 - 110	P	06/17/2025	20:17	LB136187
	Copper	1270	1250	101	90 - 110	P	06/17/2025	20:17	LB136187
	Iron	4770	5000	95	90 - 110	P	06/17/2025	20:17	LB136187
	Lead	4870	5000	98	90 - 110	P	06/17/2025	20:17	LB136187
	Magnesium	24100	25000	96	90 - 110	P	06/17/2025	20:17	LB136187
	Manganese	2390	2500	96	90 - 110	P	06/17/2025	20:17	LB136187
	Nickel	2440	2500	97	90 - 110	P	06/17/2025	20:17	LB136187
	Potassium	26000	25000	104	90 - 110	P	06/17/2025	20:17	LB136187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4940	5000	99	90 - 110	P	06/17/2025	20:17	LB136187
	Silver	1230	1250	98	90 - 110	P	06/17/2025	20:17	LB136187
	Sodium	25700	25000	103	90 - 110	P	06/17/2025	20:17	LB136187
	Thallium	4970	5000	99	90 - 110	P	06/17/2025	20:17	LB136187
	Vanadium	2410	2500	96	90 - 110	P	06/17/2025	20:17	LB136187
	Zinc	2440	2500	98	90 - 110	P	06/17/2025	20:17	LB136187