



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** Q2303
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	105	70 - 130	CV	06/16/2025	13:01	LB136159
CRI01	Aluminum	104	100	104	65 - 135	P	06/16/2025	14:53	lb136164
	Antimony	53.0	50.0	106	65 - 135	P	06/16/2025	14:53	lb136164
	Arsenic	16.3	20.0	82	65 - 135	P	06/16/2025	14:53	lb136164
	Barium	101	100	101	65 - 135	P	06/16/2025	14:53	lb136164
	Beryllium	6.39	6.0	106	65 - 135	P	06/16/2025	14:53	lb136164
	Cadmium	5.66	6.0	94	65 - 135	P	06/16/2025	14:53	lb136164
	Calcium	2030	2000	101	65 - 135	P	06/16/2025	14:53	lb136164
	Chromium	9.79	10.0	98	65 - 135	P	06/16/2025	14:53	lb136164
	Cobalt	29.9	30.0	100	65 - 135	P	06/16/2025	14:53	lb136164
	Copper	23.0	20.0	115	65 - 135	P	06/16/2025	14:53	lb136164
	Iron	110	100	110	65 - 135	P	06/16/2025	14:53	lb136164
	Lead	11.5	12.0	96	65 - 135	P	06/16/2025	14:53	lb136164
	Magnesium	2080	2000	104	65 - 135	P	06/16/2025	14:53	lb136164
	Manganese	21.2	20.0	106	65 - 135	P	06/16/2025	14:53	lb136164
	Nickel	38.8	40.0	97	65 - 135	P	06/16/2025	14:53	lb136164
	Potassium	1610	2000	81	65 - 135	P	06/16/2025	14:53	lb136164
	Selenium	17.6	20.0	88	65 - 135	P	06/16/2025	14:53	lb136164
	Silver	11.0	10.0	110	65 - 135	P	06/16/2025	14:53	lb136164
	Sodium	1830	2000	91	65 - 135	P	06/16/2025	14:53	lb136164
	Thallium	39.1	40.0	98	65 - 135	P	06/16/2025	14:53	lb136164
	Vanadium	37.5	40.0	94	65 - 135	P	06/16/2025	14:53	lb136164
	Zinc	41.7	40.0	104	65 - 135	P	06/16/2025	14:53	lb136164
CRI01	Aluminum	114	100	114	65 - 135	P	06/17/2025	12:50	LB136187
	Antimony	52.5	50.0	105	65 - 135	P	06/17/2025	12:50	LB136187
	Arsenic	19.4	20.0	97	65 - 135	P	06/17/2025	12:50	LB136187
	Barium	104	100	104	65 - 135	P	06/17/2025	12:50	LB136187
	Beryllium	6.43	6.0	107	65 - 135	P	06/17/2025	12:50	LB136187
	Cadmium	5.77	6.0	96	65 - 135	P	06/17/2025	12:50	LB136187
	Calcium	2040	2000	102	65 - 135	P	06/17/2025	12:50	LB136187
	Chromium	9.38	10.0	94	65 - 135	P	06/17/2025	12:50	LB136187
	Cobalt	30.0	30.0	100	65 - 135	P	06/17/2025	12:50	LB136187
	Copper	23.2	20.0	116	65 - 135	P	06/17/2025	12:50	LB136187
	Iron	103	100	103	65 - 135	P	06/17/2025	12:50	LB136187
	Lead	11.7	12.0	97	65 - 135	P	06/17/2025	12:50	LB136187

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** Q2303

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Magnesium	2150	2000	108	65 - 135	P	06/17/2025	12:50	LB136187
	Manganese	20.6	20.0	103	65 - 135	P	06/17/2025	12:50	LB136187
	Nickel	39.1	40.0	98	65 - 135	P	06/17/2025	12:50	LB136187
	Potassium	1660	2000	83	65 - 135	P	06/17/2025	12:50	LB136187
	Selenium	18.0	20.0	90	65 - 135	P	06/17/2025	12:50	LB136187
	Silver	10.8	10.0	108	65 - 135	P	06/17/2025	12:50	LB136187
	Sodium	1870	2000	94	65 - 135	P	06/17/2025	12:50	LB136187
	Thallium	39.8	40.0	99	65 - 135	P	06/17/2025	12:50	LB136187
	Vanadium	43.2	40.0	108	65 - 135	P	06/17/2025	12:50	LB136187
	Zinc	41.8	40.0	104	65 - 135	P	06/17/2025	12:50	LB136187