



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2303

Contract: PORT06

Lab Code: CHEM

Case No.: Q2303

SAS No.: Q2303

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB39	Mercury	0.076	+/-0.2	U	0.20	CV	06/16/2025	12:54	LB136159

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2303

Contract: PORT06

Lab Code: CHEM

Case No.: Q2303

SAS No.: Q2303

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB45	Mercury	0.076	+/-0.2	U	0.20	CV	06/16/2025	12:58	LB136159
CCB46	Mercury	0.076	+/-0.2	U	0.20	CV	06/16/2025	13:33	LB136159
CCB47	Mercury	0.076	+/-0.2	U	0.20	CV	06/16/2025	13:52	LB136159
CCB48	Mercury	0.076	+/-0.2	U	0.20	CV	06/16/2025	14:14	LB136159

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q2303							
Contract: PORT06	Lab Code: CHEM	Case No.: Q2303	SAS No.: Q2303						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	39.0	+/-50	J	100	P	06/16/2025	14:49	lb136164
	Antimony	6.76	+/-25	U	50.0	P	06/16/2025	14:49	lb136164
	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	14:49	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	14:49	lb136164
	Beryllium	0.56	+/-3	U	6.00	P	06/16/2025	14:49	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	14:49	lb136164
	Calcium	234	+/-1000	U	2000	P	06/16/2025	14:49	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	14:49	lb136164
	Cobalt	2.26	+/-15	U	30.0	P	06/16/2025	14:49	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	14:49	lb136164
	Iron	49.2	+/-50	J	100	P	06/16/2025	14:49	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	14:49	lb136164
	Magnesium	244	+/-1000	U	2000	P	06/16/2025	14:49	lb136164
	Manganese	5.94	+/-10	U	20.0	P	06/16/2025	14:49	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	14:49	lb136164
	Potassium	918	+/-1000	U	2000	P	06/16/2025	14:49	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	14:49	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	14:49	lb136164
	Sodium	868	+/-1000	U	2000	P	06/16/2025	14:49	lb136164
	Thallium	4.38	+/-20	U	40.0	P	06/16/2025	14:49	lb136164
	Vanadium	6.26	+/-20	U	40.0	P	06/16/2025	14:49	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	14:49	lb136164

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q2303							
Contract: PORT06	Lab Code: CHEM	Case No.: Q2303	SAS No.: Q2303						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	45.3	+/-50	J	100	P	06/16/2025	15:30	lb136164
	Antimony	7.24	+/-25	J	50.0	P	06/16/2025	15:30	lb136164
	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	15:30	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	15:30	lb136164
	Beryllium	0.56	+/-3	U	6.00	P	06/16/2025	15:30	lb136164
	Cadmium	1.35	+/-3	J	6.00	P	06/16/2025	15:30	lb136164
	Calcium	234	+/-1000	U	2000	P	06/16/2025	15:30	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	15:30	lb136164
	Cobalt	2.26	+/-15	U	30.0	P	06/16/2025	15:30	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	15:30	lb136164
	Iron	26.0	+/-50	J	100	P	06/16/2025	15:30	lb136164
	Lead	3.27	+/-6	J	12.0	P	06/16/2025	15:30	lb136164
	Magnesium	244	+/-1000	U	2000	P	06/16/2025	15:30	lb136164
	Manganese	5.94	+/-10	U	20.0	P	06/16/2025	15:30	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	15:30	lb136164
	Potassium	918	+/-1000	U	2000	P	06/16/2025	15:30	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	15:30	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	15:30	lb136164
	Sodium	868	+/-1000	U	2000	P	06/16/2025	15:30	lb136164
	Thallium	4.48	+/-20	J	40.0	P	06/16/2025	15:30	lb136164
	Vanadium	6.26	+/-20	U	40.0	P	06/16/2025	15:30	lb136164
	Zinc	5.12	+/-20	J	40.0	P	06/16/2025	15:30	lb136164
CCB02	Aluminum	11.3	+/-50	U	100	P	06/16/2025	16:34	lb136164
	Antimony	6.76	+/-25	U	50.0	P	06/16/2025	16:34	lb136164
	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	16:34	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	16:34	lb136164
	Beryllium	0.56	+/-3	U	6.00	P	06/16/2025	16:34	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	16:34	lb136164
	Calcium	234	+/-1000	U	2000	P	06/16/2025	16:34	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	16:34	lb136164
	Cobalt	2.26	+/-15	U	30.0	P	06/16/2025	16:34	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	16:34	lb136164
	Iron	37.5	+/-50	J	100	P	06/16/2025	16:34	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	16:34	lb136164
	Magnesium	244	+/-1000	U	2000	P	06/16/2025	16:34	lb136164
	Manganese	5.94	+/-10	U	20.0	P	06/16/2025	16:34	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	16:34	lb136164
	Potassium	918	+/-1000	U	2000	P	06/16/2025	16:34	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	16:34	lb136164

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q2303							
Contract: PORT06	Lab Code: CHEM	Case No.: Q2303	SAS No.: Q2303						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	06/16/2025	16:34	lb136164
	Sodium	868	+/-1000	U	2000	P	06/16/2025	16:34	lb136164
	Thallium	4.38	+/-20	U	40.0	P	06/16/2025	16:34	lb136164
	Vanadium	6.26	+/-20	U	40.0	P	06/16/2025	16:34	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	16:34	lb136164
CCB03	Aluminum	11.3	+/-50	U	100	P	06/16/2025	17:49	lb136164
	Antimony	6.76	+/-25	U	50.0	P	06/16/2025	17:49	lb136164
	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	17:49	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	17:49	lb136164
	Beryllium	0.56	+/-3	U	6.00	P	06/16/2025	17:49	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	17:49	lb136164
	Calcium	234	+/-1000	U	2000	P	06/16/2025	17:49	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	17:49	lb136164
	Cobalt	2.26	+/-15	U	30.0	P	06/16/2025	17:49	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	17:49	lb136164
	Iron	23.4	+/-50	U	100	P	06/16/2025	17:49	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	17:49	lb136164
	Magnesium	244	+/-1000	U	2000	P	06/16/2025	17:49	lb136164
	Manganese	5.94	+/-10	U	20.0	P	06/16/2025	17:49	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	17:49	lb136164
	Potassium	918	+/-1000	U	2000	P	06/16/2025	17:49	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	17:49	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	17:49	lb136164
	Sodium	868	+/-1000	U	2000	P	06/16/2025	17:49	lb136164
	Thallium	4.38	+/-20	U	40.0	P	06/16/2025	17:49	lb136164
	Vanadium	6.26	+/-20	U	40.0	P	06/16/2025	17:49	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	17:49	lb136164
CCB04	Aluminum	22.0	+/-50	J	100	P	06/16/2025	18:33	lb136164
	Antimony	6.76	+/-25	U	50.0	P	06/16/2025	18:33	lb136164
	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	18:33	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	18:33	lb136164
	Beryllium	0.56	+/-3	U	6.00	P	06/16/2025	18:33	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	18:33	lb136164
	Calcium	234	+/-1000	U	2000	P	06/16/2025	18:33	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	18:33	lb136164
	Cobalt	2.26	+/-15	U	30.0	P	06/16/2025	18:33	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	18:33	lb136164
	Iron	23.4	+/-50	U	100	P	06/16/2025	18:33	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	18:33	lb136164

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2303

Contract: PORT06

Lab Code: CHEM

Case No.: Q2303

SAS No.: Q2303

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	06/16/2025	18:33	lb136164
	Manganese	5.94	+/-10	U	20.0	P	06/16/2025	18:33	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	18:33	lb136164
	Potassium	918	+/-1000	U	2000	P	06/16/2025	18:33	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	18:33	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	18:33	lb136164
	Sodium	868	+/-1000	U	2000	P	06/16/2025	18:33	lb136164
	Thallium	4.38	+/-20	U	40.0	P	06/16/2025	18:33	lb136164
	Vanadium	6.26	+/-20	U	40.0	P	06/16/2025	18:33	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	18:33	lb136164
CCB05	Aluminum	11.3	+/-50	U	100	P	06/16/2025	19:21	lb136164
	Antimony	12.3	+/-25	J	50.0	P	06/16/2025	19:21	lb136164
	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	19:21	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	19:21	lb136164
	Beryllium	0.56	+/-3	U	6.00	P	06/16/2025	19:21	lb136164
	Cadmium	3.56	+/-3	J*	6.00	P	06/16/2025	19:21	lb136164
	Calcium	234	+/-1000	U	2000	P	06/16/2025	19:21	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	19:21	lb136164
	Cobalt	3.66	+/-15	J	30.0	P	06/16/2025	19:21	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	19:21	lb136164
	Iron	23.5	+/-50	J	100	P	06/16/2025	19:21	lb136164
	Lead	5.87	+/-6	J	12.0	P	06/16/2025	19:21	lb136164
	Magnesium	244	+/-1000	U	2000	P	06/16/2025	19:21	lb136164
	Manganese	5.94	+/-10	U	20.0	P	06/16/2025	19:21	lb136164
	Nickel	3.47	+/-20	J	40.0	P	06/16/2025	19:21	lb136164
	Potassium	918	+/-1000	U	2000	P	06/16/2025	19:21	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	19:21	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	19:21	lb136164
	Sodium	868	+/-1000	U	2000	P	06/16/2025	19:21	lb136164
	Thallium	7.90	+/-20	J	40.0	P	06/16/2025	19:21	lb136164
	Vanadium	6.26	+/-20	U	40.0	P	06/16/2025	19:21	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	19:21	lb136164
CCB06	Aluminum	11.3	+/-50	U	100	P	06/16/2025	20:08	lb136164
	Antimony	6.76	+/-25	U	50.0	P	06/16/2025	20:08	lb136164
	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	20:08	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	20:08	lb136164
	Beryllium	0.56	+/-3	U	6.00	P	06/16/2025	20:08	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	20:08	lb136164
	Calcium	234	+/-1000	U	2000	P	06/16/2025	20:08	lb136164

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q2303							
Contract: PORT06	Lab Code: CHEM	Case No.: Q2303	SAS No.: Q2303						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	20:08	lb136164
	Cobalt	2.26	+/-15	U	30.0	P	06/16/2025	20:08	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	20:08	lb136164
	Iron	23.4	+/-50	U	100	P	06/16/2025	20:08	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	20:08	lb136164
	Magnesium	244	+/-1000	U	2000	P	06/16/2025	20:08	lb136164
	Manganese	5.94	+/-10	U	20.0	P	06/16/2025	20:08	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	20:08	lb136164
	Potassium	918	+/-1000	U	2000	P	06/16/2025	20:08	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	20:08	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	20:08	lb136164
	Sodium	868	+/-1000	U	2000	P	06/16/2025	20:08	lb136164
	Thallium	4.38	+/-20	U	40.0	P	06/16/2025	20:08	lb136164
	Vanadium	6.26	+/-20	U	40.0	P	06/16/2025	20:08	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	20:08	lb136164
CCB07	Aluminum	11.3	+/-50	U	100	P	06/16/2025	20:56	lb136164
	Antimony	6.76	+/-25	U	50.0	P	06/16/2025	20:56	lb136164
	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	20:56	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	20:56	lb136164
	Beryllium	0.56	+/-3	U	6.00	P	06/16/2025	20:56	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	20:56	lb136164
	Calcium	234	+/-1000	U	2000	P	06/16/2025	20:56	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	20:56	lb136164
	Cobalt	2.26	+/-15	U	30.0	P	06/16/2025	20:56	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	20:56	lb136164
	Iron	23.4	+/-50	U	100	P	06/16/2025	20:56	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	20:56	lb136164
	Magnesium	244	+/-1000	U	2000	P	06/16/2025	20:56	lb136164
	Manganese	5.94	+/-10	U	20.0	P	06/16/2025	20:56	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	20:56	lb136164
	Potassium	918	+/-1000	U	2000	P	06/16/2025	20:56	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	20:56	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	20:56	lb136164
	Sodium	868	+/-1000	U	2000	P	06/16/2025	20:56	lb136164
	Thallium	4.38	+/-20	U	40.0	P	06/16/2025	20:56	lb136164
	Vanadium	6.26	+/-20	U	40.0	P	06/16/2025	20:56	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	20:56	lb136164
CCB08	Aluminum	11.3	+/-50	U	100	P	06/16/2025	21:18	lb136164
	Antimony	6.76	+/-25	U	50.0	P	06/16/2025	21:18	lb136164

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q2303							
Contract: PORT06	Lab Code: CHEM	Case No.: Q2303	SAS No.: Q2303						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	06/16/2025	21:18	lb136164
	Barium	14.6	+/-50	U	100	P	06/16/2025	21:18	lb136164
	Beryllium	0.56	+/-3	U	6.00	P	06/16/2025	21:18	lb136164
	Cadmium	0.50	+/-3	U	6.00	P	06/16/2025	21:18	lb136164
	Calcium	234	+/-1000	U	2000	P	06/16/2025	21:18	lb136164
	Chromium	2.12	+/-5	U	10.0	P	06/16/2025	21:18	lb136164
	Cobalt	2.26	+/-15	U	30.0	P	06/16/2025	21:18	lb136164
	Copper	4.60	+/-10	U	20.0	P	06/16/2025	21:18	lb136164
	Iron	23.4	+/-50	U	100	P	06/16/2025	21:18	lb136164
	Lead	2.30	+/-6	U	12.0	P	06/16/2025	21:18	lb136164
	Magnesium	244	+/-1000	U	2000	P	06/16/2025	21:18	lb136164
	Manganese	5.94	+/-10	U	20.0	P	06/16/2025	21:18	lb136164
	Nickel	3.06	+/-20	U	40.0	P	06/16/2025	21:18	lb136164
	Potassium	918	+/-1000	U	2000	P	06/16/2025	21:18	lb136164
	Selenium	9.64	+/-10	U	20.0	P	06/16/2025	21:18	lb136164
	Silver	1.62	+/-5	U	10.0	P	06/16/2025	21:18	lb136164
	Sodium	868	+/-1000	U	2000	P	06/16/2025	21:18	lb136164
	Thallium	4.38	+/-20	U	40.0	P	06/16/2025	21:18	lb136164
	Vanadium	6.26	+/-20	U	40.0	P	06/16/2025	21:18	lb136164
	Zinc	3.50	+/-20	U	40.0	P	06/16/2025	21:18	lb136164

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q2303							
Contract: PORT06	Lab Code: CHEM	Case No.: Q2303	SAS No.: Q2303						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	06/17/2025	12:46	LB136187
	Antimony	6.76	+/-25	U	50.0	P	06/17/2025	12:46	LB136187
	Arsenic	5.12	+/-10	U	20.0	P	06/17/2025	12:46	LB136187
	Barium	14.6	+/-50	U	100	P	06/17/2025	12:46	LB136187
	Beryllium	0.56	+/-3	U	6.00	P	06/17/2025	12:46	LB136187
	Cadmium	0.50	+/-3	U	6.00	P	06/17/2025	12:46	LB136187
	Calcium	234	+/-1000	U	2000	P	06/17/2025	12:46	LB136187
	Chromium	2.12	+/-5	U	10.0	P	06/17/2025	12:46	LB136187
	Cobalt	2.26	+/-15	U	30.0	P	06/17/2025	12:46	LB136187
	Copper	4.60	+/-10	U	20.0	P	06/17/2025	12:46	LB136187
	Iron	38.5	+/-50	J	100	P	06/17/2025	12:46	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	12:46	LB136187
	Magnesium	244	+/-1000	U	2000	P	06/17/2025	12:46	LB136187
	Manganese	5.94	+/-10	U	20.0	P	06/17/2025	12:46	LB136187
	Nickel	3.06	+/-20	U	40.0	P	06/17/2025	12:46	LB136187
	Potassium	918	+/-1000	U	2000	P	06/17/2025	12:46	LB136187
	Selenium	9.64	+/-10	U	20.0	P	06/17/2025	12:46	LB136187
	Silver	1.62	+/-5	U	10.0	P	06/17/2025	12:46	LB136187
	Sodium	868	+/-1000	U	2000	P	06/17/2025	12:46	LB136187
	Thallium	4.38	+/-20	U	40.0	P	06/17/2025	12:46	LB136187
	Vanadium	6.26	+/-20	U	40.0	P	06/17/2025	12:46	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	12:46	LB136187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2303

Contract: PORT06

Lab Code: CHEM

Case No.: Q2303

SAS No.: Q2303

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	06/17/2025	13:48	LB136187
	Antimony	6.76	+/-25	U	50.0	P	06/17/2025	13:48	LB136187
	Arsenic	5.12	+/-10	U	20.0	P	06/17/2025	13:48	LB136187
	Barium	14.6	+/-50	U	100	P	06/17/2025	13:48	LB136187
	Beryllium	0.56	+/-3	U	6.00	P	06/17/2025	13:48	LB136187
	Cadmium	0.50	+/-3	U	6.00	P	06/17/2025	13:48	LB136187
	Calcium	234	+/-1000	U	2000	P	06/17/2025	13:48	LB136187
	Chromium	2.12	+/-5	U	10.0	P	06/17/2025	13:48	LB136187
	Cobalt	2.26	+/-15	U	30.0	P	06/17/2025	13:48	LB136187
	Copper	4.60	+/-10	U	20.0	P	06/17/2025	13:48	LB136187
	Iron	23.4	+/-50	U	100	P	06/17/2025	13:48	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	13:48	LB136187
	Magnesium	244	+/-1000	U	2000	P	06/17/2025	13:48	LB136187
	Manganese	5.94	+/-10	U	20.0	P	06/17/2025	13:48	LB136187
	Nickel	3.06	+/-20	U	40.0	P	06/17/2025	13:48	LB136187
	Potassium	918	+/-1000	U	2000	P	06/17/2025	13:48	LB136187
	Selenium	9.64	+/-10	U	20.0	P	06/17/2025	13:48	LB136187
	Silver	1.62	+/-5	U	10.0	P	06/17/2025	13:48	LB136187
	Sodium	868	+/-1000	U	2000	P	06/17/2025	13:48	LB136187
	Thallium	4.38	+/-20	U	40.0	P	06/17/2025	13:48	LB136187
	Vanadium	7.99	+/-20	J	40.0	P	06/17/2025	13:48	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	13:48	LB136187
CCB02	Aluminum	11.3	+/-50	U	100	P	06/17/2025	14:50	LB136187
	Antimony	6.76	+/-25	U	50.0	P	06/17/2025	14:50	LB136187
	Arsenic	5.12	+/-10	U	20.0	P	06/17/2025	14:50	LB136187
	Barium	14.6	+/-50	U	100	P	06/17/2025	14:50	LB136187
	Beryllium	0.56	+/-3	U	6.00	P	06/17/2025	14:50	LB136187
	Cadmium	0.50	+/-3	U	6.00	P	06/17/2025	14:50	LB136187
	Calcium	234	+/-1000	U	2000	P	06/17/2025	14:50	LB136187
	Chromium	2.12	+/-5	U	10.0	P	06/17/2025	14:50	LB136187
	Cobalt	2.26	+/-15	U	30.0	P	06/17/2025	14:50	LB136187
	Copper	4.60	+/-10	U	20.0	P	06/17/2025	14:50	LB136187
	Iron	23.4	+/-50	U	100	P	06/17/2025	14:50	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	14:50	LB136187
	Magnesium	244	+/-1000	U	2000	P	06/17/2025	14:50	LB136187
	Manganese	5.94	+/-10	U	20.0	P	06/17/2025	14:50	LB136187
	Nickel	3.06	+/-20	U	40.0	P	06/17/2025	14:50	LB136187
	Potassium	918	+/-1000	U	2000	P	06/17/2025	14:50	LB136187
	Selenium	9.64	+/-10	U	20.0	P	06/17/2025	14:50	LB136187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q2303							
Contract: PORT06	Lab Code: CHEM	Case No.: Q2303	SAS No.: Q2303						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	06/17/2025	14:50	LB136187
	Sodium	868	+/-1000	U	2000	P	06/17/2025	14:50	LB136187
	Thallium	4.38	+/-20	U	40.0	P	06/17/2025	14:50	LB136187
	Vanadium	6.26	+/-20	U	40.0	P	06/17/2025	14:50	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	14:50	LB136187
CCB03	Aluminum	11.3	+/-50	U	100	P	06/17/2025	15:42	LB136187
	Antimony	6.76	+/-25	U	50.0	P	06/17/2025	15:42	LB136187
	Arsenic	5.12	+/-10	U	20.0	P	06/17/2025	15:42	LB136187
	Barium	14.6	+/-50	U	100	P	06/17/2025	15:42	LB136187
	Beryllium	0.56	+/-3	U	6.00	P	06/17/2025	15:42	LB136187
	Cadmium	0.50	+/-3	U	6.00	P	06/17/2025	15:42	LB136187
	Calcium	234	+/-1000	U	2000	P	06/17/2025	15:42	LB136187
	Chromium	2.12	+/-5	U	10.0	P	06/17/2025	15:42	LB136187
	Cobalt	2.26	+/-15	U	30.0	P	06/17/2025	15:42	LB136187
	Copper	4.60	+/-10	U	20.0	P	06/17/2025	15:42	LB136187
	Iron	27.5	+/-50	J	100	P	06/17/2025	15:42	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	15:42	LB136187
	Magnesium	244	+/-1000	U	2000	P	06/17/2025	15:42	LB136187
	Manganese	5.94	+/-10	U	20.0	P	06/17/2025	15:42	LB136187
	Nickel	3.06	+/-20	U	40.0	P	06/17/2025	15:42	LB136187
	Potassium	918	+/-1000	U	2000	P	06/17/2025	15:42	LB136187
	Selenium	9.64	+/-10	U	20.0	P	06/17/2025	15:42	LB136187
	Silver	1.62	+/-5	U	10.0	P	06/17/2025	15:42	LB136187
	Sodium	868	+/-1000	U	2000	P	06/17/2025	15:42	LB136187
	Thallium	4.38	+/-20	U	40.0	P	06/17/2025	15:42	LB136187
	Vanadium	7.56	+/-20	J	40.0	P	06/17/2025	15:42	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	15:42	LB136187
CCB04	Aluminum	11.3	+/-50	U	100	P	06/17/2025	16:34	LB136187
	Antimony	6.76	+/-25	U	50.0	P	06/17/2025	16:34	LB136187
	Arsenic	5.12	+/-10	U	20.0	P	06/17/2025	16:34	LB136187
	Barium	14.6	+/-50	U	100	P	06/17/2025	16:34	LB136187
	Beryllium	0.56	+/-3	U	6.00	P	06/17/2025	16:34	LB136187
	Cadmium	0.50	+/-3	U	6.00	P	06/17/2025	16:34	LB136187
	Calcium	234	+/-1000	U	2000	P	06/17/2025	16:34	LB136187
	Chromium	2.12	+/-5	U	10.0	P	06/17/2025	16:34	LB136187
	Cobalt	2.26	+/-15	U	30.0	P	06/17/2025	16:34	LB136187
	Copper	4.60	+/-10	U	20.0	P	06/17/2025	16:34	LB136187
	Iron	23.4	+/-50	U	100	P	06/17/2025	16:34	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	16:34	LB136187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q2303							
Contract: PORT06	Lab Code: CHEM	Case No.: Q2303	SAS No.: Q2303						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	06/17/2025	16:34	LB136187
	Manganese	5.94	+/-10	U	20.0	P	06/17/2025	16:34	LB136187
	Nickel	3.06	+/-20	U	40.0	P	06/17/2025	16:34	LB136187
	Potassium	918	+/-1000	U	2000	P	06/17/2025	16:34	LB136187
	Selenium	9.64	+/-10	U	20.0	P	06/17/2025	16:34	LB136187
	Silver	1.62	+/-5	U	10.0	P	06/17/2025	16:34	LB136187
	Sodium	868	+/-1000	U	2000	P	06/17/2025	16:34	LB136187
	Thallium	4.38	+/-20	U	40.0	P	06/17/2025	16:34	LB136187
	Vanadium	6.26	+/-20	U	40.0	P	06/17/2025	16:34	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	16:34	LB136187
CCB05	Aluminum	19.3	+/-50	J	100	P	06/17/2025	17:41	LB136187
	Antimony	6.76	+/-25	U	50.0	P	06/17/2025	17:41	LB136187
	Arsenic	5.12	+/-10	U	20.0	P	06/17/2025	17:41	LB136187
	Barium	14.6	+/-50	U	100	P	06/17/2025	17:41	LB136187
	Beryllium	0.56	+/-3	U	6.00	P	06/17/2025	17:41	LB136187
	Cadmium	1.06	+/-3	J	6.00	P	06/17/2025	17:41	LB136187
	Calcium	234	+/-1000	U	2000	P	06/17/2025	17:41	LB136187
	Chromium	2.12	+/-5	U	10.0	P	06/17/2025	17:41	LB136187
	Cobalt	2.26	+/-15	U	30.0	P	06/17/2025	17:41	LB136187
	Copper	4.60	+/-10	U	20.0	P	06/17/2025	17:41	LB136187
	Iron	23.4	+/-50	U	100	P	06/17/2025	17:41	LB136187
	Lead	2.94	+/-6	J	12.0	P	06/17/2025	17:41	LB136187
	Magnesium	244	+/-1000	U	2000	P	06/17/2025	17:41	LB136187
	Manganese	5.94	+/-10	U	20.0	P	06/17/2025	17:41	LB136187
	Nickel	3.06	+/-20	U	40.0	P	06/17/2025	17:41	LB136187
	Potassium	918	+/-1000	U	2000	P	06/17/2025	17:41	LB136187
	Selenium	9.64	+/-10	U	20.0	P	06/17/2025	17:41	LB136187
	Silver	1.62	+/-5	U	10.0	P	06/17/2025	17:41	LB136187
	Sodium	868	+/-1000	U	2000	P	06/17/2025	17:41	LB136187
	Thallium	4.38	+/-20	U	40.0	P	06/17/2025	17:41	LB136187
	Vanadium	6.26	+/-20	U	40.0	P	06/17/2025	17:41	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	17:41	LB136187
CCB06	Aluminum	11.3	+/-50	U	100	P	06/17/2025	18:42	LB136187
	Antimony	6.76	+/-25	U	50.0	P	06/17/2025	18:42	LB136187
	Arsenic	5.12	+/-10	U	20.0	P	06/17/2025	18:42	LB136187
	Barium	14.6	+/-50	U	100	P	06/17/2025	18:42	LB136187
	Beryllium	0.56	+/-3	U	6.00	P	06/17/2025	18:42	LB136187
	Cadmium	0.50	+/-3	U	6.00	P	06/17/2025	18:42	LB136187
	Calcium	234	+/-1000	U	2000	P	06/17/2025	18:42	LB136187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q2303							
Contract: PORT06	Lab Code: CHEM	Case No.: Q2303	SAS No.: Q2303						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	06/17/2025	18:42	LB136187
	Cobalt	2.26	+/-15	U	30.0	P	06/17/2025	18:42	LB136187
	Copper	4.60	+/-10	U	20.0	P	06/17/2025	18:42	LB136187
	Iron	23.4	+/-50	U	100	P	06/17/2025	18:42	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	18:42	LB136187
	Magnesium	244	+/-1000	U	2000	P	06/17/2025	18:42	LB136187
	Manganese	5.94	+/-10	U	20.0	P	06/17/2025	18:42	LB136187
	Nickel	3.06	+/-20	U	40.0	P	06/17/2025	18:42	LB136187
	Potassium	918	+/-1000	U	2000	P	06/17/2025	18:42	LB136187
	Selenium	9.64	+/-10	U	20.0	P	06/17/2025	18:42	LB136187
	Silver	1.62	+/-5	U	10.0	P	06/17/2025	18:42	LB136187
	Sodium	868	+/-1000	U	2000	P	06/17/2025	18:42	LB136187
	Thallium	4.38	+/-20	U	40.0	P	06/17/2025	18:42	LB136187
	Vanadium	12.6	+/-20	J	40.0	P	06/17/2025	18:42	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	18:42	LB136187
CCB07	Aluminum	11.3	+/-50	U	100	P	06/17/2025	19:30	LB136187
	Antimony	6.76	+/-25	U	50.0	P	06/17/2025	19:30	LB136187
	Arsenic	5.12	+/-10	U	20.0	P	06/17/2025	19:30	LB136187
	Barium	14.6	+/-50	U	100	P	06/17/2025	19:30	LB136187
	Beryllium	0.56	+/-3	U	6.00	P	06/17/2025	19:30	LB136187
	Cadmium	0.50	+/-3	U	6.00	P	06/17/2025	19:30	LB136187
	Calcium	234	+/-1000	U	2000	P	06/17/2025	19:30	LB136187
	Chromium	2.12	+/-5	U	10.0	P	06/17/2025	19:30	LB136187
	Cobalt	2.26	+/-15	U	30.0	P	06/17/2025	19:30	LB136187
	Copper	4.60	+/-10	U	20.0	P	06/17/2025	19:30	LB136187
	Iron	27.8	+/-50	J	100	P	06/17/2025	19:30	LB136187
	Lead	2.30	+/-6	U	12.0	P	06/17/2025	19:30	LB136187
	Magnesium	244	+/-1000	U	2000	P	06/17/2025	19:30	LB136187
	Manganese	5.94	+/-10	U	20.0	P	06/17/2025	19:30	LB136187
	Nickel	3.06	+/-20	U	40.0	P	06/17/2025	19:30	LB136187
	Potassium	918	+/-1000	U	2000	P	06/17/2025	19:30	LB136187
	Selenium	9.64	+/-10	U	20.0	P	06/17/2025	19:30	LB136187
	Silver	1.62	+/-5	U	10.0	P	06/17/2025	19:30	LB136187
	Sodium	868	+/-1000	U	2000	P	06/17/2025	19:30	LB136187
	Thallium	4.38	+/-20	U	40.0	P	06/17/2025	19:30	LB136187
	Vanadium	9.70	+/-20	J	40.0	P	06/17/2025	19:30	LB136187
	Zinc	3.50	+/-20	U	40.0	P	06/17/2025	19:30	LB136187
CCB08	Aluminum	31.0	+/-50	J	100	P	06/17/2025	20:22	LB136187
	Antimony	6.76	+/-25	U	50.0	P	06/17/2025	20:22	LB136187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q2303							
Contract: PORT06	Lab Code: CHEM	Case No.: Q2303	SAS No.: Q2303						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	06/17/2025	20:22	LB136187
	Barium	37.6	+/-50	J	100	P	06/17/2025	20:22	LB136187
	Beryllium	0.56	+/-3	U	6.00	P	06/17/2025	20:22	LB136187
	Cadmium	2.24	+/-3	J	6.00	P	06/17/2025	20:22	LB136187
	Calcium	234	+/-1000	U	2000	P	06/17/2025	20:22	LB136187
	Chromium	2.12	+/-5	U	10.0	P	06/17/2025	20:22	LB136187
	Cobalt	2.26	+/-15	U	30.0	P	06/17/2025	20:22	LB136187
	Copper	4.60	+/-10	U	20.0	P	06/17/2025	20:22	LB136187
	Iron	41.7	+/-50	J	100	P	06/17/2025	20:22	LB136187
	Lead	4.53	+/-6	J	12.0	P	06/17/2025	20:22	LB136187
	Magnesium	244	+/-1000	U	2000	P	06/17/2025	20:22	LB136187
	Manganese	9.42	+/-10	J	20.0	P	06/17/2025	20:22	LB136187
	Nickel	3.06	+/-20	U	40.0	P	06/17/2025	20:22	LB136187
	Potassium	918	+/-1000	U	2000	P	06/17/2025	20:22	LB136187
	Selenium	9.64	+/-10	U	20.0	P	06/17/2025	20:22	LB136187
	Silver	2.69	+/-5	J	10.0	P	06/17/2025	20:22	LB136187
	Sodium	868	+/-1000	U	2000	P	06/17/2025	20:22	LB136187
	Thallium	6.25	+/-20	J	40.0	P	06/17/2025	20:22	LB136187
	Vanadium	6.26	+/-20	U	40.0	P	06/17/2025	20:22	LB136187
	Zinc	4.70	+/-20	J	40.0	P	06/17/2025	20:22	LB136187