

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	<u>Portal Partners Tri-Venture</u>	SDG No.:	<u>Q2303</u>
Contract:	<u>PORT06</u>	Lab Code:	<u>CHEM</u>
ICS Source:	<u>EPA</u>	Case No.:	<u>Q2303</u>
		SAS No.:	<u>Q2303</u>
		Instrument ID:	<u>P5</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	257000	250000	103	216000	294000	06/16/2025	14:59	lb136164
	Antimony	23.3			-50	50	06/16/2025	14:59	lb136164
	Arsenic	-2.78			-20	20	06/16/2025	14:59	lb136164
	Barium	7.67	6.0	128	-94	106	06/16/2025	14:59	lb136164
	Beryllium	1.43			-6	6	06/16/2025	14:59	lb136164
	Cadmium	-1.53	1.0	153	-5	7	06/16/2025	14:59	lb136164
	Calcium	248000	240000	103	208000	282000	06/16/2025	14:59	lb136164
	Chromium	51.2	52.0	98	42	62	06/16/2025	14:59	lb136164
	Cobalt	22.8			-30	30	06/16/2025	14:59	lb136164
	Copper	-9.05	2.0	452	-18	22	06/16/2025	14:59	lb136164
	Iron	106000	100000	106	85600	116500	06/16/2025	14:59	lb136164
	Lead	4.50			-12	12	06/16/2025	14:59	lb136164
	Magnesium	254000	260000	98	216000	294000	06/16/2025	14:59	lb136164
	Manganese	5.70	7.0	81	-13	27	06/16/2025	14:59	lb136164
	Nickel	1.27	2.0	64	-38	42	06/16/2025	14:59	lb136164
	Potassium	1.45			0	0	06/16/2025	14:59	lb136164
	Selenium	-0.94			-20	20	06/16/2025	14:59	lb136164
	Silver	5.28			-10	10	06/16/2025	14:59	lb136164
	Sodium	116			0	0	06/16/2025	14:59	lb136164
	Thallium	19.4			-40	40	06/16/2025	14:59	lb136164
	Vanadium	-4.49			-40	40	06/16/2025	14:59	lb136164
	Zinc	11.4			-40	40	06/16/2025	14:59	lb136164
ICSAB01	Aluminum	253000	250000	101	209000	285000	06/16/2025	15:06	lb136164
	Antimony	638	620	103	525	711	06/16/2025	15:06	lb136164
	Arsenic	94.4	100	94	88.4	120	06/16/2025	15:06	lb136164
	Barium	539	540	100	437	637	06/16/2025	15:06	lb136164
	Beryllium	533	500	107	420	570	06/16/2025	15:06	lb136164
	Cadmium	1030	970	106	826	1120	06/16/2025	15:06	lb136164
	Calcium	246000	230000	107	199000	271000	06/16/2025	15:06	lb136164
	Chromium	563	540	104	460	624	06/16/2025	15:06	lb136164
	Cobalt	524	480	109	404	548	06/16/2025	15:06	lb136164
	Copper	495	510	97	434	588	06/16/2025	15:06	lb136164
	Iron	105000	99000	106	84400	114500	06/16/2025	15:06	lb136164
	Lead	52.7	49.0	108	37	61	06/16/2025	15:06	lb136164
	Magnesium	253000	250000	101	210000	286000	06/16/2025	15:06	lb136164
	Manganese	519	510	102	430	584	06/16/2025	15:06	lb136164
	Nickel	1010	950	106	810	1100	06/16/2025	15:06	lb136164
	Potassium	-3.59			0	0	06/16/2025	15:06	lb136164
	Selenium	55.7	46.0	121	26	66	06/16/2025	15:06	lb136164
	Silver	173	200	86	170	232	06/16/2025	15:06	lb136164
	Sodium	103			0	0	06/16/2025	15:06	lb136164
	Thallium	104	110	94	68	148	06/16/2025	15:06	lb136164

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture **SDG No.:** Q2303
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	507	490	104	417	565	06/16/2025	15:06	lb136164
	Zinc	1070	950	113	809	1095	06/16/2025	15:06	lb136164
ICSA01	Aluminum	247000	250000	99	216000	294000	06/17/2025	13:11	LB136187
	Antimony	21.2			-50	50	06/17/2025	13:11	LB136187
	Arsenic	-3.71			-20	20	06/17/2025	13:11	LB136187
	Barium	5.50	6.0	92	-94	106	06/17/2025	13:11	LB136187
	Beryllium	0.28			-6	6	06/17/2025	13:11	LB136187
	Cadmium	-1.65	1.0	165	-5	7	06/17/2025	13:11	LB136187
	Calcium	238000	240000	99	208000	282000	06/17/2025	13:11	LB136187
	Chromium	48.2	52.0	93	42	62	06/17/2025	13:11	LB136187
	Cobalt	22.0			-30	30	06/17/2025	13:11	LB136187
	Copper	-8.46	2.0	423	-18	22	06/17/2025	13:11	LB136187
	Iron	103000	100000	103	85600	116500	06/17/2025	13:11	LB136187
	Lead	5.50			-12	12	06/17/2025	13:11	LB136187
	Magnesium	245000	260000	94	216000	294000	06/17/2025	13:11	LB136187
	Manganese	5.99	7.0	86	-13	27	06/17/2025	13:11	LB136187
	Nickel	2.10	2.0	105	-38	42	06/17/2025	13:11	LB136187
	Potassium	0.58			0	0	06/17/2025	13:11	LB136187
	Selenium	0.58			-20	20	06/17/2025	13:11	LB136187
	Silver	4.61			-10	10	06/17/2025	13:11	LB136187
	Sodium	116			0	0	06/17/2025	13:11	LB136187
	Thallium	14.1			-40	40	06/17/2025	13:11	LB136187
	Vanadium	1.78			-40	40	06/17/2025	13:11	LB136187
	Zinc	10.3			-40	40	06/17/2025	13:11	LB136187
ICSAB01	Aluminum	251000	250000	100	209000	285000	06/17/2025	13:22	LB136187
	Antimony	630	620	102	525	711	06/17/2025	13:22	LB136187
	Arsenic	94.8	100	95	88.4	120	06/17/2025	13:22	LB136187
	Barium	523	540	97	437	637	06/17/2025	13:22	LB136187
	Beryllium	524	500	105	420	570	06/17/2025	13:22	LB136187
	Cadmium	1020	970	105	826	1120	06/17/2025	13:22	LB136187
	Calcium	242000	230000	105	199000	271000	06/17/2025	13:22	LB136187
	Chromium	564	540	104	460	624	06/17/2025	13:22	LB136187
	Cobalt	522	480	109	404	548	06/17/2025	13:22	LB136187
	Copper	492	510	96	434	588	06/17/2025	13:22	LB136187
	Iron	104000	99000	105	84400	114500	06/17/2025	13:22	LB136187
	Lead	53.1	49.0	108	37	61	06/17/2025	13:22	LB136187
	Magnesium	249000	250000	100	210000	286000	06/17/2025	13:22	LB136187
	Manganese	509	510	100	430	584	06/17/2025	13:22	LB136187
	Nickel	1000	950	105	810	1100	06/17/2025	13:22	LB136187
	Potassium	-2.54			0	0	06/17/2025	13:22	LB136187
	Selenium	50.7	46.0	110	26	66	06/17/2025	13:22	LB136187
	Silver	173	200	86	170	232	06/17/2025	13:22	LB136187

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture **SDG No.:** Q2303
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2303 **SAS No.:** Q2303
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Sodium	122			0	0	06/17/2025	13:22	LB136187
	Thallium	97.4	110	88	68	148	06/17/2025	13:22	LB136187
	Vanadium	503	490	103	417	565	06/17/2025	13:22	LB136187
	Zinc	1050	950	110	809	1095	06/17/2025	13:22	LB136187