

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Continuing Calib Date: 06/13/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 20:20 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1248-1       | (1) | 5.65       | 5.66      | 5.56      | 5.76  | 0.01       |
| Aroclor-1248-2       | (2) | 5.92       | 5.93      | 5.83      | 6.03  | 0.01       |
| Aroclor-1248-3       | (3) | 6.12       | 6.13      | 6.03      | 6.23  | 0.01       |
| Aroclor-1248-4       | (4) | 6.52       | 6.53      | 6.43      | 6.63  | 0.01       |
| Aroclor-1248-5       | (5) | 6.56       | 6.57      | 6.47      | 6.67  | 0.01       |
| Tetrachloro-m-xylene |     | 4.50       | 4.50      | 4.40      | 4.60  | 0.00       |
| Decachlorobiphenyl   |     | 10.19      | 10.22     | 10.12     | 10.32 | 0.03       |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Continuing Calib Date: 06/13/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 20:20 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.87       | 4.88      | 4.78      | 4.98 | 0.01       |
| Aroclor-1248-2       | (2) | 5.11       | 5.12      | 5.02      | 5.22 | 0.01       |
| Aroclor-1248-3       | (3) | 5.15       | 5.16      | 5.06      | 5.26 | 0.01       |
| Aroclor-1248-4       | (4) | 5.32       | 5.33      | 5.23      | 5.43 | 0.01       |
| Aroclor-1248-5       | (5) | 5.71       | 5.73      | 5.63      | 5.83 | 0.02       |
| Tetrachloro-m-xylene |     | 3.79       | 3.80      | 3.70      | 3.90 | 0.01       |
| Decachlorobiphenyl   |     | 8.80       | 8.83      | 8.73      | 8.93 | 0.03       |