

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072914.D Time Analyzed: 14:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.648	5.552	5.752	421.180	500.000	-15.8
Aroclor-1016-2	5.670	5.574	5.774	442.270	500.000	-11.5
Aroclor-1016-3	5.732	5.636	5.836	440.440	500.000	-11.9
Aroclor-1016-4	5.829	5.734	5.934	458.080	500.000	-8.4
Aroclor-1016-5	6.121	6.027	6.227	455.740	500.000	-8.9
Aroclor-1260-1	7.238	7.147	7.347	424.190	500.000	-15.2
Aroclor-1260-2	7.492	7.401	7.601	406.910	500.000	-18.6
Aroclor-1260-3	7.850	7.760	7.960	419.380	500.000	-16.1
Aroclor-1260-4	8.073	7.985	8.185	409.950	500.000	-18.0
Aroclor-1260-5	8.391	8.303	8.503	432.570	500.000	-13.5
Decachlorobiphenyl	10.189	10.113	10.313	46.160	50.000	-7.7
Tetrachloro-m-xylene	4.497	4.396	4.596	48.930	50.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072914.D Time Analyzed: 14:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.870	4.777	4.977	511.970	500.000	2.4
Aroclor-1016-2	4.889	4.795	4.995	521.930	500.000	4.4
Aroclor-1016-3	5.065	4.973	5.173	527.470	500.000	5.5
Aroclor-1016-4	5.107	5.015	5.215	526.980	500.000	5.4
Aroclor-1016-5	5.320	5.229	5.429	511.470	500.000	2.3
Aroclor-1260-1	6.352	6.264	6.464	520.920	500.000	4.2
Aroclor-1260-2	6.541	6.453	6.653	487.420	500.000	-2.5
Aroclor-1260-3	6.692	6.606	6.806	502.650	500.000	0.5
Aroclor-1260-4	7.162	7.077	7.277	515.240	500.000	3.0
Aroclor-1260-5	7.404	7.319	7.519	503.750	500.000	0.8
Decachlorobiphenyl	8.800	8.723	8.923	55.070	50.000	10.1
Tetrachloro-m-xylene	3.790	3.691	3.891	52.880	50.000	5.8