

Cover Page

Order ID : Q2303

Project ID : Amtrak Sawtooth Bridges 2025

Client : Portal Partners Tri-Venture

Lab Sample Number

Q2303-01
Q2303-02

Client Sample Number

B-165-SB01
B-170-SB03

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 6/20/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2303

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 06/20/2025

LAB CHRONICLE

OrderID:	Q2303	OrderDate:	6/12/2025 12:09:00 PM
Client:	Portal Partners Tri-Venture	Project:	Amtrak Sawtooth Bridges 2025
Contact:	Joseph Krupansky	Location:	D41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2303-01	B-165-SB01	SOIL	PCB	8082A	06/11/25	06/13/25	06/13/25	06/11/25
Q2303-02	B-170-SB03	SOIL	PCB	8082A	06/11/25	06/13/25	06/13/25	06/11/25

Hit Summary Sheet
SW-846

SDG No.: Q2303

Order ID: Q2303

Client: Portal Partners Tri-Venture

Project ID: Amtrak Sawtooth Bridges 2025

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : Q2303-01	B-165-SB01 B-165-SB01	SOIL	Aroclor-1254		209	3.80	20.3	ug/kg
Total Concentration:				209.000				



QC SUMMARY

Surrogate Summary

SDG No.: **Q2303**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP072163.D	PIBLK-PP072163.D	Tetrachloro-m-xylene	1	20	16.3	82		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	16.7	84		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	16.6	83		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.0	85		70 (60)	130 (140)
I.BLK-PP072903.D	PIBLK-PP072903.D	Tetrachloro-m-xylene	1	20	19.4	97		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.3	102		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.8	104		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.6	118		70 (60)	130 (140)
PB168459BL	PB168459BL	Tetrachloro-m-xylene	1	20	19.9	100		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.7	104		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.1	106		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	25.4	127		30 (32)	150 (175)
PB168459BS	PB168459BS	Tetrachloro-m-xylene	1	20	20.0	100		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.5	108		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.0	100		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	25.3	126		30 (32)	150 (175)
Q2303-01	B-165-SB01	Tetrachloro-m-xylene	1	20	19.4	97		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.1	95		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.3	107		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	24.0	120		30 (32)	150 (175)
Q2303-02	B-170-SB03	Tetrachloro-m-xylene	1	20	18.5	93		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	18.1	91		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	22.6	113		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.6	108		30 (32)	150 (175)
Q2305-01MS	TR-04-06122025MS	Tetrachloro-m-xylene	1	20	16.5	83		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.0	95		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	24.0	120		30 (32)	150 (175)
Q2305-01MSD	TR-04-06122025MSD	Tetrachloro-m-xylene	1	20	19.2	96		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	17.6	88		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.1	105		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.2	111		30 (32)	150 (175)
I.BLK-PP072918.D	PIBLK-PP072918.D	Tetrachloro-m-xylene	1	20	18.8	94		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.9	90		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.4	102		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.7	113		70 (60)	130 (140)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2303</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Portal Partners Tri-Venture</u>	DataFile :	<u>PP072911.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2305-01MS (Column 1)	TR-04-06122025MS AR1016	169.6	0	118	ug/kg	70				40 (55)	140 (146)	
	AR1260	169.6	0	113	ug/kg	67				40 (54)	140 (119)	
Client Sample ID: Q2305-01MS (Column 2)	TR-04-06122025MS AR1016	169.6	0	132	ug/kg	78				40 (55)	140 (146)	
	AR1260	169.6	0	130	ug/kg	77				40 (54)	140 (119)	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2303</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Portal Partners Tri-Venture</u>	DataFile :	<u>PP072912.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2305-01MSD (Column 1)	TR-04-06122025MSD AR1016	169.4	0	123	ug/kg	73		4		40 (55)	140 (146)	30 (15)
	AR1260	169.4	0	104	ug/kg	61		9		40 (54)	140 (119)	30 (15)
Client Sample ID: Q2305-01MSD (Column 2)	TR-04-06122025MSD AR1016	169.4	0	133	ug/kg	79		1		40 (55)	140 (146)	30 (15)
	AR1260	169.4	0	125	ug/kg	74		4		40 (54)	140 (119)	30 (15)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2303</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Portal Partners Tri-Venture</u>	Datafile :	<u>PP072905.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168459BS (Column 1)	AR1016	166.5	128	ug/kg	77				40 (71)	140 (120)	
	AR1260	166.5	126	ug/kg	76				40 (65)	140 (130)	
PB168459BS (Column 2)	AR1016	166.5	135	ug/kg	81				40 (71)	140 (120)	
	AR1260	166.5	150	ug/kg	90				40 (65)	140 (130)	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168459BL

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM Case No.: Q2303

SAS No.: Q2303 SDG NO.: Q2303

Lab Sample ID: PB168459BL

Lab File ID: PP072904.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 06/13/2025

Date Analyzed (1): 06/13/2025

Date Analyzed (2): 06/13/2025

Time Analyzed (1): 11:22

Time Analyzed (2): 11:22

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB168459BS	PB168459BS	PP072905.D	06/13/2025	06/13/2025
B-165-SB01	Q2303-01	PP072906.D	06/13/2025	06/13/2025
B-170-SB03	Q2303-02	PP072907.D	06/13/2025	06/13/2025
TR-04-06122025MS	Q2305-01MS	PP072911.D	06/13/2025	06/13/2025
TR-04-06122025MSD	Q2305-01MSD	PP072912.D	06/13/2025	06/13/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	06/11/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	06/11/25
Client Sample ID:	B-165-SB01	SDG No.:	Q2303
Lab Sample ID:	Q2303-01	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	83.6
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072906.D	1	06/13/25 08:05	06/13/25 11:55	PB168459

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.3	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.3	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.3	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.3	ug/kg
12672-29-6	Aroclor-1248	7.10	U	7.10	20.3	ug/kg
11097-69-1	Aroclor-1254	209		3.80	20.3	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.3	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.3	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.3		30 (32) - 150 (144)	107%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.0		30 (32) - 150 (175)	120%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
 Data File : PP072906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2025 11:55
 Operator : YP\AJ
 Sample : Q2303-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 B-165-SB01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/16/2025
 Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 12:51:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.494	3.791	38980456	34064797	19.386	21.312
2) SA Decachlor...	10.187	8.802	31047979	25375286	19.058m	23.971m#
Target Compounds						
26) L6 AR-1254-1	6.494	5.674	41941782	58709320	487.677m	587.362
27) L6 AR-1254-2	6.711	5.821	68357929	42603847	523.105	493.632
28) L6 AR-1254-3	7.073	6.223	60379228	63772747	457.408	495.820
29) L6 AR-1254-4	7.355	6.451	65433479	44890144	501.365	526.284
30) L6 AR-1254-5	7.769	6.868	53680353	59269719	482.535m	522.884

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 11:55
Operator : YP\AJ
Sample : Q2303-01
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
B-165-SB01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 12:51:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.494 min
 Delta R.T.: -0.004 min
 Response: 38980456
 Conc: 19.39 ng/ml

Instrument :
 ECD_P
 ClientSampleId :
 B-165-SB01


Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/16/2025
 Supervised By :mohammad ahmed 06/17/2025




#1 Tetrachloro-m-xylene

R.T.: 3.791 min
 Delta R.T.: -0.002 min
 Response: 34064797
 Conc: 21.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.187 min
 Delta R.T.: -0.025 min
 Response: 31047979
 Conc: 19.06 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.802 min
 Delta R.T.: -0.019 min
 Response: 25375286
 Conc: 23.97 ng/ml m

Response_ Signal: PP072906.D\ECD1A.ch



#26 AR-1254-1


R.T.: 6.494 min
Delta R.T.: -0.014 min
Response: 41941782
Conc: 487.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
B-165-SB01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025


Response_ Signal: PP072906.D\ECD2B.ch



#26 AR-1254-1

R.T.: 5.674 min
Delta R.T.: -0.011 min
Response: 58709320
Conc: 587.36 ng/ml


Response_ Signal: PP072906.D\ECD1A.ch



#27 AR-1254-2


R.T.: 6.711 min
Delta R.T.: -0.013 min
Response: 68357929
Conc: 523.10 ng/ml

Response_ Signal: PP072906.D\ECD2B.ch



#27 AR-1254-2

R.T.: 5.821 min
Delta R.T.: -0.013 min
Response: 42603847
Conc: 493.63 ng/ml




#28 AR-1254-3

R.T.: 7.073 min
Delta R.T.: -0.015 min
Response: 60379228
Conc: 457.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
B-165-SB01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025




#28 AR-1254-3

R.T.: 6.223 min
Delta R.T.: -0.013 min
Response: 63772747
Conc: 495.82 ng/ml




#29 AR-1254-4

R.T.: 7.355 min
Delta R.T.: -0.015 min
Response: 65433479
Conc: 501.36 ng/ml



#29 AR-1254-4

R.T.: 6.451 min
Delta R.T.: -0.014 min
Response: 44890144
Conc: 526.28 ng/ml




#30 AR-1254-5

R.T.: 7.769 min
Delta R.T.: -0.017 min
Response: 53680353
Conc: 482.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
B-165-SB01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025



#30 AR-1254-5

R.T.: 6.868 min
Delta R.T.: -0.014 min
Response: 59269719
Conc: 522.88 ng/ml

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	06/11/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	06/11/25
Client Sample ID:	B-170-SB03	SDG No.:	Q2303
Lab Sample ID:	Q2303-02	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	92.7
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072907.D	1	06/13/25 08:05	06/13/25 12:11	PB168459

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.30	U	4.30	18.3	ug/kg
11104-28-2	Aroclor-1221	4.30	U	4.30	18.3	ug/kg
11141-16-5	Aroclor-1232	4.00	U	4.00	18.3	ug/kg
53469-21-9	Aroclor-1242	4.30	U	4.30	18.3	ug/kg
12672-29-6	Aroclor-1248	6.40	U	6.40	18.3	ug/kg
11097-69-1	Aroclor-1254	3.50	U	3.50	18.3	ug/kg
37324-23-5	Aroclor-1262	5.40	U	5.40	18.3	ug/kg
11100-14-4	Aroclor-1268	3.90	U	3.90	18.3	ug/kg
11096-82-5	Aroclor-1260	3.50	U	3.50	18.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.6		30 (32) - 150 (144)	113%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.6		30 (32) - 150 (175)	108%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
 Data File : PP072907.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2025 12:11
 Operator : YP\AJ
 Sample : Q2303-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 B-170-SB03

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/16/2025
 Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 12:51:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.495	3.790	37260199	36113405	18.530m	22.594
2) SA Decachlor...	10.190	8.801	29569398	22856546	18.150m	21.591m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 12:11
Operator : YP\AJ
Sample : Q2303-02
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
B-170-SB03


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 12:51:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.495 min
Delta R.T.: -0.003 min
Response: 37260199
Conc: 18.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
B-170-SB03


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025




#1 Tetrachloro-m-xylene

R.T.: 3.790 min
Delta R.T.: -0.003 min
Response: 36113405
Conc: 22.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.190 min
Delta R.T.: -0.023 min
Response: 29569398
Conc: 18.15 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.801 min
Delta R.T.: -0.020 min
Response: 22856546
Conc: 21.59 ng/ml m



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	<u>PORT06</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2303</u>	SAS No.:	<u>Q2303</u>	SDG NO.:	<u>Q2303</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):		<u>05/19/2025</u>	<u>05/19/2025</u>		
		Calibration Times:		<u>09:42</u>	<u>18:25</u>		

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 =	<u>PP072164.D</u>	RT 750 =	<u>PP072165.D</u>
	RT 500 =	PP072166.D	RT 250 =	PP072167.D
			RT 050 =	PP072168.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.22	10.22	10.22	10.22	10.22	10.22	10.12	10.32
Tetrachloro-m-xylene	4.50	4.50	4.50	4.50	4.50	4.50	4.40	4.60
Aroclor-1268-1 (1)	8.72	8.72	8.72	8.72	8.72	8.72	8.62	8.82
Aroclor-1268-2 (2)	8.81	8.82	8.81	8.81	8.81	8.81	8.71	8.91
Aroclor-1268-3 (3)	9.04	9.05	9.04	9.04	9.05	9.04	8.94	9.14
Aroclor-1268-4 (4)	9.46	9.46	9.46	9.46	9.46	9.46	9.36	9.56
Aroclor-1268-5 (5)	9.88	9.88	9.88	9.88	9.88	9.88	9.78	9.98
Decachlorobiphenyl	10.22	10.22	10.22	10.21	10.22	10.22	10.12	10.32
Tetrachloro-m-xylene	4.50	4.51	4.50	4.50	4.50	4.50	4.40	4.60

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	<u>PORT06</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2303</u>	SAS No.:	<u>Q2303</u>	SDG NO.:	<u>Q2303</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):		<u>05/19/2025</u>	<u>05/19/2025</u>		
		Calibration Times:		<u>09:42</u>	<u>18:25</u>		

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 =	<u>PP072164.D</u>	RT 750 =	<u>PP072165.D</u>
	RT 500 =	PP072166.D	RT 250 =	PP072167.D
			RT 050 =	PP072168.D

[illegible]

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PORT06
Lab Code: CHEM **Case No.:** Q2303 **SAS No.:** Q2303 **SDG NO.:** Q2303
Instrument ID: ECD_P
Calibration Date(s): 05/19/2025 05/19/2025
Calibration Times: 09:42 18:25

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP072164.D</u>		CF 750 = <u>PP072165.D</u>			
CF 500 = <u>PP072166.D</u>		CF 250 = <u>PP072167.D</u>		CF 050 = <u>PP072168.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	66878491	69828851	76010044	83211968	79488900	75083651
Aroclor-1016-2	(2)	99468528	104467549	110518222	118579572	101924480	106991670
Aroclor-1016-3	(3)	61412187	64234915	67714060	73550892	60912680	65564947
Aroclor-1016-4	(4)	50363803	53404009	55311866	58220708	49049560	53269989
Aroclor-1016-5	(5)	45437165	47941741	49914606	51189784	50062640	48909187
Aroclor-1260-1	(1)	85553757	90346417	93957158	98060904	100055640	93594775
Aroclor-1260-2	(2)	135650987	137755076	143609204	149400496	151032820	143489717
Aroclor-1260-3	(3)	109042220	112821207	117388622	121236852	108865120	113870804
Aroclor-1260-4	(4)	102101329	107107725	110940768	114591832	101771400	107302611
Aroclor-1260-5	(5)	228876095	233706888	240834266	249937120	229956720	236662218
Decachlorobiphenyl		1538266220	1617591187	1654764640	1712186920	1622812600	1629124313
Tetrachloro-m-xylene		1944699760	2026562000	2072807240	2135906680	1874026800	2010800496
Aroclor-1232-1	(1)	44075653	45896815	48002922	48247276	49319600	47108453
Aroclor-1232-2	(2)	21717379	22928696	23244388	23029808	25495660	23283186
Aroclor-1232-3	(3)	46772523	50148184	51844656	54382920	47167260	50063109
Aroclor-1232-4	(4)	23238504	24146713	24804390	26183196	27122440	25099049
Aroclor-1232-5	(5)	15752983	16470224	16442412	18109320	20822780	17519544
Decachlorobiphenyl		1545279190	1600827560	1672901480	1703701720	1499866000	1604515190
Tetrachloro-m-xylene		1886122400	2000973440	2054641840	2122290120	1923172600	1997440080
Aroclor-1242-1	(1)	56211553	57323072	62083360	65843868	71524840	62597339
Aroclor-1242-2	(2)	82973080	85971496	91883334	97584776	75486540	86779845
Aroclor-1242-3	(3)	51559127	52121972	55496172	57080948	59071260	55065896
Aroclor-1242-4	(4)	41207603	43030657	45948576	46781836	44776460	44349026
Aroclor-1242-5	(5)	47977894	48696228	50879554	52435600	54934800	50984815
Decachlorobiphenyl		1526214180	1567133680	1645977640	1728993240	1487418400	1591147428
Tetrachloro-m-xylene		1827808490	1928927520	1991275140	2114690400	1713029600	1915146230
Aroclor-1248-1	(1)	44038397	47472432	47912646	51445364	51108500	48395468
Aroclor-1248-2	(2)	56984732	59400859	61165396	63476680	62172680	60640069
Aroclor-1248-3	(3)	64677132	65813748	68767932	70638012	61003720	66180109
Aroclor-1248-4	(4)	81606824	85304428	88412180	92196600	89440360	87392078
Aroclor-1248-5	(5)	77587697	81003728	83211240	86851844	86474060	83025714
Decachlorobiphenyl		1573368300	1599269360	1663196880	1714969280	1478724200	1605905604
Tetrachloro-m-xylene		1880240370	1909383627	1990709860	2112434560	1856348600	1949823403
Aroclor-1254-1	(1)	79792123	82783284	86922424	91684028	88834440	86003260

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1254-2	(2)	120142096	124850076	131105324	134547092	142741880	130677294	7
Aroclor-1254-3	(3)	124307316	127832729	133782854	136984312	137108200	132003082	4
Aroclor-1254-4	(4)	116304358	118550812	124418220	136069712	157210520	130510724	13
Aroclor-1254-5	(5)	107938622	112056604	116315218	117138120	102784720	111246657	5
Decachlorobiphenyl		1576500070	1614032093	1688035380	1708021160	1475084400	1612334621	6
Tetrachloro-m-xylene		1869066560	1948875347	2032435420	2069900400	1617536400	1907562825	9
Aroclor-1268-1	(1)	315692357	322966827	337626016	351356792	338115380	333151474	4
Aroclor-1268-2	(2)	265886956	272371317	285215132	296487980	279979560	279988189	4
Aroclor-1268-3	(3)	227961300	235081659	247250176	255747588	241284760	241465097	4
Aroclor-1268-4	(4)	101392203	104765767	107968348	108360296	102254740	104948271	3
Aroclor-1268-5	(5)	651541042	667177075	686960040	706355632	672967440	677000246	3
Decachlorobiphenyl		2733685450	2815599213	2937966700	3038738360	2636754200	2832548785	6
Tetrachloro-m-xylene		1849618000	1900065360	1995699220	2054630680	1744336400	1908869932	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PORT06
Lab Code: CHEM **Case No.:** Q2303 **SAS No.:** Q2303 **SDG NO.:** Q2303
Instrument ID: ECD_P **Calibration Date(s):** 05/19/2025 05/19/2025
Calibration Times: 09:42 18:25
GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP072164.D</u>		CF 750 = <u>PP072165.D</u>			
CF 500 = <u>PP072166.D</u>		CF 250 = <u>PP072167.D</u>		CF 050 = <u>PP072168.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	54525663	58849481	59632288	64515432	69624760	61429525
Aroclor-1016-2	(2)	79772717	86154421	85288758	92453296	95277560	87789350
Aroclor-1016-3	(3)	42641318	46352395	47330430	50989192	50959120	47654491
Aroclor-1016-4	(4)	33654246	36754908	37811202	41521120	40485160	38045327
Aroclor-1016-5	(5)	43791070	47744144	49105368	53701332	54780200	49824423
Aroclor-1260-1	(1)	72244144	75307759	80373306	84844232	85999160	79753720
Aroclor-1260-2	(2)	87664004	90365408	98490088	104924488	125666920	101422182
Aroclor-1260-3	(3)	79909091	81326363	88937536	92494016	91844780	86902357
Aroclor-1260-4	(4)	64349922	67685321	72647720	77711116	78060640	72090944
Aroclor-1260-5	(5)	161315528	166937577	177635780	187517596	174713540	173624004
Decachlorobiphenyl		920208980	1044964960	1093136160	1163579000	1071118200	1058601460
Tetrachloro-m-xylene		1547649930	1584243107	1593704220	1639561360	1626633000	1598358323
Aroclor-1232-1	(1)	37616932	40592059	45290150	47667724	45555860	43344545
Aroclor-1232-2	(2)	39358353	42024527	44532022	47402620	47585980	44180700
Aroclor-1232-3	(3)	20796080	22347549	23808552	25172964	24790340	23383097
Aroclor-1232-4	(4)	17897124	19415229	20680720	21689292	23918420	20720157
Aroclor-1232-5	(5)	19472370	21237964	22534450	24219412	28307940	23154427
Decachlorobiphenyl		991061470	1069964400	1120483020	1147415120	1151969600	1096178722
Tetrachloro-m-xylene		1503650850	1560846907	1685979140	1671096560	1578260800	1599966851
Aroclor-1242-1	(1)	46123670	49902476	52843000	57237560	54435320	52108405
Aroclor-1242-2	(2)	66003789	71039361	75471286	80940564	74026140	73496228
Aroclor-1242-3	(3)	34394227	38777200	41064482	44714476	36677280	39125533
Aroclor-1242-4	(4)	32146418	36854119	38826456	43192656	36894340	37582798
Aroclor-1242-5	(5)	41776015	46613400	48754570	54226988	49917340	48257663
Decachlorobiphenyl		967171680	1074707920	1046164720	1163123680	1078294400	1065892480
Tetrachloro-m-xylene		1545621850	1670448987	1600956100	1623877080	1377345600	1563649923
Aroclor-1248-1	(1)	38942472	38080832	42980810	46736940	49867020	43321615
Aroclor-1248-2	(2)	50685144	49386624	55814958	61079164	65420760	56477330
Aroclor-1248-3	(3)	53612101	51937139	58301874	64007240	67677000	59107071
Aroclor-1248-4	(4)	63542852	61081784	68799052	75621092	79219440	69652844
Aroclor-1248-5	(5)	63751699	62374025	69241268	74501112	78829180	69739457
Decachlorobiphenyl		1077930460	1066864453	1129616840	1218519160	1187538600	1136093903
Tetrachloro-m-xylene		1585370290	1544149787	1686359520	1807980440	1717995000	1668371007
Aroclor-1254-1	(1)	88394708	92134828	103452628	109580892	106208260	99954263

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1254-2	(2)	75158218	78808232	90151838	93786952	93629340	86306916	10
Aroclor-1254-3	(3)	118588697	123091027	137573488	140955076	122895140	128620686	8
Aroclor-1254-4	(4)	76184182	79817879	90049888	92363636	88066400	85296397	8
Aroclor-1254-5	(5)	103321222	111630831	120507404	120857332	110441220	113351602	7
Decachlorobiphenyl		1048618240	1081081133	1110321040	1211638680	1138223000	1117976419	6
Tetrachloro-m-xylene		1515554540	1677026120	1758978840	1703246040	1394379800	1609837068	9
Aroclor-1268-1	(1)	212587803	222334836	235771254	243664628	256729280	234217560	7
Aroclor-1268-2	(2)	188526781	196925783	208389682	217661736	222594360	206819668	7
Aroclor-1268-3	(3)	155195263	161035703	174669460	176674312	179279020	169370752	6
Aroclor-1268-4	(4)	67370219	71019111	74855072	76562592	74977520	72956903	5
Aroclor-1268-5	(5)	431064337	442819956	457037218	461770552	492119380	456962289	5
Decachlorobiphenyl		1753973260	1820277787	1902401960	1966032280	2130248600	1914586777	8
Tetrachloro-m-xylene		1581658390	1604094987	1649636500	1662342440	1498230000	1599192463	4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Instrument ID: ECD_P Date(s) Analyzed: 05/19/2025 05/19/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.70	4.60	4.80	25133000
		2	4.79	4.69	4.89	17981400
		3	4.86	4.76	4.96	58509600
		4	0.00			0
		5	0.00			0
Aroclor-1262	500	1	8.09	7.99	8.19	133245000
		2	8.41	8.31	8.51	272386000
		3	8.72	8.62	8.82	185530000
		4	8.81	8.71	8.91	134281000
		5	9.46	9.36	9.56	92460600



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Instrument ID: ECD_P Date(s) Analyzed: 05/19/2025 05/19/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.01	3.91	4.11	23124400
		2	4.09	3.99	4.19	17698400
		3	4.17	4.07	4.27	52270400
		4	0.00			0
		5	0.00			0
Aroclor-1262	500	1	6.92	6.82	7.02	113028000
		2	7.18	7.08	7.28	97127000
		3	7.70	7.60	7.80	81531200
		4	7.77	7.67	7.87	140624000
		5	8.27	8.17	8.37	65083800

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072164.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 09:42
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 10:58:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 10:54:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.498	3.791	194.5E6	154.8E6	96.811	98.534
2) SA Decachlor...	10.214	8.822	153.8E6	92020898	96.351	91.411
Target Compounds						
3) L1 AR-1016-1	5.653	4.877	66878491	54525663	936.093	955.267
4) L1 AR-1016-2	5.675	4.895	99468528	79772717	947.379	966.582
5) L1 AR-1016-3	5.737	5.072	61412187	42641318	951.196	947.882
6) L1 AR-1016-4	5.835	5.114	50363803	33654246	953.177	941.833
7) L1 AR-1016-5	6.128	5.329	45437165	43791070	953.043	942.793
31) L7 AR-1260-1	7.249	6.364	85553757	72244144	953.187	946.735
32) L7 AR-1260-2	7.503	6.553	135.7E6	87664004	971.503	941.843
33) L7 AR-1260-3	7.861	6.705	109.0E6	79909091	963.139	946.529
34) L7 AR-1260-4	8.085	7.177	102.1E6	64349922	958.508	939.431
35) L7 AR-1260-5	8.404	7.419	228.9E6	161.3E6	974.541	951.851



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072164.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 09:42
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 10:58:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 10:54:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072165.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 09:59
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 11:01:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 10:54:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.789	152.0E6	118.8E6	75.442	75.431
2) SA Decachlor...	10.213	8.822	121.3E6	78372372	75.657	76.878
Target Compounds						
3) L1 AR-1016-1	5.652	4.876	52371638	44137111	738.609	765.351
4) L1 AR-1016-2	5.673	4.894	78350662	64615816	747.492	771.637
5) L1 AR-1016-3	5.736	5.071	48176186	34764296	747.454	765.036
6) L1 AR-1016-4	5.833	5.114	40053007	27566181	755.339	764.168
7) L1 AR-1016-5	6.127	5.328	35956306	35808108	752.783	763.822
31) L7 AR-1260-1	7.247	6.363	67759813	56480819	753.285	743.413
32) L7 AR-1260-2	7.500	6.552	103.3E6	67774056	743.256	735.291
33) L7 AR-1260-3	7.860	6.705	84615905	60994772	748.257	731.431
34) L7 AR-1260-4	8.084	7.176	80330794	50763991	752.749	744.038
35) L7 AR-1260-5	8.404	7.418	175.3E6	125.2E6	747.551	742.474



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 09:59
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 11:01:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 10:54:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072166.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 10:15
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 10:55:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 10:54:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.791	103.6E6	79685211	50.000	50.000
2) SA Decachlor...	10.213	8.823	82738232	54656808	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.652	4.877	38005022	29816144	500.000	500.000
4) L1 AR-1016-2	5.674	4.895	55259111	42644379	500.000	500.000
5) L1 AR-1016-3	5.736	5.073	33857030	23665215	500.000	500.000
6) L1 AR-1016-4	5.834	5.115	27655933	18905601	500.000	500.000
7) L1 AR-1016-5	6.127	5.329	24957303	24552684	500.000	500.000
31) L7 AR-1260-1	7.247	6.364	46978579	40186653	500.000	500.000
32) L7 AR-1260-2	7.501	6.553	71804602	49245044	500.000	500.000
33) L7 AR-1260-3	7.860	6.706	58694311	44468768	500.000	500.000
34) L7 AR-1260-4	8.085	7.177	55470384	36323860	500.000	500.000
35) L7 AR-1260-5	8.403	7.419	120.4E6	88817890	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072166.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 10:15
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 10:55:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 10:54:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072167.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 10:32
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 11:04:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 10:54:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.792	53397667	40989034	26.111	25.758
2) SA Decachlor...	10.213	8.823	42804673	29089475	26.249	27.561
Target Compounds						
3) L1 AR-1016-1	5.655	4.878	20802992	16128858	281.189	271.618
4) L1 AR-1016-2	5.676	4.896	29644893	23113324	273.834	269.018
5) L1 AR-1016-3	5.739	5.073	18387723	12747298	275.562	272.213
6) L1 AR-1016-4	5.837	5.115	14555177	10380280	267.927	277.285
7) L1 AR-1016-5	6.130	5.330	12797446	13425333	263.209	276.324
31) L7 AR-1260-1	7.248	6.364	24515226	21211058	266.529	271.268
32) L7 AR-1260-2	7.504	6.553	37350124	26231122	263.765	275.072
33) L7 AR-1260-3	7.862	6.705	30309213	23123504	263.279	269.924
34) L7 AR-1260-4	8.086	7.177	28647958	19427779	263.586	275.187
35) L7 AR-1260-5	8.406	7.418	62484280	46879399	262.166	270.430



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072167.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 10:32
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 11:04:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 10:54:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072168.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 10:48
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/20/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 11:09:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 10:54:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.498	3.793	9370134	8133165	4.660	5.088
2) SA Decachlor...	10.212	8.821	8114063	5355591	4.981	5.059
Target Compounds						
3) L1 AR-1016-1	5.650	4.878	3974445	3481238	53.538m	56.670
4) L1 AR-1016-2	5.672	4.896	5096224	4763878	47.816m	54.265
5) L1 AR-1016-3	5.734	5.073	3045634	2547956	47.236m	53.467
6) L1 AR-1016-4	5.832	5.115	2452478	2024258	45.948m	53.206
7) L1 AR-1016-5	6.127	5.330	2503132	2739010	51.179	54.973
31) L7 AR-1260-1	7.247	6.364	5002782	4299958	53.452	53.915
32) L7 AR-1260-2	7.500	6.553	7551641	6283346	52.628	61.952
33) L7 AR-1260-3	7.859	6.705	5443256	4592239	47.802	52.844
34) L7 AR-1260-4	8.083	7.176	5088570	3903032	47.423	54.140
35) L7 AR-1260-5	8.403	7.418	11497836	8735677	48.583	50.314

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072168.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 10:48
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/20/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 11:09:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 10:54:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072169.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 11:05
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 11:34:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 11:34:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.794	100.3E6	78966679	50.000	50.000
2) SA Decachlor...	10.214	8.824	80876112	53101134	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.703	4.005	12566489	11562215	500.000	500.000
9) L2 AR-1221-2	4.788	4.091	8990719	8849205	500.000	500.000
10) L2 AR-1221-3	4.864	4.167	29254822	26135247	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072169.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 11:05
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 11:34:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 11:34:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072170.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 11:21
 Operator : YP\AJ
 Sample : AR1232ICC1000
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 12:10:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 12:05:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.795	188.6E6	150.4E6	95.724	94.284
2) SA Decachlor...	10.213	8.823	154.5E6	99106147	96.034	93.871
Target Compounds						
11) L3 AR-1232-1	4.863	4.168	44075653	37616932	957.349	907.448
12) L3 AR-1232-2	5.389	4.898	21717379	39358353	966.038	938.328
13) L3 AR-1232-3	5.675	5.074	46772523	20796080	948.567	932.463
14) L3 AR-1232-4	5.835	5.159	23238504	17897124	967.407	927.845
15) L3 AR-1232-5	5.925	5.331	15752983	19472370	978.586	927.105



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072170.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 11:21
Operator : YP\AJ
Sample : AR1232ICC1000
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 12:10:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 12:05:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072171.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 11:38
 Operator : YP\AJ
 Sample : AR1232ICC750
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 12:15:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 12:12:58 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.796	150.1E6	117.1E6	75.772	73.927
2) SA Decachlor...	10.215	8.824	120.1E6	80247330	74.743	75.669
Target Compounds						
11) L3 AR-1232-1	4.864	4.168	34422611	30444044	748.451	739.537
12) L3 AR-1232-2	5.390	4.898	17196522	31518395	759.894	750.945
13) L3 AR-1232-3	5.676	5.075	37611138	16760662	758.466	751.013
14) L3 AR-1232-4	5.836	5.159	18110035	14561422	752.603	753.267
15) L3 AR-1232-5	5.926	5.332	12352668	15928473	761.482	755.563



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072171.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 11:38
Operator : YP\AJ
Sample : AR1232ICC750
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 12:15:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 12:12:58 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072172.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 11:54
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 12:07:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 12:05:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.795	102.7E6	84298957	50.000	50.000
2) SA Decachlor...	10.214	8.823	83645074	56024151	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.864	4.168	24001461	22645075	500.000	500.000
12) L3 AR-1232-2	5.390	4.897	11622194	22266011	500.000	500.000
13) L3 AR-1232-3	5.676	5.075	25922328	11904276	500.000	500.000
14) L3 AR-1232-4	5.836	5.159	12402195	10340360	500.000	500.000
15) L3 AR-1232-5	5.926	5.331	8221206	11267225	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072172.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 11:54
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 12:07:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 12:05:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072173.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 12:10
 Operator : YP\AJ
 Sample : AR1232ICC250
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 12:57:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 12:57:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.795	53057253	41777414	26.318	26.023
2) SA Decachlor...	10.217	8.824	42592543	28685378	26.120	26.506
Target Compounds						
11) L3 AR-1232-1	4.865	4.167	12061819	11916931	259.084	278.487
12) L3 AR-1232-2	5.391	4.897	5757452	11850655	253.297	273.502
13) L3 AR-1232-3	5.678	5.074	13595730	6293241	267.701	273.247
14) L3 AR-1232-4	5.838	5.158	6545799	5422323	266.163	272.197
15) L3 AR-1232-5	5.928	5.331	4527330	6054853	271.199	276.907



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072173.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 12:10
Operator : YP\AJ
Sample : AR1232ICC250
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 12:57:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 12:57:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072174.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 12:27
 Operator : YP\AJ
 Sample : AR1232ICC050
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/21/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 13:01:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 13:01:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.499	3.796	9615863	7891304	4.814	4.932
2) SA Decachlor...	10.211	8.824	7499330	5759848	4.674	5.254
Target Compounds						
11) L3 AR-1232-1	4.863	4.168	2465980	2277793	52.347	52.551
12) L3 AR-1232-2	5.389	4.898	1274783	2379299	54.751	53.854
13) L3 AR-1232-3	5.674	5.075	2358363	1239517	46.317m	53.009
14) L3 AR-1232-4	5.832	5.160	1356122	1195921	55.264m	57.718
15) L3 AR-1232-5	5.924	5.331	1041139	1415397	59.427	61.129

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072174.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 12:27
Operator : YP\AJ
Sample : AR1232ICC050
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1232ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/21/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 13:01:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 13:01:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072175.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 12:43
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 13:42:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 13:38:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.797	182.8E6	154.6E6	95.720	98.241
2) SA Decachlor...	10.216	8.825	152.6E6	96717168	96.225	96.077
Target Compounds						
16) L4 AR-1242-1	5.656	4.881	56211553	46123670	950.363	932.105
17) L4 AR-1242-2	5.678	4.898	82973080	66003789	949.042	933.080
18) L4 AR-1242-3	5.740	5.076	51559127	34394227	963.224	911.604
19) L4 AR-1242-4	5.837	5.160	41207603	32146418	945.604	905.879
20) L4 AR-1242-5	6.568	5.683	47977894	41776015	970.648	922.915



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072175.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 12:43
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 13:42:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 13:38:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072176.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 12:59
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 13:45:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 13:38:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.797	144.7E6	125.3E6	75.506	78.026
2) SA Decachlor...	10.215	8.825	117.5E6	80603094	74.400	78.305
Target Compounds						
16) L4 AR-1242-1	5.656	4.881	42992304	37426857	734.417	754.223
17) L4 AR-1242-2	5.677	4.899	64478622	53279521	741.623	752.130
18) L4 AR-1242-3	5.739	5.076	39091479	29082900	736.754	763.759
19) L4 AR-1242-4	5.837	5.161	32272993	27640589	743.693	769.026
20) L4 AR-1242-5	6.568	5.684	36522171	34960050	742.554	764.745



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072176.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 12:59
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 13:45:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 13:38:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072177.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 13:15
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 13:39:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 13:38:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.505	3.798	99563757	80047805	50.000	50.000
2) SA Decachlor...	10.221	8.827	82298882	52308236	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.659	4.883	31041680	26421500	500.000	500.000
17) L4 AR-1242-2	5.681	4.901	45941667	37735643	500.000	500.000
18) L4 AR-1242-3	5.743	5.078	27748086	20532241	500.000	500.000
19) L4 AR-1242-4	5.841	5.162	22974288	19413228	500.000	500.000
20) L4 AR-1242-5	6.572	5.685	25439777	24377285	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072177.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 13:15
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 13:39:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 13:38:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072178.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 13:32
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 13:48:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 13:38:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.797	52867260	40596927	26.895	25.212
2) SA Decachlor...	10.214	8.825	43224831	29078092	26.730	27.360
Target Compounds						
16) L4 AR-1242-1	5.654	4.881	16460967	14309390	272.688	277.708
17) L4 AR-1242-2	5.675	4.899	24396194	20235141	272.269	275.819
18) L4 AR-1242-3	5.738	5.076	14270237	11178619	263.948	281.311
19) L4 AR-1242-4	5.836	5.161	11695459	10798164	264.351	286.007
20) L4 AR-1242-5	6.566	5.684	13108900	13556747	262.192	283.361



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072178.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 13:32
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 13:48:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 13:38:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072179.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 13:48
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/20/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 14:11:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 14:10:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	8565148	6886728	4.472	4.404
2) SA Decachlor...	10.216	8.824	7437092	5391472	4.674	5.058
Target Compounds						
16) L4 AR-1242-1	5.654	4.882	3576242	2721766	56.040m	52.233
17) L4 AR-1242-2	5.677	4.900	3774327	3701307	42.861m	50.361
18) L4 AR-1242-3	5.740	5.076	2953563	1833864	53.891m	46.871
19) L4 AR-1242-4	5.836	5.161	2238823	1844717	51.018m	49.084
20) L4 AR-1242-5	6.568	5.683	2746740	2495867	53.225m	52.475m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072179.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 13:48
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/20/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 14:11:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 14:10:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 14:04
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 14:59:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 14:54:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.506	3.798	188.0E6	158.5E6	97.146	96.913
2) SA Decachlor...	10.219	8.825	157.3E6	107.8E6	97.225	97.659
Target Compounds						
21) L5 AR-1248-1	5.659	4.881	44038397	38942472	957.866	950.706
22) L5 AR-1248-2	5.931	5.118	56984732	50685144	964.616	951.833
23) L5 AR-1248-3	6.134	5.161	64677132	53612101	969.345	958.095
24) L5 AR-1248-4	6.533	5.333	81606824	63542852	959.973	960.283
25) L5 AR-1248-5	6.572	5.725	77587697	63751699	965.027	958.723



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072180.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 14:04
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 14:59:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 14:54:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072181.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 14:20
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 15:02:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 14:54:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.799	143.2E6	115.8E6	74.323	72.143
2) SA Decachlor...	10.214	8.826	119.9E6	80014834	74.410	73.309
Target Compounds						
21) L5 AR-1248-1	5.655	4.882	35604324	28560624	766.105	713.991
22) L5 AR-1248-2	5.927	5.119	44550644	37039968	752.752	712.825
23) L5 AR-1248-3	6.130	5.161	49360311	38952854	743.159	713.200
24) L5 AR-1248-4	6.529	5.333	63978321	45811338	751.733	710.534
25) L5 AR-1248-5	6.568	5.726	60752796	46780519	753.748	718.348



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072181.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 14:20
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 15:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 14:54:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072182.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 14:37
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 14:56:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 14:54:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.798	99535493	84317976	50.000	50.000
2) SA Decachlor...	10.215	8.825	83159844	56480842	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.656	4.881	23956323	21490405	500.000	500.000
22) L5 AR-1248-2	5.928	5.119	30582698	27907479	500.000	500.000
23) L5 AR-1248-3	6.131	5.161	34383966	29150937	500.000	500.000
24) L5 AR-1248-4	6.530	5.333	44206090	34399526	500.000	500.000
25) L5 AR-1248-5	6.569	5.726	41605620	34620634	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072182.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 14:37
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 14:56:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 14:54:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072183.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 14:53
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 15:05:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 14:54:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.798	52810864	45199511	26.764	27.295
2) SA Decachlor...	10.217	8.825	42874232	30462979	26.180	27.121
Target Compounds						
21) L5 AR-1248-1	5.657	4.882	12861341	11684235	269.533	280.297
22) L5 AR-1248-2	5.929	5.119	15869170	15269791	263.358	281.515
23) L5 AR-1248-3	6.132	5.161	17659503	16001810	261.722	280.908
24) L5 AR-1248-4	6.531	5.333	23049150	18905273	265.299	281.073
25) L5 AR-1248-5	6.570	5.726	21712961	18625278	264.265	276.065



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072183.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 14:53
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 15:05:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 14:54:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072184.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 15:25
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/20/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 15:39:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 15:20:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.797	9281743	8589975	4.760	5.149
2) SA Decachlor...	10.214	8.824	7393621	5937693	4.604	5.226
Target Compounds						
21) L5 AR-1248-1	5.652	4.881	2555425	2493351	52.477m	57.554
22) L5 AR-1248-2	5.926	5.118	3108634	3271038	51.264	57.918
23) L5 AR-1248-3	6.128	5.160	3050186	3383850	46.089	57.249
24) L5 AR-1248-4	6.528	5.332	4472018	3960972	51.172	56.978m
25) L5 AR-1248-5	6.567	5.725	4323703	3941459	52.077	56.517

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072184.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 15:25
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/20/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 15:39:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 15:20:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072185.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 15:42
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 17:56:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 17:50:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.799	186.9E6	151.6E6	95.813	92.566
2) SA Decachlor...	10.217	8.825	157.7E6	104.9E6	96.583	97.142
Target Compounds						
26) L6 AR-1254-1	6.507	5.685	79792123	88394708	957.230	921.511
27) L6 AR-1254-2	6.724	5.834	120.1E6	75158218	956.365	909.300
28) L6 AR-1254-3	7.088	6.236	124.3E6	118.6E6	963.286	925.888
29) L6 AR-1254-4	7.370	6.465	116.3E6	76184182	966.294	916.589
30) L6 AR-1254-5	7.787	6.883	107.9E6	103.3E6	962.647	923.217



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072185.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 15:42
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 17:56:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 17:50:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072186.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 15:58
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 18:00:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 17:59:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.799	146.2E6	125.8E6	74.952	76.204
2) SA Decachlor...	10.217	8.825	121.1E6	81081085	74.439	75.075
Target Compounds						
26) L6 AR-1254-1	6.507	5.685	62087463	69101121	746.549	729.987
27) L6 AR-1254-2	6.723	5.834	93637557	59106174	746.914	726.363
28) L6 AR-1254-3	7.087	6.237	95874547	92318270	745.288	730.264
29) L6 AR-1254-4	7.370	6.465	88913109	59863409	742.441	729.887
30) L6 AR-1254-5	7.786	6.883	84042453	83723123	749.686	748.732



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072186.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 15:58
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 18:00:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 17:59:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072187.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 16:14
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 17:52:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 17:50:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.798	101.6E6	87948942	50.000	50.000
2) SA Decachlor...	10.215	8.825	84401769	55516052	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.506	5.685	43461212	51726314	500.000	500.000
27) L6 AR-1254-2	6.723	5.833	65552662	45075919	500.000	500.000
28) L6 AR-1254-3	7.086	6.236	66891427	68786744	500.000	500.000
29) L6 AR-1254-4	7.369	6.465	62209110	45024944	500.000	500.000
30) L6 AR-1254-5	7.785	6.883	58157609	60253702	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 16:14
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 17:52:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 17:50:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072188.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 16:31
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 18:03:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 17:59:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.798	51747510	42581151	26.134	25.594
2) SA Decachlor...	10.215	8.825	42700529	30290967	25.932	27.218
Target Compounds						
26) L6 AR-1254-1	6.506	5.685	22921007	27395223	268.725	278.433
27) L6 AR-1254-2	6.722	5.833	33636773	23446738	263.485	277.554
28) L6 AR-1254-3	7.085	6.237	34246078	35238769	261.967	270.959
29) L6 AR-1254-4	7.369	6.465	34017428	23090909	274.698	272.930
30) L6 AR-1254-5	7.785	6.883	29284530	30214333	258.327	264.854



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072188.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 16:31
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 18:03:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 17:59:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072189.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 16:47
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/20/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 18:06:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 17:59:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	8087682	6971899	4.231m	4.331m
2) SA Decachlor...	10.218	8.826	7375422	5691115	4.574	5.091
Target Compounds						
26) L6 AR-1254-1	6.507	5.686	4441722	5310413	51.935m	53.128
27) L6 AR-1254-2	6.724	5.834	7137094	4681467	54.616	54.242
28) L6 AR-1254-3	7.088	6.237	6855410	6144757	51.934	47.774
29) L6 AR-1254-4	7.370	6.466	7860526	4403320	60.049m	52.100m
30) L6 AR-1254-5	7.788	6.883	5139236	5522061	46.197	48.716

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072189.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 16:47
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/20/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 18:06:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 17:59:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072190.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 17:03
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 02:51:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 02:50:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.798	91392014	72543788	50.000	50.000
2) SA Decachlor...	10.218	8.826	79498650	53689400	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.088	6.920	66622555	56514187	500.000	500.000
37) L8 AR-1262-2	8.408	7.179	136.2E6	48563462	500.000	500.000
38) L8 AR-1262-3	8.724	7.703	92764853	40765600	500.000	500.000
39) L8 AR-1262-4	8.809	7.767	67140472	70312088	500.000	500.000
40) L8 AR-1262-5	9.464	8.267	46230270	32541912	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072190.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 17:03
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 02:51:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 02:50:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072191.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 17:20
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:11:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:10:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.797	185.0E6	158.2E6	92.680	95.879
2) SA Decachlor...	10.215	8.825	273.4E6	175.4E6	93.047	92.198
Target Compounds						
41) L9 AR-1268-1	8.718	7.702	315.7E6	212.6E6	935.036	901.670
42) L9 AR-1268-2	8.810	7.768	265.9E6	188.5E6	932.233	904.684
43) L9 AR-1268-3	9.041	7.971	228.0E6	155.2E6	921.986	888.508
44) L9 AR-1268-4	9.460	8.265	101.4E6	67370219	939.092	900.009
45) L9 AR-1268-5	9.876	8.565	651.5E6	431.1E6	948.441	943.171



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072191.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 17:20
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:11:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:10:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072192.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 17:36
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:11:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:10:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.506	3.799	142.5E6	120.3E6	71.406	72.929
2) SA Decachlor...	10.220	8.826	211.2E6	136.5E6	71.876	71.762
Target Compounds						
41) L9 AR-1268-1	8.722	7.704	242.2E6	166.8E6	717.436	707.258
42) L9 AR-1268-2	8.815	7.770	204.3E6	147.7E6	716.226	708.741
43) L9 AR-1268-3	9.047	7.973	176.3E6	120.8E6	713.088	691.459
44) L9 AR-1268-4	9.464	8.267	78574325	53264333	727.753	711.566
45) L9 AR-1268-5	9.880	8.568	500.4E6	332.1E6	728.402	726.669



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072192.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 17:36
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:11:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:10:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072193.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 17:53
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:11:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:10:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.799	99784961	82481825	50.000	50.000
2) SA Decachlor...	10.215	8.826	146.9E6	95120098	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.719	7.703	168.8E6	117.9E6	500.000	500.000
42) L9 AR-1268-2	8.812	7.769	142.6E6	104.2E6	500.000	500.000
43) L9 AR-1268-3	9.042	7.972	123.6E6	87334730	500.000	500.000
44) L9 AR-1268-4	9.461	8.267	53984174	37427536	500.000	500.000
45) L9 AR-1268-5	9.876	8.566	343.5E6	228.5E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072193.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 17:53
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:11:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:10:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072194.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 18:09
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:12:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:10:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.797	51365767	41558561	25.738	25.193
2) SA Decachlor...	10.214	8.824	75968459	49150807	25.857	25.836
Target Compounds						
41) L9 AR-1268-1	8.717	7.702	87839198	60916157	260.167	258.370
42) L9 AR-1268-2	8.811	7.767	74121995	54415434	259.881	261.123
43) L9 AR-1268-3	9.040	7.971	63936897	44168578	258.592	252.869
44) L9 AR-1268-4	9.459	8.265	27090074	19140648	250.908	255.703
45) L9 AR-1268-5	9.875	8.565	176.6E6	115.4E6	257.058	252.589



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072194.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 18:09
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:12:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:10:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072195.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 18:25
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/20/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:12:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:10:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	8721682	7491150	4.370m	4.541
2) SA Decachlor...	10.218	8.823	13183771	10651243	4.487	5.599
Target Compounds						
41) L9 AR-1268-1	8.720	7.701	16905769	12836464	50.072m	54.445
42) L9 AR-1268-2	8.814	7.768	13998978	11129718	49.082	53.408
43) L9 AR-1268-3	9.045	7.971	12064238	8963951	48.794	51.320
44) L9 AR-1268-4	9.464	8.265	5112737	3748876	47.354	50.082
45) L9 AR-1268-5	9.878	8.565	33648372	24605969	48.982	53.838

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072195.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 18:25
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/20/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:12:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:10:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 18:42
 Operator : YP\AJ
 Sample : PP051925ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:36:18 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.798	95813360	75129300	47.649	47.004
2) SA Decachlor...	10.216	8.825	78897811	55159923	48.430	52.106
Target Compounds						
3) L1 AR-1016-1	5.656	4.882	34275021	28415442	456.491	462.570
4) L1 AR-1016-2	5.677	4.900	52043054	40409419	486.422	460.300
5) L1 AR-1016-3	5.739	5.077	30972988	22233290	472.402	466.552
6) L1 AR-1016-4	5.837	5.120	26002231	17703522	488.122	465.327
7) L1 AR-1016-5	6.130	5.333	22892152	23633981	468.054	474.345
31) L7 AR-1260-1	7.250	6.367	45618683	37214670	487.406	466.620
32) L7 AR-1260-2	7.503	6.556	68966981	45298985	480.641	446.638
33) L7 AR-1260-3	7.861	6.709	56600304	41944639	497.057	482.664
34) L7 AR-1260-4	8.086	7.179	54236593	34998898	505.455	485.483
35) L7 AR-1260-5	8.405	7.421	116.3E6	86272130	491.560	496.891



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072196.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 18:42
Operator : YP\AJ
Sample : PP051925ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:36:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072197.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 19:14
 Operator : YP\AJ
 Sample : AR1232ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925AR1232

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:37:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.798	99088078	76822954	49.278	48.064
2) SA Decachlor...	10.217	8.824	80208843	53006313	49.234	50.072
Target Compounds						
11) L3 AR-1232-1	4.867	4.169	22860124	20991950	485.266	484.304
12) L3 AR-1232-2	5.392	4.899	11426080	20405975	490.744	461.875
13) L3 AR-1232-3	5.679	5.076	24392843	11207066	487.242	479.281
14) L3 AR-1232-4	5.839	5.159	11984156	9670115	477.474	466.701
15) L3 AR-1232-5	5.929	5.332	8016086	10482162	457.551	452.707



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072197.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 19:14
Operator : YP\AJ
Sample : AR1232ICV500
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925AR1232

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:37:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072198.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 19:47
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:37:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.798	96787545	76508111	48.134	47.867
2) SA Decachlor...	10.213	8.825	81177144	55473433	49.829	52.403
Target Compounds						
16) L4 AR-1242-1	5.655	4.882	30009936	25988486	479.412	498.739
17) L4 AR-1242-2	5.676	4.900	44873042	37823387	517.091	514.630
18) L4 AR-1242-3	5.739	5.077	26969959	20511520	489.776	524.249
19) L4 AR-1242-4	5.836	5.161	22146348	19605635	499.365	521.665
20) L4 AR-1242-5	6.567	5.684	24841123	25039100	487.226	518.863



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072198.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 19:47
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:37:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072199.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 20:19
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:38:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.798	99024800	84569869	49.246	52.910
2) SA Decachlor...	10.212	8.824	83034822	56507020	50.969	53.379
Target Compounds						
21) L5 AR-1248-1	5.656	4.881	23569496	21026666	487.019	485.362
22) L5 AR-1248-2	5.927	5.119	30714899	27827558	506.512	492.721
23) L5 AR-1248-3	6.130	5.161	34359439	29174304	519.181	493.584
24) L5 AR-1248-4	6.529	5.333	44131890	34133080	504.987	490.046
25) L5 AR-1248-5	6.568	5.725	42053839	34467498	506.516	494.232



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072199.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 20:19
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:38:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072200.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 20:52
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:39:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	100.3E6	87142783	49.887	54.520
2) SA Decachlor...	10.214	8.823	83987433	56189590	51.554	53.079
Target Compounds						
26) L6 AR-1254-1	6.506	5.684	43414677	50851971	504.803	508.752
27) L6 AR-1254-2	6.722	5.832	65079227	43668465	498.015	505.967
28) L6 AR-1254-3	7.085	6.235	66155023	67680378	501.163	526.201
29) L6 AR-1254-4	7.367	6.463	61672158	43977452	472.545	515.584
30) L6 AR-1254-5	7.784	6.881	58110470	59298415	522.357	523.137



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072200.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 20:52
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:39:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072201.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 21:25
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:20:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:19:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.798	100.4E6	87828193	52.612	54.920
2) SA Decachlor...	10.216	8.824	147.5E6	102.0E6	52.061	53.249
Target Compounds						
41) L9 AR-1268-1	8.720	7.702	171.0E6	124.0E6	513.323	529.555
42) L9 AR-1268-2	8.812	7.768	144.0E6	108.9E6	514.169	526.775
43) L9 AR-1268-3	9.044	7.971	123.7E6	91750131	512.196	541.712
44) L9 AR-1268-4	9.463	8.266	54675516	40276953	520.976	552.065
45) L9 AR-1268-5	9.877	8.565	344.9E6	244.7E6	509.466	535.577



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072201.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 21:25
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:20:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:19:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Continuing Calib Date: 06/13/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 09:44 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.65	5.65	5.55	5.75	0.00
Aroclor-1016-2	(2)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-3	(3)	5.73	5.74	5.64	5.84	0.01
Aroclor-1016-4	(4)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-5	(5)	6.12	6.13	6.03	6.23	0.01
Aroclor-1260-1	(1)	7.24	7.25	7.15	7.35	0.01
Aroclor-1260-2	(2)	7.49	7.50	7.40	7.60	0.01
Aroclor-1260-3	(3)	7.85	7.86	7.76	7.96	0.01
Aroclor-1260-4	(4)	8.07	8.09	7.99	8.19	0.02
Aroclor-1260-5	(5)	8.39	8.40	8.30	8.50	0.01
Tetrachloro-m-xylene		4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl		10.19	10.21	10.11	10.31	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Continuing Calib Date: 06/13/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 09:44 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.11	5.12	5.02	5.22	0.01
Aroclor-1016-5	(5)	5.32	5.33	5.23	5.43	0.01
Aroclor-1260-1	(1)	6.35	6.36	6.26	6.46	0.01
Aroclor-1260-2	(2)	6.54	6.55	6.45	6.65	0.01
Aroclor-1260-3	(3)	6.69	6.71	6.61	6.81	0.02
Aroclor-1260-4	(4)	7.16	7.18	7.08	7.28	0.02
Aroclor-1260-5	(5)	7.41	7.42	7.32	7.52	0.01
Tetrachloro-m-xylene		3.79	3.79	3.69	3.89	0.00
Decachlorobiphenyl		8.80	8.82	8.72	8.92	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072898.D Time Analyzed: 09:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.648	5.552	5.752	433.430	500.000	-13.3
Aroclor-1016-2	5.669	5.574	5.774	480.470	500.000	-3.9
Aroclor-1016-3	5.731	5.636	5.836	475.320	500.000	-4.9
Aroclor-1016-4	5.829	5.734	5.934	490.530	500.000	-1.9
Aroclor-1016-5	6.121	6.027	6.227	488.970	500.000	-2.2
Aroclor-1260-1	7.239	7.147	7.347	476.790	500.000	-4.6
Aroclor-1260-2	7.492	7.401	7.601	459.490	500.000	-8.1
Aroclor-1260-3	7.850	7.760	7.960	479.490	500.000	-4.1
Aroclor-1260-4	8.074	7.985	8.185	477.040	500.000	-4.6
Aroclor-1260-5	8.393	8.303	8.503	502.580	500.000	0.5
Decachlorobiphenyl	10.191	10.113	10.313	53.180	50.000	6.4
Tetrachloro-m-xylene	4.496	4.396	4.596	50.560	50.000	1.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072898.D Time Analyzed: 09:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.870	4.777	4.977	527.350	500.000	5.5
Aroclor-1016-2	4.888	4.795	4.995	537.000	500.000	7.4
Aroclor-1016-3	5.065	4.973	5.173	534.350	500.000	6.9
Aroclor-1016-4	5.106	5.015	5.215	541.100	500.000	8.2
Aroclor-1016-5	5.320	5.229	5.429	560.140	500.000	12.0
Aroclor-1260-1	6.352	6.264	6.464	543.650	500.000	8.7
Aroclor-1260-2	6.540	6.453	6.653	519.640	500.000	3.9
Aroclor-1260-3	6.692	6.606	6.806	541.410	500.000	8.3
Aroclor-1260-4	7.163	7.077	7.277	544.390	500.000	8.9
Aroclor-1260-5	7.405	7.319	7.519	549.520	500.000	9.9
Decachlorobiphenyl	8.802	8.723	8.923	60.430	50.000	20.9
Tetrachloro-m-xylene	3.790	3.691	3.891	53.970	50.000	7.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
 Data File : PP072898.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2025 09:44
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/16/2025
 Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 12:04:18 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.790	101.7E6	86269124	50.565	53.974
2) SA Decachlor...	10.191	8.802	86631646	63971218	53.177m	60.430m
Target Compounds						
3) L1 AR-1016-1	5.648	4.870	32543719	32394963	433.433	527.352
4) L1 AR-1016-2	5.669	4.888	51405842	47142876	480.466	537.000
5) L1 AR-1016-3	5.731	5.065	31164452	25464219	475.322	534.351
6) L1 AR-1016-4	5.829	5.106	26130667	20586425	490.533	541.103
7) L1 AR-1016-5	6.121	5.320	23915337	27908814	488.974	560.143
31) L7 AR-1260-1	7.239	6.352	44625451	43357786	476.794	543.646
32) L7 AR-1260-2	7.492	6.540	65932493	52702903	459.493	519.639
33) L7 AR-1260-3	7.850	6.692	54600223	47049936	479.493	541.412m
34) L7 AR-1260-4	8.074	7.163	51187446	39245721	477.038	544.392
35) L7 AR-1260-5	8.393	7.405	118.9E6	95410259	502.575	549.522

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 09:44
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 12:04:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Continuing Calib Date: 06/13/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 14:48 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.65	5.65	5.55	5.75	0.00
Aroclor-1016-2	(2)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-3	(3)	5.73	5.74	5.64	5.84	0.01
Aroclor-1016-4	(4)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-5	(5)	6.12	6.13	6.03	6.23	0.01
Aroclor-1260-1	(1)	7.24	7.25	7.15	7.35	0.01
Aroclor-1260-2	(2)	7.49	7.50	7.40	7.60	0.01
Aroclor-1260-3	(3)	7.85	7.86	7.76	7.96	0.01
Aroclor-1260-4	(4)	8.07	8.09	7.99	8.19	0.02
Aroclor-1260-5	(5)	8.39	8.40	8.30	8.50	0.01
Tetrachloro-m-xylene		4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl		10.19	10.21	10.11	10.31	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Continuing Calib Date: 06/13/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 14:48 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.11	5.12	5.02	5.22	0.01
Aroclor-1016-5	(5)	5.32	5.33	5.23	5.43	0.01
Aroclor-1260-1	(1)	6.35	6.36	6.26	6.46	0.01
Aroclor-1260-2	(2)	6.54	6.55	6.45	6.65	0.01
Aroclor-1260-3	(3)	6.69	6.71	6.61	6.81	0.02
Aroclor-1260-4	(4)	7.16	7.18	7.08	7.28	0.02
Aroclor-1260-5	(5)	7.40	7.42	7.32	7.52	0.02
Tetrachloro-m-xylene		3.79	3.79	3.69	3.89	0.00
Decachlorobiphenyl		8.80	8.82	8.72	8.92	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072914.D Time Analyzed: 14:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.648	5.552	5.752	421.180	500.000	-15.8
Aroclor-1016-2	5.670	5.574	5.774	442.270	500.000	-11.5
Aroclor-1016-3	5.732	5.636	5.836	440.440	500.000	-11.9
Aroclor-1016-4	5.829	5.734	5.934	458.080	500.000	-8.4
Aroclor-1016-5	6.121	6.027	6.227	455.740	500.000	-8.9
Aroclor-1260-1	7.238	7.147	7.347	424.190	500.000	-15.2
Aroclor-1260-2	7.492	7.401	7.601	406.910	500.000	-18.6
Aroclor-1260-3	7.850	7.760	7.960	419.380	500.000	-16.1
Aroclor-1260-4	8.073	7.985	8.185	409.950	500.000	-18.0
Aroclor-1260-5	8.391	8.303	8.503	432.570	500.000	-13.5
Decachlorobiphenyl	10.189	10.113	10.313	46.160	50.000	-7.7
Tetrachloro-m-xylene	4.497	4.396	4.596	48.930	50.000	-2.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072914.D Time Analyzed: 14:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.870	4.777	4.977	511.970	500.000	2.4
Aroclor-1016-2	4.889	4.795	4.995	521.930	500.000	4.4
Aroclor-1016-3	5.065	4.973	5.173	527.470	500.000	5.5
Aroclor-1016-4	5.107	5.015	5.215	526.980	500.000	5.4
Aroclor-1016-5	5.320	5.229	5.429	511.470	500.000	2.3
Aroclor-1260-1	6.352	6.264	6.464	520.920	500.000	4.2
Aroclor-1260-2	6.541	6.453	6.653	487.420	500.000	-2.5
Aroclor-1260-3	6.692	6.606	6.806	502.650	500.000	0.5
Aroclor-1260-4	7.162	7.077	7.277	515.240	500.000	3.0
Aroclor-1260-5	7.404	7.319	7.519	503.750	500.000	0.8
Decachlorobiphenyl	8.800	8.723	8.923	55.070	50.000	10.1
Tetrachloro-m-xylene	3.790	3.691	3.891	52.880	50.000	5.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
 Data File : PP072914.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2025 14:48
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/16/2025
 Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 15:10:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.497	3.790	98383530	84515442	48.928	52.876
2) SA Decachlor...	10.189	8.800	75201256	58300562	46.161m	55.073m
Target Compounds						
3) L1 AR-1016-1	5.648	4.870	31623978	31450081	421.183	511.970
4) L1 AR-1016-2	5.670	4.889	47319678	45819527	442.274	521.926
5) L1 AR-1016-3	5.732	5.065	28877713	25136324	440.444	527.470
6) L1 AR-1016-4	5.829	5.107	24401841	20049098	458.079	526.979
7) L1 AR-1016-5	6.121	5.320	22289720	25483621	455.737	511.468
31) L7 AR-1260-1	7.238	6.352	39702111	41545583	424.192	520.923
32) L7 AR-1260-2	7.492	6.541	58387342	49434898	406.910	487.417
33) L7 AR-1260-3	7.850	6.692	47754870	43681495	419.378	502.650m
34) L7 AR-1260-4	8.073	7.162	43988543	37144382	409.948	515.243 #
35) L7 AR-1260-5	8.391	7.404	102.4E6	87463594	432.573	503.753

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072914.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 14:48
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 15:10:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: Portal Partners Tri-Venture	SDG No.: Q2303
Project: Amtrak Sawtooth Bridges 2025	Instrument ID: ECD_P
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 05/19/2025 05/19/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	05/19/2025	09:26	PP072163.D	10.21	4.50
AR1660ICC1000	AR1660ICC1000	05/19/2025	09:42	PP072164.D	10.21	4.50
AR1660ICC750	AR1660ICC750	05/19/2025	09:59	PP072165.D	10.21	4.50
AR1660ICC500	AR1660ICC500	05/19/2025	10:15	PP072166.D	10.21	4.50
AR1660ICC250	AR1660ICC250	05/19/2025	10:32	PP072167.D	10.21	4.50
AR1660ICC050	AR1660ICC050	05/19/2025	10:48	PP072168.D	10.21	4.50
AR1221ICC500	AR1221ICC500	05/19/2025	11:05	PP072169.D	10.21	4.50
AR1232ICC1000	AR1232ICC1000	05/19/2025	11:21	PP072170.D	10.21	4.50
AR1232ICC750	AR1232ICC750	05/19/2025	11:38	PP072171.D	10.22	4.50
AR1232ICC500	AR1232ICC500	05/19/2025	11:54	PP072172.D	10.21	4.50
AR1232ICC250	AR1232ICC250	05/19/2025	12:10	PP072173.D	10.22	4.50
AR1232ICC050	AR1232ICC050	05/19/2025	12:27	PP072174.D	10.21	4.50
AR1242ICC1000	AR1242ICC1000	05/19/2025	12:43	PP072175.D	10.22	4.50
AR1242ICC750	AR1242ICC750	05/19/2025	12:59	PP072176.D	10.22	4.50
AR1242ICC500	AR1242ICC500	05/19/2025	13:15	PP072177.D	10.22	4.51
AR1242ICC250	AR1242ICC250	05/19/2025	13:32	PP072178.D	10.21	4.50
AR1242ICC050	AR1242ICC050	05/19/2025	13:48	PP072179.D	10.22	4.50
AR1248ICC1000	AR1248ICC1000	05/19/2025	14:04	PP072180.D	10.22	4.51
AR1248ICC750	AR1248ICC750	05/19/2025	14:20	PP072181.D	10.21	4.50
AR1248ICC500	AR1248ICC500	05/19/2025	14:37	PP072182.D	10.22	4.50
AR1248ICC250	AR1248ICC250	05/19/2025	14:53	PP072183.D	10.22	4.50
AR1248ICC050	AR1248ICC050	05/19/2025	15:25	PP072184.D	10.21	4.50
AR1254ICC1000	AR1254ICC1000	05/19/2025	15:42	PP072185.D	10.22	4.50
AR1254ICC750	AR1254ICC750	05/19/2025	15:58	PP072186.D	10.22	4.50
AR1254ICC500	AR1254ICC500	05/19/2025	16:14	PP072187.D	10.22	4.50
AR1254ICC250	AR1254ICC250	05/19/2025	16:31	PP072188.D	10.22	4.50
AR1254ICC050	AR1254ICC050	05/19/2025	16:47	PP072189.D	10.22	4.50
AR1262ICC500	AR1262ICC500	05/19/2025	17:03	PP072190.D	10.22	4.50
AR1268ICC1000	AR1268ICC1000	05/19/2025	17:20	PP072191.D	10.22	4.50
AR1268ICC750	AR1268ICC750	05/19/2025	17:36	PP072192.D	10.22	4.51
AR1268ICC500	AR1268ICC500	05/19/2025	17:53	PP072193.D	10.22	4.50
AR1268ICC250	AR1268ICC250	05/19/2025	18:09	PP072194.D	10.21	4.50
AR1268ICC050	AR1268ICC050	05/19/2025	18:25	PP072195.D	10.22	4.50
AR1660CCC500	AR1660CCC500	06/13/2025	09:44	PP072898.D	10.19	4.50
IBLK	IBLK	06/13/2025	11:06	PP072903.D	10.19	4.50
PB168459BL	PB168459BL	06/13/2025	11:22	PP072904.D	10.19	4.50
PB168459BS	PB168459BS	06/13/2025	11:39	PP072905.D	10.19	4.50
B-165-SB01	Q2303-01	06/13/2025	11:55	PP072906.D	10.19	4.49
B-170-SB03	Q2303-02	06/13/2025	12:11	PP072907.D	10.19	4.50
TR-04-06122025MS	Q2305-01MS	06/13/2025	13:16	PP072911.D	10.19	4.50
TR-04-06122025MSD	Q2305-01MSD	06/13/2025	13:33	PP072912.D	10.19	4.50
AR1660CCC500	AR1660CCC500	06/13/2025	14:48	PP072914.D	10.19	4.50

Analytical Sequence

1.BLK	1.BLK	06/13/2025	15:54	PP072918.D	10.19	4.50
-------	-------	------------	-------	------------	-------	------

Analytical Sequence

Client: Portal Partners Tri-Venture	SDG No.: Q2303
Project: Amtrak Sawtooth Bridges 2025	Instrument ID: ECD_P
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 05/19/2025 05/19/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	05/19/2025	09:26	PP072163.D	8.82	3.79
AR1660ICC1000	AR1660ICC1000	05/19/2025	09:42	PP072164.D	8.82	3.79
AR1660ICC750	AR1660ICC750	05/19/2025	09:59	PP072165.D	8.82	3.79
AR1660ICC500	AR1660ICC500	05/19/2025	10:15	PP072166.D	8.82	3.79
AR1660ICC250	AR1660ICC250	05/19/2025	10:32	PP072167.D	8.82	3.79
AR1660ICC050	AR1660ICC050	05/19/2025	10:48	PP072168.D	8.82	3.79
AR1221ICC500	AR1221ICC500	05/19/2025	11:05	PP072169.D	8.82	3.79
AR1232ICC1000	AR1232ICC1000	05/19/2025	11:21	PP072170.D	8.82	3.80
AR1232ICC750	AR1232ICC750	05/19/2025	11:38	PP072171.D	8.82	3.80
AR1232ICC500	AR1232ICC500	05/19/2025	11:54	PP072172.D	8.82	3.80
AR1232ICC250	AR1232ICC250	05/19/2025	12:10	PP072173.D	8.82	3.80
AR1232ICC050	AR1232ICC050	05/19/2025	12:27	PP072174.D	8.82	3.80
AR1242ICC1000	AR1242ICC1000	05/19/2025	12:43	PP072175.D	8.83	3.80
AR1242ICC750	AR1242ICC750	05/19/2025	12:59	PP072176.D	8.83	3.80
AR1242ICC500	AR1242ICC500	05/19/2025	13:15	PP072177.D	8.83	3.80
AR1242ICC250	AR1242ICC250	05/19/2025	13:32	PP072178.D	8.83	3.80
AR1242ICC050	AR1242ICC050	05/19/2025	13:48	PP072179.D	8.82	3.80
AR1248ICC1000	AR1248ICC1000	05/19/2025	14:04	PP072180.D	8.83	3.80
AR1248ICC750	AR1248ICC750	05/19/2025	14:20	PP072181.D	8.83	3.80
AR1248ICC500	AR1248ICC500	05/19/2025	14:37	PP072182.D	8.83	3.80
AR1248ICC250	AR1248ICC250	05/19/2025	14:53	PP072183.D	8.83	3.80
AR1248ICC050	AR1248ICC050	05/19/2025	15:25	PP072184.D	8.82	3.80
AR1254ICC1000	AR1254ICC1000	05/19/2025	15:42	PP072185.D	8.83	3.80
AR1254ICC750	AR1254ICC750	05/19/2025	15:58	PP072186.D	8.83	3.80
AR1254ICC500	AR1254ICC500	05/19/2025	16:14	PP072187.D	8.83	3.80
AR1254ICC250	AR1254ICC250	05/19/2025	16:31	PP072188.D	8.83	3.80
AR1254ICC050	AR1254ICC050	05/19/2025	16:47	PP072189.D	8.83	3.80
AR1262ICC500	AR1262ICC500	05/19/2025	17:03	PP072190.D	8.83	3.80
AR1268ICC1000	AR1268ICC1000	05/19/2025	17:20	PP072191.D	8.83	3.80
AR1268ICC750	AR1268ICC750	05/19/2025	17:36	PP072192.D	8.83	3.80
AR1268ICC500	AR1268ICC500	05/19/2025	17:53	PP072193.D	8.83	3.80
AR1268ICC250	AR1268ICC250	05/19/2025	18:09	PP072194.D	8.82	3.80
AR1268ICC050	AR1268ICC050	05/19/2025	18:25	PP072195.D	8.82	3.80
AR1660CCC500	AR1660CCC500	06/13/2025	09:44	PP072898.D	8.80	3.79
IBLK	IBLK	06/13/2025	11:06	PP072903.D	8.80	3.79
PB168459BL	PB168459BL	06/13/2025	11:22	PP072904.D	8.80	3.79
PB168459BS	PB168459BS	06/13/2025	11:39	PP072905.D	8.80	3.79
B-165-SB01	Q2303-01	06/13/2025	11:55	PP072906.D	8.80	3.79
B-170-SB03	Q2303-02	06/13/2025	12:11	PP072907.D	8.80	3.79
TR-04-06122025MS	Q2305-01MS	06/13/2025	13:16	PP072911.D	8.80	3.79
TR-04-06122025MSD	Q2305-01MSD	06/13/2025	13:33	PP072912.D	8.80	3.79
AR1660CCC500	AR1660CCC500	06/13/2025	14:48	PP072914.D	8.80	3.79

Analytical Sequence

1.BLK	1.BLK	06/13/2025	15:54	PP072918.D	8.80	3.79
-------	-------	------------	-------	------------	------	------

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB168459BS

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Lab Sample ID: PB168459BS Date(s) Analyzed: 06/13/2025 06/13/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072905.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.649	5.599	5.699	118	128	5.32
COLUMN 1	2	5.67	5.62	5.72	131		
	3	5.732	5.682	5.782	128		
	4	5.83	5.78	5.88	133		
	5	6.122	6.072	6.172	127		
	1	4.872	4.822	4.922	134		
COLUMN 2	2	4.89	4.84	4.94	138	135	
	3	5.066	5.016	5.116	136		
	4	5.108	5.058	5.158	136		
	5	5.322	5.272	5.372	132		
Aroclor-1260	1	7.239	7.189	7.289	135	126	17.39
COLUMN 1	2	7.493	7.443	7.543	133		
	3	7.85	7.8	7.9	115		
	4	8.074	8.024	8.124	125		
	5	8.392	8.342	8.442	124		
	1	6.353	6.303	6.403	153		
COLUMN 2	2	6.542	6.492	6.592	153	150	
	3	6.693	6.643	6.743	158		
	4	7.163	7.113	7.213	141		
	5	7.406	7.356	7.456	144		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

B-165-SB01

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Lab Sample ID: Q2303-01 Date(s) Analyzed: 06/13/2025 06/13/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072906.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	6.494	6.444	6.544	194	195	6.93
COLUMN 1	2	6.711	6.661	6.761	208		
	3	7.073	7.023	7.123	182		
	4	7.355	7.305	7.405	199		
	5	7.769	7.719	7.819	192		
	1	5.674	5.624	5.724	234		
COLUMN 2	2	5.821	5.771	5.871	196	209	
	3	6.223	6.173	6.273	197		
	4	6.451	6.401	6.501	209		
	5	6.868	6.818	6.918	208		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

TR-04-06122025MS

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Lab Sample ID: Q2305-01MS Date(s) Analyzed: 06/13/2025 06/13/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072911.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.649	5.599	5.699	109	118	11.2
COLUMN 1	2	5.67	5.62	5.72	111		
	3	5.733	5.683	5.783	113		
	4	5.829	5.779	5.879	143		
	5	6.121	6.071	6.171	112		
COLUMN 2	1	4.871	4.821	4.921	128	132	
	2	4.889	4.839	4.939	134		
	3	5.065	5.015	5.115	132		
	4	5.107	5.057	5.157	134		
	5	5.321	5.271	5.371	131		
Aroclor-1260	1	7.239	7.189	7.289	122	113	13.99
COLUMN 1	2	7.492	7.442	7.542	121		
	3	7.851	7.801	7.901	97.7		
	4	8.074	8.024	8.124	114		
	5	8.393	8.343	8.443	111		
COLUMN 2	1	6.352	6.302	6.402	138	130	
	2	6.541	6.491	6.591	129		
	3	6.692	6.642	6.742	140		
	4	7.163	7.113	7.213	121		
	5	7.405	7.355	7.455	123		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

TR-04-06122025MSD

Contract: PORT06

Lab Code: CHEM Case No.: Q2303 SAS No.: Q2303 SDG NO.: Q2303

Lab Sample ID: Q2305-01MSD Date(s) Analyzed: 06/13/2025 06/13/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072912.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.648	5.598	5.698	107	123	7.81
COLUMN 1	2	5.67	5.62	5.72	114		
	3	5.731	5.681	5.781	113		
	4	5.829	5.779	5.879	142		
	5	6.12	6.07	6.17	139		
	1	4.871	4.821	4.921	135		
COLUMN 2	2	4.889	4.839	4.939	136	133	
	3	5.065	5.015	5.115	133		
	4	5.107	5.057	5.157	135		
	5	5.321	5.271	5.371	129		
Aroclor-1260	1	7.238	7.188	7.288	116	104	18.34
COLUMN 1	2	7.492	7.442	7.542	111		
	3	7.849	7.799	7.899	91.5		
	4	8.072	8.022	8.122	98.6		
	5	8.392	8.342	8.442	104		
	1	6.353	6.303	6.403	135		
COLUMN 2	2	6.542	6.492	6.592	125	125	
	3	6.693	6.643	6.743	133		
	4	7.163	7.113	7.213	114		
	5	7.405	7.355	7.455	116		



QC SAMPLE DATA

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:		
Project:	Amtrak Sawtooth Bridges 2025		Date Received:		
Client Sample ID:	PB168459BL		SDG No.:	Q2303	
Lab Sample ID:	PB168459BL		Matrix:	SOIL	
Analytical Method:	8082A		% Solid:	100	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072904.D	1	06/13/25 08:05	06/13/25 11:22	PB168459

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.1		30 (32) - 150 (144)	106%	SPK: 20
2051-24-3	Decachlorobiphenyl	25.4		30 (32) - 150 (175)	127%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072904.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 11:22
Operator : YP\AJ
Sample : PB168459BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB168459BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 12:08:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.497	3.792	40020638	33746247	19.903	21.113
2) SA Decachlor...	10.190	8.802	33785867	26885290	20.739m	25.397m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072904.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 11:22
Operator : YP\AJ
Sample : PB168459BL
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PB168459BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 12:08:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.497 min
Delta R.T.: -0.001 min
Response: 40020638
Conc: 19.90 ng/ml

Instrument :

ECD_P

ClientSampleId :


PB168459BL

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 06/16/2025

Supervised By :mohammad ahmed 06/17/2025




#1 Tetrachloro-m-xylene

R.T.: 3.792 min
Delta R.T.: -0.002 min
Response: 33746247
Conc: 21.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.190 min
Delta R.T.: -0.022 min
Response: 33785867
Conc: 20.74 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.802 min
Delta R.T.: -0.019 min
Response: 26885290
Conc: 25.40 ng/ml m

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	05/19/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	05/19/25
Client Sample ID:	PIBLK-PP072163.D	SDG No.:	Q2303
Lab Sample ID:	I.BLK-PP072163.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072163.D	1		05/19/25	PP051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.3		70 (60) - 130 (140)	82%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.7		70 (60) - 130 (140)	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072163.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 09:26
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:35:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.495	3.788	32787184	26552565	16.306	16.612
2) SA Decachlor...	10.213	8.822	27208113	18027788	16.701	17.030

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072163.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 09:26
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:35:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.495 min
Delta R.T.: -0.003 min
Response: 32787184
Conc: 16.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.788 min
Delta R.T.: -0.004 min
Response: 26552565
Conc: 16.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.213 min
Delta R.T.: 0.000 min
Response: 27208113
Conc: 16.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.822 min
Delta R.T.: 0.000 min
Response: 18027788
Conc: 17.03 ng/ml

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	06/13/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	06/13/25
Client Sample ID:	PIBLK-PP072903.D	SDG No.:	Q2303
Lab Sample ID:	I.BLK-PP072903.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072903.D	1		06/13/25	PP061325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.4		70 (60) - 130 (140)	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.3		70 (60) - 130 (140)	102%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
 Data File : PP072903.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2025 11:06
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/16/2025
 Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 12:07:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.791	38967063	33197340	19.379	20.770
2) SA Decachlor...	10.190	8.802	33116462	25031097	20.328m	23.645m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 11:06
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 12:07:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.496 min
Delta R.T.: -0.002 min
Response: 38967063
Conc: 19.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025




#1 Tetrachloro-m-xylene

R.T.: 3.791 min
Delta R.T.: -0.002 min
Response: 33197340
Conc: 20.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.190 min
Delta R.T.: -0.022 min
Response: 33116462
Conc: 20.33 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.802 min
Delta R.T.: -0.019 min
Response: 25031097
Conc: 23.65 ng/ml m

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	06/13/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	06/13/25	
Client Sample ID:	PIBLK-PP072918.D		SDG No.:	Q2303	
Lab Sample ID:	I.BLK-PP072918.D		Matrix:	WATER	
Analytical Method:	8082A		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	5030				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072918.D	1		06/13/25	pp061325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.8		70 (60) - 130 (140)	94%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.9		70 (60) - 130 (140)	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
 Data File : PP072918.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2025 15:54
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/16/2025
 Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 22:34:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.498	3.792	37808692	32562628	18.803	20.373
2) SA Decachlor...	10.190	8.802	29208123	23984693	17.929m	22.657m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072918.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 15:54
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
I.BLK

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 22:34:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP072918.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 4.498 min
Delta R.T.: 0.000 min
Response: 37808692
Conc: 18.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025


Response_ Signal: PP072918.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.792 min
Delta R.T.: -0.002 min
Response: 32562628
Conc: 20.37 ng/ml


Response_ Signal: PP072918.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.190 min
Delta R.T.: -0.022 min
Response: 29208123
Conc: 17.93 ng/ml m

Response_ Signal: PP072918.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.802 min
Delta R.T.: -0.019 min
Response: 23984693
Conc: 22.66 ng/ml m

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	PB168459BS	SDG No.:	Q2303
Lab Sample ID:	PB168459BS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072905.D	1	06/13/25 08:05	06/13/25 11:39	PB168459

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	135		3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	150		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.0		30 (32) - 150 (144)	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	25.3		30 (32) - 150 (175)	126%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
 Data File : PP072905.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2025 11:39
 Operator : YP\AJ
 Sample : PB168459BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB168459BS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/16/2025
 Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 12:09:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.498	3.792	40169134	31903952	19.977	19.960
2) SA Decachlor...	10.190	8.802	35021247	26754384	21.497m	25.273m
Target Compounds						
3) L1 AR-1016-1	5.649	4.872	26680767	24789041	355.347	403.536
4) L1 AR-1016-2	5.670	4.890	41958741	36314017	392.168	413.649
5) L1 AR-1016-3	5.732	5.066	25171664	19413373	383.920	407.378
6) L1 AR-1016-4	5.830	5.108	21345981	15484270	400.713	406.995
7) L1 AR-1016-5	6.122	5.322	18695921	19722797	382.258	395.846
31) L7 AR-1260-1	7.239	6.353	37913213	36618592	405.078	459.146
32) L7 AR-1260-2	7.493	6.542	57299372	46492286	399.327	458.404
33) L7 AR-1260-3	7.850	6.693	39291721	41259896	345.055	474.785m#
34) L7 AR-1260-4	8.074	7.163	40158502	30609464	374.255	424.595
35) L7 AR-1260-5	8.392	7.406	88448947	74931980	373.735	431.576

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 11:39
Operator : YP\AJ
Sample : PB168459BS
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PB168459BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 12:09:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	06/12/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	06/12/25
Client Sample ID:	TR-04-06122025MS	SDG No.:	Q2303
Lab Sample ID:	Q2305-01MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	98.2
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072911.D	1	06/13/25 08:05	06/13/25 13:16	PB168459

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	132		4.00	17.3	ug/kg
11104-28-2	Aroclor-1221	4.10	U	4.10	17.3	ug/kg
11141-16-5	Aroclor-1232	3.80	U	3.80	17.3	ug/kg
53469-21-9	Aroclor-1242	4.10	U	4.10	17.3	ug/kg
12672-29-6	Aroclor-1248	6.00	U	6.00	17.3	ug/kg
11097-69-1	Aroclor-1254	3.30	U	3.30	17.3	ug/kg
37324-23-5	Aroclor-1262	5.10	U	5.10	17.3	ug/kg
11100-14-4	Aroclor-1268	3.70	U	3.70	17.3	ug/kg
11096-82-5	Aroclor-1260	130		3.30	17.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.6		30 (32) - 150 (144)	98%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.0		30 (32) - 150 (175)	120%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
 Data File : PP072911.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2025 13:16
 Operator : YP\AJ
 Sample : Q2305-01MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TR-04-06122025MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/16/2025
 Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 15:08:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.791	33234834	31260874	16.528m	19.558
2) SA Decachlor...	10.192	8.803	30916738	25429954	18.978m	24.022m#
Target Compounds						
3) L1 AR-1016-1	5.649	4.871	24121167	23232465	321.257	378.197
4) L1 AR-1016-2	5.670	4.889	35134623	34759090	328.387	395.937
5) L1 AR-1016-3	5.733	5.065	21864376	18552285	333.477	389.308
6) L1 AR-1016-4	5.829	5.107	22423834	14978581	420.947	393.704
7) L1 AR-1016-5	6.121	5.321	16211630	19173658	331.464m	384.824
31) L7 AR-1260-1	7.239	6.352	33593566	32486654	358.926	407.337
32) L7 AR-1260-2	7.492	6.541	51253106	38595759	357.190	380.546
33) L7 AR-1260-3	7.851	6.692	32803804	35825236	288.079	412.247m#
34) L7 AR-1260-4	8.074	7.163	36057354	25712760	336.034	356.671
35) L7 AR-1260-5	8.393	7.405	77142773	62873730	325.962	362.126

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 13:16
Operator : YP\AJ
Sample : Q2305-01MS
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
TR-04-06122025MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 15:08:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	06/12/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	06/12/25
Client Sample ID:	TR-04-06122025MSD	SDG No.:	Q2303
Lab Sample ID:	Q2305-01MSD	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	98.2
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072912.D	1	06/13/25 08:05	06/13/25 13:33	PB168459

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	133		4.00	17.3	ug/kg
11104-28-2	Aroclor-1221	4.10	U	4.10	17.3	ug/kg
11141-16-5	Aroclor-1232	3.80	U	3.80	17.3	ug/kg
53469-21-9	Aroclor-1242	4.10	U	4.10	17.3	ug/kg
12672-29-6	Aroclor-1248	6.00	U	6.00	17.3	ug/kg
11097-69-1	Aroclor-1254	3.30	U	3.30	17.3	ug/kg
37324-23-5	Aroclor-1262	5.10	U	5.10	17.3	ug/kg
11100-14-4	Aroclor-1268	3.70	U	3.70	17.3	ug/kg
11096-82-5	Aroclor-1260	125		3.30	17.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.1		30 (32) - 150 (144)	105%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.2		30 (32) - 150 (175)	111%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
 Data File : PP072912.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2025 13:33
 Operator : YP\AJ
 Sample : Q2305-01MSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TR-04-06122025MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/16/2025
 Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 15:09:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.791	38571068	33642517	19.182	21.048
2) SA Decachlor...	10.192	8.803	28726646	23458074	17.633m	22.159m#
Target Compounds						
3) L1 AR-1016-1	5.648	4.871	23727991	24449219	316.021	398.004 #
4) L1 AR-1016-2	5.670	4.889	36059409	35231377	337.030	401.317
5) L1 AR-1016-3	5.731	5.065	21910989	18712170	334.188	392.663
6) L1 AR-1016-4	5.829	5.107	22267482	15146081	418.012	398.106
7) L1 AR-1016-5	6.120	5.321	20113007	18927275	411.232m	379.879
31) L7 AR-1260-1	7.238	6.353	31951789	31831883	341.384	399.127
32) L7 AR-1260-2	7.492	6.542	47146577	37437875	328.571	369.129
33) L7 AR-1260-3	7.849	6.693	30755476	34196276	270.091	393.502m#
34) L7 AR-1260-4	8.072	7.163	31218931	24357595	290.943m	337.873
35) L7 AR-1260-5	8.392	7.405	72968053	59684207	308.322	343.756

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061325\
Data File : PP072912.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2025 13:33
Operator : YP\AJ
Sample : Q2305-01MSD
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
TR-04-06122025MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/16/2025
Supervised By :mohammad ahmed 06/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 15:09:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PP051925	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC050	PP072168.D	AR-1016-1	yogesh	5/20/2025 7:28:08 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1660ICC050	PP072168.D	AR-1016-2	yogesh	5/20/2025 7:28:08 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1660ICC050	PP072168.D	AR-1016-3	yogesh	5/20/2025 7:28:08 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1660ICC050	PP072168.D	AR-1016-4	yogesh	5/20/2025 7:28:08 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1232ICC050	PP072174.D	AR-1232-3	yogesh	5/21/2025 8:50:30 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1232ICC050	PP072174.D	AR-1232-4	yogesh	5/21/2025 8:50:30 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-1	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-2	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-3	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-4	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-5	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-5 #2	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1248ICC050	PP072184.D	AR-1248-1	yogesh	5/20/2025 7:28:12 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software

Manual Integration Report

Sequence:	PP051925	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1248ICC050	PP072184.D	AR-1248-4 #2	yogesh	5/20/2025 7:28:12 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1254ICC050	PP072189.D	AR-1254-1	yogesh	5/20/2025 7:28:14 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1254ICC050	PP072189.D	AR-1254-4	yogesh	5/20/2025 7:28:14 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1254ICC050	PP072189.D	AR-1254-4 #2	yogesh	5/20/2025 7:28:14 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1254ICC050	PP072189.D	Tetrachloro-m-xylene	yogesh	5/20/2025 7:28:14 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1254ICC050	PP072189.D	Tetrachloro-m-xylene #2	yogesh	5/20/2025 7:28:14 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1268ICC050	PP072195.D	AR-1268-1	yogesh	5/20/2025 7:28:15 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1268ICC050	PP072195.D	Tetrachloro-m-xylene	yogesh	5/20/2025 7:28:15 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software

Manual Integration Report

Sequence:	PP061325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP072898.D	AR-1260-3 #2	yogesh	6/16/2025 8:15:52 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072898.D	Decachlorobiphenyl	yogesh	6/16/2025 8:15:52 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072898.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:15:52 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072899.D	Decachlorobiphenyl	yogesh	6/16/2025 8:15:53 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072899.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:15:53 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072900.D	Decachlorobiphenyl	yogesh	6/16/2025 8:15:56 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072900.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:15:56 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072901.D	Decachlorobiphenyl	yogesh	6/16/2025 8:15:58 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072901.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:15:58 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1268CCC500	PP072902.D	AR-1268-4	yogesh	6/16/2025 8:16:00 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1268CCC500	PP072902.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:00 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1268CCC500	PP072902.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:00 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
I.BLK	PP072903.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:02 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software

Manual Integration Report

Sequence:	PP061325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PP072903.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:02 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
PB168459BL	PP072904.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:03 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
PB168459BL	PP072904.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:03 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
PB168459BS	PP072905.D	AR-1260-3 #2	yogesh	6/16/2025 8:16:05 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
PB168459BS	PP072905.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:05 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
PB168459BS	PP072905.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:05 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2303-01	PP072906.D	AR-1254-5	yogesh	6/16/2025 8:16:07 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2303-01	PP072906.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:07 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2303-01	PP072906.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:07 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2303-02	PP072907.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:08 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2303-02	PP072907.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:08 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2303-02	PP072907.D	Tetrachloro-m-xylene	yogesh	6/16/2025 8:16:08 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2305-01MS	PP072911.D	AR-1016-5	yogesh	6/16/2025 8:16:16 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software

Manual Integration Report

Sequence:	PP061325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q2305-01MS	PP072911.D	AR-1260-3 #2	yogesh	6/16/2025 8:16:16 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2305-01MS	PP072911.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:16 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2305-01MS	PP072911.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:16 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2305-01MS	PP072911.D	Tetrachloro-m-xylene	yogesh	6/16/2025 8:16:16 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2305-01MSD	PP072912.D	AR-1016-5	yogesh	6/16/2025 8:16:17 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2305-01MSD	PP072912.D	AR-1260-3 #2	yogesh	6/16/2025 8:16:17 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2305-01MSD	PP072912.D	AR-1260-4	yogesh	6/16/2025 8:16:17 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2305-01MSD	PP072912.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:17 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
Q2305-01MSD	PP072912.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:17 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072914.D	AR-1260-3 #2	yogesh	6/16/2025 8:16:21 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072914.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:21 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072914.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:21 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072915.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:23 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software

Manual Integration Report

Sequence:	PP061325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP072915.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:23 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072916.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:25 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072916.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:25 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072917.D	AR-1254-5	yogesh	6/16/2025 8:16:26 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072917.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:26 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072917.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:26 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
I.BLK	PP072918.D	Decachlorobiphenyl	yogesh	6/16/2025 8:16:28 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
I.BLK	PP072918.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:16:28 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072928.D	AR-1260-3 #2	yogesh	6/16/2025 8:17:34 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072928.D	Decachlorobiphenyl	yogesh	6/16/2025 8:17:34 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072928.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:17:34 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072929.D	Decachlorobiphenyl	yogesh	6/16/2025 8:17:36 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072929.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:17:36 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software

Manual Integration Report

Sequence:	PP061325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1248CCC500	PP072930.D	AR-1248-1	yogesh	6/16/2025 8:17:38 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072930.D	AR-1248-1 #2	yogesh	6/16/2025 8:17:38 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072930.D	Decachlorobiphenyl	yogesh	6/16/2025 8:17:38 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072930.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:17:38 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072931.D	Decachlorobiphenyl	yogesh	6/16/2025 8:17:40 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072931.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:17:40 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
I.BLK	PP072932.D	Decachlorobiphenyl	yogesh	6/16/2025 8:17:42 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
I.BLK	PP072932.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:17:42 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072943.D	AR-1260-3 #2	yogesh	6/16/2025 8:18:00 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072943.D	Decachlorobiphenyl	yogesh	6/16/2025 8:18:00 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072943.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:18:00 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072944.D	AR-1242-1 #2	yogesh	6/16/2025 8:18:03 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072944.D	Decachlorobiphenyl	yogesh	6/16/2025 8:18:03 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software

Manual Integration Report

Sequence:	PP061325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP072944.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:18:03 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072945.D	AR-1248-1 #2	yogesh	6/16/2025 8:18:05 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072945.D	Decachlorobiphenyl	yogesh	6/16/2025 8:18:05 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072945.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:18:05 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072946.D	AR-1254-1	yogesh	6/16/2025 8:18:07 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072946.D	AR-1254-5	yogesh	6/16/2025 8:18:07 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072946.D	Decachlorobiphenyl	yogesh	6/16/2025 8:18:07 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072946.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:18:07 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
I.BLK	PP072947.D	Decachlorobiphenyl	yogesh	6/16/2025 8:18:08 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
I.BLK	PP072947.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:18:08 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072952.D	AR-1016-5 #2	yogesh	6/16/2025 8:18:17 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072952.D	AR-1260-3 #2	yogesh	6/16/2025 8:18:17 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1660CCC500	PP072952.D	Decachlorobiphenyl	yogesh	6/16/2025 8:18:17 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software

Manual Integration Report

Sequence:	PP061325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP072952.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:18:17 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072953.D	AR-1242-1 #2	yogesh	6/16/2025 8:18:19 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072953.D	AR-1242-5 #2	yogesh	6/16/2025 8:18:19 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072953.D	Decachlorobiphenyl	yogesh	6/16/2025 8:18:19 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1242CCC500	PP072953.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:18:19 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072954.D	AR-1248-1	yogesh	6/16/2025 8:18:21 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072954.D	AR-1248-1 #2	yogesh	6/16/2025 8:18:21 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072954.D	AR-1248-4 #2	yogesh	6/16/2025 8:18:21 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072954.D	Decachlorobiphenyl	yogesh	6/16/2025 8:18:21 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1248CCC500	PP072954.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:18:21 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072955.D	AR-1254-4 #2	yogesh	6/16/2025 8:18:23 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072955.D	Decachlorobiphenyl	yogesh	6/16/2025 8:18:23 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
AR1254CCC500	PP072955.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:18:23 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PP061325

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PP072956.D	Decachlorobiphenyl	yogesh	6/16/2025 8:18:25 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software
I.BLK	PP072956.D	Decachlorobiphenyl #2	yogesh	6/16/2025 8:18:25 AM	mohammad	6/17/2025 3:41:07	Peak Integrated by Software

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344, PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359, PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP072162.D	19 May 2025 09:09	YP\AJ	Ok
2	I.BLK	PP072163.D	19 May 2025 09:26	YP\AJ	Ok
3	AR1660ICC1000	PP072164.D	19 May 2025 09:42	YP\AJ	Ok
4	AR1660ICC750	PP072165.D	19 May 2025 09:59	YP\AJ	Ok
5	AR1660ICC500	PP072166.D	19 May 2025 10:15	YP\AJ	Ok
6	AR1660ICC250	PP072167.D	19 May 2025 10:32	YP\AJ	Ok
7	AR1660ICC050	PP072168.D	19 May 2025 10:48	YP\AJ	Ok,M
8	AR1221ICC500	PP072169.D	19 May 2025 11:05	YP\AJ	Ok
9	AR1232ICC1000	PP072170.D	19 May 2025 11:21	YP\AJ	Ok
10	AR1232ICC750	PP072171.D	19 May 2025 11:38	YP\AJ	Ok
11	AR1232ICC500	PP072172.D	19 May 2025 11:54	YP\AJ	Ok
12	AR1232ICC250	PP072173.D	19 May 2025 12:10	YP\AJ	Ok
13	AR1232ICC050	PP072174.D	19 May 2025 12:27	YP\AJ	Ok,M
14	AR1242ICC1000	PP072175.D	19 May 2025 12:43	YP\AJ	Ok
15	AR1242ICC750	PP072176.D	19 May 2025 12:59	YP\AJ	Ok
16	AR1242ICC500	PP072177.D	19 May 2025 13:15	YP\AJ	Ok
17	AR1242ICC250	PP072178.D	19 May 2025 13:32	YP\AJ	Ok
18	AR1242ICC050	PP072179.D	19 May 2025 13:48	YP\AJ	Ok,M
19	AR1248ICC1000	PP072180.D	19 May 2025 14:04	YP\AJ	Ok
20	AR1248ICC750	PP072181.D	19 May 2025 14:20	YP\AJ	Ok
21	AR1248ICC500	PP072182.D	19 May 2025 14:37	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1248ICC250	PP072183.D	19 May 2025 14:53	YP\AJ	Ok
23	AR1248ICC050	PP072184.D	19 May 2025 15:25	YP\AJ	Ok,M
24	AR1254ICC1000	PP072185.D	19 May 2025 15:42	YP\AJ	Ok
25	AR1254ICC750	PP072186.D	19 May 2025 15:58	YP\AJ	Ok
26	AR1254ICC500	PP072187.D	19 May 2025 16:14	YP\AJ	Ok
27	AR1254ICC250	PP072188.D	19 May 2025 16:31	YP\AJ	Ok
28	AR1254ICC050	PP072189.D	19 May 2025 16:47	YP\AJ	Ok,M
29	AR1262ICC500	PP072190.D	19 May 2025 17:03	YP\AJ	Ok
30	AR1268ICC1000	PP072191.D	19 May 2025 17:20	YP\AJ	Ok
31	AR1268ICC750	PP072192.D	19 May 2025 17:36	YP\AJ	Ok
32	AR1268ICC500	PP072193.D	19 May 2025 17:53	YP\AJ	Ok
33	AR1268ICC250	PP072194.D	19 May 2025 18:09	YP\AJ	Ok
34	AR1268ICC050	PP072195.D	19 May 2025 18:25	YP\AJ	Ok,M
35	PP051925ICV500	PP072196.D	19 May 2025 18:42	YP\AJ	Ok
36	AR1232ICV500	PP072197.D	19 May 2025 19:14	YP\AJ	Ok
37	AR1242ICV500	PP072198.D	19 May 2025 19:47	YP\AJ	Ok
38	AR1248ICV500	PP072199.D	19 May 2025 20:19	YP\AJ	Ok
39	AR1254ICV500	PP072200.D	19 May 2025 20:52	YP\AJ	Ok
40	AR1268ICV500	PP072201.D	19 May 2025 21:25	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP061325

Review By	yogesh	Review On	6/13/2025 1:22:51 PM
Supervise By	mohammad	Supervise On	6/17/2025 3:41:07 AM
SubDirectory	PP061325	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP072897.D	13 Jun 2025 09:28	YP\AJ	Ok
2	AR1660CCC500	PP072898.D	13 Jun 2025 09:44	YP\AJ	Ok,M
3	AR1242CCC500	PP072899.D	13 Jun 2025 10:01	YP\AJ	Ok,M
4	AR1248CCC500	PP072900.D	13 Jun 2025 10:17	YP\AJ	Ok,M
5	AR1254CCC500	PP072901.D	13 Jun 2025 10:33	YP\AJ	Ok,M
6	AR1268CCC500	PP072902.D	13 Jun 2025 10:50	YP\AJ	Ok,M
7	I.BLK	PP072903.D	13 Jun 2025 11:06	YP\AJ	Ok,M
8	PB168459BL	PP072904.D	13 Jun 2025 11:22	YP\AJ	Ok,M
9	PB168459BS	PP072905.D	13 Jun 2025 11:39	YP\AJ	Ok,M
10	Q2303-01	PP072906.D	13 Jun 2025 11:55	YP\AJ	Ok,M
11	Q2303-02	PP072907.D	13 Jun 2025 12:11	YP\AJ	Ok,M
12	Q2298-01	PP072908.D	13 Jun 2025 12:28	YP\AJ	Ok,M
13	Q2301-02	PP072909.D	13 Jun 2025 12:44	YP\AJ	Ok,M
14	Q2305-01	PP072910.D	13 Jun 2025 13:00	YP\AJ	Ok,M
15	Q2305-01MS	PP072911.D	13 Jun 2025 13:16	YP\AJ	Ok,M
16	Q2305-01MSD	PP072912.D	13 Jun 2025 13:33	YP\AJ	Ok,M
17	Q2307-01	PP072913.D	13 Jun 2025 13:49	YP\AJ	Ok,M
18	AR1660CCC500	PP072914.D	13 Jun 2025 14:48	YP\AJ	Ok,M
19	AR1242CCC500	PP072915.D	13 Jun 2025 15:05	YP\AJ	Ok,M
20	AR1248CCC500	PP072916.D	13 Jun 2025 15:21	YP\AJ	Ok,M
21	AR1254CCC500	PP072917.D	13 Jun 2025 15:37	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP061325

Review By	yogesh	Review On	6/13/2025 1:22:51 PM
Supervise By	mohammad	Supervise On	6/17/2025 3:41:07 AM
SubDirectory	PP061325	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	I.BLK	PP072918.D	13 Jun 2025 15:54	YP\AJ	Ok,M
23	PB168461BL	PP072919.D	13 Jun 2025 16:10	YP\AJ	Ok,M
24	PB168461BS	PP072920.D	13 Jun 2025 16:26	YP\AJ	Ok,M
25	Q2308-01	PP072921.D	13 Jun 2025 16:42	YP\AJ	Ok,M
26	Q2310-01	PP072922.D	13 Jun 2025 16:59	YP\AJ	Ok,M
27	Q2304-01	PP072923.D	13 Jun 2025 17:15	YP\AJ	Ok,M
28	Q2304-02	PP072924.D	13 Jun 2025 17:31	YP\AJ	Ok,M
29	Q2304-03	PP072925.D	13 Jun 2025 17:48	YP\AJ	Ok,M
30	Q2304-04	PP072926.D	13 Jun 2025 18:04	YP\AJ	Ok,M
31	Q2304-05	PP072927.D	13 Jun 2025 18:20	YP\AJ	Ok,M
32	AR1660CCC500	PP072928.D	13 Jun 2025 19:31	YP\AJ	Ok,M
33	AR1242CCC500	PP072929.D	13 Jun 2025 20:03	YP\AJ	Ok,M
34	AR1248CCC500	PP072930.D	13 Jun 2025 20:20	YP\AJ	Ok,M
35	AR1254CCC500	PP072931.D	13 Jun 2025 20:36	YP\AJ	Ok,M
36	I.BLK	PP072932.D	13 Jun 2025 20:52	YP\AJ	Ok,M
37	PB168480BL	PP072933.D	13 Jun 2025 21:08	YP\AJ	Ok,M
38	PB168480BS	PP072934.D	13 Jun 2025 21:25	YP\AJ	Ok,M
39	Q2326-01	PP072935.D	13 Jun 2025 21:41	YP\AJ	Ok,M
40	Q2326-02	PP072936.D	13 Jun 2025 21:57	YP\AJ	Ok,M
41	Q2326-03	PP072937.D	13 Jun 2025 22:13	YP\AJ	Ok,M
42	Q2326-04	PP072938.D	13 Jun 2025 22:30	YP\AJ	Ok,M
43	Q2326-05	PP072939.D	13 Jun 2025 22:46	YP\AJ	Ok,M
44	Q2326-06	PP072940.D	13 Jun 2025 23:02	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP061325

Review By	yogesh	Review On	6/13/2025 1:22:51 PM
Supervise By	mohammad	Supervise On	6/17/2025 3:41:07 AM
SubDirectory	PP061325	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	Q2326-07	PP072941.D	13 Jun 2025 23:19	YP\AJ	Ok,M
46	Q2326-08	PP072942.D	13 Jun 2025 23:35	YP\AJ	Ok,M
47	AR1660CCC500	PP072943.D	14 Jun 2025 02:07	YP\AJ	Ok,M
48	AR1242CCC500	PP072944.D	14 Jun 2025 02:40	YP\AJ	Ok,M
49	AR1248CCC500	PP072945.D	14 Jun 2025 02:56	YP\AJ	Ok,M
50	AR1254CCC500	PP072946.D	14 Jun 2025 03:12	YP\AJ	Ok,M
51	I.BLK	PP072947.D	14 Jun 2025 03:28	YP\AJ	Ok,M
52	Q2327-01	PP072948.D	14 Jun 2025 03:45	YP\AJ	Ok,M
53	Q2327-02	PP072949.D	14 Jun 2025 04:01	YP\AJ	Ok,M
54	Q2327-03	PP072950.D	14 Jun 2025 04:17	YP\AJ	Ok,M
55	Q2327-04	PP072951.D	14 Jun 2025 04:34	YP\AJ	Ok,M
56	AR1660CCC500	PP072952.D	14 Jun 2025 08:24	YP\AJ	Ok,M
57	AR1242CCC500	PP072953.D	14 Jun 2025 08:40	YP\AJ	Ok,M
58	AR1248CCC500	PP072954.D	14 Jun 2025 08:56	YP\AJ	Ok,M
59	AR1254CCC500	PP072955.D	14 Jun 2025 09:12	YP\AJ	Ok,M
60	I.BLK	PP072956.D	14 Jun 2025 09:29	YP\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072162.D	19 May 2025 09:09		YP\AJ	Ok
2	I.BLK	I.BLK	PP072163.D	19 May 2025 09:26		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP072164.D	19 May 2025 09:42		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP072165.D	19 May 2025 09:59		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP072166.D	19 May 2025 10:15		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP072167.D	19 May 2025 10:32		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP072168.D	19 May 2025 10:48		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP072169.D	19 May 2025 11:05		YP\AJ	Ok
9	AR1232ICC1000	AR1232ICC1000	PP072170.D	19 May 2025 11:21		YP\AJ	Ok
10	AR1232ICC750	AR1232ICC750	PP072171.D	19 May 2025 11:38		YP\AJ	Ok
11	AR1232ICC500	AR1232ICC500	PP072172.D	19 May 2025 11:54		YP\AJ	Ok
12	AR1232ICC250	AR1232ICC250	PP072173.D	19 May 2025 12:10		YP\AJ	Ok
13	AR1232ICC050	AR1232ICC050	PP072174.D	19 May 2025 12:27		YP\AJ	Ok,M
14	AR1242ICC1000	AR1242ICC1000	PP072175.D	19 May 2025 12:43		YP\AJ	Ok
15	AR1242ICC750	AR1242ICC750	PP072176.D	19 May 2025 12:59		YP\AJ	Ok
16	AR1242ICC500	AR1242ICC500	PP072177.D	19 May 2025 13:15		YP\AJ	Ok
17	AR1242ICC250	AR1242ICC250	PP072178.D	19 May 2025 13:32		YP\AJ	Ok
18	AR1242ICC050	AR1242ICC050	PP072179.D	19 May 2025 13:48		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC1000	AR1248ICC1000	PP072180.D	19 May 2025 14:04		YPIAJ	Ok
20	AR1248ICC750	AR1248ICC750	PP072181.D	19 May 2025 14:20		YPIAJ	Ok
21	AR1248ICC500	AR1248ICC500	PP072182.D	19 May 2025 14:37		YPIAJ	Ok
22	AR1248ICC250	AR1248ICC250	PP072183.D	19 May 2025 14:53		YPIAJ	Ok
23	AR1248ICC050	AR1248ICC050	PP072184.D	19 May 2025 15:25		YPIAJ	Ok,M
24	AR1254ICC1000	AR1254ICC1000	PP072185.D	19 May 2025 15:42		YPIAJ	Ok
25	AR1254ICC750	AR1254ICC750	PP072186.D	19 May 2025 15:58		YPIAJ	Ok
26	AR1254ICC500	AR1254ICC500	PP072187.D	19 May 2025 16:14		YPIAJ	Ok
27	AR1254ICC250	AR1254ICC250	PP072188.D	19 May 2025 16:31		YPIAJ	Ok
28	AR1254ICC050	AR1254ICC050	PP072189.D	19 May 2025 16:47		YPIAJ	Ok,M
29	AR1262ICC500	AR1262ICC500	PP072190.D	19 May 2025 17:03		YPIAJ	Ok
30	AR1268ICC1000	AR1268ICC1000	PP072191.D	19 May 2025 17:20		YPIAJ	Ok
31	AR1268ICC750	AR1268ICC750	PP072192.D	19 May 2025 17:36		YPIAJ	Ok
32	AR1268ICC500	AR1268ICC500	PP072193.D	19 May 2025 17:53		YPIAJ	Ok
33	AR1268ICC250	AR1268ICC250	PP072194.D	19 May 2025 18:09		YPIAJ	Ok
34	AR1268ICC050	AR1268ICC050	PP072195.D	19 May 2025 18:25		YPIAJ	Ok,M
35	PP051925ICV500	ICVPP051925	PP072196.D	19 May 2025 18:42		YPIAJ	Ok
36	AR1232ICV500	ICVPP051925AR1232	PP072197.D	19 May 2025 19:14		YPIAJ	Ok
37	AR1242ICV500	ICVPP051925AR1242	PP072198.D	19 May 2025 19:47		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

38	AR1248ICV500	ICVPP051925AR1248	PP072199.D	19 May 2025 20:19		YPIAJ	Ok
39	AR1254ICV500	ICVPP051925AR1254	PP072200.D	19 May 2025 20:52		YPIAJ	Ok
40	AR1268ICV500	ICVPP051925AR1268	PP072201.D	19 May 2025 21:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP061325

Review By	yogesh	Review On	6/13/2025 1:22:51 PM
Supervise By	mohammad	Supervise On	6/17/2025 3:41:07 AM
SubDirectory	PP061325	HP Acquire Method	HP Processing Method PP051925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072897.D	13 Jun 2025 09:28		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP072898.D	13 Jun 2025 09:44	DCB high in 2nd column	YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP072899.D	13 Jun 2025 10:01		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP072900.D	13 Jun 2025 10:17		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP072901.D	13 Jun 2025 10:33		YP\AJ	Ok,M
6	AR1268CCC500	AR1268CCC500	PP072902.D	13 Jun 2025 10:50		YP\AJ	Ok,M
7	I.BLK	I.BLK	PP072903.D	13 Jun 2025 11:06		YP\AJ	Ok,M
8	PB168459BL	PB168459BL	PP072904.D	13 Jun 2025 11:22		YP\AJ	Ok,M
9	PB168459BS	PB168459BS	PP072905.D	13 Jun 2025 11:39		YP\AJ	Ok,M
10	Q2303-01	B-165-SB01	PP072906.D	13 Jun 2025 11:55	AR1254 Hit	YP\AJ	Ok,M
11	Q2303-02	B-170-SB03	PP072907.D	13 Jun 2025 12:11		YP\AJ	Ok,M
12	Q2298-01	AU-05-061125	PP072908.D	13 Jun 2025 12:28		YP\AJ	Ok,M
13	Q2301-02	WC-URBAN-FILL-C	PP072909.D	13 Jun 2025 12:44		YP\AJ	Ok,M
14	Q2305-01	TR-04-06122025	PP072910.D	13 Jun 2025 13:00		YP\AJ	Ok,M
15	Q2305-01MS	TR-04-06122025MS	PP072911.D	13 Jun 2025 13:16		YP\AJ	Ok,M
16	Q2305-01MSD	TR-04-06122025MSD	PP072912.D	13 Jun 2025 13:33		YP\AJ	Ok,M
17	Q2307-01	LINDEN-SAA	PP072913.D	13 Jun 2025 13:49		YP\AJ	Ok,M
18	AR1660CCC500	AR1660CCC500	PP072914.D	13 Jun 2025 14:48		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP061325

Review By	yogesh	Review On	6/13/2025 1:22:51 PM
Supervise By	mohammad	Supervise On	6/17/2025 3:41:07 AM
SubDirectory	PP061325	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1242CCC500	AR1242CCC500	PP072915.D	13 Jun 2025 15:05		YPIAJ	Ok,M
20	AR1248CCC500	AR1248CCC500	PP072916.D	13 Jun 2025 15:21		YPIAJ	Ok,M
21	AR1254CCC500	AR1254CCC500	PP072917.D	13 Jun 2025 15:37	1st peak low in 1st column	YPIAJ	Ok,M
22	I.BLK	I.BLK	PP072918.D	13 Jun 2025 15:54		YPIAJ	Ok,M
23	PB168461BL	PB168461BL	PP072919.D	13 Jun 2025 16:10		YPIAJ	Ok,M
24	PB168461BS	PB168461BS	PP072920.D	13 Jun 2025 16:26		YPIAJ	Ok,M
25	Q2308-01	EO-02-06122025	PP072921.D	13 Jun 2025 16:42		YPIAJ	Ok,M
26	Q2310-01	TP-7	PP072922.D	13 Jun 2025 16:59		YPIAJ	Ok,M
27	Q2304-01	RBR200057-1	PP072923.D	13 Jun 2025 17:15	AR1248 Hit	YPIAJ	Ok,M
28	Q2304-02	RBR200057-2	PP072924.D	13 Jun 2025 17:31	AR1248 Hit	YPIAJ	Ok,M
29	Q2304-03	VNJ239-3	PP072925.D	13 Jun 2025 17:48	AR1248 Hit	YPIAJ	Ok,M
30	Q2304-04	VNJ239-4	PP072926.D	13 Jun 2025 18:04	AR1248 Hit	YPIAJ	Ok,M
31	Q2304-05	VNJ239-5	PP072927.D	13 Jun 2025 18:20	AR1248 Hit , Confirm corrected typo	YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP072928.D	13 Jun 2025 19:31		YPIAJ	Ok,M
33	AR1242CCC500	AR1242CCC500	PP072929.D	13 Jun 2025 20:03		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP072930.D	13 Jun 2025 20:20		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP072931.D	13 Jun 2025 20:36	DCB high in 2nd column	YPIAJ	Ok,M
36	I.BLK	I.BLK	PP072932.D	13 Jun 2025 20:52		YPIAJ	Ok,M
37	PB168480BL	PB168480BL	PP072933.D	13 Jun 2025 21:08		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP061325

Review By	yogesh	Review On	6/13/2025 1:22:51 PM
Supervise By	mohammad	Supervise On	6/17/2025 3:41:07 AM
SubDirectory	PP061325	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24332,PP24347,PP24352,PP24357 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

38	PB168480BS	PB168480BS	PP072934.D	13 Jun 2025 21:25		YPIAJ	Ok,M
39	Q2326-01	BC274770-1-1	PP072935.D	13 Jun 2025 21:41	DCB high in 1st column	YPIAJ	Ok,M
40	Q2326-02	BC274770-1-2	PP072936.D	13 Jun 2025 21:57	DCB high in 1st column	YPIAJ	Ok,M
41	Q2326-03	BC274770-2-1	PP072937.D	13 Jun 2025 22:13	DCB high in 1st column	YPIAJ	Ok,M
42	Q2326-04	BC274770-2-2	PP072938.D	13 Jun 2025 22:30		YPIAJ	Ok,M
43	Q2326-05	BC274770-3-1	PP072939.D	13 Jun 2025 22:46	DCB high in 1st column	YPIAJ	Ok,M
44	Q2326-06	BC274770-3-2	PP072940.D	13 Jun 2025 23:02		YPIAJ	Ok,M
45	Q2326-07	ECA345N-1-1	PP072941.D	13 Jun 2025 23:19		YPIAJ	Ok,M
46	Q2326-08	ECA345N-1-2	PP072942.D	13 Jun 2025 23:35		YPIAJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PP072943.D	14 Jun 2025 02:07		YPIAJ	Ok,M
48	AR1242CCC500	AR1242CCC500	PP072944.D	14 Jun 2025 02:40		YPIAJ	Ok,M
49	AR1248CCC500	AR1248CCC500	PP072945.D	14 Jun 2025 02:56		YPIAJ	Ok,M
50	AR1254CCC500	AR1254CCC500	PP072946.D	14 Jun 2025 03:12	DCB high in 2nd column	YPIAJ	Ok,M
51	I.BLK	I.BLK	PP072947.D	14 Jun 2025 03:28		YPIAJ	Ok,M
52	Q2327-01	HEH700H-1-1	PP072948.D	14 Jun 2025 03:45	AR1254 Hit	YPIAJ	Ok,M
53	Q2327-02	HEH700H-1-2	PP072949.D	14 Jun 2025 04:01		YPIAJ	Ok,M
54	Q2327-03	HEH700H-2-1	PP072950.D	14 Jun 2025 04:17	AR1254 Hit	YPIAJ	Ok,M
55	Q2327-04	HEH700H-2-2	PP072951.D	14 Jun 2025 04:34	AR1254 Hit	YPIAJ	Ok,M
56	AR1660CCC500	AR1660CCC500	PP072952.D	14 Jun 2025 08:24		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP061325

Review By	yogesh	Review On	6/13/2025 1:22:51 PM
Supervise By	mohammad	Supervise On	6/17/2025 3:41:07 AM
SubDirectory	PP061325	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

57	AR1242CCC500	AR1242CCC500	PP072953.D	14 Jun 2025 08:40		YPIAJ	Ok,M
58	AR1248CCC500	AR1248CCC500	PP072954.D	14 Jun 2025 08:56		YPIAJ	Ok,M
59	AR1254CCC500	AR1254CCC500	PP072955.D	14 Jun 2025 09:12		YPIAJ	Ok,M
60	I.BLK	I.BLK	PP072956.D	14 Jun 2025 09:29		YPIAJ	Ok,M

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/13/2025

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 06/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 06/13/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136126

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2287-01	CONCRETE-SLAB	1	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
Q2296-01	WC-1	2	1.18	10.56	11.74	9.74	81.1	
Q2296-02	WC-1-EPH	3	1.18	10.23	11.41	9.55	81.8	
Q2296-03	WC-1-VOC	4	1.19	10.57	11.76	9.63	79.8	
Q2296-05	WC-2	5	1.19	10.58	11.77	10.52	88.2	
Q2296-06	WC-2-EPH	6	1.19	10.30	11.49	10.5	90.4	
Q2296-07	WC-2-VOC	7	1.19	10.21	11.4	10.62	92.4	
Q2296-09	WC-3	8	1.15	10.78	11.93	10.33	85.2	
Q2296-10	WC-3-EPH	9	1.19	10.60	11.79	10.1	84.1	
Q2296-11	WC-3-VOC	10	1.18	10.81	11.99	10.36	84.9	
Q2296-13	WC-4	11	1.15	10.77	11.92	9.95	81.7	
Q2296-14	WC-4-EPH	12	1.13	10.48	11.61	9.88	83.5	
Q2296-15	WC-4-VOC	13	1.14	10.72	11.86	9.71	79.9	
Q2296-17	WC-5	14	1.17	10.49	11.66	10.63	90.2	
Q2296-18	WC-5-EPH	15	1.17	10.50	11.67	10.32	87.1	
Q2296-19	WC-5-VOC	16	1.15	10.83	11.98	10.97	90.7	
Q2296-21	WC-6	19	1.13	10.37	11.5	10.2	87.5	
Q2296-22	WC-6-EPH	17	1.14	10.44	11.58	10.65	91.1	
Q2296-23	WC-6-VOC	18	1.17	10.82	11.99	10.36	84.9	
Q2297-01	TP-3	20	1.15	10.96	12.11	10.2	82.6	
Q2297-02	TP-3-EPH	21	1.14	10.85	11.99	11.04	91.2	
Q2297-03	TP-3-VOC	22	1.12	10.16	11.28	10.3	90.4	
Q2298-01	AU-05-061125	23	1.15	11.63	12.78	11.7	90.7	
Q2298-02	AU-05-061125	24	1.15	11.63	12.78	11.7	90.7	
Q2301-02	WC-URBAN-FILL-C	25	1.16	10.37	11.53	9.77	83.0	
Q2303-01	B-165-SB01	26	1.14	10.74	11.88	10.12	83.6	
Q2303-02	B-170-SB03	27	1.13	10.37	11.5	10.74	92.7	
Q2304-01	RBR200057-1	28	1.00	1.00	2.00	2.00	100.0	wipe sample



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/13/2025

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 06/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 06/13/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136126

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2304-02	RBR200057-2	29	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2304-03	VNJ239-3	30	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2304-04	VNJ239-4	31	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2304-05	VNJ239-5	32	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2305-01	TR-04-06122025	33	1.18	10.17	11.35	11.17	98.2	
Q2305-02	TR-04-06122025-E2	34	1.13	10.69	11.82	11.24	94.6	
Q2307-01	LINDEN-SAA	35	1.14	10.54	11.68	8.87	73.3	
Q2307-02	LINDEN-SAA	36	1.13	10.70	11.83	8.87	72.3	
Q2308-01	EO-02-06122025	37	1.12	10.24	11.36	10.38	90.4	
Q2308-02	EO-02-06122025-E2	38	1.19	10.75	11.94	10.98	91.1	
Q2310-01	TP-7	39	1.18	10.42	11.6	10.42	88.7	
Q2310-02	TP-7-EPH	40	1.15	11.29	12.44	11.08	88.0	
Q2310-03	TP-7-VOC	41	1.12	10.91	12.03	10.99	90.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

136126

WorkList Name : %1-061225

WorkList ID : 190126

Department : Wet-Chemistry

Date : 06-12-2025 08:16:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2287-01	CONCRETE-SLAB	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/11/2025	Chemtech -SO
Q2296-01	WC-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-02	WC-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-03	WC-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-05	WC-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-06	WC-2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-07	WC-2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-09	WC-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-10	WC-3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-11	WC-3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-13	WC-4	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-14	WC-4-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-15	WC-4-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-17	WC-5	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-18	WC-5-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-19	WC-5-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-21	WC-6	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-22	WC-6-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2296-23	WC-6-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	N22	06/11/2025	Chemtech -SO
Q2297-01	TP-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	N11	06/11/2025	Chemtech -SO
Q2297-02	TP-3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	N11	06/11/2025	Chemtech -SO

Date/Time 06-12-25 15:20

Raw Sample Received by: 88 woc

Raw Sample Relinquished by: CP SM

Date/Time

06-12-25

Raw Sample Received by:

Raw Sample Relinquished by:

WORKLIST(Hardcopy Internal Chain)

136126

WorkList Name : %1-061225

WorkList ID : 190126

Department : Wet-Chemistry

Date : 06-12-2025 08:16:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2297-03	TP-3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	N11	06/11/2025	Chemtech -SO
Q2298-01	AU-05-061125	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	06/11/2025	Chemtech -SO
Q2298-02	AU-05-061125	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	06/11/2025	Chemtech -SO
Q2301-02	WC-URBAN-FILL-C	Solid	Percent Solids	Cool 4 deg C	ENTA05	D41	06/11/2025	Chemtech -SO
Q2303-01	B-165-SB01	Solid	Percent Solids	Cool 4 deg C	PORT06	D41	06/11/2025	Chemtech -SO
Q2303-02	B-170-SB03	Solid	Percent Solids	Cool 4 deg C	PORT06	D41	06/11/2025	Chemtech -SO
Q2304-01	RBR200057-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	06/11/2025	Chemtech -SO
Q2304-02	RBR200057-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	06/12/2025	Chemtech -SO
Q2304-03	VNJ239-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	06/12/2025	Chemtech -SO
Q2304-04	VNJ239-4	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	06/12/2025	Chemtech -SO
Q2304-05	VNJ239-5	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	06/12/2025	Chemtech -SO
Q2305-01	TR-04-06122025	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	06/12/2025	Chemtech -SO
Q2305-02	TR-04-06122025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	06/12/2025	Chemtech -SO
Q2307-01	LINDEN-SAA	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/12/2025	Chemtech -SO
Q2307-02	LINDEN-SAA	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	06/12/2025	Chemtech -SO
Q2308-01	EO-02-06122025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D51	06/12/2025	Chemtech -SO
Q2308-02	EO-02-06122025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D51	06/12/2025	Chemtech -SO
Q2310-01	TP-7	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/12/2025	Chemtech -SO
Q2310-02	TP-7-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/12/2025	Chemtech -SO
Q2310-03	TP-7-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	06/12/2025	Chemtech -SO

Date/Time 06/12/25 15:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/12/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

SOP ID: M3541-ASE Extraction-14

Clean Up SOP #: Acid Cleanup

Extraction Start Date : 06/13/2025

Matrix : Solid

Extraction Start Time : 08:05

Weigh By: EH

Extraction By: RJ

Extraction End Date : 06/13/2025

Balance check: RJ

Filter By: RJ

Extraction End Time : 11:00

Balance ID: EX-SC-2

pH Meter ID: N/A

Concentration By: EH

pH Strip Lot#: N/A

Hood ID: 3,7

Supervisor By : RUPESH

Extraction Method: ☐ Separatory Funnel

☐ Continous Liquid/Liquid

☐ Sonication

☐ Waste Dilution

☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24461
Surrogate	1.0ML	200 PPB	PP24597
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2613
Baked Na2SO4	N/A	EP2620
Sand	N/A	E2865
Hexane	N/A	E3938
H2SO4 1:1	N/A	EP2610
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS723.

KD Bath ID: N/A

Envap ID: NEVAP-02

KD Bath Temperature: N/A

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
06/13/25	RJ (Sat 204)	T.P. Pest / RUP.
11:05	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 06/13/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB168459BL	ABLK459	PCB	30.01	N/A	ritesh	Evelyn	10			U3-1
PB168459BS	ALCS459	PCB	30.03	N/A	ritesh	Evelyn	10			2
Q2298-01	AU-05-061125	PCB	30.07	N/A	ritesh	Evelyn	10	D		3
Q2301-02	WC-URBAN-FILL-C	PCB	30.03	N/A	ritesh	Evelyn	10	D		4
Q2303-01	B-165-SB01	PCB	30.08	N/A	ritesh	Evelyn	10			5
Q2303-02	B-170-SB03	PCB	30.05	N/A	ritesh	Evelyn	10			6
Q2305-01	TR-04-06122025	PCB	30.04	N/A	ritesh	Evelyn	10	D		U6-1
Q2305-01MS	TR-04-06122025MS	PCB	30.02	N/A	ritesh	Evelyn	10	D		2
Q2305-01MS D	TR-04-06122025MSD	PCB	30.06	N/A	ritesh	Evelyn	10	D		3
Q2307-01	LINDEN-SAA	PCB	30.07	N/A	ritesh	Evelyn	10	C		4
Q2308-01	EO-02-06122025	PCB	30.03	N/A	ritesh	Evelyn	10	D		5
Q2310-01	TP-7	PCB	30.01	N/A	ritesh	Evelyn	10	D		6

* Extracts relinquished on the same date as received.

6/13/25

1596151
8:05

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2298 WorkList ID : 190173 Department : Extraction Date : 06-13-2025 07:59:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2298-01	AU-05-061125	Solid	PCB	Cool 4 deg C	PSEG05	D41	06/11/2025	8082A
Q2301-02	WC-URBAN-FILL-C	Solid	PCB	Cool 4 deg C	ENTA05	D41	06/11/2025	8082A
Q2303-01	B-165-SB01	Solid	PCB	Cool 4 deg C	PORT06	D41	06/11/2025	8082A
Q2303-02	B-170-SB03	Solid	PCB	Cool 4 deg C	PORT06	D41	06/11/2025	8082A
Q2305-01	TR-04-06122025	Solid	PCB	Cool 4 deg C	PSEG05	D41	06/12/2025	8082A
Q2307-01	LINDEN-SAA	Solid	PCB	Cool 4 deg C	PSEG03	D51	06/12/2025	8082A
Q2308-01	EO-02-06122025	Solid	PCB	Cool 4 deg C	PSEG05	D51	06/12/2025	8082A
Q2310-01	TP-7	Solid	PCB	Cool 4 deg C	PSEG03	D41	06/12/2025	8082A

Date/Time 06/13/25 8:00
Raw Sample Received by: Ed (let lab)
Raw Sample Relinquished by: Ed Sm

Date/Time 06/13/25 8:25
Raw Sample Received by: CP Sm
Raw Sample Relinquished by: B Let (u)

Prep Standard - Chemical Standard Summary

Order ID : Q2303

Test : PCB

Prepbatch ID : PB168459,

Sequence ID/Qc Batch ID: PP061325,

Standard ID :

EP2610,EP2613,EP2620,PP24329,PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369,PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387,PP24461,PP24597,

Chemical ID :

E2865,E3551,E3804,E3877,E3917,E3932,E3933,E3938,M6157,P 11522,P12699,P12702,P12931,P12936,P12949,P12955,P12957,P13356,P13357,P13373,P13381,P13589,P13591,P13697,P13702,P13830,P13878,P13883,W 3112,W3177,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025
<u>FROM</u>	1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2613	05/09/2025	11/05/2025	RUPESHKUMAR SHAH	None	None	Riteshkumar Patel 05/09/2025
<u>FROM</u> 4000.00000ml of E3932 + 4000.00000ml of E3933 = Final Quantity: 8000.000 ml								

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2620	05/30/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/30/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
203	AR1660 750 PPB STD	PP24331	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
204	AR1660 500 PPB STD	PP24332	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
205	AR1660 250 PPB STD	PP24333	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
206	AR1660 50 PPB STD	PP24334	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24332 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
213	AR1221 1000 PPB WORKING SOLUTION	PP24335	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13702 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1079	AR1221 750 PPB STD	PP24336	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
222	AR1221 500 PPB STD	PP24337	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24335 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1080	AR1221 250 PPB STD	PP24338	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1081	AR1221 50 PPB STD	PP24339	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24337 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
214	AR1232 1000 PPB WORKING SOLUTION	PP24340	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13878 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1063	AR1232 750 PPB STD	PP24341	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
223	AR1232 500 PPB STD	PP24342	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24340 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1064	AR1232 250 PPB STD	PP24343	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1065	AR1232 50 PPB STD	PP24344	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24342 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
215	AR1242 1000 PPB WORKING STD	PP24345	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12931 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1067	AR1242 750 PPB STD	PP24346	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.75000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
224	AR1242 500 PPB STD	PP24347	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24345 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1068	AR1242 250 PPB STD	PP24348	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1069	AR1242 50 PPB STD	PP24349	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24347 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
216	AR1248 1000 PPB WORKING STD	PP24350	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12936 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1075	AR1248 750 PPB STD	PP24351	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
225	AR1248 500 PPB STD	PP24352	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24350 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1076	AR1248 250 PPB STD	PP24353	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1077	AR1248 50 PPB STD	PP24354	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24352 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
217	AR1254 1000 PPB WORKING STD	PP24355	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13830 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1071	AR1254 750 PPB STD	PP24356	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
226	AR1254 500 PPB STD	PP24357	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24355 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1072	AR1254 250 PPB STD	PP24358	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1073	AR1254 50 PPB STD	PP24359	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24357 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1529	AR1262 1000 PPB Working Solution	PP24360	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13883 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3753	AR1262 750 PPB STD	PP24361	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1530	AR1262 500 PPB STD	PP24362	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24360 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3754	AR1262 250 PPB STD	PP24363	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3755	AR1262 50 PPB STD	PP24364	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24362 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1532	AR1268 1000 PPB Working Solution	PP24365	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13381 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3820	AR1268 750 PPB STD	PP24366	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1533	AR1268 500 PPB STD	PP24367	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24365 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3821	AR1268 250 PPB STD	PP24368	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3822	AR1268 50 PPB STD	PP24369	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24367 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
404	AR1660 100 PPM Stock Solution 2nd Source	PP24370	03/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12949 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
405	AR1660 1000/100 PPB ICV STD	PP24371	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 98.50000ml of W3177 + 0.50000ml of PP24329 + 1.00000ml of PP24370 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
406	AR1660 500 PPB ICV	PP24372	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24371 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24373	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13373 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1886	AR1221 500 PPB ICV	PP24374	03/18/2025	08/12/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of E3877 + 0.50000ml of W3177 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24375	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12699 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1888	AR1232 500 PPB ICV	PP24376	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24375 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24377	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13589 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1891	AR1242 500 PPB ICV	PP24378	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24377 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24379	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13591 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1892	AR1248 500 PPB ICV	PP24380	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24379 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1894	AR1254 500 PPB ICV	PP24382	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3757	AR1262 1000 PPB Working Solution second source	PP24384	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12702 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3758	AR1262 500 PPB STD ICV	PP24385	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24384 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24386	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P11522 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3823	AR1268 500 PPB STD ICV	PP24387	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24386 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24461	04/11/2025	10/03/2025	Abdul Mirza	None	None	Yogesh Patel
								04/16/2025

FROM 0.50000ml of P12955 + 99.50000ml of E3917 = Final Quantity: 100.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24597	05/20/2025	11/05/2025	Abdul Mirza	None	None	Yogesh Patel 05/22/2025
<u>FROM</u>	1.00000ml of P13357 + 999.00000ml of E3932 = Final Quantity: 1000.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	0000243821	06/30/2025	04/30/2020 / RAJESH	04/28/2020 / RAJESH	E2865

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	12/04/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	11/05/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3917

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	11/05/2025	05/05/2025 / RUPESH	04/23/2025 / RUPESH	E3932

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	11/05/2025	05/05/2025 / RUPESH	04/23/2025 / RUPESH	E3933

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	/	05/14/2025 / RUPESH	E3938

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	91867 / Aroclor 1232 100 ug/mL	020823	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12699

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	x9166 / Aroclor 1262 100 ug/mL	060523	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12702

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	a0203672	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12931

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	a0202803	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12936

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	09/18/2025	03/18/2025 / yogesh	12/20/2023 / Yogesh	P12949

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	10/11/2025	04/11/2025 / Abdul	12/20/2023 / Yogesh	P12955

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	11/20/2025	05/20/2025 / Abdul	04/22/2024 / Abdul	P13357

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0207475	09/18/2025	03/18/2025 / yogesh	05/03/2024 / Abdul	P13381

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13589

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13591

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0215270	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0217391	09/18/2025	03/18/2025 / yogesh	12/09/2024 / Ankita	P13830

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0220950	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13883

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Sand
Purified
Washed and Ignited



Material No.: 3382-05
Batch No.: 0000243821
Manufactured Date: 2018/04/09
Retest Date: 2025/04/07
Revision No: 1


Certificate of Analysis

Test	Specification	Result
Substances Soluble in HCl	$\leq 0.16 \%$	0.01

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC

E 2865


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.5	0.1
Water (H ₂ O)	≤ 0.5 %	0.1 %
Solubility in H ₂ O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities – Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities – Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities – Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities – Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

Recd by RP on 9/25/24

E 3804

>>> Continued on page 2 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
------	---------------	--------

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales
Sr. Manager, Quality Assurance

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

 E3877

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 03/31/25

E3917

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

 **avantor™**



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

RS

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3932



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantors**TM



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E3933



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantors**



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E3938



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™



MG157
MS

Material No.: 9673-33

Batch No.: 24I1262013

Manufactured Date: 2024-08-07

Retest Date: 2029-08-06

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO ₂)	<= 2 ppm	<2 ppm
Ammonium (NH ₄)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO ₃)	<= 0.2 ppm	0.1 ppm
Phosphate (PO ₄)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

 **avantors**™



Material No.: 9673-33

Batch No.: 2411262013

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 1.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	1.1 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	< 1.0 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak

Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



Certificate of Analysis

P11518
↓
P11522
AJ
02/21/22

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1

Lot Issue Date: 09-Feb-2021

Lot Number: 0006587800

Expiration Date: 31-Mar-2029

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.


Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937



CERTIFIED WEIGHT REPORT

Part Number: 91867
Lot Number: 020823
Description: WP 037 - Aroclor 1232
 PCB Technical Mixture
Expiration Date: 020833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB

Solvent:
 Aceton


Weight(s) shown below were combined and diluted to (mL): 100.0
 5E-05 Balance Uncertainty
 0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments

GC3-M1 Analysis by Melissa Stonier
 Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
 Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
 Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
 Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
 Rate = 8°C/min, Total run time = 35 min
 Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
 Standard injection = 1.5µL, Range=3







110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32009 **Lot No.:** A0203672

Description: Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

p12928

→

p12932

AJ

12/27/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	—%	1,004.7 µg/mL	+/- 55.7515

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Russ Boothamer

Russ Boothamer - Operations Technician I

Date Mixed: 26-Oct-2023

Balance Serial #

B442140311

Jennifer Polino

Jennifer Polino - Operations Tech III - ARM QC

Date Passed: 06-Nov-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32010 **Lot No.:** A0202803

Description: Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P1219697
P1219697
AF
12104123

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	13897600	---%	1,001.7 µg/mL	+/- 55.5850

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente
Laith Clemente - Operations Technician I

Date Mixed: 03-Oct-2023

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397



Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

Weight(s) shown below were combined and diluted to (mL):
200.0

Balance Uncertainty
0.010

Flask Uncertainty
0.010

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

Weight(s) shown below were combined and diluted to (mL):
200.0


Balance Uncertainty
0.010

Flask Uncertainty
0.010

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty (%)	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20084

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty

Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Part Number:
20084

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty


Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





CERTIFIED WEIGHT REPORT

Part Number: 99139
Lot Number: 121823
Description: Aroclor 1254

Expiration Date: 121833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB

Volume(s) shown below were combined and diluted to (mL): 20.0
Note: Aroclor 1254 is a mix of isomers.

Solvent(s): Iso-octane
Lot#: 82227

Formulated By: Anthony Mahoney	121823	DATE
Reviewed By: Pedro L. Rentas	121823	DATE


P12956 / Y.P.
12/18/23
P12957

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Initial Uncertainty	Pipette (mL)	Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information		
											(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA) LD50
1. Aroclor 1254	79100	121823	0.10	2.00	0.017	0.017	1003.3	100.1	11097-69-1	1.8	0.5mg/m3 (skin)		or-rat 1295mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150 °C (Time 1 = 4 min), Temp 2 = 280 °C (Time 2 = 13.5 min)
Rate = 8 °C/min, Total run time = 35 min
Injector temp. = 200 °C, FID Temp. = 300 °C, FID Signal = EDAQ Channel 1
Standard injection = 1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name: Aroclor 1221 Standard
Product Number: PP-292-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006783205
Lot Issue Date: 20-Feb-2024
Expiration Date: 31-Mar-2032

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1221	100.3 ±	0.5 µg/mL	011104-28-2	NT01017

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13342
AJ
05106124
P13343

Page: 1 of 2

CSD-QA-015.2

ISO 17025
Cert No. AT-1937


250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034
Cert No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.2

ISO 17025
Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410 **Lot No.:** A0207475

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	10947000	---%	1,000.0 µg/mL	+/- 55.4925

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13380
↓
P 13381
↓
SAZU
05/6/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 500:1

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 06-Feb-2024

Balance Serial # B442140311

Dylan Murphy
Dylan Murphy - Operations Technician I


Date Passed: 09-Feb-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13380 } (2)
↓
P13381
1

[Signature]
05/6/2024

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

Product Name: Aroclor 1242 Standard

Lot Number: 0006665550

Product Number: PP-312-1

Lot Issue Date: 08-Feb-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Jan-2027

Component Name	CERTIFIED VALUES		CAS#	Analyte Lot
	Concentration	Expanded Uncertainty		
Aroclor 1242	100.4	± 0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590AJ
10/14/24

ISO 17034



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 95T215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1



ISO 17025
Cert No. AT-

Reference Material Certificate
Product Information Sheet

Product Name: Aroclor 1248 Standard

Lot Number: 0006726317

Product Number: PP-342-1

Lot Issue Date: 27-Jan-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 28-Feb-2031

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3 ±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AS
10/14/2024



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1

ISO 17025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix
Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697
↓
P13701 } Y.P.
10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B442140311


Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0215270

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13902
P13703 } Y.P.
10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	14969200	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 16-Aug-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 20-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
AJ
12/09/24

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024

Balance Serial # C322230531

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008

Lot No.: A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	---%	1,007.0 µg/mL	+/- 55.8810

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13878



P13880

AJ
01/28/25

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye

Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024

Balance Serial #

C322230531

Brittany Federhik

Brittany Federhiko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32409 **Lot No.:** A0220950

Description: Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: April 30, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,002.0 µg/mL	+/- 55.6035

Solvent: Hexane

CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13882

↓

AJ
01/28/25

P13883

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckal - Mix Technician

Date Mixed: 09-Jan-2025 **Balance Serial #** C322230531



Britiany Federinko - Operations Tech I

Date Passed: 14-Jan-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Grannett Fleming
ADDRESS: 1010 Adams Ave
CITY Audobon STATE: PA ZIP: 19405
ATTENTION: Joe Krupansky
PHONE: 610-301-8342 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Amtrak Replacement of SB
PROJECT NO.: 9500000818 LOCATION: Kearny, NJ
PROJECT MANAGER: Joe Krupansky
e-mail: QAQC@BEMSYS.com
PHONE: 610-301-8342 FAX:

CLIENT BILLING INFORMATION

BILL TO: Alliance PO#:
ADDRESS: 284 Sheffield St
CITY Mountainside STATE: NJ ZIP: 07093
ATTENTION: Samuela Beatty PHONE: 908-788-3148

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): 10 DAYS*
EDD: 10 DAYS*

*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT BEM EDD

PCBs
TAL Metals
Cr(VI) Cr(III)
1 2 3 4 5 6 7 8 9

PRESERVATIVES

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HNO3 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	B-165-SB01	S		X	6/11/25	11:10	1	X	X	X							
2.	B-170-SB03	S		X	6/11/25	11:15	1	X	X	X							
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>Jessy Martinez</u>	DATE/TIME: <u>6/11/25</u>	RECEIVED BY: 1. <u>[Signature]</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>27</u> °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	Comments: <u>B-165-SB01 & B-170-SB03 have limited volume</u>
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	Page <u>1</u> of <u>1</u> CLIENT: ☒ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488