

SOP ID : MSM4500-NH3 B,G-Ammonia-18

SDG No : N/A

Start Digest Date: 06/13/2025 Time : 10:10 Temp : 150 °C

Matrix : WATER

End Digest Date: 06/13/2025 Time : 11:10 Temp : 157 °C

Pipette ID : WC

Balance ID : N/A

Hood ID : HOOD#2

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : WC-DIST-BLOCK-1

Filter paper ID : N/A

Prep Technician Signature: 12

Weigh By : IZ

pH Meter ID : N/A

Supervisor Signature:

Standard Name	MLS USED	STD REF. # FROM LOG
LCSW	1.0ML	WP113449
MS/MSD SPIKE SOL.	1.0ML	WP113450
PBW	50.0ML	W3112
RL CHECK	0.1ML	WP113450
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
BORATE BUFFER	2.5ML	WP111325
NAOH 6N	0.5-2.0ML	WP111318
H2SO4 0.04N	5.0ML	WP112828
pH strip-Ammonia	N/A	W3133
KI-starch paper	N/A	W3155
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

ALL GLASSWEAR ARE STEAMED OUT AND THERE WERE NO TRACE OF AMMONIA USING NESLER REAGENT WP111604. Due to bad matrix and client history 1ML was taken as an initial volume for Q2309-01 and Q2309-05

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
06/13/25 11:25	12 (WC)	12 (WC)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB168427BL	PBW427	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
PB168427BS	LCS427	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2286-03DUP	LAW-25-0084DUP	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2286-03MS	LAW-25-0084MS	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2286-03MSD	LAW-25-0084MSD	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2286-03	LAW-25-0084	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2309-01	EFFLUENT	1	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2309-05	INFLUENT	1	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A