

## Initial and Continuing Calibration Verification

**Client:** Holland Manufacturing Co.

**SDG No.:** Q2309

**Project:** Pre Treatment Plant 2025

**RunNo.:** LB136152

| Analyte                                       | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|-----------------------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV</b><br>Orthophosphate as P  | mg/L  | 0.505  | 0.50       | 101        | 90-110                 | 06/13/2025    |
| Sample ID: <b>CCV1</b><br>Orthophosphate as P | mg/L  | 0.519  | 0.5        | 104        | 90-110                 | 06/13/2025    |
| Sample ID: <b>CCV2</b><br>Orthophosphate as P | mg/L  | 0.505  | 0.5        | 101        | 90-110                 | 06/13/2025    |

## Initial and Continuing Calibration Verification

**Client:** Holland Manufacturing Co.

**SDG No.:** Q2309

**Project:** Pre Treatment Plant 2025

**RunNo.:** LB136153

| Analyte                                     | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|---------------------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV</b><br>Phosphorus, Total  | mg/L  | 0.494  | 0.50       | 99         | 90-110                 | 06/13/2025    |
| Sample ID: <b>CCV1</b><br>Phosphorus, Total | mg/L  | 0.514  | 0.50       | 103        | 90-110                 | 06/13/2025    |
| Sample ID: <b>CCV2</b><br>Phosphorus, Total | mg/L  | 0.496  | 0.50       | 99         | 90-110                 | 06/13/2025    |
| Sample ID: <b>CCV3</b><br>Phosphorus, Total | mg/L  | 0.502  | 0.50       | 100        | 90-110                 | 06/13/2025    |

## Initial and Continuing Calibration Verification

**Client:** Holland Manufacturing Co.

**SDG No.:** Q2309

**Project:** Pre Treatment Plant 2025

**RunNo.:** LB136156

| Analyte                                | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|----------------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV1</b><br>Ammonia as N | mg/L  | 0.97   | 1          | 97         | 90-110                 | 06/13/2025    |
| Sample ID: <b>CCV1</b><br>Ammonia as N | mg/L  | 1      | 1          | 100        | 90-110                 | 06/13/2025    |
| Sample ID: <b>CCV2</b><br>Ammonia as N | mg/L  | 0.95   | 1          | 95         | 90-110                 | 06/13/2025    |
| Sample ID: <b>CCV3</b><br>Ammonia as N | mg/L  | 1      | 1          | 100        | 90-110                 | 06/13/2025    |

## Initial and Continuing Calibration Verification

**Client:** Holland Manufacturing Co.

**SDG No.:** Q2309

**Project:** Pre Treatment Plant 2025

**RunNo.:** LB136156

| Analyte | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|---------|-------|--------|------------|---------------|---------------------------|------------------|
|---------|-------|--------|------------|---------------|---------------------------|------------------|