

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2309

Project: Pre Treatment Plant 2025

Run No.: LB136135

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136135BS							
BOD5	mg/L	198	213		107	1	84.6-115.4	06/12/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2309

Project: Pre Treatment Plant 2025

Run No.: LB136143

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136143BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	06/13/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2309

Project: Pre Treatment Plant 2025

Run No.: LB136145

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136145BS							
TSS	mg/L	550	532		97	1	90-110	06/13/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2309

Project: Pre Treatment Plant 2025

Run No.: LB136152

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136152BS							
Orthophosphate as P	mg/L	0.5	0.50		99	1	90-110	06/13/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2309

Project: Pre Treatment Plant 2025

Run No.: LB136156

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168427BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	06/13/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2309

Project: Pre Treatment Plant 2025

Run No.: LB136153

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168478BS							
Phosphorus, Total	mg/L	0.50	0.49		98	1	90-110	06/13/2025