

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2314 SAS No.: Q2314 SDG No.: Q2314

Instrument ID: BNA_N Calibration Date/Time: 06/14/2025 22:31

Lab File ID: BN037273.D Init. Calib. Date(s): 06/13/2025 06/13/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 13:33 17:11

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.537	0.562		4.7	20.0
Fluoranthene-d10	1.047	1.055		0.9	20.0
2-Fluorophenol	0.982	0.959		-2.3	20.0
Phenol-d6	1.035	1.017		-1.7	20.0
Nitrobenzene-d5	0.395	0.415		5.1	20.0
2-Fluorobiphenyl	1.681	1.763		4.9	20.0
2,4,6-Tribromophenol	0.166	0.170		2.4	20.0
Terphenyl-d14	0.904	0.900		-0.4	20.0
1,4-Dioxane	0.549	0.499		-9.1	20.0

All other compounds must meet a minimum RRF of 0.010.