

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061325\
 Data File : BN037273.D
 Acq On : 14 Jun 2025 22:31
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 16 02:45:45 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 13 18:43:34 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	0.00
2	1,4-Dioxane	0.400	0.364	9.0	103	0.00
3	n-Nitrosodimethylamine	0.400	0.429	-7.2	113	0.00
4 S	2-Fluorophenol	0.400	0.390	2.5	110	0.00
5 S	Phenol-d6	0.400	0.393	1.8	114	0.00
6	bis(2-Chloroethyl)ether	0.400	0.436	-9.0	125	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	-0.01
8 S	Nitrobenzene-d5	0.400	0.420	-5.0	115	0.00
9	Naphthalene	0.400	0.397	0.8	108	0.00
10	Hexachlorobutadiene	0.400	0.409	-2.2	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.419	-4.7	107	0.00
12	2-Methylnaphthalene	0.400	0.392	2.0	104	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.400	0.409	-2.2	100	-0.01
15 S	2-Fluorobiphenyl	0.400	0.420	-5.0	105	0.00
16	Acenaphthylene	0.400	0.386	3.5	102	-0.01
17	Acenaphthene	0.400	0.390	2.5	101	0.00
18	Fluorene	0.400	0.388	3.0	100	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	101	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.430	-7.5	126	-0.01
21	4-Bromophenyl-phenylether	0.400	0.409	-2.2	110	0.00
22	Hexachlorobenzene	0.400	0.395	1.3	97	0.00
23	Atrazine	0.400	0.384	4.0	101	-0.01
24	Pentachlorophenol	0.400	0.382	4.5	115	0.00
25	Phenanthrene	0.400	0.388	3.0	105	-0.01
26	Anthracene	0.400	0.377	5.8	102	0.00
27 SURR	Fluoranthene-d10	0.400	0.403	-0.8	101	0.00
28	Fluoranthene	0.400	0.370	7.5	98	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	107	0.00
30	Pyrene	0.400	0.393	1.8	101	0.00
31 S	Terphenyl-d14	0.400	0.398	0.5	102	0.00
32	Benzo(a)anthracene	0.400	0.393	1.8	116	0.00
33	Chrysene	0.400	0.403	-0.8	108	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.376	6.0	99	0.00
35 I	Perylene-d12	0.400	0.400	0.0	117	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.423	-5.7	132	-0.02
37	Benzo(b)fluoranthene	0.400	0.387	3.3	120	0.00
38	Benzo(k)fluoranthene	0.400	0.375	6.3	114	0.00
39 C	Benzo(a)pyrene	0.400	0.397	0.8	124	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.412	-3.0	141	-0.01
41	Benzo(g,h,i)perylene	0.400	0.406	-1.5	123	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4