

Surrogate Summary

SW-846

SDG No.: Q2314

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168476BL	PB168476BL	2-Methylnaphthalene-d10	0.4	0.33	83		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.29	74		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.38	95		58	132
PB168476BS	PB168476BS	2-Methylnaphthalene-d10	0.4	0.43	108		30	150
		Fluoranthene-d10	0.4	0.34	85		30	150
		Nitrobenzene-d5	0.4	0.38	94		55	111
		2-Fluorobiphenyl	0.4	0.39	97		53	106
		Terphenyl-d14	0.4	0.38	95		58	132
PB168476BSD	PB168476BSD	2-Methylnaphthalene-d10	0.4	0.40	99		30	150
		Fluoranthene-d10	0.4	0.32	81		30	150
		Nitrobenzene-d5	0.4	0.34	85		55	111
		2-Fluorobiphenyl	0.4	0.36	89		53	106
		Terphenyl-d14	0.4	0.36	90		58	132
Q2314-04	BP-VPB-182-GW-880-882	2-Methylnaphthalene-d10	0.4	0.31	76		30	150
		Fluoranthene-d10	0.4	0.36	89		30	150
		Nitrobenzene-d5	0.4	0.28	70		55	111
		2-Fluorobiphenyl	0.4	0.32	79		53	106
		Terphenyl-d14	0.4	0.39	97		58	132
Q2314-05	BP-VPB-182-EB-20250612	2-Methylnaphthalene-d10	0.4	0.29	72		30	150
		Fluoranthene-d10	0.4	0.35	87		30	150
		Nitrobenzene-d5	0.4	0.25	63		55	111
		2-Fluorobiphenyl	0.4	0.30	74		53	106
		Terphenyl-d14	0.4	0.36	91		58	132
Q2314-06	VPB182-HYD-20250612	2-Methylnaphthalene-d10	0.4	0.027	7	*	30	150
		Fluoranthene-d10	0.4	0.29	73		30	150
		Nitrobenzene-d5	0.4	0.28	69		55	111
		2-Fluorobiphenyl	0.4	0.33	82		53	106
		Terphenyl-d14	0.4	0.40	99		58	132