

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Maresh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134155
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247
CCC	VP134156
Internal Standard/PEM	
ICV/I.BLK	VP134248
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086861.D	06 Jun 2025 07:59		JC\MD	Ok
2	VSTDIC001	VSTDIC001	VN086862.D	06 Jun 2025 12:44	Method failed for com.#13	JC\MD	Ok,M
3	VSTDIC005	VSTDIC005	VN086863.D	06 Jun 2025 13:17		JC\MD	Ok,M
4	VSTDIC020	VSTDIC020	VN086864.D	06 Jun 2025 13:40		JC\MD	Ok,M
5	VSTDIC050	VSTDIC050	VN086865.D	06 Jun 2025 14:03		JC\MD	Ok,M
6	VSTDIC100	VSTDIC100	VN086866.D	06 Jun 2025 14:26		JC\MD	Ok,M
7	VSTDIC150	VSTDIC150	VN086867.D	06 Jun 2025 14:49		JC\MD	Ok,M
8	IBLK	IBLK	VN086868.D	06 Jun 2025 15:12		JC\MD	Ok
9	VSTDICV050	ICVVN060625	VN086869.D	06 Jun 2025 15:54		JC\MD	Ok,M
10	VN0606WBL01	VN0606WBL01	VN086870.D	06 Jun 2025 16:47		JC\MD	Ok
11	VN0606WBL02	VN0606WBL02	VN086871.D	06 Jun 2025 17:10		JC\MD	Ok
12	VN0606WBS01	VN0606WBS01	VN086872.D	06 Jun 2025 17:33		JC\MD	Ok,M
13	VN0606WBSD01	VN0606WBSD01	VN086873.D	06 Jun 2025 17:56		JC\MD	Ok,M
14	Q2254-01	BP-VPB-182-GW-810-8	VN086874.D	06 Jun 2025 18:19	vial A pH<2 endccc out of tune	JC\MD	Not Ok
15	Q2237-02	TW-WTS-10	VN086875.D	06 Jun 2025 18:42	vial A pH<2	JC\MD	Ok
16	Q2216-02	3887	VN086876.D	06 Jun 2025 19:05	vial A pH<2	JC\MD	Ok
17	Q2216-03	3888	VN086877.D	06 Jun 2025 19:28	vial A pH<2	JC\MD	Ok
18	Q2216-04	3864	VN086878.D	06 Jun 2025 19:51	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Maresh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134155 VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC Internal Standard/PEM	VP134156		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134248		

19	Q2216-05	3865	VN086879.D	06 Jun 2025 20:13	vial A pH<2 Out of Tune	JC\MD	Not Ok
20	Q2216-06	3851	VN086880.D	06 Jun 2025 20:36	vial A pH<2 Out of Tune	JC\MD	Not Ok
21	Q2206-04	TP-1	VN086881.D	06 Jun 2025 20:59	vial A pH<2 Out of Tune	JC\MD	Not Ok
22	Q2242-04	TP09-MHJ	VN086882.D	06 Jun 2025 21:21	vial A pH<2 Out of Tune	JC\MD	Not Ok
23	Q2192-01	SB-1	VN086883.D	06 Jun 2025 21:44	vial A pH<2 Out of Tune	JC\MD	Not Ok
24	Q2198-02	B-202-SB02	VN086884.D	06 Jun 2025 22:07	vial A pH<2 Out of Tune	JC\MD	Not Ok
25	Q2198-04	B-207-SB02	VN086885.D	06 Jun 2025 22:29	vial A pH<2 Out of Tune	JC\MD	Not Ok
26	VSTDCCC050	VSTDCCC050EC	VN086886.D	06 Jun 2025 22:52	Out of Tune	JC\MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061325

Review By	John Carlone	Review On	6/16/2025 9:41:28 AM
Supervise By	Maresh Dadoda	Supervise On	6/16/2025 5:03:41 PM
SubDirectory	VN061325	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134329
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134330,VP134331

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086995.D	13 Jun 2025 10:26		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086996.D	13 Jun 2025 12:52	pH#Lot#V12668	JC\MD	Ok,M
3	VN0613MBL01	VN0613MBL01	VN086997.D	13 Jun 2025 13:26		JC\MD	Ok
4	VN0613WBL01	VN0613WBL01	VN086998.D	13 Jun 2025 13:47		JC\MD	Ok
5	Q2282-04	251233	VN086999.D	13 Jun 2025 14:10	vial A pH<2	JC\MD	Ok
6	VN0613WBS01	VN0613WBS01	VN087000.D	13 Jun 2025 14:32	BS Failed High For Com.#17,69	JC\MD	Ok,M
7	VN0613WBSD01	VN0613WBSD01	VN087001.D	13 Jun 2025 15:06	BSD Failed High For Comp.#17,52,88	JC\MD	Ok,M
8	Q2314-06	VPB182-HYD-2025061	VN087002.D	13 Jun 2025 15:28	vial A pH<2	JC\MD	Ok
9	Q2314-01	BP-VPB-182-TB-20250	VN087003.D	13 Jun 2025 15:50	vial A pH<2 TB	JC\MD	Ok
10	Q2314-05	BP-VPB-182-EB-20250	VN087004.D	13 Jun 2025 16:12	vial A pH<2 EB,Hit of com.#16,25	JC\MD	ReRun
11	Q2314-04	BP-VPB-182-GW-880-8	VN087005.D	13 Jun 2025 16:34	vial A pH<2	JC\MD	Ok
12	Q2314-02	BP-VPB-182-GW-820-8	VN087006.D	13 Jun 2025 16:56	vial A pH<2	JC\MD	Ok
13	Q2314-05	BP-VPB-182-EB-20250	VN087007.D	13 Jun 2025 17:18	vial B pH<2 EB	JC\MD	Ok
14	IBLK	IBLK	VN087008.D	13 Jun 2025 17:40		JC\MD	Ok
15	IBLK	IBLK	VN087009.D	13 Jun 2025 18:02		JC\MD	Ok
16	Q2268-03DL	MW-2-20250605DL	VN087010.D	13 Jun 2025 18:24	vial B pH<2	JC\MD	Ok
17	Q2268-06DL	MW-2-20250605-ADL	VN087011.D	13 Jun 2025 18:46	vial B pH<2	JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN061325

Review By	John Carlone	Review On	6/16/2025 9:41:28 AM
Supervise By	Maresh Dadoda	Supervise On	6/16/2025 5:03:41 PM
SubDirectory	VN061325	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134329		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134330,VP134331		

18	Q2268-07DL	MW-6-20250605DL	VN087012.D	13 Jun 2025 19:07	BS,BSD Failed High	JC\MD	ReRun
19	Q2268-08DL	MW-3-20250605DL	VN087013.D	13 Jun 2025 19:29	BS,BSD Failed High	JC\MD	ReRun
20	Q2302-01RE	250611078-02-VOARE	VN087014.D	13 Jun 2025 19:51	vial C pH#6.0	JC\MD	Confirms
21	VSTDCCC050	VSTDCCC050EC	VN087015.D	13 Jun 2025 20:13		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060225

Review By	Mahesh Dadoda	Review On	6/3/2025 2:38:36 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/3/2025 2:39:16 PM
SubDirectory	VY060225	HP Acquire Method	MSVOA_Y
		HP Processing Method	82Y060225S.M

STD. NAME	STD REF.#
Tune/Reschk	VP134084
Initial Calibration Stds	VP134085,VP134086,VP134087,VP134088,VP134089,VP134090
CCC	VP134092,VP134093
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134091
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022488.D	02 Jun 2025 08:31		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022489.D	02 Jun 2025 09:08	Need ICAL	SY/MD	Not Ok
3	VY0602SBL01	VY0602SBL01	VY022490.D	02 Jun 2025 10:38		SY/MD	Not Ok
4	VSTDICC005	VSTDICC005	VY022491.D	02 Jun 2025 11:46		SY/MD	Ok,M
5	VSTDICC010	VSTDICC010	VY022492.D	02 Jun 2025 12:09		SY/MD	Ok,M
6	VSTDICC020	VSTDICC020	VY022493.D	02 Jun 2025 12:32		SY/MD	Ok,M
7	VSTDICCC050	VSTDICCC050	VY022494.D	02 Jun 2025 12:54		SY/MD	Ok,M
8	VSTDICC100	VSTDICC100	VY022495.D	02 Jun 2025 13:17		SY/MD	Ok,M
9	VSTDICC150	VSTDICC150	VY022496.D	02 Jun 2025 13:39		SY/MD	Ok,M
10	VIBLK	VIBLK	VY022497.D	02 Jun 2025 14:03		SY/MD	Ok
11	VSTDICV050	ICVVY060225	VY022498.D	02 Jun 2025 14:59	Icv Fail For DOD For Com.# 18	SY/MD	Ok,M
12	VY0602SBL02	VY0602SBL02	VY022499.D	02 Jun 2025 16:00		SY/MD	Ok
13	VY0602SBS01	VY0602SBS01	VY022500.D	02 Jun 2025 16:24		SY/MD	Ok,M
14	VY0602SBSD01	VY0602SBSD01	VY022501.D	02 Jun 2025 16:47		SY/MD	Ok,M
15	Q2160-02	TP04-MHG-VOC	VY022502.D	02 Jun 2025 17:10	vial A	SY/MD	Ok
16	Q2152-01	OK-02-05292025	VY022503.D	02 Jun 2025 17:34	Not Purged,vial B	SY/MD	Not Ok
17	Q2153-01RE	TR-04-0592025RE	VY022504.D	02 Jun 2025 17:57	Surrogate Fail;Internal Standard Fail, vial B	SY/MD	Confirms

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060225

Review By	Mahesh Dadoda	Review On	6/3/2025 2:38:36 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/3/2025 2:39:16 PM		
SubDirectory	VY060225	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y060225S.M
STD. NAME	STD REF.#				
Tune/Reschk	VP134084				
Initial Calibration Stds	VP134085,VP134086,VP134087,VP134088,VP134089,VP134090				
CCC	VP134092,VP134093				
Internal Standard/PEM	VP133934				
ICV/I.BLK	VP134091				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2147-02	VOC	VY022505.D	02 Jun 2025 18:21	Not purge,vial B	SY/MD	Not Ok
19	Q2150-02	TP-42	VY022506.D	02 Jun 2025 18:44	vial B	SY/MD	Ok
20	Q2150-05RE	TP-47RE	VY022507.D	02 Jun 2025 19:08	Internal Standard Fail,vial B	SY/MD	Confirms
21	Q2161-01	B27-SOIL-SAMPLE	VY022508.D	02 Jun 2025 19:31	vial-B,Internal Standard Fail; Surrogate Fail	SY/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VY022509.D	02 Jun 2025 19:54		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY061625

Review By	Semsettin Yesilyurt	Review On	6/17/2025 4:00:42 PM
Supervise By	Maresh Dadoda	Supervise On	6/17/2025 4:09:50 PM
SubDirectory	VY061625	HP Acquire Method	MSVOA_Y HP Processing Method 82y060225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134372
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134373,VP134374 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022694.D	16 Jun 2025 07:59		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022695.D	16 Jun 2025 08:31		SY/MD	Ok,M
3	VY0616SBL01	VY0616SBL01	VY022696.D	16 Jun 2025 09:40		SY/MD	Ok
4	VY0616SBS01	VY0616SBS01	VY022697.D	16 Jun 2025 10:12		SY/MD	Ok,M
5	VY0616SBSD01	VY0616SBSD01	VY022698.D	16 Jun 2025 10:34		SY/MD	Ok,M
6	Q2328-01	CHRT25653	VY022699.D	16 Jun 2025 11:22	vial-A Internal Standard Fail	SY/MD	Ok,M
7	Q2322-01	CL-01-061325	VY022700.D	16 Jun 2025 11:46	vial-A Internal Standard Fail	SY/MD	ReRun
8	Q2314-08	BP-VPB-182-GW-840-8	VY022701.D	16 Jun 2025 12:09	vial-A	SY/MD	Ok
9	Q2324-01	HD-01-6132025	VY022702.D	16 Jun 2025 12:33	vial-A Not purged	SY/MD	Not Ok
10	Q2323-01	PL-01-06132025	VY022703.D	16 Jun 2025 12:56	vial-A	SY/MD	Ok
11	Q2312-03	TP-1-VOC	VY022704.D	16 Jun 2025 14:12	vial-A	SY/MD	Ok
12	Q2319-03	MH-B-VOC	VY022705.D	16 Jun 2025 14:36	vial-A	SY/MD	Ok,M
13	Q2305-01RE	TR-04-06122025RE	VY022706.D	16 Jun 2025 14:59	vial-B Internal Standard Fail;Surrogate fail	SY/MD	Confirms
14	Q2325-03	TP-8-VOC	VY022707.D	16 Jun 2025 15:22	vial-A	SY/MD	Ok
15	Q2328-01RE	CHRT25653RE	VY022708.D	16 Jun 2025 15:46	vial-B Not purged	SY/MD	Not Ok
16	VIBLK	VIBLK	VY022709.D	16 Jun 2025 16:09		SY/MD	Ok
17	VSTDCCC050	VSTDCCC050EC	VY022710.D	16 Jun 2025 16:55		SY/MD	Ok,M

M : Manual Integration