

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q2314 SAS No.: Q2314 SDG NO.: Q2314
 Lab File ID: VN086996.D Date Analyzed: 06/13/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:52
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	154974	13.794				
UPPER LIMIT	309948	14.294				
LOWER LIMIT	77487	13.294				
EPA SAMPLE NO.						
BP-VPB-182-TB-20250610	169487	13.79				
BP-VPB-182-GW-820-822	167434	13.79				
BP-VPB-182-GW-880-882	177372	13.79				
BP-VPB-182-EB-20250612	177791	13.79				
VPB182-HYD-20250612	173766	13.79				
VN0613WBL01	258118	13.79				
VN0613WBS01	150656	13.79				
VN0613WBSD01	143189	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.