

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q2314 SAS No.: Q2314 SDG NO.: Q2314
 Lab File ID: VY022695.D Date Analyzed: 06/16/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	115228	13.346				
UPPER LIMIT	230456	13.846				
LOWER LIMIT	57614	12.846				
EPA SAMPLE NO.						
BP-VPB-182-GW-840-842	65272	13.35				
VY0616SBL01	63063	13.35				
VY0616SBS01	109254	13.35				
VY0616SBSD01	103858	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.