

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2314 SAS No.: Q2314 SDG NO.: Q2314
Lab File ID: VN086861.D BFB Injection Date: 06/06/2025
Instrument ID: MSVOA_N BFB Injection Time: 07:59
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.3
75	30.0 - 60.0% of mass 95	48.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	66.6
175	5.0 - 9.0% of mass 174	4.7 (7.1) 1
176	95.0 - 101.0% of mass 174	65.3 (98.1) 1
177	5.0 - 9.0% of mass 176	4.4 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VN086862.D	06/06/2025	12:44
VSTDIC005	VSTDIC005	VN086863.D	06/06/2025	13:17
VSTDIC020	VSTDIC020	VN086864.D	06/06/2025	13:40
VSTDIC050	VSTDIC050	VN086865.D	06/06/2025	14:03
VSTDIC100	VSTDIC100	VN086866.D	06/06/2025	14:26
VSTDIC150	VSTDIC150	VN086867.D	06/06/2025	14:49