

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tetra Tech NUS, Inc. **SDG No.:** Q2316  
**Contract:** TETR06 **Lab Code:** CHEM **Case No.:** Q2316 **SAS No.:** Q2316  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| ICV01     | Iron    | 3960           | 4000       | 99            | 90 - 110                  | P | 06/17/2025       | 12:25            | LB136187      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tetra Tech NUS, Inc. **SDG No.:** Q2316  
**Contract:** TETR06 **Lab Code:** CHEM **Case No.:** Q2316 **SAS No.:** Q2316  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| LLICV01   | Iron    | 110            | 100        | 110           | 80 - 120                  | P | 06/17/2025       | 12:41            | LB136187      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tetra Tech NUS, Inc. **SDG No.:** Q2316  
**Contract:** TETR06 **Lab Code:** CHEM **Case No.:** Q2316 **SAS No.:** Q2316  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Iron    | 4850           | 5000       | 97            | 90 - 110                  | P | 06/17/2025       | 13:40            | LB136187      |
| CCV02     | Iron    | 4990           | 5000       | 100           | 90 - 110                  | P | 06/17/2025       | 14:43            | LB136187      |
| CCV03     | Iron    | 4870           | 5000       | 97            | 90 - 110                  | P | 06/17/2025       | 15:38            | LB136187      |
| CCV04     | Iron    | 4980           | 5000       | 100           | 90 - 110                  | P | 06/17/2025       | 16:30            | LB136187      |
| CCV05     | Iron    | 4900           | 5000       | 98            | 90 - 110                  | P | 06/17/2025       | 17:37            | LB136187      |
| CCV06     | Iron    | 4900           | 5000       | 98            | 90 - 110                  | P | 06/17/2025       | 18:38            | LB136187      |
| CCV07     | Iron    | 4900           | 5000       | 98            | 90 - 110                  | P | 06/17/2025       | 19:25            | LB136187      |
| CCV08     | Iron    | 4770           | 5000       | 95            | 90 - 110                  | P | 06/17/2025       | 20:17            | LB136187      |