



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Tetra Tech NUS, Inc.				SDG No.:	Q2316				
Contract:	TETR06	Lab Code:	CHEM	Case No.:	Q2316	SAS No.:	Q2316			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Iron	38.5	+/-50	J	80.0	100	P	06/17/2025	12:46	LB136187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc. SDG No.: Q2316
Contract: TETR06 Lab Code: CHEM Case No.: Q2316 SAS No.: Q2316

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Iron	23.4	+/-50	U	80.0	100	P	06/17/2025	13:48	LB136187
CCB02	Iron	23.4	+/-50	U	80.0	100	P	06/17/2025	14:50	LB136187
CCB03	Iron	27.5	+/-50	J	80.0	100	P	06/17/2025	15:42	LB136187
CCB04	Iron	23.4	+/-50	U	80.0	100	P	06/17/2025	16:34	LB136187
CCB05	Iron	23.4	+/-50	U	80.0	100	P	06/17/2025	17:41	LB136187
CCB06	Iron	23.4	+/-50	U	80.0	100	P	06/17/2025	18:42	LB136187
CCB07	Iron	27.8	+/-50	J	80.0	100	P	06/17/2025	19:30	LB136187
CCB08	Iron	41.7	+/-50	J	80.0	100	P	06/17/2025	20:22	LB136187