

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2316 **SAS No.:** Q2316
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	103000	100000	103	85600	116500	06/17/2025	13:11	LB136187
ICSAB01	Iron	104000	99000	105	84400	114500	06/17/2025	13:22	LB136187