

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** Q2316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2316 **SAS No.:** Q2316
Matrix: Water **Sample ID:** Q2316-02 **Client ID:** RW8-SP303-20250612DUP
Percent Solids for Sample: NA **Duplicate ID** Q2316-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Iron	ug/L	20	43.3	J	47.9	J	10		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** Q2316
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2316 **SAS No.:** Q2316
Matrix: Water **Sample ID:** Q2316-02MS **Client ID:** RW8-SP303-20250612MSD
Percent Solids for Sample: NA **Duplicate ID** Q2316-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Iron	ug/L	20	1620		1570		3		P