

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2316 SAS No.: Q2316 SDG NO.: Q2316

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/14/2025

Lab File ID: BN037273.D Time Analyzed: 22:31

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1389	7.575	3410	10.35	1759	14.22
UPPER LIMIT	2778	8.075	6820	10.851	3518	14.724
LOWER LIMIT	694.5	7.075	1705	9.851	879.5	13.724
EPA SAMPLE NO.						
01 PB168476BL	1111	7.58	2537	10.36	1345	14.22
02 PB168476BS	1101	7.58	2637	10.35	1350	14.22
03 PB168476BSD	1070	7.58	2630	10.35	1332	14.22
04 RW8-SP100-20250612	1039	7.58	2587	10.35	1385	14.22
05 RW8-SP303-20250612	1035	7.58	2641	10.35	1419	14.22

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.