

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2316

SDG NO.: Q2316

Lab File ID: BN037223.D

DFTPP Injection Date: 06/13/2025

Instrument ID: BNA_N

DFTPP Injection Time: 11:34

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.1) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.3 (0.4) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
365	Greater than 1% of mass 198	5.3
441	Present, but less than mass 443	90.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	12.6 (20.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037225.D	06/13/2025	13:33
SSTDICC0.2	SSTDICC0.2	BN037226.D	06/13/2025	14:10
SSTDICCC0.4	SSTDICCC0.4	BN037227.D	06/13/2025	14:46
SSTDICC0.8	SSTDICC0.8	BN037228.D	06/13/2025	15:22
SSTDICC1.6	SSTDICC1.6	BN037229.D	06/13/2025	15:59
SSTDICC3.2	SSTDICC3.2	BN037230.D	06/13/2025	16:35
SSTDICC5.0	SSTDICC5.0	BN037231.D	06/13/2025	17:11
SSTDCCC0.4EC	SSTDCCC0.4	BN037238.D	06/13/2025	22:01