

Process Report

Order ID : Q2317

Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
06/13/2025	B2505058	Q2317-04	-8.0	10601	6 L	10202		C2504014	VL042485.D Seq :VL050525 TUNE : VL042468.D ICAL : VL042466.D CCAL : VL042469.D MB : VL042470.D
06/13/2025	B2505058	Q2317-01	-5.9	10588	6 L	10713		C2504014	VL042485.D Seq :VL050525 TUNE : VL042468.D ICAL : VL042466.D CCAL : VL042469.D MB : VL042470.D
06/13/2025	B2505058	Q2317-02	-4.7	10492	6 L	10236		C2504014	VL042485.D Seq :VL050525 TUNE : VL042468.D ICAL : VL042466.D CCAL : VL042469.D MB : VL042470.D
06/13/2025	B2505058	Q2317-05	-6.7	10315	6 L	10241		C2504014	VL042485.D Seq :VL050525 TUNE : VL042468.D ICAL : VL042466.D CCAL : VL042469.D MB : VL042470.D
06/13/2025	B2505058	Q2317-03	-1.8	10198	6 L	10207		C2504007	VL042330.D Seq :VL041525 TUNE : VL042325.D ICAL : VL042177.D CCAL : VL042326.D MB : VL042327.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2504014

Type of Canister : 6 L

Cleaning Date : 04/28/2025

Analyst : SAM

Oven Temperature : 125.0 c

Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10266	-30	04/29/2025	05/05/2025	-30	04/30/2025		10492	VL042485.D
10269	-30	04/29/2025	05/05/2025	-30	04/30/2025		10492	VL042485.D
10601	-30	04/29/2025	05/05/2025	-30	04/30/2025	Q2317-04	10492	VL042485.D
10492	-30	04/29/2025	05/05/2025	-30	04/30/2025	Q2317-02	10492	VL042485.D
10060	-30	04/29/2025	05/05/2025	-30	04/30/2025		10492	VL042485.D
10588	-30	04/29/2025	05/05/2025	-30	04/30/2025	Q2317-01	10492	VL042485.D
10443	-30	04/29/2025	05/05/2025	-30	04/30/2025		10492	VL042485.D
10315	-30	04/29/2025	05/05/2025	-30	04/30/2025	Q2317-05	10492	VL042485.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2504007

Type of Canister : 6 L

Cleaning Date : 04/14/2025

Analyst : SAM

Oven Temperature : 125.0 c



Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10599	-30	04/15/2025	04/15/2025	-30	04/16/2025	Q2011-02	10401	VL042330.D
10595	-30	04/15/2025	04/15/2025	-30	04/16/2025	Q1908-02	10401	VL042330.D
10268	-30	04/15/2025	04/15/2025	-30	04/16/2025	Q1880-02	10401	VL042330.D
10323	-30	04/15/2025	04/15/2025	-30	04/16/2025	Q1880-03	10401	VL042330.D
10258	-30	04/15/2025	04/15/2025	-30	04/16/2025	Q1880-01	10401	VL042330.D
10198	-30	04/15/2025	04/15/2025	-30	04/16/2025	Q2317-03	10401	VL042330.D
10053	-30	04/15/2025	04/15/2025	-30	04/16/2025		10401	VL042330.D
10401	-30	04/15/2025	04/15/2025	-30	04/16/2025	Q1880-05	10401	VL042330.D

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042171.D
Acq On : 27 Mar 2025 07:44
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri Mar 28 02:00:52 2025



AutoFind: Scans 3177, 3178, 3179; Background Corrected with Scan 3150

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	23.0	87995	PASS
75	95	30	66	56.3	215296	PASS
95	95	100	100	100.0	382268	PASS
96	95	5	9	6.6	25383	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	63.8	243904	PASS
175	174	4	9	8.0	19539	PASS
176	174	93	101	95.7	233493	PASS
177	176	5	9	6.8	15877	PASS

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL032725AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Fri Mar 28 02:00:52 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL042177.D 0.1 =VL042176.D 0.5 =VL042175.D 1 =VL042174.D 2 =VL042173.D 10 =VL042172.D 15 =VL042178.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.602	1.514	1.482	1.404	1.380	1.476	6.03
3) Chlorodifluoro...			1.622	1.541	1.514	1.457	1.408	1.508	5.41
4) Chloromethane			0.660	0.597	0.618	0.573	0.570	0.603	6.17
5) T Vinyl Chloride	0.882	0.705	0.635	0.576	0.584	0.549	0.546	0.639	18.84
6) T Bromomethane			0.327	0.268	0.266	0.256	0.250	0.273	11.20
7) Chloroethane			0.286	0.248	0.223	0.227	0.226	0.242	10.92
8) T Dichlorotetra...			1.306	1.283	1.274	1.214	1.211	1.257	3.40
9) T Propene			0.734	0.736	0.698	0.660	0.630	0.691	6.69
10) T Heptane			1.847	1.940	1.897	1.792	1.719	1.839	4.73
11) T Trichlorofluor...			1.562	1.457	1.480	1.339	1.346	1.437	6.59
12) T 1,1,2-Trichlor...			1.247	1.215	1.188	1.120	1.135	1.181	4.52
13) Ethanol			0.110	0.086	0.073	0.029	0.028	0.065	55.48
14) T Bromoethene			0.529	0.447	0.460	0.417	0.409	0.452	10.47
15) T Acetone			1.699	1.573	1.469	1.113	1.139	1.399	18.73
16) T 1,3-Butadiene			0.952	0.705	0.608	0.573	0.577	0.683	23.38
17) tert-Butyl alc...			1.935	1.799	1.756	1.618	1.601	1.742	7.91
18) T 1,1-Dichloroet...			0.624	0.565	0.560	0.504	0.507	0.552	8.99
19) T Isopropyl Alcohol			1.016	0.800	0.738	0.702	0.689	0.789	16.99
20) T Methylene Chlo...			0.624	0.563	0.545	0.438	0.435	0.521	15.85
21) T Allyl Chloride			1.067	1.022	1.020	0.902	0.921	0.986	7.25
22) T trans-1,2-Dich...			0.704	0.701	0.621	0.590	0.575	0.638	9.56
23) T Vinyl Acetate			0.754	0.692	0.550	0.498	0.532	0.605	18.40
24) T 1,1-Dichloroet...			1.247	1.208	1.187	1.122	1.149	1.183	4.17
25) T Ethyl Acetate			4.017	3.619	3.703	3.457	3.353	3.630	7.06
26) T Hexane			1.507	1.462	1.383	1.333	1.325	1.402	5.71
27) T Carbon Disulfide			1.608	1.512	1.538	1.436	1.423	1.503	5.07
28) T Methyl tert-Bu...			0.921	0.886	0.894	0.831	0.832	0.873	4.56
29) T Chloroform			2.233	2.068	2.071	1.932	1.956	2.052	5.81
30) T Cyclohexane			1.337	1.238	1.211	1.162	1.140	1.217	6.33
31) T cis-1,2-Dichlo...			1.507	1.460	1.426	1.371	1.353	1.423	4.44
32) T 1,1,1-Trichlor...	3.151	2.495	2.279	2.201	2.152	2.062	2.105	2.349	16.23

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.754	0.771	0.770	0.667	0.664	0.725	7.58
35) T Carbon Tetrach...	1.069	0.782	0.709	0.732	0.748	0.703	0.725	0.781	16.60
36) T Benzene			1.049	1.027	1.013	0.934	0.934	0.991	5.47
37) T 1,2-Dichloroet...			0.613	0.569	0.563	0.537	0.556	0.568	4.95
38) T Trichloroethene	0.463	0.437	0.438	0.408	0.398	0.374	0.378	0.414	8.00
39) T 1,2-Dichloropr...			0.386	0.373	0.365	0.337	0.344	0.361	5.68

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\										
Method File : VL032725AIR.M										
40) T	1,4-Dioxane			0.234	0.197	0.191	0.169	0.167	0.192	14.08
41) T	Tetrahydrofuran			0.439	0.425	0.432	0.398	0.393	0.418	4.98
42) T	Bromodichlorom...			0.778	0.779	0.802	0.769	0.785	0.782	1.58
43) T	Methyl Methacr...			0.442	0.451	0.408	0.403	0.394	0.420	6.03
44) T	2,2,4-Trimethy...			1.781	1.689	1.723	1.608	1.576	1.675	5.00
45) T	t-1,3-Dichloro...			0.474	0.447	0.453	0.439	0.440	0.450	3.15
46) T	cis-1,3-Dichlo...			0.549	0.560	0.566	0.550	0.535	0.552	2.12
47) T	1,1,2-Trichlor...			0.419	0.382	0.379	0.364	0.365	0.382	5.88
48) T	Dibromochlorom...			0.651	0.641	0.663	0.636	0.644	0.647	1.62
49) T	Bromoform			0.566	0.588	0.597	0.582	0.570	0.581	2.18
50) T	4-Methyl-2-Pen...			1.286	1.214	1.211	1.027	1.008	1.149	10.79
51) T	2-Hexanone			0.999	1.018	1.031	0.844	0.827	0.944	10.56
52) T	Tetrachloroethene	0.572	0.359	0.401	0.361	0.371	0.343	0.345	0.393	20.64
53) T	Toluene			1.239	1.203	1.202	1.125	1.108	1.175	4.76
54) T	1,2-Dibromoethane	0.574		0.543	0.538	0.520	0.508	0.510	0.532	4.74
-----ISTD-----										
55) I	Chlorobenzene-d5									
56) T	1,1,1,2-Tetrac...			0.576	0.575	0.576	0.510	0.515	0.550	6.30
57) T	Chlorobenzene			1.024	0.979	0.994	0.875	0.871	0.949	7.46
58) T	Ethyl Benzene			1.850	1.822	1.894	1.671	1.645	1.776	6.28
59) T	m/p-Xylene			1.505	1.472	1.484	1.321	1.309	1.418	6.71
60) T	o-Xylene			1.521	1.475	1.499	1.328	1.287	1.422	7.50
61) T	Styrene			0.709	0.722	0.735	0.683	0.668	0.703	3.92
62) T	Isopropylbenzene			2.290	2.199	2.261	1.962	1.920	2.126	8.14
63) T	1,1,2,2-Tetrac...	0.986	0.800	0.907	0.889	0.893	0.796	0.777	0.864	8.78
64) T	n-propylbenzene			0.578	0.581	0.576	0.499	0.502	0.547	7.81
65) T	tert-Butylbenzene			2.190	2.071	2.144	1.736	1.691	1.966	11.96
66) T	Benzyl Chloride			0.217	0.213	0.221	0.203	0.189	0.209	6.06
67) T	sec-Butylbenzene			3.005	2.994	2.947	2.404	2.316	2.733	12.54
68) S	1-Bromo-4-Fluo...	0.820	0.827	0.832	0.810	0.828	0.813	0.820	0.821	0.98
69) T	p-Isopropyltol...			2.554	2.551	2.560	2.078	2.025	2.354	11.74
70) T	n-Butylbenzene			2.536	2.598	2.682	2.137	2.082	2.407	11.51
71) T	2-Chlorotoluene			1.717	1.728	1.759	1.538	1.510	1.650	7.07
72) T	4-Ethyltoluene			1.834	1.859	1.929	1.629	1.601	1.770	8.26
73) T	1,3,5-Trimethy...			1.616	1.671	1.677	1.387	1.381	1.546	9.71
74) T	1,2,4-Trimethy...			1.994	1.958	1.966	1.510	1.478	1.781	14.74
75) T	1,3-Dichlorobe...			1.036	1.022	1.063	0.882	0.882	0.977	8.99
76) T	1,4-Dichlorobe...			1.030	1.017	1.040	0.888	0.886	0.972	8.05
77) T	1,2-Dichlorobe...			1.012	1.031	1.014	0.850	0.850	0.951	9.78
78) T	Hexachloro-1,3...			1.057	1.113	1.086	0.802	0.784	0.968	16.65
79) T	Naphthalene	1.637		1.979	2.181	2.314	1.806	1.773	1.948	13.31
80) T	Naphthalene,2-...			0.721	0.891	1.163	0.990	1.013	0.955	17.09
81) T	1,2,4-Trichlor...			0.821	0.941	1.028	0.859	0.843	0.899	9.51

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042172.D
 Acq On : 27 Mar 2025 08:26
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Quant Time: Mar 28 01:31:57 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	187002	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	545068	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	519536	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.371	95	422410	9.901	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.000%
Target Compounds	Qvalue					
2) Dichlorodifluoromethane	1.496	85	262496	9.507	ppbv	100
3) Chlorodifluoromethane	1.473	51	272522	9.662	ppbv	100
4) Chloromethane	1.531	50	107099	9.491	ppbv	100
5) Vinyl Chloride	1.577	62	102699	8.925	ppbv	100
6) Bromomethane	1.664	94	47950	9.377	ppbv	100
7) Chloroethane	1.703	64	42516	9.387	ppbv	100
8) Dichlorotetrafluoroethane	1.551	85	227041	9.656	ppbv	100
9) Propene	1.483	41	123386	9.709	ppbv	100
10) Heptane	4.972	43	335086	9.744	ppbv	100
11) Trichlorofluoromethane	1.874	101	250349	9.318	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.137	101	209365	9.481	ppbv	100
13) Ethanol	1.719	45	5338	5.364	ppbv	100
14) Bromoethene	1.777	108	77999	9.220	ppbv	100
15) Acetone	1.832	43	208061	7.955	ppbv	100
16) 1,3-Butadiene	1.606	39	107104	9.105	ppbv	100
17) tert-Butyl alcohol	2.036	59	302616	9.290	ppbv	100
18) 1,1-Dichloroethene	2.030	96	94236	9.128	ppbv	100
19) Isopropyl Alcohol	1.884	45	131359	8.902	ppbv	100
20) Methylene Chloride	2.059	84	81828	8.401	ppbv	100
21) Allyl Chloride	2.091	41	168602	9.141	ppbv	100
22) trans-1,2-Dichloroethene	2.337	96	110371	9.149	ppbv	100
23) Vinyl Acetate	2.460	43	93080	8.563	ppbv	100
24) 1,1-Dichloroethane	2.405	63	209753	9.484	ppbv	100
25) Ethyl Acetate	2.833	43	646501	9.524	ppbv	100
26) Hexane	2.823	57	249331	9.510	ppbv	100
27) Carbon Disulfide	2.146	76	268507	9.550	ppbv	100
28) Methyl tert-Butyl Ether	2.438	73	155398	9.520	ppbv	100
29) Chloroform	2.845	83	361264	9.415	ppbv	100
30) Cyclohexane	3.852	84	217215	9.537	ppbv	100
31) cis-1,2-Dichloroethene	2.716	61	256425	9.634	ppbv	100
32) 1,1,1-Trichloroethane	3.363	97	385628	8.784	ppbv	100
34) 2-Butanone	2.554	43	363735	9.203	ppbv	100
35) Carbon Tetrachloride	3.758	117	382925	9.129	ppbv	100
36) Benzene	3.651	78	508860	9.417	ppbv	100
37) 1,2-Dichloroethane	3.214	62	292937	9.469	ppbv	100
38) Trichloroethene	4.545	130	204027	9.045	ppbv	100
39) 1,2-Dichloropropane	4.312	63	183439	9.321	ppbv	100
40) 1,4-Dioxane	4.577	88	92327	9.042	ppbv #	100
41) Tetrahydrofuran	3.049	42	216851	9.643	ppbv	100
42) Bromodichloromethane	4.483	83	418887	9.822	ppbv	100
43) Methyl Methacrylate	4.872	69	219399	9.618	ppbv	100
44) 2,2,4-Trimethylpentane	4.635	57	876649	9.600	ppbv	100
45) t-1,3-Dichloropropene	6.406	75	239180	9.746	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042172.D
 Acq On : 27 Mar 2025 08:26
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Quant Time: Mar 28 01:31:57 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	300021	9.973	ppbv	100
47) 1,1,2-Trichloroethane	6.574	97	198176	9.526	ppbv	100
48) Dibromochloromethane	7.380	129	346499	9.828	ppbv	100
49) Bromoform	9.477	173	317189	10.021	ppbv	100
50) 4-Methyl-2-Pentanone	5.745	43	559800	8.934	ppbv	100
51) 2-Hexanone	7.432	43	460013	8.848	ppbv #	100
52) Tetrachloroethene	8.218	164	186912	8.722	ppbv	100
53) Toluene	6.927	91	613378	9.576	ppbv	100
54) 1,2-Dibromoethane	7.649	107	276797	9.536	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.917	131	265037	9.268	ppbv	100
57) Chlorobenzene	8.917	112	454511	9.223	ppbv	100
58) Ethyl Benzene	9.351	91	868003	9.406	ppbv	100
59) m/p-Xylene	9.549	91	1372806m	18.607	ppbv	
60) o-Xylene	9.960	91	689933	9.340	ppbv	100
61) Styrene	9.866	104	354697	9.706	ppbv	100
62) Isopropylbenzene	10.536	105	1019255	9.227	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.960	83	413392	9.210	ppbv	100
64) n-propylbenzene	10.986	120	259405	9.122	ppbv	100
65) tert-Butylbenzene	11.529	119	901951	8.828	ppbv	100
66) Benzyl Chloride	11.610	91	105411	9.724	ppbv	100
67) sec-Butylbenzene	11.766	105	1249106	8.796	ppbv	100
69) p-Isopropyltoluene	11.924	119	1079675	8.830	ppbv	100
70) n-Butylbenzene	12.277	91	1110351	8.879	ppbv	100
71) 2-Chlorotoluene	10.898	91	799102	9.320	ppbv	100
72) 4-Ethyltoluene	11.122	105	846232	9.200	ppbv	100
73) 1,3,5-Trimethylbenzene	11.199	105	720566	8.970	ppbv	100
74) 1,2,4-Trimethylbenzene	11.536	105	784745	8.480	ppbv	100
75) 1,3-Dichlorobenzene	11.604	146	458226	9.028	ppbv	100
76) 1,4-Dichlorobenzene	11.669	146	461371	9.134	ppbv	100
77) 1,2-Dichlorobenzene	11.944	146	441473	8.933	ppbv	100
78) Hexachloro-1,3-Butadiene	13.879	225	416735	8.284	ppbv	100
79) Naphthalene	13.510	128	938470	9.271	ppbv	100
80) Naphthalene,2-methyl-	14.455	142	514394	10.436	ppbv	100
81) 1,2,4-Trichlorobenzene	13.439	180	446420	9.561	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042172.D
Acq On : 27 Mar 2025 08:26
Operator : SY/MD
Sample : VSTDICCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:31:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042173.D
 Acq On : 27 Mar 2025 08:59
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:32:52 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	188395	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	546394	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	504538	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.370 95 417532 10.077 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.800%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	55859	2.008	ppbv	99
3) Chlorodifluoromethane	1.473	51	57045	2.007	ppbv	99
4) Chloromethane	1.531	50	23293	2.049	ppbv	94
5) Vinyl Chloride	1.576	62	21989	1.897	ppbv	99
6) Bromomethane	1.664	94	10027	1.946	ppbv	98
7) Chloroethane	1.703	64	8413	1.844	ppbv	92
8) Dichlorotetrafluoroethane	1.551	85	48008	2.027	ppbv	98
9) Propene	1.483	41	26295	2.054	ppbv	94
10) Heptane	4.972	43	71466	2.063	ppbv	99
11) Trichlorofluoromethane	1.874	101	55754	2.060	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.136	101	44761	2.012	ppbv	98
13) Ethanol	1.719	45	2748m	2.741	ppbv	
14) Bromoethene	1.780	108	17323	2.033	ppbv #	89
15) Acetone	1.832	43	55365	2.101	ppbv	95
16) 1,3-Butadiene	1.606	39	22894	1.932	ppbv	95
17) tert-Butyl alcohol	2.039	59	66174	2.016	ppbv	100
18) 1,1-Dichloroethene	2.030	96	21103	2.029	ppbv	95
19) Isopropyl Alcohol	1.884	45	27804	1.870	ppbv #	35
20) Methylene Chloride	2.059	84	20518	2.091	ppbv	96
21) Allyl Chloride	2.094	41	38425	2.068	ppbv	99
22) trans-1,2-Dichloroethene	2.337	96	23384	1.924	ppbv	92
23) Vinyl Acetate	2.460	43	20736	1.894	ppbv #	97
24) 1,1-Dichloroethane	2.405	63	44740	2.008	ppbv	99
25) Ethyl Acetate	2.832	43	139516	2.040	ppbv	99
26) Hexane	2.823	57	52095	1.972	ppbv	92
27) Carbon Disulfide	2.146	76	57969	2.047	ppbv #	76
28) Methyl tert-Butyl Ether	2.437	73	33703	2.049	ppbv	99
29) Chloroform	2.845	83	78043	2.019	ppbv	95
30) Cyclohexane	3.852	84	45641	1.989	ppbv	99
31) cis-1,2-Dichloroethene	2.719	61	53713	2.003	ppbv	96
32) 1,1,1-Trichloroethane	3.363	97	81070	1.833	ppbv	99
34) 2-Butanone	2.554	43	84157	2.124	ppbv	99
35) Carbon Tetrachloride	3.758	117	81735	1.944	ppbv	96
36) Benzene	3.651	78	110727	2.044	ppbv	97
37) 1,2-Dichloroethane	3.214	62	61476	1.982	ppbv	97
38) Trichloroethene	4.541	130	43501	1.924	ppbv	97
39) 1,2-Dichloropropane	4.308	63	39912	2.023	ppbv	93
40) 1,4-Dioxane	4.587	88	20884m	2.040	ppbv	
41) Tetrahydrofuran	3.056	42	47262	2.097	ppbv	99
42) Bromodichloromethane	4.483	83	87640	2.050	ppbv	99
43) Methyl Methacrylate	4.878	69	44624	1.952	ppbv	97
44) 2,2,4-Trimethylpentane	4.632	57	188282	2.057	ppbv	100
45) t-1,3-Dichloropropene	6.406	75	49461	2.011	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042173.D
 Acq On : 27 Mar 2025 08:59
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:32:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 01:31:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
46) cis-1,3-Dichloropropene	5.593	75	61829	2.050	ppbv	#	91
47) 1,1,2-Trichloroethane	6.577	97	41387	1.985	ppbv		95
48) Dibromochloromethane	7.383	129	72441	2.050	ppbv		100
49) Bromoform	9.477	173	65221	2.056	ppbv		98
50) 4-Methyl-2-Pentanone	5.752	43	132328	2.107	ppbv		99
51) 2-Hexanone	7.435	43	112681	2.162	ppbv	#	99
52) Tetrachloroethene	8.221	164	40558	1.888	ppbv		98
53) Toluene	6.923	91	131314	2.045	ppbv		99
54) 1,2-Dibromoethane	7.648	107	56799	1.952	ppbv		90
56) 1,1,1,2-Tetrachloroethane	8.920	131	58141	2.093	ppbv		97
57) Chlorobenzene	8.914	112	100308	2.096	ppbv	#	94
58) Ethyl Benzene	9.351	91	191114	2.132	ppbv		96
59) m/p-Xylene	9.548	91	299531m	4.181	ppbv		
60) o-Xylene	9.959	91	151292	2.109	ppbv		98
61) Styrene	9.865	104	74134	2.089	ppbv		98
62) Isopropylbenzene	10.535	105	228191	2.127	ppbv		99
63) 1,1,2,2-Tetrachloroethane	9.956	83	90114	2.067	ppbv		99
64) n-propylbenzene	10.985	120	58105	2.104	ppbv		92
65) tert-Butylbenzene	11.529	119	216336	2.180	ppbv		99
66) Benzyl Chloride	11.610	91	22272	2.116	ppbv		95
67) sec-Butylbenzene	11.762	105	297381	2.156	ppbv		99
69) p-Isopropyltoluene	11.921	119	258357	2.176	ppbv		100
70) n-Butylbenzene	12.277	91	270616	2.228	ppbv		99
71) 2-Chlorotoluene	10.898	91	177487	2.132	ppbv		100
72) 4-Ethyltoluene	11.121	105	194660	2.179	ppbv		99
73) 1,3,5-Trimethylbenzene	11.199	105	169221	2.169	ppbv		96
74) 1,2,4-Trimethylbenzene	11.532	105	198351	2.207	ppbv		96
75) 1,3-Dichlorobenzene	11.604	146	107217	2.175	ppbv		97
76) 1,4-Dichlorobenzene	11.668	146	104925	2.139	ppbv		99
77) 1,2-Dichlorobenzene	11.943	146	102310	2.132	ppbv		99
78) Hexachloro-1,3-Butadiene	13.879	225	109539	2.242	ppbv		99
79) Naphthalene	13.510	128	233451	2.375	ppbv		97
80) Naphthalene,2-methyl-	14.455	142	117313	2.451	ppbv		99
81) 1,2,4-Trichlorobenzene	13.435	180	103765	2.288	ppbv		97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042174.D
 Acq On : 27 Mar 2025 09:30
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Quant Time: Mar 28 01:33:49 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	186244	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	537244	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.879	117	499830	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.371	95	404753	9.861	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.600%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.496	85	28203	1.026	ppbv	99
3) Chlorodifluoromethane	1.473	51	28693	1.021	ppbv	94
4) Chloromethane	1.531	50	11111	0.989	ppbv	99
5) Vinyl Chloride	1.577	62	10723	0.936	ppbv	98
6) Bromomethane	1.667	94	4984	0.979	ppbv	97
7) Chloroethane	1.703	64	4622	1.025	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	23886	1.020	ppbv	98
9) Propene	1.483	41	13699m	1.082	ppbv	
10) Heptane	4.975	43	36132	1.055	ppbv	100
11) Trichlorofluoromethane	1.875	101	27136	1.014	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.137	101	22629	1.029	ppbv	99
13) Ethanol	1.719	45	1610m	1.624	ppbv	
14) Bromoethene	1.781	108	8330	0.989	ppbv #	82
15) Acetone	1.836	43	29294	1.125	ppbv	95
16) 1,3-Butadiene	1.606	39	13133	1.121	ppbv	97
17) tert-Butyl alcohol	2.040	59	33500	1.033	ppbv	97
18) 1,1-Dichloroethene	2.033	96	10532	1.024	ppbv	89
19) Isopropyl Alcohol	1.884	45	14903	1.014	ppbv #	38
20) Methylene Chloride	2.059	84	10481	1.080	ppbv	93
21) Allyl Chloride	2.095	41	19039	1.036	ppbv	99
22) trans-1,2-Dichloroethene	2.341	96	13051m	1.086	ppbv	
23) Vinyl Acetate	2.464	43	12883m	1.190	ppbv	
24) 1,1-Dichloroethane	2.409	63	22505	1.022	ppbv	99
25) Ethyl Acetate	2.836	43	67402	0.997	ppbv #	96
26) Hexane	2.823	57	27223	1.043	ppbv	98
27) Carbon Disulfide	2.146	76	28167	1.006	ppbv #	47
28) Methyl tert-Butyl Ether	2.441	73	16501	1.015	ppbv	95
29) Chloroform	2.845	83	38516	1.008	ppbv	96
30) Cyclohexane	3.852	84	23052	1.016	ppbv	97
31) cis-1,2-Dichloroethene	2.716	61	27187	1.026	ppbv	89
32) 1,1,1-Trichloroethane	3.363	97	40993	0.938	ppbv	99
34) 2-Butanone	2.557	43	41423	1.063	ppbv	98
35) Carbon Tetrachloride	3.758	117	39345	0.952	ppbv	93
36) Benzene	3.651	78	55156	1.036	ppbv	94
37) 1,2-Dichloroethane	3.218	62	30543	1.002	ppbv	98
38) Trichloroethene	4.545	130	21906	0.985	ppbv	93
39) 1,2-Dichloropropane	4.312	63	20020	1.032	ppbv	99
40) 1,4-Dioxane	4.593	88	10592m	1.052	ppbv	
41) Tetrahydrofuran	3.056	42	22818	1.029	ppbv	97
42) Bromodichloromethane	4.483	83	41864	0.996	ppbv	90
43) Methyl Methacrylate	4.875	69	24233m	1.078	ppbv	
44) 2,2,4-Trimethylpentane	4.632	57	90727	1.008	ppbv	96
45) t-1,3-Dichloropropene	6.403	75	23993m	0.992	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042174.D
 Acq On : 27 Mar 2025 09:30
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:33:49 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	30063	1.014	ppbv	93
47) 1,1,2-Trichloroethane	6.577	97	20546	1.002	ppbv	96
48) Dibromochloromethane	7.383	129	34425	0.991	ppbv	100
49) Bromoform	9.481	173	31610	1.013	ppbv	98
50) 4-Methyl-2-Pentanone	5.752	43	65226m	1.056	ppbv	
51) 2-Hexanone	7.438	43	54680	1.067	ppbv #	100
52) Tetrachloroethene	8.222	164	19415	0.919	ppbv	95
53) Toluene	6.927	91	64612	1.023	ppbv	99
54) 1,2-Dibromoethane	7.652	107	28879	1.009	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.924	131	28764	1.045	ppbv #	87
57) Chlorobenzene	8.917	112	48930	1.032	ppbv	93
58) Ethyl Benzene	9.348	91	91086	1.026	ppbv	97
59) m/p-Xylene	9.529	91	147162m	2.073	ppbv	
60) o-Xylene	9.960	91	73702	1.037	ppbv	96
61) Styrene	9.866	104	36110	1.027	ppbv	97
62) Isopropylbenzene	10.536	105	109897	1.034	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.960	83	44431	1.029	ppbv	99
64) n-propylbenzene	10.986	120	29054	1.062	ppbv	98
65) tert-Butylbenzene	11.526	119	103520	1.053	ppbv	96
66) Benzyl Chloride	11.614	91	10660	1.022	ppbv #	93
67) sec-Butylbenzene	11.762	105	149660	1.095	ppbv	97
69) p-Isopropyltoluene	11.921	119	127509	1.084	ppbv	99
70) n-Butylbenzene	12.280	91	129860	1.079	ppbv	99
71) 2-Chlorotoluene	10.898	91	86356	1.047	ppbv	100
72) 4-Ethyltoluene	11.122	105	92912	1.050	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	83497	1.080	ppbv	92
74) 1,2,4-Trimethylbenzene	11.533	105	97858	1.099	ppbv	94
75) 1,3-Dichlorobenzene	11.604	146	51096	1.046	ppbv	99
76) 1,4-Dichlorobenzene	11.669	146	50841	1.046	ppbv	99
77) 1,2-Dichlorobenzene	11.944	146	51523	1.084	ppbv	99
78) Hexachloro-1,3-Butadiene	13.879	225	55611	1.149	ppbv	97
79) Naphthalene	13.510	128	109031	1.120	ppbv	99
80) Naphthalene,2-methyl-	14.455	142	44518	0.939	ppbv	99
81) 1,2,4-Trichlorobenzene	13.436	180	47040	1.047	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042174.D
Acq On : 27 Mar 2025 09:30
Operator : SY/MD
Sample : VSTDIC001
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:33:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042175.D
 Acq On : 27 Mar 2025 10:01
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:34:42 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	180502	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	523832	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	485078	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.370 95 403513 10.130 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 101.300%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	14455	0.542	ppbv	95
3) Chlorodifluoromethane	1.473	51	14637	0.538	ppbv #	89
4) Chloromethane	1.531	50	5956	0.547	ppbv	94
5) Vinyl Chloride	1.577	62	5732	0.516	ppbv	92
6) Bromomethane	1.664	94	2949	0.597	ppbv	95
7) Chloroethane	1.703	64	2582	0.591	ppbv #	87
8) Dichlorotetrafluoroethane	1.551	85	11783	0.519	ppbv	98
9) Propene	1.483	41	6622m	0.540	ppbv	
10) Heptane	4.972	43	16673	0.502	ppbv	89
11) Trichlorofluoromethane	1.874	101	14099	0.544	ppbv	92
12) 1,1,2-Trichlorotrifluo...	2.136	101	11252	0.528	ppbv	99
13) Ethanol	1.722	45	997	1.038	ppbv #	36
14) Bromoethene	1.780	108	4770	0.584	ppbv #	86
15) Acetone	1.835	43	15331	0.607	ppbv #	85
16) 1,3-Butadiene	1.606	39	8594	0.757	ppbv	82
17) tert-Butyl alcohol	2.039	59	17467	0.556	ppbv	97
18) 1,1-Dichloroethene	2.030	96	5634	0.565	ppbv #	92
19) Isopropyl Alcohol	1.884	45	9171	0.644	ppbv #	39
20) Methylene Chloride	2.059	84	5634	0.599	ppbv #	93
21) Allyl Chloride	2.091	41	9632	0.541	ppbv	89
22) trans-1,2-Dichloroethene	2.334	96	6353	0.546	ppbv	83
23) Vinyl Acetate	2.457	43	6809	0.649	ppbv #	91
24) 1,1-Dichloroethane	2.405	63	11258	0.527	ppbv	97
25) Ethyl Acetate	2.836	43	36258	0.553	ppbv	98
26) Hexane	2.823	57	13605	0.538	ppbv	92
27) Carbon Disulfide	2.146	76	14510	0.535	ppbv #	1
28) Methyl tert-Butyl Ether	2.437	73	8310	0.527	ppbv #	89
29) Chloroform	2.845	83	20149	0.544	ppbv	98
30) Cyclohexane	3.852	84	12063m	0.549	ppbv	
31) cis-1,2-Dichloroethene	2.719	61	13601	0.529	ppbv	92
32) 1,1,1-Trichloroethane	3.360	97	20567	0.485	ppbv	100
34) 2-Butanone	2.557	43	19739	0.520	ppbv	96
35) Carbon Tetrachloride	3.755	117	18558	0.460	ppbv	93
36) Benzene	3.651	78	27488	0.529	ppbv #	87
37) 1,2-Dichloroethane	3.214	62	16062	0.540	ppbv	100
38) Trichloroethene	4.538	130	11460m	0.529	ppbv	
39) 1,2-Dichloropropane	4.308	63	10123	0.535	ppbv	92
40) 1,4-Dioxane	4.593	88	6121m	0.624	ppbv	
41) Tetrahydrofuran	3.059	42	11509m	0.533	ppbv	
42) Bromodichloromethane	4.483	83	20374	0.497	ppbv #	98
43) Methyl Methacrylate	4.875	69	11583m	0.528	ppbv	
44) 2,2,4-Trimethylpentane	4.632	57	46646	0.532	ppbv	95
45) t-1,3-Dichloropropene	6.406	75	12404	0.526	ppbv #	79

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042175.D
 Acq On : 27 Mar 2025 10:01
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
 Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:34:42 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	14374m	0.497	ppbv	
47) 1,1,2-Trichloroethane	6.574	97	10972m	0.549	ppbv	
48) Dibromochloromethane	7.380	129	17042	0.503	ppbv	99
49) Bromoform	9.480	173	14828	0.487	ppbv	98
50) 4-Methyl-2-Pentanone	5.758	43	33674	0.559	ppbv	97
51) 2-Hexanone	7.438	43	26154	0.523	ppbv #	100
52) Tetrachloroethene	8.218	164	10502	0.510	ppbv	94
53) Toluene	6.923	91	32444	0.527	ppbv	98
54) 1,2-Dibromoethane	7.648	107	14220	0.510	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.920	131	13959	0.523	ppbv #	1
57) Chlorobenzene	8.914	112	24826	0.540	ppbv #	91
58) Ethyl Benzene	9.348	91	44862	0.521	ppbv	96
59) m/p-Xylene	9.542	91	73028m	1.060	ppbv	
60) o-Xylene	9.959	91	36879	0.535	ppbv	98
61) Styrene	9.866	104	17198	0.504	ppbv	96
62) Isopropylbenzene	10.536	105	55531	0.538	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.959	83	21989	0.525	ppbv	96
64) n-propylbenzene	10.985	120	14030	0.528	ppbv	95
65) tert-Butylbenzene	11.526	119	53115	0.557	ppbv	100
66) Benzyl Chloride	11.613	91	5263	0.520	ppbv	97
67) sec-Butylbenzene	11.762	105	72879	0.550	ppbv	98
69) p-Isopropyltoluene	11.921	119	61933	0.542	ppbv	100
70) n-Butylbenzene	12.277	91	61507	0.527	ppbv	99
71) 2-Chlorotoluene	10.895	91	41637	0.520	ppbv	95
72) 4-Ethyltoluene	11.121	105	44473	0.518	ppbv #	94
73) 1,3,5-Trimethylbenzene	11.199	105	39190	0.523	ppbv	97
74) 1,2,4-Trimethylbenzene	11.532	105	48370	0.560	ppbv	97
75) 1,3-Dichlorobenzene	11.600	146	25122	0.530	ppbv	100
76) 1,4-Dichlorobenzene	11.668	146	24991	0.530	ppbv	97
77) 1,2-Dichlorobenzene	11.943	146	24548	0.532	ppbv	96
78) Hexachloro-1,3-Butadiene	13.879	225	25636	0.546	ppbv	99
79) Naphthalene	13.510	128	47990	0.508	ppbv	95
80) Naphthalene,2-methyl-	14.458	142	17486	0.380	ppbv	98
81) 1,2,4-Trichlorobenzene	13.436	180	19924	0.457	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042175.D
Acq On : 27 Mar 2025 10:01
Operator : SY/MD
Sample : VSTDICC0.5
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:34:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042176.D
 Acq On : 27 Mar 2025 10:32
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
 Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:35:34 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	180740	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	528028	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	481506	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	398069	10.067	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.700%
Target Compounds						
5) Vinyl Chloride	1.580	62	1274	0.115	ppbv #	42
32) 1,1,1-Trichloroethane	3.366	97	4509m	0.106	ppbv	
35) Carbon Tetrachloride	3.768	117	4131m	0.102	ppbv	
38) Trichloroethene	4.551	130	2306m	0.106	ppbv	
52) Tetrachloroethene	8.225	164	1895m	0.091	ppbv	
54) 1,2-Dibromoethane	7.652	107	3033m	0.108	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.966	83	3852m	0.093	ppbv	
79) Naphthalene	13.513	128	7882	0.084	ppbv #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042176.D
Acq On : 27 Mar 2025 10:32
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:35:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042177.D
 Acq On : 27 Mar 2025 11:03
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
 Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:36:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 01:31:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	178058	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	512915	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	475617	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	389935	9.983	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.800%
Target Compounds						
5) Vinyl Chloride	1.580	62	471m	0.043	ppbv	Qvalue
32) 1,1,1-Trichloroethane	3.366	97	1683m	0.040	ppbv	
35) Carbon Tetrachloride	3.758	117	1645m	0.042	ppbv	
38) Trichloroethene	4.541	130	712m	0.034	ppbv	
52) Tetrachloroethene	8.208	164	880m	0.044	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.963	83	1407m	0.034	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042177.D
Acq On : 27 Mar 2025 11:03
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Mar 28 01:36:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
Supervised By : Mahesh Dadoda 04/03/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042178.D
 Acq On : 27 Mar 2025 11:36
 Operator : SY/MD
 Sample : VSTDICC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Quant Time: Mar 28 01:37:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 01:31:23 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	173372	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	499360	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	473077	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	387777	9.981	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.800%

Target Compounds	Qvalue					
2) Dichlorodifluoromethane	1.495	85	358962	14.023	ppbv	99
3) Chlorodifluoromethane	1.473	51	366164	14.002	ppbv	98
4) Chloromethane	1.531	50	148133	14.160	ppbv	98
5) Vinyl Chloride	1.576	62	141938	13.304	ppbv	97
6) Bromomethane	1.667	94	65113	13.734	ppbv	99
7) Chloroethane	1.703	64	58820	14.007	ppbv	99
8) Dichlorotetrafluoroethane	1.550	85	314834	14.442	ppbv	98
9) Propene	1.482	41	163799	13.902	ppbv	95
10) Heptane	4.972	43	446949	14.019	ppbv	99
11) Trichlorofluoromethane	1.874	101	349967	14.050	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.136	101	295217	14.419	ppbv	99
13) Ethanol	1.719	45	7382	8.001	ppbv	92
14) Bromoethene	1.780	108	106436	13.571	ppbv	98
15) Acetone	1.832	43	296254	12.218	ppbv	98
16) 1,3-Butadiene	1.605	39	150122	13.766	ppbv	100
17) tert-Butyl alcohol	2.039	59	416362	13.787	ppbv	99
18) 1,1-Dichloroethene	2.029	96	131720	13.762	ppbv	99
19) Isopropyl Alcohol	1.884	45	179117	13.093	ppbv	100
20) Methylene Chloride	2.059	84	113183	12.533	ppbv	95
21) Allyl Chloride	2.094	41	239440	14.002	ppbv	99
22) trans-1,2-Dichloroethene	2.337	96	149455	13.362	ppbv	98
23) Vinyl Acetate	2.460	43	138419	13.736	ppbv #	96
24) 1,1-Dichloroethane	2.405	63	298760	14.570	ppbv	99
25) Ethyl Acetate	2.832	43	871882	13.855	ppbv	100
26) Hexane	2.822	57	344636	14.178	ppbv	98
27) Carbon Disulfide	2.146	76	370022	14.196	ppbv	97
28) Methyl tert-Butyl Ether	2.437	73	216463	14.303	ppbv	95
29) Chloroform	2.845	83	508614	14.297	ppbv	100
30) Cyclohexane	3.852	84	296458	14.040	ppbv	99
31) cis-1,2-Dichloroethene	2.719	61	351945	14.262	ppbv	97
32) 1,1,1-Trichloroethane	3.363	97	547329	13.448	ppbv	99
34) 2-Butanone	2.554	43	497020	13.726	ppbv	99
35) Carbon Tetrachloride	3.758	117	543385	14.140	ppbv	100
36) Benzene	3.651	78	699549	14.131	ppbv	98
37) 1,2-Dichloroethane	3.214	62	416544	14.697	ppbv	99
38) Trichloroethene	4.541	130	283430	13.715	ppbv	98
39) 1,2-Dichloropropane	4.308	63	257891	14.304	ppbv	99
40) 1,4-Dioxane	4.577	88	124925	13.354	ppbv #	91
41) Tetrahydrofuran	3.049	42	294634	14.301	ppbv	100
42) Bromodichloromethane	4.483	83	587772	15.043	ppbv	99
43) Methyl Methacrylate	4.874	69	295300	14.130	ppbv	99
44) 2,2,4-Trimethylpentane	4.635	57	1180275	14.108	ppbv	99
45) t-1,3-Dichloropropene	6.405	75	329360	14.649	ppbv	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042178.D
 Acq On : 27 Mar 2025 11:36
 Operator : SY/MD
 Sample : VSTDICC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:37:01 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	400813	14.542	ppbv	98
47) 1,1,2-Trichloroethane	6.577	97	273137	14.331	ppbv	98
48) Dibromochloromethane	7.383	129	482431	14.936	ppbv	100
49) Bromoform	9.480	173	427143	14.730	ppbv	99
50) 4-Methyl-2-Pentanone	5.745	43	755146	13.155	ppbv	99
51) 2-Hexanone	7.431	43	619419	13.005	ppbv #	99
52) Tetrachloroethene	8.221	164	258356	13.159	ppbv	95
53) Toluene	6.926	91	829648	14.137	ppbv	100
54) 1,2-Dibromoethane	7.645	107	381823	14.359	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.920	131	365410	14.032	ppbv	99
57) Chlorobenzene	8.917	112	618201	13.777	ppbv	100
58) Ethyl Benzene	9.351	91	1167096	13.889	ppbv	97
59) m/p-Xylene	9.548	91	1857561m	27.651	ppbv	
60) o-Xylene	9.962	91	913234	13.577	ppbv	99
61) Styrene	9.865	104	474195	14.250	ppbv	99
62) Isopropylbenzene	10.535	105	1362416	13.544	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.959	83	551386	13.491	ppbv	99
64) n-propylbenzene	10.985	120	356118	13.753	ppbv	94
65) tert-Butylbenzene	11.529	119	1200149	12.901	ppbv	98
66) Benzyl Chloride	11.613	91	134397	13.615	ppbv	100
67) sec-Butylbenzene	11.765	105	1643629	12.711	ppbv	99
69) p-Isopropyltoluene	11.924	119	1437016	12.906	ppbv	98
70) n-Butylbenzene	12.280	91	1477443	12.975	ppbv	99
71) 2-Chlorotoluene	10.898	91	1071432	13.724	ppbv	99
72) 4-Ethyltoluene	11.121	105	1136415	13.569	ppbv	99
73) 1,3,5-Trimethylbenzene	11.202	105	979783	13.395	ppbv	99
74) 1,2,4-Trimethylbenzene	11.535	105	1048874	12.447	ppbv	92
75) 1,3-Dichlorobenzene	11.607	146	626015	13.545	ppbv	99
76) 1,4-Dichlorobenzene	11.668	146	628694	13.668	ppbv	100
77) 1,2-Dichlorobenzene	11.947	146	602844	13.396	ppbv	100
78) Hexachloro-1,3-Butadiene	13.882	225	556392	12.147	ppbv	99
79) Naphthalene	13.513	128	1258479	13.653	ppbv	98
80) Naphthalene,2-methyl-	14.455	142	718900	16.018	ppbv	98
81) 1,2,4-Trichlorobenzene	13.439	180	598416	14.075	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042178.D
Acq On : 27 Mar 2025 11:36
Operator : SY/MD
Sample : VSTDIC015
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:37:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042179.D
 Acq On : 27 Mar 2025 12:23
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL032725

Quant Time: Mar 28 02:05:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	174350	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	498105	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	473357	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	379929	9.774	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.700%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.496	85	245714	9.545	ppbv	100
3) Chlorodifluoromethane	1.473	51	247046	9.394	ppbv	97
4) Chloromethane	1.531	50	98527	9.365	ppbv	98
5) Vinyl Chloride	1.577	62	95578	8.573	ppbv	97
6) Bromomethane	1.667	94	41950	8.799	ppbv	97
7) Chloroethane	1.703	64	37400	8.856	ppbv	94
8) Dichlorotetrafluoroethane	1.551	85	213819	9.753	ppbv	95
9) Propene	1.483	41	105338	8.739	ppbv	99
10) Heptane	4.972	43	298768	9.318	ppbv	100
11) Trichlorofluoromethane	1.874	101	236793	9.453	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.136	101	193965	9.421	ppbv	98
13) Ethanol	1.719	45	5773	5.066	ppbv	97
14) Bromoethene	1.780	108	73572	9.328	ppbv	97
15) Acetone	1.832	43	198300	8.132	ppbv	98
16) 1,3-Butadiene	1.606	39	97840	8.216	ppbv	99
17) tert-Butyl alcohol	2.039	59	276195	9.094	ppbv	99
18) 1,1-Dichloroethene	2.033	96	87137	9.053	ppbv	89
19) Isopropyl Alcohol	1.884	45	122768	8.923	ppbv	98
20) Methylene Chloride	2.059	84	72813	8.018	ppbv	97
21) Allyl Chloride	2.094	41	157418	9.154	ppbv	99
22) trans-1,2-Dichloroethene	2.337	96	101141	9.092	ppbv	93
23) Vinyl Acetate	2.460	43	90260	8.553	ppbv #	97
24) 1,1-Dichloroethane	2.405	63	197968	9.600	ppbv	99
25) Ethyl Acetate	2.832	43	592585	9.364	ppbv	99
26) Hexane	2.823	57	231125	9.455	ppbv	99
27) Carbon Disulfide	2.146	76	248834	9.493	ppbv	97
28) Methyl tert-Butyl Ether	2.437	73	146470	9.624	ppbv	95
29) Chloroform	2.845	83	344106	9.619	ppbv	100
30) Cyclohexane	3.852	84	197897	9.323	ppbv	98
31) cis-1,2-Dichloroethene	2.719	61	234487	9.449	ppbv	96
32) 1,1,1-Trichloroethane	3.363	97	365364	8.921	ppbv	99
34) 2-Butanone	2.554	43	331236	9.171	ppbv	99
35) Carbon Tetrachloride	3.758	117	363392	9.339	ppbv	98
36) Benzene	3.651	78	461261	9.341	ppbv	99
37) 1,2-Dichloroethane	3.217	62	273559	9.676	ppbv	99
38) Trichloroethene	4.541	130	191746	9.306	ppbv	97
39) 1,2-Dichloropropane	4.311	63	170839	9.500	ppbv	95
40) 1,4-Dioxane	4.577	88	80592	8.443	ppbv #	97
41) Tetrahydrofuran	3.052	42	196163	9.431	ppbv	99
42) Bromodichloromethane	4.486	83	384006	9.853	ppbv	100
43) Methyl Methacrylate	4.878	69	195339	9.344	ppbv	99
44) 2,2,4-Trimethylpentane	4.635	57	787400	9.436	ppbv	99
45) t-1,3-Dichloropropene	6.409	75	212335	9.467	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042179.D
 Acq On : 27 Mar 2025 12:23
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL032725

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 02:05:15 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	265101	9.643	ppbv	98
47) 1,1,2-Trichloroethane	6.577	97	183228	9.638	ppbv	98
48) Dibromochloromethane	7.383	129	322766	10.018	ppbv	100
49) Bromoform	9.480	173	285149	9.858	ppbv	99
50) 4-Methyl-2-Pentanone	5.749	43	500791	8.749	ppbv	99
51) 2-Hexanone	7.432	43	410889	8.741	ppbv #	99
52) Tetrachloroethene	8.221	164	173420	8.855	ppbv	93
53) Toluene	6.927	91	556897	9.514	ppbv	99
54) 1,2-Dibromoethane	7.645	107	256930	9.695	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.920	131	249270	9.567	ppbv	99
57) Chlorobenzene	8.917	112	422531	9.411	ppbv	98
58) Ethyl Benzene	9.351	91	791360	9.412	ppbv	98
59) m/p-Xylene	9.548	91	1260340m	18.772	ppbv	
60) o-Xylene	9.959	91	632918	9.404	ppbv	99
61) Styrene	9.866	104	315910	9.488	ppbv	99
62) Isopropylbenzene	10.539	105	944938	9.388	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.959	83	382635	9.357	ppbv	99
64) n-propylbenzene	10.985	120	244547	9.439	ppbv	98
65) tert-Butylbenzene	11.529	119	843828	9.065	ppbv	98
66) Benzyl Chloride	11.613	91	86804	8.789	ppbv	97
67) sec-Butylbenzene	11.765	105	1172636	9.063	ppbv	100
69) p-Isopropyltoluene	11.924	119	1019694	9.153	ppbv	100
70) n-Butylbenzene	12.280	91	1038548	9.115	ppbv	100
71) 2-Chlorotoluene	10.898	91	732874	9.382	ppbv	100
72) 4-Ethyltoluene	11.125	105	785185	9.370	ppbv	99
73) 1,3,5-Trimethylbenzene	11.202	105	673541	9.203	ppbv	98
74) 1,2,4-Trimethylbenzene	11.536	105	729155	8.648	ppbv	94
75) 1,3-Dichlorobenzene	11.607	146	426573	9.224	ppbv	99
76) 1,4-Dichlorobenzene	11.672	146	426318	9.263	ppbv	100
77) 1,2-Dichlorobenzene	11.947	146	412180	9.154	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	356386	7.776	ppbv	100
79) Naphthalene	13.513	128	773311	8.385	ppbv	99
80) Naphthalene,2-methyl-	14.455	142	323443	7.151	ppbv	99
81) 1,2,4-Trichlorobenzene	13.439	180	363863	8.553	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042179.D
Acq On : 27 Mar 2025 12:23
Operator : SY/MD
Sample : VSTDICV010
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
ICVVL032725

Quant Time: Mar 28 02:05:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 02:00:52 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
Supervised By :Mahesh Dadoda 04/03/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042179.D
 Acq On : 27 Mar 2025 12:23
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL032725

Quant Time: Mar 28 02:05:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	1.476	1.409	4.5	94	0.00
3	Chlorodifluoromethane	1.508	1.417	6.0	91	0.00
4	Chloromethane	0.603	0.565	6.3	92	0.00
5 T	Vinyl Chloride	0.639	0.548	14.2	93	0.00
6 T	Bromomethane	0.273	0.241	11.7	87	0.00
7	Chloroethane	0.242	0.215	11.2	88	0.00
8 T	Dichlorotetrafluoroethane	1.257	1.226	2.5	94	0.00
9 T	Propene	0.691	0.604	12.6	85	0.00
10 T	Heptane	1.839	1.714	6.8	89	0.00
11 T	Trichlorofluoromethane	1.437	1.358	5.5	95	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.181	1.113	5.8	93	0.00
13	Ethanol	0.065	0.033#	49.2#	108	0.00
14 T	Bromoethene	0.452	0.422	6.6	94	0.00
15 T	Acetone	1.399	1.137	18.7	95	0.00
16 T	1,3-Butadiene	0.683	0.561	17.9	91	0.00
17	tert-Butyl alcohol	1.742	1.584	9.1	91	0.00
18 T	1,1-Dichloroethene	0.552	0.500	9.4	92	0.00
19 T	Isopropyl Alcohol	0.789	0.704	10.8	93	0.00
20 T	Methylene Chloride	0.521	0.418	19.8	89	0.00
21 T	Allyl Chloride	0.986	0.903	8.4	93	0.00
22 T	trans-1,2-Dichloroethene	0.638	0.580	9.1	92	0.00
23 T	Vinyl Acetate	0.605	0.518	14.4	97	0.00
24 T	1,1-Dichloroethane	1.183	1.135	4.1	94	0.00
25 T	Ethyl Acetate	3.630	3.399	6.4	92	0.00
26 T	Hexane	1.402	1.326	5.4	93	0.00
27 T	Carbon Disulfide	1.503	1.427	5.1	93	0.00
28 T	Methyl tert-Butyl Ether	0.873	0.840	3.8	94	0.00
29 T	Chloroform	2.052	1.974	3.8	95	0.00
30 T	Cyclohexane	1.217	1.135	6.7	91	0.00
31 T	cis-1,2-Dichloroethene	1.423	1.345	5.5	91	0.00
32 T	1,1,1-Trichloroethane	2.349	2.096	10.8	95	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
34 T	2-Butanone	0.725	0.665	8.3	91	0.00
35 T	Carbon Tetrachloride	0.781	0.730	6.5	95	0.00
36 T	Benzene	0.991	0.926	6.6	91	0.00
37 T	1,2-Dichloroethane	0.568	0.549	3.3	93	0.00
38 T	Trichloroethene	0.414	0.385	7.0	94	0.00
39 T	1,2-Dichloropropane	0.361	0.343	5.0	93	0.00
40 T	1,4-Dioxane	0.192	0.162	15.6	87	0.00
41 T	Tetrahydrofuran	0.418	0.394	5.7	90	0.00
42 T	Bromodichloromethane	0.782	0.771	1.4	92	0.00
43	Methyl Methacrylate	0.420	0.392	6.7	89	0.00
44 T	2,2,4-Trimethylpentane	1.675	1.581	5.6	90	0.00
45 T	t-1,3-Dichloropropene	0.450	0.426	5.3	89	0.00
46 T	cis-1,3-Dichloropropene	0.552	0.532	3.6	88	0.00
47 T	1,1,2-Trichloroethane	0.382	0.368	3.7	92	0.00
48 T	Dibromochloromethane	0.647	0.648	-0.2	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042179.D
 Acq On : 27 Mar 2025 12:23
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL032725

Quant Time: Mar 28 02:05:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.581	0.572	1.5	90	0.00
50 T	4-Methyl-2-Pentanone	1.149	1.005	12.5	89	0.00
51 T	2-Hexanone	0.944	0.825	12.6	89	0.00
52 T	Tetrachloroethene	0.393	0.348	11.5	93	0.00
53 T	Toluene	1.175	1.118	4.9	91	0.00
54 T	1,2-Dibromoethane	0.532	0.516	3.0	93	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
56	1,1,1,2-Tetrachloroethane	0.550	0.527	4.2	94	0.00
57 T	Chlorobenzene	0.949	0.893	5.9	93	0.00
58 T	Ethyl Benzene	1.776	1.672	5.9	91	0.00
59 T	m/p-Xylene	1.418	1.331	6.1	92	0.00
60 T	o-Xylene	1.422	1.337	6.0	92	0.00
61 T	Styrene	0.703	0.667	5.1	89	0.00
62	Isopropylbenzene	2.126	1.996	6.1	93	0.00
63 T	1,1,2,2-Tetrachloroethane	0.864	0.808	6.5	93	0.00
64	n-propylbenzene	0.547	0.517	5.5	94	0.00
65	tert-Butylbenzene	1.966	1.783	9.3	94	0.00
66 T	Benzyl Chloride	0.209	0.183	12.4	82	0.00
67	sec-Butylbenzene	2.733	2.477	9.4	94	0.00
68 S	1-Bromo-4-Fluorobenzene	0.821	0.803	2.2	90	0.00
69	p-Isopropyltoluene	2.354	2.154	8.5	94	0.00
70	n-Butylbenzene	2.407	2.194	8.8	94	0.00
71	2-Chlorotoluene	1.650	1.548	6.2	92	0.00
72 T	4-Ethyltoluene	1.770	1.659	6.3	93	0.00
73 T	1,3,5-Trimethylbenzene	1.546	1.423	8.0	93	0.00
74 T	1,2,4-Trimethylbenzene	1.781	1.540	13.5	93	0.00
75 T	1,3-Dichlorobenzene	0.977	0.901	7.8	93	0.00
76 T	1,4-Dichlorobenzene	0.972	0.901	7.3	92	0.00
77 T	1,2-Dichlorobenzene	0.951	0.871	8.4	93	0.00
78 T	Hexachloro-1,3-Butadiene	0.968	0.753	22.2	86	0.00
79 T	Naphthalene	1.948	1.634	16.1	82	0.00
80 T	Naphthalene,2-methyl-	0.955	0.683	28.5	63	0.00
81 T	1,2,4-Trichlorobenzene	0.899	0.769	14.5	82	0.00

(#) = Out of Range



SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
Data File : VL042325.D
Acq On : 15 Apr 2025 07:45
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri Mar 28 02:00:52 2025



AutoFind: Scans 3185, 3186, 3187; Background Corrected with Scan 3171

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	24.0	62363	PASS
75	95	30	66	55.5	144299	PASS
95	95	100	100	100.0	260117	PASS
96	95	5	9	6.7	17331	PASS
173	174	0.00	2	1.3	2038	PASS
174	95	50	120	62.4	162261	PASS
175	174	4	9	6.8	11092	PASS
176	174	93	101	94.8	153792	PASS
177	176	5	9	6.2	9603	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042326.D
 Acq On : 15 Apr 2025 11:30
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/16/2025
 Supervised By : Mahesh Dadoda 04/16/2025

Quant Time: Apr 16 01:29:10 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.806	49	172263	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	3.988	114	457227	10.000	ppbv	0.04
55) Chlorobenzene-d5	8.911	117	411493	10.000	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.403 95 325960 9.646 ppbv 0.03
 Spiked Amount 10.000 Range 65 - 135 Recovery = 96.500%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.505	85	214541	8.435	ppbv	99
3) Chlorodifluoromethane	1.483	51	235250	9.054	ppbv	98
4) Chloromethane	1.541	50	89653	8.625	ppbv	98
5) Vinyl Chloride	1.589	62	83204	7.554	ppbv	96
6) Bromomethane	1.677	94	36297	7.705	ppbv	99
7) Chloroethane	1.716	64	32145	7.704	ppbv	97
8) Dichlorotetrafluoroethane	1.564	85	182884	8.443	ppbv	94
9) Propene	1.492	41	101034	8.483	ppbv	98
10) Heptane	5.020	43	276748	8.736	ppbv	98
11) Trichlorofluoromethane	1.887	101	202387	8.178	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.153	101	157030	7.719	ppbv	98
13) Ethanol	1.732	45	4440	3.944	ppbv	93
14) Bromoethene	1.793	108	59588	7.646	ppbv	99
15) Acetone	1.845	43	183439	7.614	ppbv	96
16) 1,3-Butadiene	1.619	39	90954	7.731	ppbv	100
17) tert-Butyl alcohol	2.052	59	209737	6.990	ppbv	100
18) 1,1-Dichloroethene	2.046	96	72322	7.605	ppbv	96
19) Isopropyl Alcohol	1.897	45	104424	7.682	ppbv	98
20) Methylene Chloride	2.075	84	60337	6.725	ppbv	90
21) Allyl Chloride	2.111	41	132956	7.825	ppbv	98
22) trans-1,2-Dichloroethene	2.357	96	81572	7.422	ppbv	96
23) Vinyl Acetate	2.480	43	72160	6.920	ppbv	97
24) 1,1-Dichloroethane	2.425	63	169411	8.315	ppbv	100
25) Ethyl Acetate	2.855	43	564611	9.030	ppbv	99
26) Hexane	2.845	57	213782	8.851	ppbv	98
27) Carbon Disulfide	2.166	76	195452	7.547	ppbv	97
28) Methyl tert-Butyl Ether	2.457	73	116301	7.734	ppbv	96
29) Chloroform	2.868	83	314164	8.888	ppbv	100
30) Cyclohexane	3.884	84	173360	8.266	ppbv	98
31) cis-1,2-Dichloroethene	2.739	61	208160	8.490	ppbv	96
32) 1,1,1-Trichloroethane	3.392	97	311758	7.704	ppbv	98
34) 2-Butanone	2.577	43	317117	9.565	ppbv	97
35) Carbon Tetrachloride	3.790	117	317942	8.902	ppbv	99
36) Benzene	3.684	78	424254	9.360	ppbv	94
37) 1,2-Dichloroethane	3.243	62	239556	9.231	ppbv	100
38) Trichloroethene	4.583	130	172146	9.102	ppbv	96
39) 1,2-Dichloropropane	4.350	63	151707	9.190	ppbv	96
40) 1,4-Dioxane	4.622	88	72075	8.226	ppbv #	98
41) Tetrahydrofuran	3.075	42	181471	9.505	ppbv	95
42) Bromodichloromethane	4.525	83	333167	9.313	ppbv	97
43) Methyl Methacrylate	4.923	69	155399	8.098	ppbv	96
44) 2,2,4-Trimethylpentane	4.677	57	728571	9.511	ppbv	99
45) t-1,3-Dichloropropene	6.461	75	151063	7.338	ppbv	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042326.D
 Acq On : 15 Apr 2025 11:30
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/16/2025
 Supervised By : Mahesh Dadoda 04/16/2025

Quant Time: Apr 16 01:29:10 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.648	75	204339	8.097	ppbv	98
47) 1,1,2-Trichloroethane	6.626	97	163132	9.348	ppbv	97
48) Dibromochloromethane	7.425	129	272592	9.217	ppbv	99
49) Bromoform	9.513	173	238880	8.997	ppbv	100
50) 4-Methyl-2-Pentanone	5.800	43	442851	8.428	ppbv	97
51) 2-Hexanone	7.474	43	353471	8.192	ppbv #	97
52) Tetrachloroethene	8.260	164	144481	8.037	ppbv	98
53) Toluene	6.975	91	461854	8.595	ppbv	100
54) 1,2-Dibromoethane	7.687	107	220075	9.047	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.956	131	214175	9.455	ppbv	99
57) Chlorobenzene	8.953	112	375357	9.617	ppbv	97
58) Ethyl Benzene	9.383	91	639857	8.754	ppbv	99
59) m/p-Xylene	9.581	91	1063509m	18.222	ppbv	
60) o-Xylene	9.992	91	528896	9.040	ppbv	95
61) Styrene	9.898	104	245300	8.474	ppbv	97
62) Isopropylbenzene	10.565	105	794408	9.079	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.992	83	343786	9.671	ppbv	100
64) n-propylbenzene	11.015	120	212879	9.452	ppbv	95
65) tert-Butylbenzene	11.555	119	721078	8.911	ppbv	97
66) Benzyl Chloride	11.639	91	54427	6.339	ppbv	98
67) sec-Butylbenzene	11.791	105	1026079	9.123	ppbv	99
69) p-Isopropyltoluene	11.947	119	865385	8.935	ppbv	99
70) n-Butylbenzene	12.303	91	884265	8.928	ppbv	99
71) 2-Chlorotoluene	10.927	91	599592	8.830	ppbv	98
72) 4-Ethyltoluene	11.147	105	657765	9.029	ppbv	99
73) 1,3,5-Trimethylbenzene	11.228	105	557593	8.764	ppbv	99
74) 1,2,4-Trimethylbenzene	11.558	105	623563	8.507	ppbv #	86
75) 1,3-Dichlorobenzene	11.630	146	381572	9.492	ppbv	98
76) 1,4-Dichlorobenzene	11.694	146	380035	9.499	ppbv	98
77) 1,2-Dichlorobenzene	11.969	146	379896	9.706	ppbv	97
78) Hexachloro-1,3-Butadiene	13.902	225	311900	7.828	ppbv	99
79) Naphthalene	13.533	128	760876	9.490	ppbv	99
80) Naphthalene,2-methyl-	14.478	142	365398	9.294	ppbv	98
81) 1,2,4-Trichlorobenzene	13.461	180	337032	9.114	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042326.D
 Acq On : 15 Apr 2025 11:30
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 16 01:29:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	92	0.02
2 T	Dichlorodifluoromethane	1.476	1.245	15.7	82	0.00
3	Chlorodifluoromethane	1.508	1.366	9.4	86	0.00
4	Chloromethane	0.603	0.520	13.8	84	0.00
5 T	Vinyl Chloride	0.639	0.483	24.4	81	0.01
6 T	Bromomethane	0.273	0.211	22.7	76	0.01
7	Chloroethane	0.242	0.187	22.7	76	0.01
8 T	Dichlorotetrafluoroethane	1.257	1.062	15.5	81	0.01
9 T	Propene	0.691	0.587	15.1	82	0.00
10 T	Heptane	1.839	1.607	12.6	83	0.05
11 T	Trichlorofluoromethane	1.437	1.175	18.2	81	0.01
12 T	1,1,2-Trichlorotrifluoroeth	1.181	0.912	22.8	75	0.02
13	Ethanol	0.065	0.026#	60.0#	83	0.01
14 T	Bromoethene	0.452	0.346	23.5	76	0.02
15 T	Acetone	1.399	1.065	23.9	88	0.01
16 T	1,3-Butadiene	0.683	0.528	22.7	85	0.01
17	tert-Butyl alcohol	1.742	1.218	30.1#	69	0.02
18 T	1,1-Dichloroethene	0.552	0.420	23.9	77	0.02
19 T	Isopropyl Alcohol	0.789	0.606	23.2	79	0.01
20 T	Methylene Chloride	0.521	0.350	32.8#	74	0.02
21 T	Allyl Chloride	0.986	0.772	21.7	79	0.02
22 T	trans-1,2-Dichloroethene	0.638	0.474	25.7	74	0.02
23 T	Vinyl Acetate	0.605	0.419	30.7#	78	0.02
24 T	1,1-Dichloroethane	1.183	0.983	16.9	81	0.02
25 T	Ethyl Acetate	3.630	3.278	9.7	87	0.02
26 T	Hexane	1.402	1.241	11.5	86	0.02
27 T	Carbon Disulfide	1.503	1.135	24.5	73	0.02
28 T	Methyl tert-Butyl Ether	0.873	0.675	22.7	75	0.02
29 T	Chloroform	2.052	1.824	11.1	87	0.02
30 T	Cyclohexane	1.217	1.006	17.3	80	0.03
31 T	cis-1,2-Dichloroethene	1.423	1.208	15.1	81	0.02
32 T	1,1,1-Trichloroethane	2.349	1.810	22.9	81	0.03
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.04
34 T	2-Butanone	0.725	0.694	4.3	87	0.02
35 T	Carbon Tetrachloride	0.781	0.695	11.0	83	0.03
36 T	Benzene	0.991	0.928	6.4	83	0.03
37 T	1,2-Dichloroethane	0.568	0.524	7.7	82	0.03
38 T	Trichloroethene	0.414	0.377	8.9	84	0.04
39 T	1,2-Dichloropropane	0.361	0.332	8.0	83	0.04
40 T	1,4-Dioxane	0.192	0.158	17.7	78	0.05
41 T	Tetrahydrofuran	0.418	0.397	5.0	84	0.03
42 T	Bromodichloromethane	0.782	0.729	6.8	80	0.04
43	Methyl Methacrylate	0.420	0.340	19.0	71	0.05
44 T	2,2,4-Trimethylpentane	1.675	1.593	4.9	83	0.04
45 T	t-1,3-Dichloropropene	0.450	0.330	26.7	63	0.05
46 T	cis-1,3-Dichloropropene	0.552	0.447	19.0	68	0.05
47 T	1,1,2-Trichloroethane	0.382	0.357	6.5	82	0.05
48 T	Dibromochloromethane	0.647	0.596	7.9	79	0.05

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042326.D
 Acq On : 15 Apr 2025 11:30
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 16 01:29:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.581	0.522	10.2	75	0.04
50 T	4-Methyl-2-Pentanone	1.149	0.969	15.7	79	0.05
51 T	2-Hexanone	0.944	0.773	18.1	77	0.04
52 T	Tetrachloroethene	0.393	0.316	19.6	77	0.04
53 T	Toluene	1.175	1.010	14.0	75	0.05
54 T	1,2-Dibromoethane	0.532	0.481	9.6	80	0.04
55 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.04
56	1,1,1,2-Tetrachloroethane	0.550	0.520	5.5	81	0.04
57 T	Chlorobenzene	0.949	0.912	3.9	83	0.04
58 T	Ethyl Benzene	1.776	1.555	12.4	74	0.03
59 T	m/p-Xylene	1.418	1.292	8.9	77	0.03
60 T	o-Xylene	1.422	1.285	9.6	77	0.03
61 T	Styrene	0.703	0.596	15.2	69	0.03
62	Isopropylbenzene	2.126	1.931	9.2	78	0.03
63 T	1,1,2,2-Tetrachloroethane	0.864	0.835	3.4	83	0.03
64	n-propylbenzene	0.547	0.517	5.5	82	0.03
65	tert-Butylbenzene	1.966	1.752	10.9	80	0.03
66 T	Benzyl Chloride	0.209	0.132	36.8#	52	0.03
67	sec-Butylbenzene	2.733	2.494	8.7	82	0.03
68 S	1-Bromo-4-Fluorobenzene	0.821	0.792	3.5	77	0.03
69	p-Isopropyltoluene	2.354	2.103	10.7	80	0.02
70	n-Butylbenzene	2.407	2.149	10.7	80	0.03
71	2-Chlorotoluene	1.650	1.457	11.7	75	0.03
72 T	4-Ethyltoluene	1.770	1.598	9.7	78	0.03
73 T	1,3,5-Trimethylbenzene	1.546	1.355	12.4	77	0.03
74 T	1,2,4-Trimethylbenzene	1.781	1.515	14.9	79	0.02
75 T	1,3-Dichlorobenzene	0.977	0.927	5.1	83	0.03
76 T	1,4-Dichlorobenzene	0.972	0.924	4.9	82	0.03
77 T	1,2-Dichlorobenzene	0.951	0.923	2.9	86	0.03
78 T	Hexachloro-1,3-Butadiene	0.968	0.758	21.7	75	0.02
79 T	Naphthalene	1.948	1.849	5.1	81	0.02
80 T	Naphthalene,2-methyl-	0.955	0.888	7.0	71	0.02
81 T	1,2,4-Trichlorobenzene	0.899	0.819	8.9	75	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042327.D
 Acq On : 15 Apr 2025 12:12
 Operator : SY/MD
 Sample : VL0415ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0415ABL01

Quant Time: Apr 16 01:30:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.813	49	173729	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	3.994	114	447320	10.000	ppbv	0.04
55) Chlorobenzene-d5	8.917	117	385318	10.000	ppbv	0.04

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.406	95	292759	9.252	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	92.500%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
Data File : VL042327.D
Acq On : 15 Apr 2025 12:12
Operator : SY/MD
Sample : VL0415ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0415ABL01

Quant Time: Apr 16 01:30:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Mar 28 02:00:52 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042328.D
 Acq On : 15 Apr 2025 12:54
 Operator : SY/MD
 Sample : VL0415ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0415ABS01

Quant Time: Apr 16 01:30:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 04/16/2025
 Supervised By : Mahesh Dadoda 04/16/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.810	49	171986	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	3.994	114	444595	10.000	ppbv	0.04
55) Chlorobenzene-d5	8.917	117	399604	10.000	ppbv	0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.406	95	318902	9.718	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.200%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.509	85	214241	8.437	ppbv	99
3) Chlorodifluoromethane	1.486	51	234966	9.058	ppbv	97
4) Chloromethane	1.541	50	88230	8.502	ppbv	98
5) Vinyl Chloride	1.589	62	82221	7.476	ppbv	94
6) Bromomethane	1.680	94	36050	7.665	ppbv	100
7) Chloroethane	1.716	64	33039	7.931	ppbv	97
8) Dichlorotetrafluoroethane	1.564	85	178173	8.239	ppbv	95
9) Propene	1.492	41	94136	7.917	ppbv	99
10) Heptane	5.027	43	267288	8.451	ppbv	99
11) Trichlorofluoromethane	1.890	101	201665	8.162	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.156	101	154306	7.598	ppbv	98
13) Ethanol	1.732	45	4407	3.921	ppbv	97
14) Bromoethene	1.793	108	57635	7.408	ppbv	99
15) Acetone	1.848	43	178724	7.430	ppbv	96
16) 1,3-Butadiene	1.619	39	89479	7.617	ppbv	96
17) tert-Butyl alcohol	2.055	59	202421	6.757	ppbv	99
18) 1,1-Dichloroethene	2.049	96	69013	7.269	ppbv	97
19) Isopropyl Alcohol	1.900	45	100289	7.390	ppbv	97
20) Methylene Chloride	2.075	84	60054	6.704	ppbv	88
21) Allyl Chloride	2.111	41	127800	7.534	ppbv	98
22) trans-1,2-Dichloroethene	2.360	96	78991	7.198	ppbv	98
23) Vinyl Acetate	2.483	43	64777	6.222	ppbv	98
24) 1,1-Dichloroethane	2.428	63	168373	8.277	ppbv	99
25) Ethyl Acetate	2.861	43	545140	8.732	ppbv	99
26) Hexane	2.848	57	205629	8.528	ppbv	97
27) Carbon Disulfide	2.166	76	191388	7.402	ppbv	99
28) Methyl tert-Butyl Ether	2.460	73	107667	7.172	ppbv	96
29) Chloroform	2.874	83	309262	8.763	ppbv	100
30) Cyclohexane	3.891	84	165914	7.924	ppbv	98
31) cis-1,2-Dichloroethene	2.742	61	197651	8.074	ppbv	99
32) 1,1,1-Trichloroethane	3.395	97	304143	7.528	ppbv	97
34) 2-Butanone	2.580	43	300576	9.323	ppbv	98
35) Carbon Tetrachloride	3.797	117	314849	9.065	ppbv	99
36) Benzene	3.690	78	407959	9.256	ppbv	98
37) 1,2-Dichloroethane	3.247	62	235609	9.337	ppbv	99
38) Trichloroethene	4.590	130	164664	8.954	ppbv	96
39) 1,2-Dichloropropane	4.354	63	149260	9.299	ppbv	97
40) 1,4-Dioxane	4.625	88	68773	8.072	ppbv #	95
41) Tetrahydrofuran	3.082	42	173776	9.361	ppbv	95
42) Bromodichloromethane	4.532	83	327497	9.414	ppbv	99
43) Methyl Methacrylate	4.930	69	151531	8.121	ppbv	98
44) 2,2,4-Trimethylpentane	4.680	57	707832	9.503	ppbv	98
45) t-1,3-Dichloropropene	6.470	75	143714	7.179	ppbv	91

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042328.D
 Acq On : 15 Apr 2025 12:54
 Operator : SY/MD
 Sample : VL0415ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0415ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/16/2025
 Supervised By :Mahesh Dadoda 04/16/2025

Quant Time: Apr 16 01:30:27 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 02:00:52 2025


Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.658	75	192419	7.841	ppbv	96
47) 1,1,2-Trichloroethane	6.635	97	163251	9.621	ppbv	98
48) Dibromochloromethane	7.428	129	266192	9.256	ppbv	100
49) Bromoform	9.519	173	236686	9.168	ppbv	99
50) 4-Methyl-2-Pentanone	5.807	43	429988	8.416	ppbv	95
51) 2-Hexanone	7.480	43	344172	8.203	ppbv #	96
52) Tetrachloroethene	8.263	164	139845	8.000	ppbv	95
53) Toluene	6.978	91	447178	8.559	ppbv	99
54) 1,2-Dibromoethane	7.694	107	214982	9.089	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.959	131	216770	9.855	ppbv	98
57) Chlorobenzene	8.956	112	369132	9.739	ppbv	95
58) Ethyl Benzene	9.390	91	618669	8.716	ppbv	100
59) m/p-Xylene	9.584	91	1049238m	18.512	ppbv	
60) o-Xylene	9.995	91	517471	9.107	ppbv	96
61) Styrene	9.904	104	229601	8.168	ppbv	97
62) Isopropylbenzene	10.568	105	785673	9.247	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.995	83	351308	10.176	ppbv	99
64) n-propylbenzene	11.018	120	210294	9.615	ppbv	98
65) tert-Butylbenzene	11.558	119	718948	9.149	ppbv	97
66) Benzyl Chloride	11.642	91	55150	6.614	ppbv	99
67) sec-Butylbenzene	11.795	105	1018861	9.328	ppbv	100
69) p-Isopropyltoluene	11.950	119	863264	9.179	ppbv	99
70) n-Butylbenzene	12.306	91	891883	9.273	ppbv	99
71) 2-Chlorotoluene	10.930	91	590408	8.953	ppbv	99
72) 4-Ethyltoluene	11.150	105	652962	9.230	ppbv	98
73) 1,3,5-Trimethylbenzene	11.231	105	559451	9.055	ppbv	100
74) 1,2,4-Trimethylbenzene	11.565	105	632271	8.883	ppbv	97
75) 1,3-Dichlorobenzene	11.633	146	381882	9.782	ppbv	98
76) 1,4-Dichlorobenzene	11.697	146	383116	9.861	ppbv	98
77) 1,2-Dichlorobenzene	11.973	146	386768	10.175	ppbv	98
78) Hexachloro-1,3-Butadiene	13.905	225	314860	8.138	ppbv	100
79) Naphthalene	13.539	128	777722	9.989	ppbv	99
80) Naphthalene,2-methyl-	14.481	142	375541	9.836	ppbv	98
81) 1,2,4-Trichlorobenzene	13.465	180	343011	9.551	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Quant Time: Apr 16 01:30:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 02:00:52 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042330.D
 Acq On : 15 Apr 2025 13:57
 Operator : SY/MD
 Sample : QC041425 CAN 10401
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC041425 CAN 10401

Quant Time: Apr 16 01:31:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.813	49	162884	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	3.997	114	405181	10.000	ppbv	0.05
55) Chlorobenzene-d5	8.917	117	353699	10.000	ppbv	0.04

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.406	95	273327	9.410	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.100%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
Data File : VL042330.D
Acq On : 15 Apr 2025 13:57
Operator : SY/MD
Sample : QC041425 CAN 10401
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041425 CAN 10401

Quant Time: Apr 16 01:31:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Mar 28 02:00:52 2025
Response via : Initial Calibration





Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042457.D
Acq On : 01 May 2025 08:19
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri May 02 03:16:31 2025



AutoFind: Scans 3178, 3179, 3180; Background Corrected with Scan 3165

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	25.5	33424	PASS
75	95	30	66	62.8	82325	PASS
95	95	100	100	100.0	131013	PASS
96	95	5	9	6.8	8845	PASS
173	174	0.00	2	0.6	483	PASS
174	95	50	120	57.4	75195	PASS
175	174	4	9	7.5	5628	PASS
176	174	93	101	97.9	73643	PASS
177	176	5	9	6.3	4652	PASS

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL050125AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Fri May 02 03:16:31 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL042466.D 0.1 =VL042462.D 0.5 =VL042461.D 1 =VL042460.D 2 =VL042459.D 10 =VL042458.D 15 =VL042465.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.887	1.650	1.629	1.470	1.639	1.655	9.01
3) Chlorodifluoro...			1.886	1.871	1.820	1.569	1.602	1.750	8.71
4) Chloromethane			0.738	0.696	0.734	0.624	0.666	0.692	6.90
5) T Vinyl Chloride	0.756	0.964	0.708	0.625	0.623	0.567	0.631	0.696	19.18
6) T Bromomethane			0.349	0.359	0.327	0.286	0.324	0.329	8.64
7) Chloroethane			0.250	0.272	0.271	0.231	0.261	0.257	6.55
8) T Dichlorotetra...			1.445	1.371	1.402	1.235	1.368	1.364	5.77
9) T Propene			0.585	0.606	0.631	0.525	0.515	0.572	8.84
10) T Heptane			1.276	1.634	1.593	1.445	1.474	1.484	9.48
11) T Trichlorofluor...			1.660	1.489	1.528	1.376	1.582	1.527	6.96
12) T 1,1,2-Trichlor...			1.318	1.234	1.245	1.099	1.201	1.219	6.52
13) Ethanol			0.138	0.136	0.143	0.080	0.084	0.116	27.19
14) T Bromoethene			0.488	0.443	0.457	0.398	0.466	0.450	7.39
15) T Acetone			2.229	2.094	2.022	1.318	1.465	1.826	22.28
16) T 1,3-Butadiene			0.883	0.874	0.836	0.681	0.751	0.805	10.77
17) tert-Butyl alc...			1.816	1.615	1.620	1.489	1.502	1.608	8.15
18) T 1,1-Dichloroet...			0.650	0.587	0.567	0.495	0.547	0.569	9.98
19) T Isopropyl Alcohol			1.221	1.087	0.976	0.876	0.891	1.010	14.33
20) T Methylene Chlo...			0.753	0.547	0.518	0.449	0.477	0.549	21.88
21) T Allyl Chloride			1.283	1.181	1.164	1.076	1.166	1.174	6.25
22) T trans-1,2-Dich...			0.741	0.742	0.645	0.552	0.644	0.665	11.96
23) T Vinyl Acetate			1.898	1.521	1.810	1.726	1.756	1.742	8.03
24) T 1,1-Dichloroet...			1.432	1.465	1.368	1.196	1.322	1.357	7.79
25) T Ethyl Acetate			3.002	3.044	2.945	2.832	2.792	2.923	3.69
26) T Hexane			1.278	1.290	1.345	1.215	1.239	1.273	3.91
27) T Carbon Disulfide			1.509	1.446	1.499	1.418	1.558	1.486	3.69
28) T Methyl tert-Bu...			0.802	0.853	0.759	0.671	0.740	0.765	8.91
29) T Chloroform			2.551	2.482	2.399	2.198	2.301	2.386	5.89
30) T Cyclohexane			1.135	0.998	1.096	1.045	1.048	1.065	4.92
31) T cis-1,2-Dichlo...			1.551	1.604	1.545	1.342	1.386	1.486	7.71
32) T 1,1,1-Trichlor...	2.980	2.842	2.567	2.501	2.511	2.288	2.433	2.589	9.28

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.802	0.732	0.774	0.659	0.650	0.723	9.32
35) T Carbon Tetrach...	1.166	0.799	0.856	0.935	0.875	0.827	0.902	0.908	13.44
36) T Benzene			0.924	0.974	1.036	0.892	0.892	0.944	6.53
37) T 1,2-Dichloroet...			0.738	0.755	0.769	0.666	0.709	0.727	5.66
38) T Trichloroethene	0.241	0.419	0.389	0.364	0.404	0.347	0.355	0.360	16.31
39) T 1,2-Dichloropr...			0.368	0.405	0.362	0.317	0.315	0.353	10.71

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\ Method File : VL050125AIR.M											
40)	T	1,4-Dioxane			0.176	0.185	0.172	0.151	0.144	0.165	10.60
41)	T	Tetrahydrofuran			0.329	0.354	0.387	0.326	0.322	0.344	7.98
42)	T	Bromodichlorom...			0.848	0.921	0.931	0.839	0.904	0.889	4.77
43)		Methyl Methacr...			0.328	0.341	0.362	0.363	0.366	0.352	4.72
44)	T	2,2,4-Trimethy...			1.490	1.590	1.633	1.437	1.498	1.530	5.23
45)	T	t-1,3-Dichloro...			0.438	0.423	0.460	0.480	0.510	0.462	7.46
46)	T	cis-1,3-Dichlo...			0.526	0.539	0.549	0.555	0.577	0.549	3.47
47)	T	1,1,2-Trichlor...			0.372	0.409	0.428	0.357	0.371	0.387	7.74
48)	T	Dibromochlorom...			0.651	0.643	0.673	0.630	0.674	0.654	2.96
49)	T	Bromoform			0.540	0.535	0.573	0.527	0.568	0.549	3.73
50)	T	4-Methyl-2-Pen...			0.852	0.980	0.985	0.870	0.887	0.915	6.92
51)	T	2-Hexanone			0.649	0.754	0.797	0.722	0.736	0.731	7.42
52)	T	Tetrachloroethene	0.509	0.380	0.325	0.339	0.338	0.305	0.318	0.359	19.49
53)	T	Toluene			0.935	0.969	1.025	1.016	1.050	0.999	4.64
54)	T	1,2-Dibromoethane		0.625	0.589	0.599	0.610	0.554	0.577	0.592	4.26
-----ISTD-----											
55)	I	Chlorobenzene-d5									
56)		1,1,1,2-Tetrac...			0.618	0.633	0.596	0.542	0.564	0.591	6.38
57)	T	Chlorobenzene			1.101	1.060	1.036	0.904	0.923	1.005	8.64
58)	T	Ethyl Benzene			1.492	1.634	1.752	1.725	1.754	1.671	6.68
59)	T	m/p-Xylene			1.212	1.378	1.506	1.429	1.479	1.401	8.31
60)	T	o-Xylene			1.197	1.393	1.519	1.423	1.481	1.403	8.90
61)	T	Styrene			0.415	0.500	0.538	0.583	0.582	0.523	13.35
62)		Isopropylbenzene			1.781	2.030	2.165	2.033	2.108	2.023	7.26
63)	T	1,1,2,2-Tetrac...	1.140	0.945	0.976	0.985	0.990	0.846	0.887	0.967	9.69
64)		n-propylbenzene			0.400	0.524	0.549	0.482	0.518	0.495	11.75
65)		tert-Butylbenzene			1.519	1.791	2.028	1.817	1.873	1.806	10.23
66)	T	Benzyl Chloride			0.158	0.204	0.195	0.232	0.255	0.209	17.73
67)		sec-Butylbenzene			2.149	2.560	2.905	2.541	2.675	2.566	10.69
68)	S	1-Bromo-4-Fluo...	0.920	0.911	0.902	0.928	0.915	0.951	0.963	0.927	2.39
69)		p-Isopropyltol...			1.805	2.166	2.406	2.150	2.283	2.162	10.39
70)		n-Butylbenzene			1.802	2.218	2.560	2.198	2.396	2.235	12.66
71)		2-Chlorotoluene			1.405	1.569	1.697	1.597	1.667	1.587	7.20
72)	T	4-Ethyltoluene			1.281	1.647	1.828	1.722	1.792	1.654	13.28
73)	T	1,3,5-Trimethy...			1.177	1.467	1.618	1.475	1.535	1.455	11.44
74)	T	1,2,4-Trimethy...			1.568	1.788	2.075	1.646	1.723	1.760	11.07
75)	T	1,3-Dichlorobe...			0.975	1.022	1.056	0.893	0.952	0.980	6.42
76)	T	1,4-Dichlorobe...			0.869	0.945	0.966	0.907	0.949	0.927	4.22
77)	T	1,2-Dichlorobe...			0.977	1.007	1.033	0.864	0.914	0.959	7.23
78)	T	Hexachloro-1,3...			1.028	1.081	1.070	0.772	0.838	0.958	14.93
79)	T	Naphthalene		0.720	1.168	1.657	2.155	1.527	1.661	1.482	33.01
80)	T	Naphthalene,2-...			0.498	0.925	1.573	1.172	1.338	1.101	37.38
81)	T	1,2,4-Trichlor...			0.599	0.798	0.947	0.778	0.828	0.790	15.84

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042458.D
 Acq On : 01 May 2025 09:03
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/02/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:39:48 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	71402	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	201290	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	174168	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 165650 10.258 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.600%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	104948	8.882	ppbv	100
3) Chlorodifluoromethane	1.476	51	112008	8.966	ppbv	100
4) Chloromethane	1.531	50	44585	9.028	ppbv	100
5) Vinyl Chloride	1.580	62	40466	8.138	ppbv	100
6) Bromomethane	1.667	94	20392	8.680	ppbv	100
7) Chloroethane	1.703	64	16528	9.002	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	88162	9.051	ppbv	100
9) Propene	1.483	41	37491	9.175	ppbv	100
10) Heptane	4.978	43	103167	9.733	ppbv	100
11) Trichlorofluoromethane	1.874	101	98215	9.320	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.136	101	78482	9.015	ppbv	100
13) Ethanol	1.719	45	5679	6.849	ppbv	100
14) Bromoethene	1.780	108	28440	8.849	ppbv	100
15) Acetone	1.835	43	94092	7.219	ppbv	100
16) 1,3-Butadiene	1.609	39	48641	8.461	ppbv	100
17) tert-Butyl alcohol	2.039	59	106304	9.257	ppbv	100
18) 1,1-Dichloroethene	2.033	96	35320	8.689	ppbv	100
19) Isopropyl Alcohol	1.884	45	62533	8.668	ppbv	100
20) Methylene Chloride	2.059	84	32039	8.176	ppbv	100
21) Allyl Chloride	2.094	41	76859	9.169	ppbv	100
22) trans-1,2-Dichloroethene	2.340	96	39427	8.305	ppbv	100
23) Vinyl Acetate	2.460	43	123215	10.146	ppbv	100
24) 1,1-Dichloroethane	2.408	63	85367	8.814	ppbv	100
25) Ethyl Acetate	2.835	43	202215	9.689	ppbv	100
26) Hexane	2.826	57	86764	9.543	ppbv	100
27) Carbon Disulfide	2.146	76	101248	9.542	ppbv #	92
28) Methyl tert-Butyl Ether	2.437	73	47941	8.777	ppbv	100
29) Chloroform	2.848	83	156976	9.213	ppbv	100
30) Cyclohexane	3.855	84	74614	9.815	ppbv	100
31) cis-1,2-Dichloroethene	2.719	61	95797	9.031	ppbv	100
32) 1,1,1-Trichloroethane	3.366	97	163334	8.836	ppbv	100
34) 2-Butanone	2.557	43	132685	9.113	ppbv	100
35) Carbon Tetrachloride	3.758	117	166490	9.511	ppbv	100
36) Benzene	3.654	78	179532	9.451	ppbv	100
37) 1,2-Dichloroethane	3.217	62	133963	9.150	ppbv	100
38) Trichloroethene	4.544	130	69846	9.642	ppbv	100
39) 1,2-Dichloropropane	4.311	63	63894	8.979	ppbv	100
40) 1,4-Dioxane	4.583	88	30303m	9.161	ppbv	
41) Tetrahydrofuran	3.052	42	65713	9.516	ppbv	100
42) Bromodichloromethane	4.486	83	168957	9.445	ppbv	100
43) Methyl Methacrylate	4.881	69	73117	10.339	ppbv	100
44) 2,2,4-Trimethylpentane	4.635	57	289161	9.420	ppbv	100
45) t-1,3-Dichloropropene	6.412	75	96692	10.385	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042458.D
 Acq On : 01 May 2025 09:03
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:39:48 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.600	75	111768	10.111	ppbv	100
47) 1,1,2-Trichloroethane	6.580	97	71805	9.207	ppbv	100
48) Dibromochloromethane	7.386	129	126754	9.628	ppbv	100
49) Bromoform	9.480	173	106108	9.610	ppbv	100
50) 4-Methyl-2-Pentanone	5.748	43	175193	9.518	ppbv	100
51) 2-Hexanone	7.435	43	145272	9.855	ppbv	100
52) Tetrachloroethene	8.224	164	61456	8.500	ppbv	100
53) Toluene	6.930	91	204494	10.171	ppbv	100
54) 1,2-Dibromoethane	7.648	107	111506	9.300	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.924	131	94346	9.172	ppbv	99
57) Chlorobenzene	8.920	112	157401	8.994	ppbv	100
58) Ethyl Benzene	9.354	91	300412	10.321	ppbv	100
59) m/p-Xylene	9.529	91	497922m	20.404	ppbv	
60) o-Xylene	9.963	91	247784	10.157	ppbv	100
61) Styrene	9.872	104	101509	11.134	ppbv	100
62) Isopropylbenzene	10.539	105	354055	10.047	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.963	83	147295	9.305	ppbv	100
64) n-propylbenzene	10.989	120	84006	9.787	ppbv	100
65) tert-Butylbenzene	11.529	119	316532	10.065	ppbv	100
66) Benzyl Chloride	11.613	91	40331	11.096	ppbv	100
67) sec-Butylbenzene	11.769	105	442600	9.903	ppbv	100
69) p-Isopropyltoluene	11.924	119	374540	9.947	ppbv	100
70) n-Butylbenzene	12.280	91	382883	9.836	ppbv	100
71) 2-Chlorotoluene	10.898	91	278215	10.065	ppbv	100
72) 4-Ethyltoluene	11.124	105	299831	10.409	ppbv	100
73) 1,3,5-Trimethylbenzene	11.202	105	256952	10.143	ppbv	100
74) 1,2,4-Trimethylbenzene	11.536	105	286615	9.351	ppbv	100
75) 1,3-Dichlorobenzene	11.607	146	155598	9.118	ppbv	100
76) 1,4-Dichlorobenzene	11.671	146	157966	9.782	ppbv	100
77) 1,2-Dichlorobenzene	11.947	146	150432	9.007	ppbv	100
78) Hexachloro-1,3-Butadiene	13.879	225	134402	8.058	ppbv	100
79) Naphthalene	13.513	128	266036	10.352	ppbv	100
80) Naphthalene,2-methyl-	14.458	142	204042	10.641	ppbv	100
81) 1,2,4-Trichlorobenzene	13.439	180	135473	9.844	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042459.D
 Acq On : 01 May 2025 09:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/02/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:40:35 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	73220	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	201951	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	174624	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 159777 9.869 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 98.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	23859	1.969	ppbv	100
3) Chlorodifluoromethane	1.476	51	26655	2.081	ppbv	95
4) Chloromethane	1.531	50	10751	2.123	ppbv	95
5) Vinyl Chloride	1.580	62	9122	1.789	ppbv	92
6) Bromomethane	1.667	94	4793	1.990	ppbv #	62
7) Chloroethane	1.703	64	3972	2.110	ppbv	94
8) Dichlorotetrafluoroethane	1.554	85	20529	2.055	ppbv	100
9) Propene	1.482	41	9234	2.204	ppbv	95
10) Heptane	4.978	43	23329m	2.146	ppbv	
11) Trichlorofluoromethane	1.877	101	22375	2.070	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.140	101	18230	2.042	ppbv	97
13) Ethanol	1.722	45	2099	2.468	ppbv	94
14) Bromoethene	1.780	108	6686	2.029	ppbv #	96
15) Acetone	1.835	43	29607	2.215	ppbv	99
16) 1,3-Butadiene	1.609	39	12245	2.077	ppbv	92
17) tert-Butyl alcohol	2.042	59	23722	2.014	ppbv #	93
18) 1,1-Dichloroethene	2.033	96	8303	1.992	ppbv	92
19) Isopropyl Alcohol	1.884	45	14288	1.931	ppbv	99
20) Methylene Chloride	2.062	84	7590	1.889	ppbv	96
21) Allyl Chloride	2.094	41	17050	1.983	ppbv	96
22) trans-1,2-Dichloroethene	2.340	96	9444	1.940	ppbv	97
23) Vinyl Acetate	2.463	43	26502	2.128	ppbv	98
24) 1,1-Dichloroethane	2.408	63	20027	2.016	ppbv	98
25) Ethyl Acetate	2.839	43	43122	2.015	ppbv	97
26) Hexane	2.822	57	19691	2.112	ppbv	95
27) Carbon Disulfide	2.149	76	21945	2.017	ppbv #	43
28) Methyl tert-Butyl Ether	2.441	73	11119	1.985	ppbv	94
29) Chloroform	2.848	83	35127	2.010	ppbv	86
30) Cyclohexane	3.858	84	16056	2.060	ppbv	95
31) cis-1,2-Dichloroethene	2.719	61	22619	2.079	ppbv	95
32) 1,1,1-Trichloroethane	3.369	97	36768	1.940	ppbv	99
34) 2-Butanone	2.557	43	31242	2.139	ppbv	95
35) Carbon Tetrachloride	3.764	117	35326	2.011	ppbv	97
36) Benzene	3.661	78	41848	2.196	ppbv #	88
37) 1,2-Dichloroethane	3.217	62	31049	2.114	ppbv	99
38) Trichloroethene	4.544	130	16311m	2.244	ppbv	
39) 1,2-Dichloropropane	4.311	63	14611	2.047	ppbv	91
40) 1,4-Dioxane	4.593	88	6957m	2.096	ppbv	
41) Tetrahydrofuran	3.059	42	15641m	2.258	ppbv	
42) Bromodichloromethane	4.489	83	37615	2.096	ppbv #	99
43) Methyl Methacrylate	4.884	69	14610m	2.059	ppbv	
44) 2,2,4-Trimethylpentane	4.638	57	65967	2.142	ppbv	99
45) t-1,3-Dichloropropene	6.409	75	18560m	1.987	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042459.D
 Acq On : 01 May 2025 09:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:40:35 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	22184	2.000	ppbv	93
47) 1,1,2-Trichloroethane	6.587	97	17307m	2.212	ppbv	
48) Dibromochloromethane	7.386	129	27174	2.057	ppbv	99
49) Bromoform	9.483	173	23126	2.088	ppbv	98
50) 4-Methyl-2-Pentanone	5.758	43	39798	2.155	ppbv	99
51) 2-Hexanone	7.441	43	32196	2.177	ppbv #	97
52) Tetrachloroethene	8.224	164	13670	1.885	ppbv	93
53) Toluene	6.930	91	41392	2.052	ppbv	97
54) 1,2-Dibromoethane	7.652	107	24650	2.049	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.927	131	20822	2.019	ppbv	95
57) Chlorobenzene	8.920	112	36187	2.062	ppbv #	88
58) Ethyl Benzene	9.354	91	61177	2.096	ppbv	91
59) m/p-Xylene	9.551	91	105185m	4.299	ppbv	
60) o-Xylene	9.963	91	53051	2.169	ppbv	99
61) Styrene	9.872	104	18776	2.054	ppbv	97
62) Isopropylbenzene	10.539	105	75626	2.140	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.963	83	34562	2.178	ppbv	98
64) n-propylbenzene	10.989	120	19160	2.226	ppbv	94
65) tert-Butylbenzene	11.532	119	70813	2.246	ppbv	97
66) Benzyl Chloride	11.613	91	6794	1.864	ppbv	97
67) sec-Butylbenzene	11.769	105	101458	2.264	ppbv	99
69) p-Isopropyltoluene	11.924	119	84026	2.226	ppbv	98
70) n-Butylbenzene	12.280	91	89400	2.291	ppbv	97
71) 2-Chlorotoluene	10.901	91	59271	2.139	ppbv	97
72) 4-Ethyltoluene	11.124	105	63833	2.210	ppbv	98
73) 1,3,5-Trimethylbenzene	11.205	105	56516	2.225	ppbv	94
74) 1,2,4-Trimethylbenzene	11.536	105	72485	2.359	ppbv	94
75) 1,3-Dichlorobenzene	11.607	146	36885	2.156	ppbv	97
76) 1,4-Dichlorobenzene	11.671	146	33728	2.083	ppbv	94
77) 1,2-Dichlorobenzene	11.947	146	36078	2.154	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	37360	2.234	ppbv	99
79) Naphthalene	13.513	128	75248	2.920	ppbv	96
80) Naphthalene,2-methyl-	14.458	142	54922	2.857	ppbv	96
81) 1,2,4-Trichlorobenzene	13.442	180	33063	2.396	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042460.D
 Acq On : 01 May 2025 10:08
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:41:24 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	71499	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	200111	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	170872	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.374 95 158600 10.011 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	11796	0.997	ppbv	96
3) Chlorodifluoromethane	1.476	51	13375	1.069	ppbv	96
4) Chloromethane	1.531	50	4975	1.006	ppbv	95
5) Vinyl Chloride	1.576	62	4469	0.898	ppbv	100
6) Bromomethane	1.667	94	2570	1.092	ppbv	86
7) Chloroethane	1.703	64	1943	1.057	ppbv	94
8) Dichlorotetrafluoroethane	1.554	85	9804	1.005	ppbv	91
9) Propene	1.482	41	4331	1.058	ppbv	88
10) Heptane	4.975	43	11681m	1.101	ppbv	
11) Trichlorofluoromethane	1.877	101	10649m	1.009	ppbv	
12) 1,1,2-Trichlorotrifluo...	2.140	101	8821	1.012	ppbv	97
13) Ethanol	1.719	45	975	1.174	ppbv #	52
14) Bromoethene	1.780	108	3164	0.983	ppbv #	91
15) Acetone	1.835	43	14974	1.147	ppbv	98
16) 1,3-Butadiene	1.609	39	6248	1.085	ppbv	96
17) tert-Butyl alcohol	2.039	59	11546	1.004	ppbv #	91
18) 1,1-Dichloroethene	2.033	96	4200	1.032	ppbv	86
19) Isopropyl Alcohol	1.884	45	7775	1.076	ppbv	99
20) Methylene Chloride	2.059	84	3909	0.996	ppbv	86
21) Allyl Chloride	2.094	41	8443	1.006	ppbv	94
22) trans-1,2-Dichloroethene	2.337	96	5305	1.116	ppbv	88
23) Vinyl Acetate	2.460	43	10876	0.894	ppbv #	94
24) 1,1-Dichloroethane	2.405	63	10477	1.080	ppbv	97
25) Ethyl Acetate	2.835	43	21761	1.041	ppbv	99
26) Hexane	2.822	57	9223	1.013	ppbv	97
27) Carbon Disulfide	2.149	76	10342	0.973	ppbv #	21
28) Methyl tert-Butyl Ether	2.441	73	6101	1.115	ppbv	93
29) Chloroform	2.848	83	17744	1.040	ppbv	97
30) Cyclohexane	3.852	84	7139	0.938	ppbv	88
31) cis-1,2-Dichloroethene	2.719	61	11472	1.080	ppbv	87
32) 1,1,1-Trichloroethane	3.366	97	17881	0.966	ppbv	98
34) 2-Butanone	2.560	43	14641	1.012	ppbv	94
35) Carbon Tetrachloride	3.755	117	18704	1.075	ppbv #	76
36) Benzene	3.654	78	19493	1.032	ppbv #	94
37) 1,2-Dichloroethane	3.214	62	15116	1.039	ppbv	98
38) Trichloroethene	4.541	130	7278m	1.011	ppbv	
39) 1,2-Dichloropropane	4.302	63	8103m	1.145	ppbv	
40) 1,4-Dioxane	4.593	88	3703m	1.126	ppbv	
41) Tetrahydrofuran	3.062	42	7091m	1.033	ppbv	
42) Bromodichloromethane	4.486	83	18427	1.036	ppbv #	81
43) Methyl Methacrylate	4.874	69	6829m	0.971	ppbv	
44) 2,2,4-Trimethylpentane	4.635	57	31820m	1.043	ppbv	
45) t-1,3-Dichloropropene	6.409	75	8461m	0.914	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042460.D
 Acq On : 01 May 2025 10:08
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:41:24 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.596	75	10781m	0.981	ppbv	
47) 1,1,2-Trichloroethane	6.583	97	8183m	1.055	ppbv	
48) Dibromochloromethane	7.386	129	12859m	0.982	ppbv	
49) Bromoform	9.483	173	10709	0.976	ppbv	97
50) 4-Methyl-2-Pentanone	5.768	43	19618m	1.072	ppbv	
51) 2-Hexanone	7.448	43	15091m	1.030	ppbv	
52) Tetrachloroethene	8.221	164	6780	0.943	ppbv	86
53) Toluene	6.930	91	19390	0.970	ppbv	97
54) 1,2-Dibromoethane	7.655	107	11980	1.005	ppbv	91
56) 1,1,1,2-Tetrachloroethane	8.920	131	10811	1.071	ppbv	94
57) Chlorobenzene	8.917	112	18111	1.055	ppbv	90
58) Ethyl Benzene	9.354	91	27914	0.978	ppbv	97
59) m/p-Xylene	9.548	91	47105m	1.968	ppbv	
60) o-Xylene	9.962	91	23806	0.995	ppbv	98
61) Styrene	9.872	104	8537	0.954	ppbv	98
62) Isopropylbenzene	10.539	105	34679	1.003	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.962	83	16837	1.084	ppbv	98
64) n-propylbenzene	10.985	120	8958	1.064	ppbv	86
65) tert-Butylbenzene	11.529	119	30604	0.992	ppbv	95
66) Benzyl Chloride	11.613	91	3486	0.978	ppbv #	96
67) sec-Butylbenzene	11.765	105	43740	0.998	ppbv	98
69) p-Isopropyltoluene	11.927	119	37011	1.002	ppbv	99
70) n-Butylbenzene	12.280	91	37892	0.992	ppbv	97
71) 2-Chlorotoluene	10.898	91	26813	0.989	ppbv	99
72) 4-Ethyltoluene	11.124	105	28149	0.996	ppbv	96
73) 1,3,5-Trimethylbenzene	11.202	105	25073	1.009	ppbv	97
74) 1,2,4-Trimethylbenzene	11.535	105	30550	1.016	ppbv	92
75) 1,3-Dichlorobenzene	11.607	146	17456	1.043	ppbv	98
76) 1,4-Dichlorobenzene	11.671	146	16151	1.019	ppbv	99
77) 1,2-Dichlorobenzene	11.947	146	17207	1.050	ppbv	94
78) Hexachloro-1,3-Butadiene	13.879	225	18472	1.129	ppbv	98
79) Naphthalene	13.513	128	28316	1.123	ppbv #	94
80) Naphthalene,2-methyl-	14.458	142	15804	0.840	ppbv	93
81) 1,2,4-Trichlorobenzene	13.439	180	13644	1.011	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042460.D
Acq On : 01 May 2025 10:08
Operator : SY/MD
Sample : VSTDIC001
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
Supervised By :Mahesh Dadoda 05/05/2025


Quant Time: May 02 01:41:24 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042461.D
 Acq On : 01 May 2025 10:40
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/02/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:42:12 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	68913	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	193701	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	166625	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 150277 9.728 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 97.300%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	6501	0.570	ppbv	94
3) Chlorodifluoromethane	1.476	51	6500	0.539	ppbv	99
4) Chloromethane	1.535	50	2542	0.533	ppbv	95
5) Vinyl Chloride	1.580	62	2441	0.509	ppbv	92
6) Bromomethane	1.667	94	1202	0.530	ppbv	97
7) Chloroethane	1.703	64	862	0.486	ppbv #	68
8) Dichlorotetrafluoroethane	1.554	85	4979	0.530	ppbv	93
9) Propene	1.486	41	2017	0.511	ppbv	96
10) Heptane	4.982	43	4397m	0.430	ppbv	
11) Trichlorofluoromethane	1.878	101	5720	0.562	ppbv	95
12) 1,1,2-Trichlorotrifluo...	2.137	101	4540	0.540	ppbv	94
13) Ethanol	1.722	45	474m	0.592	ppbv	
14) Bromoethene	1.784	108	1680	0.542	ppbv #	87
15) Acetone	1.839	43	7680	0.610	ppbv #	82
16) 1,3-Butadiene	1.609	39	3044	0.549	ppbv	86
17) tert-Butyl alcohol	2.046	59	6256	0.564	ppbv #	91
18) 1,1-Dichloroethene	2.036	96	2240	0.571	ppbv	94
19) Isopropyl Alcohol	1.887	45	4208	0.604	ppbv #	96
20) Methylene Chloride	2.062	84	2594	0.686	ppbv #	80
21) Allyl Chloride	2.098	41	4420	0.546	ppbv #	93
22) trans-1,2-Dichloroethene	2.340	96	2554	0.557	ppbv	95
23) Vinyl Acetate	2.460	43	6541	0.558	ppbv #	95
24) 1,1-Dichloroethane	2.412	63	4933	0.528	ppbv	98
25) Ethyl Acetate	2.842	43	10344	0.514	ppbv #	96
26) Hexane	2.829	57	4403	0.502	ppbv	95
27) Carbon Disulfide	2.153	76	5201	0.508	ppbv #	1
28) Methyl tert-Butyl Ether	2.441	73	2762	0.524	ppbv #	82
29) Chloroform	2.852	83	8791	0.535	ppbv	92
30) Cyclohexane	3.855	84	3910m	0.533	ppbv	
31) cis-1,2-Dichloroethene	2.726	61	5345m	0.522	ppbv	
32) 1,1,1-Trichloroethane	3.366	97	8846m	0.496	ppbv	
34) 2-Butanone	2.564	43	7766	0.554	ppbv	93
35) Carbon Tetrachloride	3.768	117	8286	0.492	ppbv	86
36) Benzene	3.655	78	8951	0.490	ppbv #	92
37) 1,2-Dichloroethane	3.224	62	7149	0.507	ppbv #	95
38) Trichloroethene	4.548	130	3769m	0.541	ppbv	
39) 1,2-Dichloropropane	4.321	63	3561m	0.520	ppbv	
40) 1,4-Dioxane	4.619	88	1701m	0.534	ppbv	
41) Tetrahydrofuran	3.069	42	3187m	0.480	ppbv	
42) Bromodichloromethane	4.496	83	8212	0.477	ppbv	93
43) Methyl Methacrylate	4.891	69	3179m	0.467	ppbv	
44) 2,2,4-Trimethylpentane	4.648	57	14435m	0.489	ppbv	
45) t-1,3-Dichloropropene	6.415	75	4238m	0.473	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042461.D
 Acq On : 01 May 2025 10:40
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:42:12 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.600	75	5091m	0.479	ppbv	
47) 1,1,2-Trichloroethane	6.584	97	3606m	0.480	ppbv	
48) Dibromochloromethane	7.390	129	6304m	0.498	ppbv	
49) Bromoform	9.484	173	5228	0.492	ppbv	97
50) 4-Methyl-2-Pentanone	5.771	43	8248m	0.466	ppbv	
51) 2-Hexanone	7.454	43	6283m	0.443	ppbv	
52) Tetrachloroethene	8.221	164	3151m	0.453	ppbv	
53) Toluene	6.933	91	9052m	0.468	ppbv	
54) 1,2-Dibromoethane	7.652	107	5705	0.494	ppbv	83
56) 1,1,1,2-Tetrachloroethane	8.924	131	5152	0.524	ppbv #	92
57) Chlorobenzene	8.924	112	9173	0.548	ppbv #	86
58) Ethyl Benzene	9.357	91	12427	0.446	ppbv	90
59) m/p-Xylene	9.552	91	20195m	0.865	ppbv	
60) o-Xylene	9.966	91	9976m	0.427	ppbv	
61) Styrene	9.872	104	3457m	0.396	ppbv	
62) Isopropylbenzene	10.545	105	14837	0.440	ppbv	95
63) 1,1,2,2-Tetrachloroethane	9.963	83	8135	0.537	ppbv	96
64) n-propylbenzene	10.992	120	3329m	0.405	ppbv	
65) tert-Butylbenzene	11.536	119	12655	0.421	ppbv	99
66) Benzyl Chloride	11.613	91	1317m	0.379	ppbv	
67) sec-Butylbenzene	11.769	105	17905	0.419	ppbv	97
69) p-Isopropyltoluene	11.927	119	15036	0.417	ppbv	94
70) n-Butylbenzene	12.283	91	15016	0.403	ppbv	93
71) 2-Chlorotoluene	10.905	91	11702	0.443	ppbv	99
72) 4-Ethyltoluene	11.128	105	10671	0.387	ppbv #	91
73) 1,3,5-Trimethylbenzene	11.206	105	9808	0.405	ppbv	89
74) 1,2,4-Trimethylbenzene	11.539	105	13060	0.445	ppbv #	83
75) 1,3-Dichlorobenzene	11.607	146	8127	0.498	ppbv	96
76) 1,4-Dichlorobenzene	11.675	146	7238	0.469	ppbv	96
77) 1,2-Dichlorobenzene	11.950	146	8141	0.509	ppbv	94
78) Hexachloro-1,3-Butadiene	13.886	225	8565	0.537	ppbv	99
79) Naphthalene	13.517	128	9735m	0.396	ppbv	
80) Naphthalene,2-methyl-	14.458	142	4153	0.226	ppbv	98
81) 1,2,4-Trichlorobenzene	13.445	180	4992	0.379	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042462.D
 Acq On : 01 May 2025 11:12
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:43:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri May 02 01:37:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	65449	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	186907	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	157618	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	143644	9.830	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.300%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	631	0.138	ppbv #	53
32) 1,1,1-Trichloroethane	3.363	97	1860m	0.110	ppbv	
35) Carbon Tetrachloride	3.758	117	1493	0.092	ppbv #	55
38) Trichloroethene	4.551	130	784m	0.117	ppbv	
52) Tetrachloroethene	8.221	164	710m	0.106	ppbv	
54) 1,2-Dibromoethane	7.665	107	1169m	0.105	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.966	83	1490	0.104	ppbv #	75
79) Naphthalene	13.517	128	1135m	0.049	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042462.D
Acq On : 01 May 2025 11:12
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:43:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 02 01:37:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042465.D
 Acq On : 01 May 2025 12:49
 Operator : SY/MD
 Sample : VSTDICC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:45:25 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	64314	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	178877	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	157673	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 151837 10.387 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 103.900%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	158069	14.852	ppbv	99
3) Chlorodifluoromethane	1.476	51	154556	13.735	ppbv	100
4) Chloromethane	1.535	50	64267	14.447	ppbv	95
5) Vinyl Chloride	1.580	62	60873	13.592	ppbv	98
6) Bromomethane	1.667	94	31246	14.766	ppbv	87
7) Chloroethane	1.703	64	25181	15.227	ppbv	92
8) Dichlorotetrafluoroethane	1.554	85	131995	15.044	ppbv	99
9) Propene	1.486	41	49647	13.489	ppbv	99
10) Heptane	4.982	43	142235	14.898	ppbv	100
11) Trichlorofluoromethane	1.878	101	152655	16.082	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.140	101	115855	14.775	ppbv	96
13) Ethanol	1.722	45	8090	10.832	ppbv	98
14) Bromoethene	1.781	108	44928	15.519	ppbv	96
15) Acetone	1.836	43	141306	12.036	ppbv	99
16) 1,3-Butadiene	1.609	39	72455	13.992	ppbv	98
17) tert-Butyl alcohol	2.043	59	144935	14.012	ppbv	99
18) 1,1-Dichloroethene	2.033	96	52813	14.424	ppbv	98
19) Isopropyl Alcohol	1.887	45	86000	13.235	ppbv #	1
20) Methylene Chloride	2.062	84	46063	13.050	ppbv	94
21) Allyl Chloride	2.098	41	112451	14.893	ppbv	99
22) trans-1,2-Dichloroethene	2.341	96	62153	14.534	ppbv	100
23) Vinyl Acetate	2.464	43	169418	15.489	ppbv	99
24) 1,1-Dichloroethane	2.409	63	127567	14.622	ppbv	97
25) Ethyl Acetate	2.836	43	269372	14.329	ppbv	100
26) Hexane	2.826	57	119552	14.598	ppbv	97
27) Carbon Disulfide	2.150	76	150258	15.722	ppbv #	92
28) Methyl tert-Butyl Ether	2.441	73	71341	14.500	ppbv	99
29) Chloroform	2.849	83	221976	14.464	ppbv	99
30) Cyclohexane	3.859	84	101146	14.771	ppbv	99
31) cis-1,2-Dichloroethene	2.723	61	133735	13.997	ppbv	93
32) 1,1,1-Trichloroethane	3.370	97	234752	14.099	ppbv	99
34) 2-Butanone	2.557	43	174521	13.488	ppbv	97
35) Carbon Tetrachloride	3.765	117	242144	15.566	ppbv	99
36) Benzene	3.658	78	239352	14.179	ppbv	98
37) 1,2-Dichloroethane	3.221	62	190179	14.618	ppbv	98
38) Trichloroethene	4.551	130	95257	14.798	ppbv	94
39) 1,2-Dichloropropane	4.312	63	84414	13.349	ppbv	93
40) 1,4-Dioxane	4.584	88	38551	13.115	ppbv #	100
41) Tetrahydrofuran	3.056	42	86296	14.063	ppbv	98
42) Bromodichloromethane	4.493	83	242556	15.259	ppbv	99
43) Methyl Methacrylate	4.885	69	98279	15.639	ppbv	100
44) 2,2,4-Trimethylpentane	4.642	57	401920	14.734	ppbv	99
45) t-1,3-Dichloropropene	6.419	75	136762	16.529	ppbv	92

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042465.D
 Acq On : 01 May 2025 12:49
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:45:25 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025


Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.600	75	154705m	15.749	ppbv	
47) 1,1,2-Trichloroethane	6.587	97	99496	14.356	ppbv	93
48) Dibromochloromethane	7.393	129	180970	15.468	ppbv	98
49) Bromoform	9.487	173	152440	15.535	ppbv	99
50) 4-Methyl-2-Pentanone	5.755	43	237985	14.550	ppbv	97
51) 2-Hexanone	7.438	43	197423	15.071	ppbv	99
52) Tetrachloroethene	8.228	164	85271	13.272	ppbv	90
53) Toluene	6.937	91	281731	15.768	ppbv	100
54) 1,2-Dibromoethane	7.655	107	154739	14.523	ppbv	94
56) 1,1,1,2-Tetrachloroethane	8.927	131	133410	14.326	ppbv	98
57) Chlorobenzene	8.927	112	218370	13.783	ppbv	99
58) Ethyl Benzene	9.358	91	414805	15.743	ppbv	98
59) m/p-Xylene	9.555	91	699743m	31.674	ppbv	
60) o-Xylene	9.966	91	350328	15.863	ppbv	100
61) Styrene	9.876	104	137720	16.686	ppbv	96
62) Isopropylbenzene	10.542	105	498619	15.629	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.966	83	209726	14.635	ppbv	99
64) n-propylbenzene	10.992	120	122452	15.759	ppbv	97
65) tert-Butylbenzene	11.536	119	443042	15.562	ppbv	97
66) Benzyl Chloride	11.617	91	60376	18.348	ppbv	97
67) sec-Butylbenzene	11.769	105	632600	15.636	ppbv	100
69) p-Isopropyltoluene	11.928	119	539860	15.837	ppbv	99
70) n-Butylbenzene	12.284	91	566780	16.084	ppbv	99
71) 2-Chlorotoluene	10.905	91	394227	15.755	ppbv	100
72) 4-Ethyltoluene	11.128	105	423746	16.250	ppbv	99
73) 1,3,5-Trimethylbenzene	11.206	105	362951	15.826	ppbv	96
74) 1,2,4-Trimethylbenzene	11.539	105	407416	14.683	ppbv	96
75) 1,3-Dichlorobenzene	11.610	146	225215	14.579	ppbv	98
76) 1,4-Dichlorobenzene	11.675	146	224505	15.357	ppbv	99
77) 1,2-Dichlorobenzene	11.950	146	216147	14.295	ppbv	99
78) Hexachloro-1,3-Butadiene	13.886	225	198112	13.121	ppbv	99
79) Naphthalene	13.517	128	392900	16.888	ppbv	98
80) Naphthalene,2-methyl-	14.459	142	316493	18.232	ppbv	99
81) 1,2,4-Trichlorobenzene	13.442	180	195942	15.728	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED

Quant Time: May 02 01:45:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri May 02 01:37:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042466.D
 Acq On : 01 May 2025 14:10
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:46:15 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.791	49	66546	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	186737	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	161050	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	148088	9.918	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.200%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	151	0.033	ppbv #	87
32) 1,1,1-Trichloroethane	3.370	97	595m	0.035	ppbv	
35) Carbon Tetrachloride	3.771	117	653m	0.040	ppbv	
38) Trichloroethene	4.551	130	135	0.020	ppbv #	40
52) Tetrachloroethene	8.231	164	285m	0.042	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.973	83	551m	0.038	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
Data File : VL042466.D
Acq On : 01 May 2025 14:10
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:46:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 02 01:37:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL050125

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 03:48:56 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	63539	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	183679	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	156982	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.380 95 150522 10.342 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 103.400%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	108782	10.346	ppbv	97
3) Chlorodifluoromethane	1.476	51	110290	9.921	ppbv	98
4) Chloromethane	1.534	50	43950	10.001	ppbv	96
5) Vinyl Chloride	1.580	62	41181	9.307	ppbv	99
6) Bromomethane	1.667	94	21021	10.055	ppbv	91
7) Chloroethane	1.706	64	16630	10.179	ppbv	90
8) Dichlorotetrafluoroethane	1.554	85	89345	10.307	ppbv	99
9) Propene	1.486	41	36920	10.153	ppbv	97
10) Heptane	4.985	43	94631	10.033	ppbv	95
11) Trichlorofluoromethane	1.878	101	100713	10.380	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.140	101	81143	10.474	ppbv	99
13) Ethanol	1.722	45	5736	7.773	ppbv	99
14) Bromoethene	1.780	108	30094	10.522	ppbv	97
15) Acetone	1.835	43	95460	8.230	ppbv	98
16) 1,3-Butadiene	1.609	39	49566	9.689	ppbv	100
17) tert-Butyl alcohol	2.043	59	102047	9.986	ppbv	100
18) 1,1-Dichloroethene	2.033	96	35784	9.892	ppbv	98
19) Isopropyl Alcohol	1.887	45	60977	9.499	ppbv	99
20) Methylene Chloride	2.062	84	31212	8.951	ppbv	94
21) Allyl Chloride	2.098	41	73365	9.835	ppbv	94
22) trans-1,2-Dichloroethene	2.340	96	42922	10.160	ppbv	96
23) Vinyl Acetate	2.463	43	113131	10.220	ppbv #	98
24) 1,1-Dichloroethane	2.408	63	89117	10.340	ppbv	97
25) Ethyl Acetate	2.839	43	186185	10.025	ppbv	99
26) Hexane	2.826	57	83132	10.275	ppbv	96
27) Carbon Disulfide	2.149	76	100196	10.612	ppbv	96
28) Methyl tert-Butyl Ether	2.441	73	48326	9.942	ppbv	100
29) Chloroform	2.852	83	152381	10.050	ppbv	96
30) Cyclohexane	3.858	84	70581	10.434	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	95793	10.148	ppbv	96
32) 1,1,1-Trichloroethane	3.370	97	161465	9.816	ppbv	99
34) 2-Butanone	2.560	43	125883	9.475	ppbv	97
35) Carbon Tetrachloride	3.764	117	166814	9.998	ppbv	97
36) Benzene	3.658	78	171384	9.888	ppbv	98
37) 1,2-Dichloroethane	3.221	62	130063	9.736	ppbv	100
38) Trichloroethene	4.551	130	65940	9.976	ppbv	96
39) 1,2-Dichloropropane	4.315	63	60010	9.248	ppbv	95
40) 1,4-Dioxane	4.583	88	28306m	9.316	ppbv	
41) Tetrahydrofuran	3.059	42	59604	9.440	ppbv	98
42) Bromodichloromethane	4.489	83	163210	9.999	ppbv	95
43) Methyl Methacrylate	4.881	69	67516	10.438	ppbv	99
44) 2,2,4-Trimethylpentane	4.642	57	277217	9.867	ppbv	100
45) t-1,3-Dichloropropene	6.415	75	89638m	10.563	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL050125

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/02/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 02 03:48:56 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.609	75	104665	10.378	ppbv	97
47) 1,1,2-Trichloroethane	6.584	97	67821	9.530	ppbv	95
48) Dibromochloromethane	7.393	129	121381	10.103	ppbv	97
49) Bromoform	9.487	173	102150	10.138	ppbv	100
50) 4-Methyl-2-Pentanone	5.752	43	159280	9.478	ppbv	99
51) 2-Hexanone	7.438	43	126396	9.407	ppbv #	98
52) Tetrachloroethene	8.228	164	58272	8.832	ppbv	96
53) Toluene	6.940	91	191922	10.461	ppbv	99
54) 1,2-Dibromoethane	7.658	107	105149	9.664	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.927	131	92337	9.959	ppbv	97
57) Chlorobenzene	8.927	112	148759	9.431	ppbv	92
58) Ethyl Benzene	9.357	91	280660	10.698	ppbv	99
59) m/p-Xylene	9.555	91	468984m	21.324	ppbv	
60) o-Xylene	9.966	91	238347	10.824	ppbv	97
61) Styrene	9.872	104	91128	11.090	ppbv	97
62) Isopropylbenzene	10.542	105	337992	10.641	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.966	83	140306	9.242	ppbv	100
64) n-propylbenzene	10.992	120	79983	10.303	ppbv	99
65) tert-Butylbenzene	11.536	119	302506	10.672	ppbv	99
66) Benzyl Chloride	11.617	91	34933	10.663	ppbv	97
67) sec-Butylbenzene	11.769	105	425444	10.562	ppbv	99
69) p-Isopropyltoluene	11.927	119	363563	10.712	ppbv	98
70) n-Butylbenzene	12.283	91	371369	10.585	ppbv	99
71) 2-Chlorotoluene	10.905	91	264063	10.599	ppbv	99
72) 4-Ethyltoluene	11.128	105	287375	11.069	ppbv	99
73) 1,3,5-Trimethylbenzene	11.206	105	247128	10.823	ppbv	93
74) 1,2,4-Trimethylbenzene	11.539	105	278306	10.074	ppbv	98
75) 1,3-Dichlorobenzene	11.610	146	150026	9.754	ppbv	99
76) 1,4-Dichlorobenzene	11.675	146	151740	10.425	ppbv	97
77) 1,2-Dichlorobenzene	11.950	146	146314	9.719	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	133688	8.893	ppbv	99
79) Naphthalene	13.516	128	251318	10.806	ppbv	98
80) Naphthalene,2-methyl-	14.458	142	198604	11.489	ppbv	99
81) 1,2,4-Trichlorobenzene	13.442	180	129406	10.433	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL050125

Quant Time: May 02 03:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	1.655	1.712	-3.4	104	0.00
3	Chlorodifluoromethane	1.750	1.736	0.8	98	0.00
4	Chloromethane	0.692	0.692	0.0	99	0.00
5 T	Vinyl Chloride	0.696	0.648	6.9	102	0.00
6 T	Bromomethane	0.329	0.331	-0.6	103	0.00
7	Chloroethane	0.257	0.262	-1.9	101	0.00
8 T	Dichlorotetrafluoroethane	1.364	1.406	-3.1	101	0.00
9 T	Propene	0.572	0.581	-1.6	98	0.00
10 T	Heptane	1.484	1.489	-0.3	92	0.00
11 T	Trichlorofluoromethane	1.527	1.585	-3.8	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.219	1.277	-4.8	103	0.00
13	Ethanol	0.116	0.090	22.4	101	0.00
14 T	Bromoethene	0.450	0.474	-5.3	106	0.00
15 T	Acetone	1.826	1.502	17.7	101	0.00
16 T	1,3-Butadiene	0.805	0.780	3.1	102	0.00
17	tert-Butyl alcohol	1.608	1.606	0.1	96	0.00
18 T	1,1-Dichloroethene	0.569	0.563	1.1	101	0.00
19 T	Isopropyl Alcohol	1.010	0.960	5.0	98	0.00
20 T	Methylene Chloride	0.549	0.491	10.6	97	0.00
21 T	Allyl Chloride	1.174	1.155	1.6	95	0.00
22 T	trans-1,2-Dichloroethene	0.665	0.676	-1.7	109	0.00
23 T	Vinyl Acetate	1.742	1.780	-2.2	92	0.00
24 T	1,1-Dichloroethane	1.357	1.403	-3.4	104	0.00
25 T	Ethyl Acetate	2.923	2.930	-0.2	92	0.00
26 T	Hexane	1.273	1.308	-2.7	96	0.00
27 T	Carbon Disulfide	1.486	1.577	-6.1	99	0.00
28 T	Methyl tert-Butyl Ether	0.765	0.761	0.5	101	0.00
29 T	Chloroform	2.386	2.398	-0.5	97	0.00
30 T	Cyclohexane	1.065	1.111	-4.3	95	0.00
31 T	cis-1,2-Dichloroethene	1.486	1.508	-1.5	100	0.00
32 T	1,1,1-Trichloroethane	2.589	2.541	1.9	99	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
34 T	2-Butanone	0.723	0.685	5.3	95	0.00
35 T	Carbon Tetrachloride	0.908	0.908	0.0	100	0.00
36 T	Benzene	0.944	0.933	1.2	95	0.00
37 T	1,2-Dichloroethane	0.727	0.708	2.6	97	0.00
38 T	Trichloroethene	0.360	0.359	0.3	94	0.00
39 T	1,2-Dichloropropane	0.353	0.327	7.4	94	0.00
40 T	1,4-Dioxane	0.165	0.154	6.7	93	0.00
41 T	Tetrahydrofuran	0.344	0.325	5.5	91	0.00
42 T	Bromodichloromethane	0.889	0.889	0.0	97	0.00
43	Methyl Methacrylate	0.352	0.368	-4.5	92	0.00
44 T	2,2,4-Trimethylpentane	1.530	1.509	1.4	96	0.00
45 T	t-1,3-Dichloropropene	0.462	0.488	-5.6	93	0.00
46 T	cis-1,3-Dichloropropene	0.549	0.570	-3.8	94	0.00
47 T	1,1,2-Trichloroethane	0.387	0.369	4.7	94	0.00
48 T	Dibromochloromethane	0.654	0.661	-1.1	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042467.D
 Acq On : 01 May 2025 15:46
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL050125

Quant Time: May 02 03:48:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.549	0.556	-1.3	96	0.00
50 T	4-Methyl-2-Pentanone	0.915	0.867	5.2	91	0.00
51 T	2-Hexanone	0.731	0.688	5.9	87	0.00
52 T	Tetrachloroethene	0.359	0.317	11.7	95	0.00
53 T	Toluene	0.999	1.045	-4.6	94	0.00
54 T	1,2-Dibromoethane	0.592	0.572	3.4	94	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
56	1,1,1,2-Tetrachloroethane	0.591	0.588	0.5	98	0.00
57 T	Chlorobenzene	1.005	0.948	5.7	95	0.00
58 T	Ethyl Benzene	1.671	1.788	-7.0	93	0.00
59 T	m/p-Xylene	1.401	1.494	-6.6	94	0.03
60 T	o-Xylene	1.403	1.518	-8.2	96	0.00
61 T	Styrene	0.523	0.580	-10.9	90	0.00
62	Isopropylbenzene	2.023	2.153	-6.4	95	0.00
63 T	1,1,2,2-Tetrachloroethane	0.967	0.894	7.5	95	0.00
64	n-propylbenzene	0.495	0.510	-3.0	95	0.00
65	tert-Butylbenzene	1.806	1.927	-6.7	96	0.00
66 T	Benzyl Chloride	0.209	0.223	-6.7	87	0.00
67	sec-Butylbenzene	2.566	2.710	-5.6	96	0.00
68 S	1-Bromo-4-Fluorobenzene	0.927	0.959	-3.5	91	0.00
69	p-Isopropyltoluene	2.162	2.316	-7.1	97	0.00
70	n-Butylbenzene	2.235	2.366	-5.9	97	0.00
71	2-Chlorotoluene	1.587	1.682	-6.0	95	0.00
72 T	4-Ethyltoluene	1.654	1.831	-10.7	96	0.00
73 T	1,3,5-Trimethylbenzene	1.455	1.574	-8.2	96	0.00
74 T	1,2,4-Trimethylbenzene	1.760	1.773	-0.7	97	0.00
75 T	1,3-Dichlorobenzene	0.980	0.956	2.4	96	0.00
76 T	1,4-Dichlorobenzene	0.927	0.967	-4.3	96	0.00
77 T	1,2-Dichlorobenzene	0.959	0.932	2.8	97	0.00
78 T	Hexachloro-1,3-Butadiene	0.958	0.852	11.1	99	0.00
79 T	Naphthalene	1.482	1.601	-8.0	94	0.00
80 T	Naphthalene,2-methyl-	1.101	1.265	-14.9	97	0.00
81 T	1,2,4-Trichlorobenzene	0.790	0.824	-4.3	96	0.00

(#) = Out of Range



SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
Data File : VL042468.D
Acq On : 05 May 2025 07:45
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri May 02 03:16:31 2025



AutoFind: Scans 3179, 3180, 3181; Background Corrected with Scan 3162

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	26.0	66184	PASS
75	95	30	66	62.7	159424	PASS
95	95	100	100	100.0	254123	PASS
96	95	5	9	6.3	15968	PASS
173	174	0.00	2	1.3	1836	PASS
174	95	50	120	55.0	139883	PASS
175	174	4	9	7.8	10890	PASS
176	174	93	101	97.0	135672	PASS
177	176	5	9	5.8	7826	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
 Data File : VL042469.D
 Acq On : 05 May 2025 08:29
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 01:38:03 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	74861	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	207859	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	180939	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.380 95 171319 10.212 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	125145	10.102	ppbv	99
3) Chlorodifluoromethane	1.476	51	122899	9.383	ppbv	99
4) Chloromethane	1.534	50	49599	9.579	ppbv	92
5) Vinyl Chloride	1.580	62	46672	8.953	ppbv	98
6) Bromomethane	1.667	94	26234	10.651	ppbv	92
7) Chloroethane	1.706	64	19310	10.032	ppbv	99
8) Dichlorotetrafluoroethane	1.554	85	104312	10.214	ppbv	99
9) Propene	1.486	41	39686	9.264	ppbv	98
10) Heptane	4.985	43	109506	9.854	ppbv	99
11) Trichlorofluoromethane	1.878	101	117806	10.305	ppbv	99
12) 1,1,2-Trichlorotrifluoro...	2.140	101	92903	10.178	ppbv	99
13) Ethanol	1.722	45	6032	6.938	ppbv	97
14) Bromoethene	1.784	108	33877	10.053	ppbv	99
15) Acetone	1.835	43	110962	8.120	ppbv	99
16) 1,3-Butadiene	1.609	39	56863	9.434	ppbv	99
17) tert-Butyl alcohol	2.043	59	124285	10.323	ppbv	100
18) 1,1-Dichloroethene	2.036	96	42276	9.919	ppbv	96
19) Isopropyl Alcohol	1.887	45	73049	9.658	ppbv	99
20) Methylene Chloride	2.062	84	36341	8.845	ppbv	96
21) Allyl Chloride	2.098	41	87203	9.922	ppbv	98
22) trans-1,2-Dichloroethene	2.340	96	47159	9.474	ppbv	92
23) Vinyl Acetate	2.463	43	134402	10.305	ppbv	98
24) 1,1-Dichloroethane	2.412	63	97677	9.619	ppbv	100
25) Ethyl Acetate	2.839	43	212175	9.697	ppbv	100
26) Hexane	2.826	57	95290	9.996	ppbv	95
27) Carbon Disulfide	2.153	76	120890	10.867	ppbv #	92
28) Methyl tert-Butyl Ether	2.444	73	54017	9.432	ppbv	96
29) Chloroform	2.852	83	173954	9.738	ppbv	97
30) Cyclohexane	3.858	84	78668	9.871	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	107420	9.659	ppbv	96
32) 1,1,1-Trichloroethane	3.370	97	184899	9.540	ppbv	99
34) 2-Butanone	2.560	43	134813	8.967	ppbv	93
35) Carbon Tetrachloride	3.765	117	191095	10.120	ppbv	97
36) Benzene	3.661	78	192975	9.838	ppbv	96
37) 1,2-Dichloroethane	3.221	62	146792	9.710	ppbv	99
38) Trichloroethene	4.554	130	76779	10.264	ppbv	89
39) 1,2-Dichloropropane	4.318	63	67421	9.181	ppbv	93
40) 1,4-Dioxane	4.590	88	29648	8.622	ppbv #	100
41) Tetrahydrofuran	3.059	42	66188	9.263	ppbv	97
42) Bromodichloromethane	4.496	83	191130	10.347	ppbv	98
43) Methyl Methacrylate	4.881	69	76368	10.433	ppbv	99
44) 2,2,4-Trimethylpentane	4.645	57	312811	9.838	ppbv	100
45) t-1,3-Dichloropropene	6.422	75	105437	10.980	ppbv	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
 Data File : VL042469.D
 Acq On : 05 May 2025 08:29
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 01:38:03 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025


Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.609	75	117796	10.321	ppbv	96
47) 1,1,2-Trichloroethane	6.590	97	77722	9.651	ppbv	90
48) Dibromochloromethane	7.393	129	143524	10.556	ppbv	97
49) Bromoform	9.490	173	118610	10.402	ppbv	99
50) 4-Methyl-2-Pentanone	5.758	43	183038	9.625	ppbv	98
51) 2-Hexanone	7.445	43	147526	9.703	ppbv #	97
52) Tetrachloroethene	8.231	164	67482	9.038	ppbv	95
53) Toluene	6.940	91	220871	10.638	ppbv	98
54) 1,2-Dibromoethane	7.658	107	119914	9.739	ppbv	96
56) 1,1,1,2-Tetrachloroethane	8.930	131	106249	9.942	ppbv	99
57) Chlorobenzene	8.927	112	168792	9.284	ppbv	96
58) Ethyl Benzene	9.361	91	317094	10.487	ppbv	100
59) m/p-Xylene	9.558	91	541542m	21.363	ppbv	
60) o-Xylene	9.969	91	267806	10.552	ppbv	100
61) Styrene	9.879	104	105698	11.160	ppbv	97
62) Isopropylbenzene	10.545	105	385788	10.538	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.966	83	159883	9.137	ppbv	99
64) n-propylbenzene	10.992	120	93317	10.429	ppbv	98
65) tert-Butylbenzene	11.536	119	349071	10.684	ppbv	99
66) Benzyl Chloride	11.620	91	50060	13.257	ppbv	96
67) sec-Butylbenzene	11.772	105	490679	10.568	ppbv	99
69) p-Isopropyltoluene	11.931	119	417116	10.663	ppbv	99
70) n-Butylbenzene	12.287	91	427083	10.561	ppbv	99
71) 2-Chlorotoluene	10.905	91	301623	10.504	ppbv	98
72) 4-Ethyltoluene	11.128	105	326323	10.905	ppbv	100
73) 1,3,5-Trimethylbenzene	11.209	105	283998	10.791	ppbv	98
74) 1,2,4-Trimethylbenzene	11.542	105	320216	10.056	ppbv	94
75) 1,3-Dichlorobenzene	11.610	146	171827	9.693	ppbv	99
76) 1,4-Dichlorobenzene	11.675	146	177262	10.566	ppbv	99
77) 1,2-Dichlorobenzene	11.950	146	167623	9.660	ppbv	99
78) Hexachloro-1,3-Butadiene	13.885	225	154431	8.913	ppbv	100
79) Naphthalene	13.517	128	294069	10.970	ppbv	98
80) Naphthalene,2-methyl-	14.462	142	236563	11.873	ppbv	99
81) 1,2,4-Trichlorobenzene	13.445	180	151252	10.580	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED

Quant Time: May 06 01:38:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
 Data File : VL042469.D
 Acq On : 05 May 2025 08:29
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 06 01:38:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	1.655	1.672	-1.0	119	0.00
3	Chlorodifluoromethane	1.750	1.642	6.2	110	0.00
4	Chloromethane	0.692	0.663	4.2	111	0.00
5 T	Vinyl Chloride	0.696	0.623	10.5	115	0.00
6 T	Bromomethane	0.329	0.350	-6.4	129	0.00
7	Chloroethane	0.257	0.258	-0.4	117	0.00
8 T	Dichlorotetrafluoroethane	1.364	1.393	-2.1	118	0.00
9 T	Propene	0.572	0.530	7.3	106	0.00
10 T	Heptane	1.484	1.463	1.4	106	0.00
11 T	Trichlorofluoromethane	1.527	1.574	-3.1	120	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.219	1.241	-1.8	118	0.00
13	Ethanol	0.116	0.081	30.2#	106	0.00
14 T	Bromoethene	0.450	0.453	-0.7	119	0.00
15 T	Acetone	1.826	1.482	18.8	118	0.00
16 T	1,3-Butadiene	0.805	0.760	5.6	117	0.00
17	tert-Butyl alcohol	1.608	1.660	-3.2	117	0.00
18 T	1,1-Dichloroethene	0.569	0.565	0.7	120	0.00
19 T	Isopropyl Alcohol	1.010	0.976	3.4	117	0.00
20 T	Methylene Chloride	0.549	0.485	11.7	113	0.00
21 T	Allyl Chloride	1.174	1.165	0.8	113	0.00
22 T	trans-1,2-Dichloroethene	0.665	0.630	5.3	120	0.00
23 T	Vinyl Acetate	1.742	1.795	-3.0	109	0.00
24 T	1,1-Dichloroethane	1.357	1.305	3.8	114	0.00
25 T	Ethyl Acetate	2.923	2.834	3.0	105	0.00
26 T	Hexane	1.273	1.273	0.0	110	0.00
27 T	Carbon Disulfide	1.486	1.615	-8.7	119	0.00
28 T	Methyl tert-Butyl Ether	0.765	0.722	5.6	113	0.00
29 T	Chloroform	2.386	2.324	2.6	111	0.00
30 T	Cyclohexane	1.065	1.051	1.3	105	0.00
31 T	cis-1,2-Dichloroethene	1.486	1.435	3.4	112	0.00
32 T	1,1,1-Trichloroethane	2.589	2.470	4.6	113	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
34 T	2-Butanone	0.723	0.649	10.2	102	0.00
35 T	Carbon Tetrachloride	0.908	0.919	-1.2	115	0.00
36 T	Benzene	0.944	0.928	1.7	107	0.00
37 T	1,2-Dichloroethane	0.727	0.706	2.9	110	0.00
38 T	Trichloroethene	0.360	0.369	-2.5	110	0.00
39 T	1,2-Dichloropropane	0.353	0.324	8.2	106	0.00
40 T	1,4-Dioxane	0.165	0.143	13.3	98	0.00
41 T	Tetrahydrofuran	0.344	0.318	7.6	101	0.00
42 T	Bromodichloromethane	0.889	0.920	-3.5	113	0.00
43	Methyl Methacrylate	0.352	0.367	-4.3	104	0.00
44 T	2,2,4-Trimethylpentane	1.530	1.505	1.6	108	0.00
45 T	t-1,3-Dichloropropene	0.462	0.507	-9.7	109	0.00
46 T	cis-1,3-Dichloropropene	0.549	0.567	-3.3	105	0.00
47 T	1,1,2-Trichloroethane	0.387	0.374	3.4	108	0.00
48 T	Dibromochloromethane	0.654	0.690	-5.5	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
 Data File : VL042469.D
 Acq On : 05 May 2025 08:29
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 06 01:38:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.549	0.571	-4.0	112	0.00
50 T	4-Methyl-2-Pentanone	0.915	0.881	3.7	104	0.00
51 T	2-Hexanone	0.731	0.710	2.9	102	0.00
52 T	Tetrachloroethene	0.359	0.325	9.5	110	0.00
53 T	Toluene	0.999	1.063	-6.4	108	0.00
54 T	1,2-Dibromoethane	0.592	0.577	2.5	108	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
56	1,1,1,2-Tetrachloroethane	0.591	0.587	0.7	113	0.00
57 T	Chlorobenzene	1.005	0.933	7.2	107	0.00
58 T	Ethyl Benzene	1.671	1.752	-4.8	106	0.00
59 T	m/p-Xylene	1.401	1.496	-6.8	109	0.03
60 T	o-Xylene	1.403	1.480	-5.5	108	0.00
61 T	Styrene	0.523	0.584	-11.7	104	0.00
62	Isopropylbenzene	2.023	2.132	-5.4	109	0.00
63 T	1,1,2,2-Tetrachloroethane	0.967	0.884	8.6	109	0.00
64	n-propylbenzene	0.495	0.516	-4.2	111	0.00
65	tert-Butylbenzene	1.806	1.929	-6.8	110	0.00
66 T	Benzyl Chloride	0.209	0.277	-32.5#	124	0.00
67	sec-Butylbenzene	2.566	2.712	-5.7	111	0.00
68 S	1-Bromo-4-Fluorobenzene	0.927	0.947	-2.2	103	0.00
69	p-Isopropyltoluene	2.162	2.305	-6.6	111	0.00
70	n-Butylbenzene	2.235	2.360	-5.6	112	0.00
71	2-Chlorotoluene	1.587	1.667	-5.0	108	0.00
72 T	4-Ethyltoluene	1.654	1.803	-9.0	109	0.00
73 T	1,3,5-Trimethylbenzene	1.455	1.570	-7.9	111	0.00
74 T	1,2,4-Trimethylbenzene	1.760	1.770	-0.6	112	0.00
75 T	1,3-Dichlorobenzene	0.980	0.950	3.1	110	0.00
76 T	1,4-Dichlorobenzene	0.927	0.980	-5.7	112	0.00
77 T	1,2-Dichlorobenzene	0.959	0.926	3.4	111	0.00
78 T	Hexachloro-1,3-Butadiene	0.958	0.853	11.0	115	0.00
79 T	Naphthalene	1.482	1.625	-9.6	111	0.00
80 T	Naphthalene,2-methyl-	1.101	1.307	-18.7	116	0.00
81 T	1,2,4-Trichlorobenzene	0.790	0.836	-5.8	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
 Data File : VL042470.D
 Acq On : 05 May 2025 09:19
 Operator : SY/MD
 Sample : VL0505ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0505ABL01

Quant Time: May 06 01:38:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	74757	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	208518	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	177829	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	161047	9.768	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.700%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
Data File : VL042470.D
Acq On : 05 May 2025 09:19
Operator : SY/MD
Sample : VL0505ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0505ABL01

Quant Time: May 06 01:38:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
 Data File : VL042471.D
 Acq On : 05 May 2025 10:07
 Operator : SY/MD
 Sample : VL0505ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0505ABS01

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 01:39:20 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	73094	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	206118	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	177063	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 168649 10.273 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	124629	10.303	ppbv	97
3) Chlorodifluoromethane	1.476	51	121193	9.477	ppbv	100
4) Chloromethane	1.531	50	48459	9.585	ppbv	97
5) Vinyl Chloride	1.580	62	48263	9.482	ppbv	97
6) Bromomethane	1.667	94	24122	10.030	ppbv	89
7) Chloroethane	1.703	64	19046	10.134	ppbv	97
8) Dichlorotetrafluoroethane	1.554	85	103007	10.330	ppbv	99
9) Propene	1.483	41	38988	9.321	ppbv	99
10) Heptane	4.985	43	105620m	9.734	ppbv	
11) Trichlorofluoromethane	1.878	101	115696	10.365	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.140	101	90819	10.191	ppbv	99
13) Ethanol	1.722	45	5973	7.037	ppbv	99
14) Bromoethene	1.780	108	33388	10.148	ppbv	98
15) Acetone	1.835	43	106709	7.997	ppbv	99
16) 1,3-Butadiene	1.609	39	57359	9.746	ppbv	99
17) tert-Butyl alcohol	2.039	59	117319	9.980	ppbv	98
18) 1,1-Dichloroethene	2.033	96	41229	9.907	ppbv	98
19) Isopropyl Alcohol	1.887	45	70522	9.550	ppbv	100
20) Methylene Chloride	2.062	84	35706	8.901	ppbv	94
21) Allyl Chloride	2.094	41	84722	9.873	ppbv	98
22) trans-1,2-Dichloroethene	2.340	96	46305	9.528	ppbv	95
23) Vinyl Acetate	2.463	43	130023	10.210	ppbv #	97
24) 1,1-Dichloroethane	2.408	63	98989	9.984	ppbv	99
25) Ethyl Acetate	2.836	43	207844	9.728	ppbv	100
26) Hexane	2.826	57	90030	9.673	ppbv	98
27) Carbon Disulfide	2.149	76	115286	10.614	ppbv	94
28) Methyl tert-Butyl Ether	2.441	73	58319	10.429	ppbv	100
29) Chloroform	2.849	83	168484	9.660	ppbv	98
30) Cyclohexane	3.858	84	78428	10.078	ppbv	97
31) cis-1,2-Dichloroethene	2.722	61	104191	9.595	ppbv	97
32) 1,1,1-Trichloroethane	3.370	97	182273	9.632	ppbv	99
34) 2-Butanone	2.557	43	133597	8.961	ppbv	95
35) Carbon Tetrachloride	3.761	117	188916	10.090	ppbv	95
36) Benzene	3.661	78	188320	9.682	ppbv	97
37) 1,2-Dichloroethane	3.221	62	145347	9.695	ppbv	100
38) Trichloroethene	4.548	130	74307m	10.017	ppbv	
39) 1,2-Dichloropropane	4.315	63	66905	9.188	ppbv	88
40) 1,4-Dioxane	4.580	88	31401	9.209	ppbv #	100
41) Tetrahydrofuran	3.056	42	66257	9.351	ppbv	97
42) Bromodichloromethane	4.489	83	189214	10.330	ppbv	93
43) Methyl Methacrylate	4.881	69	74953	10.326	ppbv	97
44) 2,2,4-Trimethylpentane	4.642	57	307008	9.737	ppbv	99
45) t-1,3-Dichloropropene	6.415	75	103231	10.841	ppbv	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
 Data File : VL042471.D
 Acq On : 05 May 2025 10:07
 Operator : SY/MD
 Sample : VL0505ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0505ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 01:39:20 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 03:16:31 2025


Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.603	75	116725	10.313	ppbv	95
47) 1,1,2-Trichloroethane	6.584	97	77049	9.648	ppbv #	92
48) Dibromochloromethane	7.390	129	138518	10.274	ppbv	97
49) Bromoform	9.487	173	114364	10.115	ppbv	99
50) 4-Methyl-2-Pentanone	5.752	43	177898	9.433	ppbv	99
51) 2-Hexanone	7.438	43	145213	9.631	ppbv	98
52) Tetrachloroethene	8.228	164	67191	9.076	ppbv	86
53) Toluene	6.936	91	214815	10.434	ppbv	98
54) 1,2-Dibromoethane	7.655	107	117449	9.619	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.927	131	103494	9.897	ppbv	98
57) Chlorobenzene	8.924	112	168941	9.495	ppbv	94
58) Ethyl Benzene	9.357	91	312698	10.568	ppbv	96
59) m/p-Xylene	9.552	91	532663m	21.473	ppbv	
60) o-Xylene	9.966	91	267198	10.758	ppbv	100
61) Styrene	9.875	104	103232	11.138	ppbv	97
62) Isopropylbenzene	10.542	105	380840	10.630	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.966	83	155457	9.079	ppbv	100
64) n-propylbenzene	10.992	120	93814	10.714	ppbv	94
65) tert-Butylbenzene	11.532	119	342014	10.697	ppbv	97
66) Benzyl Chloride	11.620	91	45711	12.370	ppbv	98
67) sec-Butylbenzene	11.769	105	480000	10.565	ppbv	99
69) p-Isopropyltoluene	11.927	119	412889	10.786	ppbv	99
70) n-Butylbenzene	12.283	91	425652	10.756	ppbv	98
71) 2-Chlorotoluene	10.905	91	301813	10.741	ppbv	100
72) 4-Ethyltoluene	11.128	105	326835	11.161	ppbv	99
73) 1,3,5-Trimethylbenzene	11.206	105	280716	10.900	ppbv	97
74) 1,2,4-Trimethylbenzene	11.539	105	316052	10.143	ppbv	99
75) 1,3-Dichlorobenzene	11.610	146	174274	10.046	ppbv	97
76) 1,4-Dichlorobenzene	11.675	146	174860	10.651	ppbv	98
77) 1,2-Dichlorobenzene	11.950	146	166423	9.801	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	153156	9.032	ppbv	99
79) Naphthalene	13.513	128	286218	10.911	ppbv	97
80) Naphthalene,2-methyl-	14.458	142	232641	11.932	ppbv	99
81) 1,2,4-Trichlorobenzene	13.442	180	147081	10.513	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Quant Time: May 06 01:39:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
 Data File : VL042485.D
 Acq On : 05 May 2025 20:26
 Operator : SY/MD
 Sample : QC042825 CAN 10492
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC042825 CAN 10492

Quant Time: May 06 01:46:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	62446	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	169909	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	147043	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	133429	9.787	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.900%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
Data File : VL042485.D
Acq On : 05 May 2025 20:26
Operator : SY/MD
Sample : QC042825 CAN 10492
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC042825 CAN 10492

Quant Time: May 06 01:46:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration



Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL032725

Review By	Maresh Dadoda	Review On	4/3/2025 8:14:07 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	4/3/2025 8:45:16 AM		
SubDirectory	VL032725	HP Acquire Method	TO15AIR.M	HP Processing Method	VL032725AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2589			
Initial Calibration Stds		AP2593,AP2595,AP2596			
CCC		AP2593			
Internal Standard/PEM		AP2589			
ICV/I.BLK		AP2594			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042171.D	27 Mar 2025 07:44	SY/MD	Ok
2	VSTDICCC010	VL042172.D	27 Mar 2025 08:26	SY/MD	Ok,M
3	VSTDICCC002	VL042173.D	27 Mar 2025 08:59	SY/MD	Ok,M
4	VSTDICCC001	VL042174.D	27 Mar 2025 09:30	SY/MD	Ok,M
5	VSTDICCC0.5	VL042175.D	27 Mar 2025 10:01	SY/MD	Ok,M
6	VSTDICCC0.1	VL042176.D	27 Mar 2025 10:32	SY/MD	Ok,M
7	VSTDICCC0.03	VL042177.D	27 Mar 2025 11:03	SY/MD	Ok,M
8	VSTDICCC015	VL042178.D	27 Mar 2025 11:36	SY/MD	Ok,M
9	VSTDICV010	VL042179.D	27 Mar 2025 12:23	SY/MD	Ok,M
10	VL0327ABL01	VL042180.D	27 Mar 2025 13:50	SY/MD	Ok
11	VL0327ABS01	VL042181.D	27 Mar 2025 14:39	SY/MD	Ok,M
12	QC032625 CAN 10448	VL042182.D	27 Mar 2025 16:41	SY/MD	Ok
13	Q1649-01	VL042183.D	27 Mar 2025 17:13	SY/MD	Not Ok
14	Q1649-02	VL042184.D	27 Mar 2025 17:45	SY/MD	Not Ok
15	Q1669-01	VL042185.D	27 Mar 2025 18:22	SY/MD	Dilution
16	Q1669-02	VL042186.D	27 Mar 2025 18:59	SY/MD	Ok,M
17	Q1669-03	VL042187.D	27 Mar 2025 19:36	SY/MD	Ok,M
18	Q1669-04	VL042188.D	27 Mar 2025 20:12	SY/MD	Ok,M
19	Q1669-09	VL042189.D	27 Mar 2025 20:48	SY/MD	Ok,M
20	Q1669-09DUP	VL042190.D	27 Mar 2025 21:24	SY/MD	Ok,M
21	Q1669-05	VL042191.D	27 Mar 2025 21:57	SY/MD	Not Ok

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL032725

Review By	Maresh Dadoda	Review On	4/3/2025 8:14:07 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	4/3/2025 8:45:16 AM		
SubDirectory	VL032725	HP Acquire Method	TO15AIR.M	HP Processing Method	VL032725AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2589			
Initial Calibration Stds		AP2593,AP2595,AP2596			
CCC		AP2593			
Internal Standard/PEM		AP2589			
ICV/I.BLK		AP2594			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q1669-07DL	VL042192.D	27 Mar 2025 22:29	SY/MD	Ok,M
23	Q1669-08	VL042193.D	27 Mar 2025 23:01	SY/MD	Not Ok
24	VIBLK	VL042194.D	27 Mar 2025 23:34	SY/MD	Ok
25	Q1669-01DL	VL042195.D	28 Mar 2025 00:06	SY/MD	Ok,M
26	Q1669-01DUP	VL042196.D	28 Mar 2025 00:44	SY/MD	Ok,M
27	Q1669-05	VL042197.D	28 Mar 2025 01:20	SY/MD	Ok,M
28	Q1669-07	VL042198.D	28 Mar 2025 01:57	SY/MD	Dilution
29	Q1669-08	VL042199.D	28 Mar 2025 02:33	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL041525

Review By	Semsettin Yesilyurt	Review On	4/16/2025 2:19:24 PM		
Supervise By	Maresh Dadoda	Supervise On	4/16/2025 3:47:46 PM		
SubDirectory	VL041525	HP Acquire Method	TO15AIR.M	HP Processing Method	VL032725AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2591			
Initial Calibration Stds		AP2593,AP2595,AP2596			
CCC		AP2593			
Internal Standard/PEM		AP2591			
ICV/I.BLK		AP2594			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042325.D	15 Apr 2025 07:45	SY/MD	Ok
2	VSTDCCC010	VL042326.D	15 Apr 2025 11:30	SY/MD	Ok,M
3	VL0415ABL01	VL042327.D	15 Apr 2025 12:12	SY/MD	Ok
4	VL0415ABS01	VL042328.D	15 Apr 2025 12:54	SY/MD	Ok,M
5	QC041425 CAN 10401	VL042329.D	15 Apr 2025 13:26	SY/MD	Not Ok
6	QC041425 CAN 10401	VL042330.D	15 Apr 2025 13:57	SY/MD	Ok
7	Q1665-18DL	VL042331.D	15 Apr 2025 14:29	SY/MD	Ok
8	Q1665-10DL2	VL042332.D	15 Apr 2025 15:01	SY/MD	Ok,M
9	Q1802-02	VL042333.D	15 Apr 2025 15:35	SY/MD	Ok,M
10	Q1802-02DUP	VL042334.D	15 Apr 2025 16:08	SY/MD	Ok,M
11	Q1802-03	VL042335.D	15 Apr 2025 16:42	SY/MD	Ok,M
12	Q1802-01	VL042336.D	15 Apr 2025 17:14	SY/MD	Dilution
13	Q1802-01DL	VL042337.D	15 Apr 2025 17:46	SY/MD	Not Ok
14	Q1802-01DL	VL042338.D	15 Apr 2025 18:21	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL050125

Review By	Semsettin Yesilyurt	Review On	5/2/2025 7:59:23 AM		
Supervise By	Mahesh Dadoda	Supervise On	5/5/2025 2:55:39 PM		
SubDirectory	VL050125	HP Acquire Method	TO15AIR.M	HP Processing Method	VL050125AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2603			
Initial Calibration Stds		AP2606,AP2607,AP2608			
CCC		AP2606			
Internal Standard/PEM		AP2603			
ICV/I.BLK		AP2609			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042457.D	01 May 2025 08:19	SY/MD	Ok
2	VSTDICCC010	VL042458.D	01 May 2025 09:03	SY/MD	Ok,M
3	VSTDICCC002	VL042459.D	01 May 2025 09:37	SY/MD	Ok,M
4	VSTDICCC001	VL042460.D	01 May 2025 10:08	SY/MD	Ok,M
5	VSTDICCC0.5	VL042461.D	01 May 2025 10:40	SY/MD	Ok,M
6	VSTDICCC0.1	VL042462.D	01 May 2025 11:12	SY/MD	Ok,M
7	VSTDICCC0.03	VL042463.D	01 May 2025 11:44	SY/MD	Not Ok
8	VSTDICCC0.03	VL042464.D	01 May 2025 12:15	SY/MD	Not Ok
9	VSTDICCC015	VL042465.D	01 May 2025 12:49	SY/MD	Ok,M
10	VSTDICCC0.03	VL042466.D	01 May 2025 14:10	SY/MD	Ok,M
11	VSTDICV010	VL042467.D	01 May 2025 15:46	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL050525

Review By	Semsettin Yesilyurt	Review On	5/6/2025 10:22:30 AM		
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:42:58 PM		
SubDirectory	VL050525	HP Acquire Method	TO15AIR.M	HP Processing Method	VL050125AIR.M
STD. NAME		STD REF.#			
Tune/Reschk	AP2603				
Initial Calibration Stds	AP2606,AP2607,AP2608				
CCC	AP2606				
Internal Standard/PEM	AP2603				
ICV/I.BLK	AP2609				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042468.D	05 May 2025 07:45	SY/MD	Ok
2	VSTDCCC010	VL042469.D	05 May 2025 08:29	SY/MD	Ok,M
3	VL0505ABL01	VL042470.D	05 May 2025 09:19	SY/MD	Ok
4	VL0505ABS01	VL042471.D	05 May 2025 10:07	SY/MD	Ok,M
5	Q1927-03DL	VL042472.D	05 May 2025 10:50	SY/MD	Not Ok
6	Q1927-04DL	VL042473.D	05 May 2025 11:22	SY/MD	Ok,M
7	Q1927-03	VL042474.D	05 May 2025 13:11	SY/MD	Ok,M
8	Q1927-04	VL042475.D	05 May 2025 14:44	SY/MD	Dilution
9	Q1927-01	VL042476.D	05 May 2025 15:18	SY/MD	Ok,M
10	Q1927-01DUP	VL042477.D	05 May 2025 15:53	SY/MD	Ok,M
11	Q1927-01DL	VL042478.D	05 May 2025 16:25	SY/MD	Not Ok
12	Q1927-02	VL042479.D	05 May 2025 16:59	SY/MD	Ok,M
13	Q1927-02DL	VL042480.D	05 May 2025 17:31	SY/MD	Not Ok
14	Q1927-05	VL042481.D	05 May 2025 18:06	SY/MD	Ok,M
15	Q1927-05DL	VL042482.D	05 May 2025 18:39	SY/MD	Not Ok
16	Q1927-06	VL042483.D	05 May 2025 19:15	SY/MD	Ok,M
17	Q1927-06DL	VL042484.D	05 May 2025 19:49	SY/MD	Not Ok
18	QC042825 CAN 10492	VL042485.D	05 May 2025 20:26	SY/MD	Ok

M : Manual Integration