

ANALYTICAL RESULTS SUMMARY

VOLATILE ORGANICS

PROJECT NAME : YELLOW CAB - TO-15

RTP ENVIRONMENTAL

239 US Highway 22 East

Green Brook, NJ - 08812

Phone No: 732-968-9600

ORDER ID : Q2317

ATTENTION : David Gabel



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	5
2.1) TO-15- Case Narrative	5
3) Qualifier Page	16
4) QA Checklist	17
5) TO-15 Data	18
6) Shipping Document	90
6.1) CHAIN OF CUSTODY	91
6.2) Lab Certificate	96
6.3) Internal COC	97
6.4) Air sample Pressure And Dilution Log Book	98
6.5) Storage Location	99

DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Laboratory Name : Alliance Technical Group LLC

Client : RTP Environmental

Project Location : Bayonne, NJ

Project Number :

Laboratory Sample ID(s) : Q2317

Sampling Date(s) : 6/12/2025

List DKQP Methods Used (e.g., 8260,8270, et Cetra) **TO-15**

1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?	☒ Yes ☐ No
1A	Were the method specified handling, preservation, and holding time requirements met?	☒ Yes ☐ No
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)	☐ Yes ☐ No ☒ N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?	☒ Yes ☐ No
3	Were samples received at an appropriate temperature (4±2° C)?	☐ Yes ☐ No ☒ N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?	☐ Yes ☒ No
5	a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt? b)Were these reporting limits met?	☒ Yes ☐ No ☒ Yes ☐ No ☐ N/A
6	For each analytical method referenced in this laboratory report package, were results reported for all constituents identified in the method-specific analyte lists presented in the DKQP documents and/or site-specific QAPP?	☒ Yes ☐ No
7	Are project-specific matrix spikes and/or laboratory duplicates included in this data set?	☒ Yes ☐ No

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."

Cover Page

Order ID : Q2317

Project ID : Yellow Cab - TO-15

Client : RTP Environmental

Lab Sample Number

Q2317-01
Q2317-02
Q2317-03
Q2317-04
Q2317-05

Client Sample Number

AA-3
115-IA-1
115-IA-2
115-SG1
115-SG2

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 6/23/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

RTP Environmental

Project Name: Yellow Cab - TO-15

Project # N/A

Order ID # Q2317

Test Name: TO-15

A. Number of Samples and Date of Receipt:

5 Air samples were received on 06/12/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: TO-15. This data package contains results for TO-15.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax, Entech 7100 A Preconcentrator. The analysis of TO-15 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD for {Q2317-01DUP} with File ID: VL042627.D met criteria except for Ethyl Benzene[200%] due to difference in results of original and DUP.

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Due to potential high concentration of target analytes, Samples 115-SG1, 115-SG2 were initially diluted.

Samples 115-IA-1, 115-IA-2 were diluted due to high concentrations.

E. Additional Comments:

The Sample #AA-3, 115-IA-1DL, 115-IA-2, 115-IA-2DL, 115-SG1 and 115-SG2 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

The Manual Integrations are performed for the followings.

Manual Integration Report			
Sequence	VL052925	Instrument	MSVOA_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICCC010	VL042579.D	Ethanol	SAM	5/30/2025 7:46:58 AM	MMDadoda	5/30/2025 12:25:20 PM	Peak Integrated by Software incorrectly
VSTDICCC010	VL042579.D	m/p-Xylene	SAM	5/30/2025 7:46:58 AM	MMDadoda	5/30/2025 12:25:20 PM	Peak Integrated by Software incorrectly
VSTDICCC002	VL042580.D	1,4-Dioxane	SAM	5/30/2025 7:47:04 AM	MMDadoda	5/30/2025 12:25:22 PM	Peak Integrated by Software incorrectly
VSTDICCC002	VL042580.D	Ethanol	SAM	5/30/2025 7:47:04 AM	MMDadoda	5/30/2025 12:25:22 PM	Peak Integrated by Software incorrectly
VSTDICCC002	VL042580.D	m/p-Xylene	SAM	5/30/2025 7:47:04 AM	MMDadoda	5/30/2025 12:25:22 PM	Peak Integrated by Software incorrectly
VSTDICCC002	VL042580.D	Methyl Methacrylate	SAM	5/30/2025 7:47:04 AM	MMDadoda	5/30/2025 12:25:22 PM	Peak Integrated by Software

							incorrectly
VSTDIC002	VL042580.D	t-1,3-Dichloropropene	SAM	5/30/2025 7:47:04 AM	MMDadoda	5/30/2025 12:25:22 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042581.D	1,1,2-Trichloroethane	SAM	5/30/2025 7:47:55 AM	MMDadoda	5/30/2025 12:25:24 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042581.D	1,4-Dioxane	SAM	5/30/2025 7:47:55 AM	MMDadoda	5/30/2025 12:25:24 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042581.D	2,2,4-Trimethylpentane	SAM	5/30/2025 7:47:55 AM	MMDadoda	5/30/2025 12:25:24 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042581.D	cis-1,3-Dichloropropene	SAM	5/30/2025 7:47:55 AM	MMDadoda	5/30/2025 12:25:24 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042581.D	Heptane	SAM	5/30/2025 7:47:55 AM	MMDadoda	5/30/2025 12:25:24 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042581.D	m/p-Xylene	SAM	5/30/2025 7:47:55 AM	MMDadoda	5/30/2025 12:25:24 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042581.D	t-1,3-Dichloropropene	SAM	5/30/2025 7:47:55 AM	MMDadoda	5/30/2025 12:25:24 PM	Peak Integrated by Software incorrectly
VSTDIC00.5	VL042582.D	1,1,2-Trichloroethane	SAM	5/30/2025 7:48:07 AM	MMDadoda	5/30/2025 12:25:26 PM	Peak Integrated by Software incorrectly

VSTDIC0.5	VL042582.D	1,2-Dichloropropane	SAM	5/30/2025 7:48:07 AM	MMDadoda	5/30/2025 12:25:26 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042582.D	1,4-Dioxane	SAM	5/30/2025 7:48:07 AM	MMDadoda	5/30/2025 12:25:26 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042582.D	4-Methyl-2-Pentanone	SAM	5/30/2025 7:48:07 AM	MMDadoda	5/30/2025 12:25:26 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042582.D	Benzyl Chloride	SAM	5/30/2025 7:48:07 AM	MMDadoda	5/30/2025 12:25:26 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042582.D	cis-1,3-Dichloropropene	SAM	5/30/2025 7:48:07 AM	MMDadoda	5/30/2025 12:25:26 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042582.D	Cyclohexane	SAM	5/30/2025 7:48:07 AM	MMDadoda	5/30/2025 12:25:26 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042582.D	Dibromochloromethane	SAM	5/30/2025 7:48:07 AM	MMDadoda	5/30/2025 12:25:26 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042582.D	Ethanol	SAM	5/30/2025 7:48:07 AM	MMDadoda	5/30/2025 12:25:26 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042582.D	m/p-Xylene	SAM	5/30/2025 7:48:07 AM	MMDadoda	5/30/2025 12:25:26 PM	Peak Integrated by Software incorrectly
VSTDIC0.1	VL042583.D	1,2-Dibromoethane	SAM	5/30/2025	MMDadoda	5/30/2025 12:25:28	Peak Integrated

				7:48:14 AM		PM	d by Software incorrectl y
VSTDIC0.1	VL042583.D	Naphthalene	SAM	5/30/2025 7:48:14 AM	MMDadoda	5/30/2025 12:25:28 PM	Peak Integrated by Software incorrectl y
VSTDIC0.1	VL042583.D	Tetrachloroethene	SAM	5/30/2025 7:48:14 AM	MMDadoda	5/30/2025 12:25:28 PM	Peak Integrated by Software incorrectl y
VSTDIC0.1	VL042583.D	Trichloroethene	SAM	5/30/2025 7:48:14 AM	MMDadoda	5/30/2025 12:25:28 PM	Peak Integrated by Software incorrectl y
VSTDIC0.03	VL042584.D	1,1,1-Trichloroethane	SAM	5/30/2025 7:47:15 AM	MMDadoda	5/30/2025 12:25:30 PM	Peak Integrated by Software incorrectl y
VSTDIC0.03	VL042584.D	Carbon Tetrachloride	SAM	5/30/2025 7:47:15 AM	MMDadoda	5/30/2025 12:25:30 PM	Peak Integrated by Software incorrectl y
VSTDIC0.03	VL042584.D	Tetrachloroethene	SAM	5/30/2025 7:47:15 AM	MMDadoda	5/30/2025 12:25:30 PM	Peak Integrated by Software incorrectl y
VSTDIC0.03	VL042584.D	Trichloroethene	SAM	5/30/2025 7:47:15 AM	MMDadoda	5/30/2025 12:25:30 PM	Peak Integrated by Software incorrectl y
VSTDIC015	VL042585.D	Ethanol	SAM	5/30/2025 7:48:19 AM	MMDadoda	5/30/2025 12:25:32 PM	Peak Integrated by Software incorrectl y
VSTDIC015	VL042585.D	m/p-Xylene	SAM	5/30/2025 7:48:19 AM	MMDadoda	5/30/2025 12:25:32 PM	Peak Integrated by Software

							incorrectly
VSTDICV010	VL042586.D	Ethanol	SAM	5/30/2025 7:48:25 AM	MMDadoda	5/30/2025 12:25:34 PM	Peak Integrated by Software incorrectly
VSTDICV010	VL042586.D	m/p-Xylene	SAM	5/30/2025 7:48:25 AM	MMDadoda	5/30/2025 12:25:34 PM	Peak Integrated by Software incorrectly
VSTDICV010	VL042586.D	t-1,3-Dichloropropene	SAM	5/30/2025 7:48:25 AM	MMDadoda	5/30/2025 12:25:34 PM	Peak Integrated by Software incorrectly

Sequence	VL061625	Instrument	MSVOA_I
----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC010	VL042622.D	m/p-Xylene	SAM	6/17/2025 8:36:27 AM	MMDadoda	6/17/2025 4:10:09 PM	Peak Integrated by Software incorrectly
VL0616ABS01	VL042624.D	m/p-Xylene	SAM	6/17/2025 8:36:32 AM	MMDadoda	6/17/2025 4:10:10 PM	Peak Integrated by Software incorrectly
Q2317-01	VL042626.D	4-Methyl-2-Pentanone	SAM	6/17/2025 8:41:32 AM	MMDadoda	6/17/2025 4:10:12 PM	Peak Integrated by Software incorrectly
Q2317-01	VL042626.D	Carbon Tetrachloride	SAM	6/17/2025 8:41:32 AM	MMDadoda	6/17/2025 4:10:12 PM	Peak Integrated by Software incorrectly
Q2317-01	VL042626.D	Chlorodifluoromethane	SAM	6/17/2025 8:41:32 AM	MMDadoda	6/17/2025 4:10:12 PM	Peak Integrated by Software

							incorrectl y
Q2317-01	VL042626. D	Ethyl Benzene	SAM	6/17/202 5 8:41:32 AM	MMDadoda	6/17/2025 4:10:12 PM	Peak Integrate d by Software incorrectl y
Q2317-01	VL042626. D	Heptane	SAM	6/17/202 5 8:41:32 AM	MMDadoda	6/17/2025 4:10:12 PM	Peak Integrate d by Software incorrectl y
Q2317-01	VL042626. D	m/p-Xylene	SAM	6/17/202 5 8:41:32 AM	MMDadoda	6/17/2025 4:10:12 PM	Peak Integrate d by Software incorrectl y
Q2317-01	VL042626. D	o-Xylene	SAM	6/17/202 5 8:41:32 AM	MMDadoda	6/17/2025 4:10:12 PM	Peak Integrate d by Software incorrectl y
Q2317-01	VL042626. D	Styrene	SAM	6/17/202 5 8:41:32 AM	MMDadoda	6/17/2025 4:10:12 PM	Peak Integrate d by Software incorrectl y
Q2317-01DUP	VL042627. D	2,2,4- Trimethylpentane	SAM	6/17/202 5 8:36:43 AM	MMDadoda	6/17/2025 4:10:14 PM	Peak Integrate d by Software incorrectl y
Q2317-01DUP	VL042627. D	4-Methyl-2- Pentanone	SAM	6/17/202 5 8:36:43 AM	MMDadoda	6/17/2025 4:10:14 PM	Peak Integrate d by Software incorrectl y
Q2317-01DUP	VL042627. D	Carbon Tetrachloride	SAM	6/17/202 5 8:36:43 AM	MMDadoda	6/17/2025 4:10:14 PM	Peak Integrate d by Software incorrectl y
Q2317-01DUP	VL042627. D	Chlorodifluorometha ne	SAM	6/17/202 5 8:36:43 AM	MMDadoda	6/17/2025 4:10:14 PM	Peak Integrate d by Software incorrectl y

Q2317-01DUP	VL042627.D	Heptane	SAM	6/17/2025 8:36:43 AM	MMDadoda	6/17/2025 4:10:14 PM	Peak Integrated by Software incorrectly
Q2317-01DUP	VL042627.D	Toluene	SAM	6/17/2025 8:36:43 AM	MMDadoda	6/17/2025 4:10:14 PM	Peak Integrated by Software incorrectly
Q2317-02	VL042628.D	1,2-Dichloroethane	SAM	6/17/2025 8:41:59 AM	MMDadoda	6/17/2025 4:10:15 PM	Peak Integrated by Software incorrectly
Q2317-02	VL042628.D	4-Methyl-2-Pentanone	SAM	6/17/2025 8:41:59 AM	MMDadoda	6/17/2025 4:10:15 PM	Peak Integrated by Software incorrectly
Q2317-02	VL042628.D	Carbon Tetrachloride	SAM	6/17/2025 8:41:59 AM	MMDadoda	6/17/2025 4:10:15 PM	Peak Integrated by Software incorrectly
Q2317-02	VL042628.D	Ethyl Benzene	SAM	6/17/2025 8:41:59 AM	MMDadoda	6/17/2025 4:10:15 PM	Peak Integrated by Software incorrectly
Q2317-02	VL042628.D	Heptane	SAM	6/17/2025 8:41:59 AM	MMDadoda	6/17/2025 4:10:15 PM	Peak Integrated by Software incorrectly
Q2317-02	VL042628.D	o-Xylene	SAM	6/17/2025 8:41:59 AM	MMDadoda	6/17/2025 4:10:15 PM	Peak Integrated by Software incorrectly
Q2317-02	VL042628.D	Propene	SAM	6/17/2025 8:41:59 AM	MMDadoda	6/17/2025 4:10:15 PM	Peak Integrated by Software incorrectly
Q2317-02	VL042628.D	Styrene	SAM	6/17/2025	MMDadoda	6/17/2025 4:10:15	Peak Integrated

				8:41:59 AM		PM	d by Software incorrectl y
Q2317-02	VL042628.D	Tetrachloroethene	SAM	6/17/2025 8:41:59 AM	MMdadoda	6/17/2025 4:10:15 PM	Peak Integrate d by Software incorrectl y
Q2317-02	VL042628.D	Trichloroethene	SAM	6/17/2025 8:41:59 AM	MMdadoda	6/17/2025 4:10:15 PM	Peak Integrate d by Software incorrectl y
Q2317-03	VL042629.D	Carbon Tetrachloride	sam	6/20/2025 2:44:20 AM	MMdadoda	6/20/2025 2:45:06 AM	Peak Integrate d by Software incorrectl y
Q2317-03	VL042629.D	Heptane	sam	6/20/2025 2:44:20 AM	MMdadoda	6/20/2025 2:45:06 AM	Peak Integrate d by Software incorrectl y
Q2317-03	VL042629.D	m/p-Xylene	sam	6/20/2025 2:44:20 AM	MMdadoda	6/20/2025 2:45:06 AM	Peak Integrate d by Software incorrectl y
Q2317-03	VL042629.D	Propene	sam	6/20/2025 2:44:20 AM	MMdadoda	6/20/2025 2:45:06 AM	Peak Integrate d by Software incorrectl y
Q2317-03	VL042629.D	tert-Butyl alcohol	sam	6/20/2025 2:44:20 AM	MMdadoda	6/20/2025 2:45:06 AM	Peak Integrate d by Software incorrectl y
Q2317-03	VL042629.D	Tetrachloroethene	sam	6/20/2025 2:44:20 AM	MMdadoda	6/20/2025 2:45:06 AM	Peak Integrate d by Software incorrectl y
Q2317-04	VL042630.D	4-Methyl-2-Pentanone	SAM	6/17/2025 8:41:54 AM	MMdadoda	6/17/2025 4:10:20 PM	Peak Integrate d by Software

							incorrectl y
Q2317-04	VL042630. D	Heptane	SAM	6/17/202 5 8:41:54 AM	MMDadoda	6/17/2025 4:10:20 PM	Peak Integrate d by Software incorrectl y
Q2317-04	VL042630. D	Tetrachloroethene	SAM	6/17/202 5 8:41:54 AM	MMDadoda	6/17/2025 4:10:20 PM	Peak Integrate d by Software incorrectl y
Q2317-05	VL042632. D	m/p-Xylene	SAM	6/17/202 5 8:42:24 AM	MMDadoda	6/17/2025 4:10:23 PM	Peak Integrate d by Software incorrectl y
Q2317-05	VL042632. D	Toluene	SAM	6/17/202 5 8:42:24 AM	MMDadoda	6/17/2025 4:10:23 PM	Peak Integrate d by Software incorrectl y
Q2317-02DL	VL042636. D	Propene	SAM	6/17/202 5 8:41:44 AM	MMDadoda	6/17/2025 4:10:26 PM	Peak Integrate d by Software incorrectl y
Q2317-02DL	VL042636. D	Toluene	SAM	6/17/202 5 8:41:44 AM	MMDadoda	6/17/2025 4:10:26 PM	Peak Integrate d by Software incorrectl y
Q2317-03DL	VL042637. D	Tetrahydrofuran	SAM	6/17/202 5 8:41:38 AM	MMDadoda	6/17/2025 4:10:28 PM	Peak Integrate d by Software incorrectl y
Q2317-03DL	VL042637. D	Toluene	SAM	6/17/202 5 8:41:38 AM	MMDadoda	6/17/2025 4:10:28 PM	Peak Integrate d by Software incorrectl y



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2317

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 06/23/2025

LAB CHRONICLE

OrderID:	Q2317	OrderDate:	6/13/2025 10:33:00 AM
Client:	RTP Environmental	Project:	Yellow Cab - TO-15
Contact:	David Gabel	Location:	D41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2317-01	AA-3	Air	TO-15	TO-15	06/12/25		06/16/25	06/12/25
Q2317-02	115-IA-1	Air	TO-15	TO-15	06/12/25		06/16/25	06/12/25
Q2317-02DL	115-IA-1DL	Air	TO-15	TO-15	06/12/25		06/16/25	06/12/25
Q2317-03	115-IA-2	Air	TO-15	TO-15	06/12/25		06/16/25	06/12/25
Q2317-03DL	115-IA-2DL	Air	TO-15	TO-15	06/12/25		06/16/25	06/12/25
Q2317-04	115-SG1	Air	TO-15	TO-15	06/12/25		06/16/25	06/12/25
Q2317-05	115-SG2	Air	TO-15	TO-15	06/12/25		06/16/25	06/12/25

Hit Summary Sheet SW-846

SDG No.: Q2317
Client: RTP Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: AA-3								
Q2317-01	AA-3	Air	Dichlorodifluoromethane	2.37	J	1.04	2.47	ug/m3
Q2317-01	AA-3	Air	Chloromethane	1.12		0.21	1.03	ug/m3
Q2317-01	AA-3	Air	Trichlorofluoromethane	1.52	J	0.90	2.81	ug/m3
Q2317-01	AA-3	Air	Heptane	0.61	J	0.45	2.05	ug/m3
Q2317-01	AA-3	Air	Acetone	24.2		0.57	1.19	ug/m3
Q2317-01	AA-3	Air	Methylene Chloride	10.4		0.66	1.74	ug/m3
Q2317-01	AA-3	Air	2-Butanone	1.92		0.35	1.47	ug/m3
Q2317-01	AA-3	Air	Carbon Tetrachloride	0.50		0.060	0.19	ug/m3
Q2317-01	AA-3	Air	2,2,4-Trimethylpentane	0.89	J	0.47	2.34	ug/m3
Q2317-01	AA-3	Air	Benzene	1.02	J	0.29	1.60	ug/m3
Q2317-01	AA-3	Air	4-Methyl-2-Pentanone	0.74	J	0.37	2.05	ug/m3
Q2317-01	AA-3	Air	Toluene	1.73	J	0.41	1.88	ug/m3
Q2317-01	AA-3	Air	m/p-Xylene	2.08	J	0.91	4.34	ug/m3
Q2317-01	AA-3	Air	o-Xylene	1.00	J	0.52	2.17	ug/m3
Q2317-01	AA-3	Air	Styrene	0.55	J	0.47	2.13	ug/m3
Q2317-01	AA-3	Air	1,2,4-Trimethylbenzene	0.84	J	0.39	2.46	ug/m3
Q2317-01	AA-3	Air	Hexane	16.9		0.39	1.76	ug/m3
Total Voc :				68.5				
Total Concentration:				68.5				
Client ID: 115-IA-1								
Q2317-02	115-IA-1	Air	Dichlorodifluoromethane	2.32	J	1.04	2.47	ug/m3
Q2317-02	115-IA-1	Air	Chloromethane	1.47		0.21	1.03	ug/m3
Q2317-02	115-IA-1	Air	Tetrahydrofuran	4.13		0.38	1.47	ug/m3
Q2317-02	115-IA-1	Air	Trichlorofluoromethane	45.5		0.90	2.81	ug/m3
Q2317-02	115-IA-1	Air	Heptane	0.82	J	0.45	2.05	ug/m3
Q2317-02	115-IA-1	Air	Acetone	91.0	E	0.57	1.19	ug/m3
Q2317-02	115-IA-1	Air	Methylene Chloride	4.52		0.66	1.74	ug/m3
Q2317-02	115-IA-1	Air	2-Butanone	6.19		0.35	1.47	ug/m3
Q2317-02	115-IA-1	Air	Carbon Tetrachloride	0.57		0.060	0.19	ug/m3
Q2317-02	115-IA-1	Air	Chloroform	2.83		0.39	2.44	ug/m3
Q2317-02	115-IA-1	Air	2,2,4-Trimethylpentane	0.75	J	0.47	2.34	ug/m3
Q2317-02	115-IA-1	Air	Benzene	1.05	J	0.29	1.60	ug/m3
Q2317-02	115-IA-1	Air	1,2-Dichloroethane	0.69	J	0.36	2.02	ug/m3
Q2317-02	115-IA-1	Air	Trichloroethene	0.27		0.11	0.16	ug/m3
Q2317-02	115-IA-1	Air	4-Methyl-2-Pentanone	1.07	J	0.37	2.05	ug/m3
Q2317-02	115-IA-1	Air	Toluene	2.83		0.41	1.88	ug/m3

Hit Summary Sheet SW-846

SDG No.: Q2317

Client: RTP Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q2317-02	115-IA-1	Air	Tetrachloroethene	0.47		0.14	0.20	ug/m3
Q2317-02	115-IA-1	Air	Ethyl Benzene	0.52	J	0.52	2.17	ug/m3
Q2317-02	115-IA-1	Air	m/p-Xylene	1.69	J	0.91	4.34	ug/m3
Q2317-02	115-IA-1	Air	o-Xylene	0.74	J	0.52	2.17	ug/m3
Q2317-02	115-IA-1	Air	Styrene	0.60	J	0.47	2.13	ug/m3
Q2317-02	115-IA-1	Air	1,2,4-Trimethylbenzene	1.38	J	0.39	2.46	ug/m3
Q2317-02	115-IA-1	Air	1,4-Dichlorobenzene	105	E	0.30	3.01	ug/m3
Q2317-02	115-IA-1	Air	Hexane	14.4		0.39	1.76	ug/m3
			Total Voc :	290				
			Total Concentration:	290				
Client ID:	115-IA-1DL							
Q2317-02DL	115-IA-1DL	Air	Chloromethane	2.02	JD	1.01	5.16	ug/m3
Q2317-02DL	115-IA-1DL	Air	Tetrahydrofuran	3.54	JD	1.92	7.37	ug/m3
Q2317-02DL	115-IA-1DL	Air	Trichlorofluoromethane	48.3	D	4.50	14.1	ug/m3
Q2317-02DL	115-IA-1DL	Air	Acetone	97.9	D	2.85	5.94	ug/m3
Q2317-02DL	115-IA-1DL	Air	Methylene Chloride	7.30	JD	3.30	8.69	ug/m3
Q2317-02DL	115-IA-1DL	Air	2-Butanone	6.19	JD	1.77	7.37	ug/m3
Q2317-02DL	115-IA-1DL	Air	Toluene	2.49	JD	2.07	9.42	ug/m3
Q2317-02DL	115-IA-1DL	Air	1,4-Dichlorobenzene	96.8	D	1.62	15.0	ug/m3
Q2317-02DL	115-IA-1DL	Air	Hexane	11.3	D	1.94	8.81	ug/m3
			Total Voc :	276				
			Total Concentration:	276				
Client ID:	115-IA-2							
Q2317-03	115-IA-2	Air	Dichlorodifluoromethane	2.47		1.04	2.47	ug/m3
Q2317-03	115-IA-2	Air	Chloromethane	1.45		0.21	1.03	ug/m3
Q2317-03	115-IA-2	Air	Tetrahydrofuran	4.42		0.38	1.47	ug/m3
Q2317-03	115-IA-2	Air	Trichlorofluoromethane	30.4		0.90	2.81	ug/m3
Q2317-03	115-IA-2	Air	tert-Butyl alcohol	1.12	J	0.58	1.52	ug/m3
Q2317-03	115-IA-2	Air	Heptane	0.86	J	0.45	2.05	ug/m3
Q2317-03	115-IA-2	Air	Acetone	76.5	E	0.57	1.19	ug/m3
Q2317-03	115-IA-2	Air	Methylene Chloride	2.05		0.66	1.74	ug/m3
Q2317-03	115-IA-2	Air	2-Butanone	6.78		0.35	1.47	ug/m3
Q2317-03	115-IA-2	Air	Carbon Tetrachloride	0.50		0.060	0.19	ug/m3
Q2317-03	115-IA-2	Air	Chloroform	1.47	J	0.39	2.44	ug/m3
Q2317-03	115-IA-2	Air	Benzene	1.09	J	0.29	1.60	ug/m3
Q2317-03	115-IA-2	Air	Toluene	2.26		0.41	1.88	ug/m3
Q2317-03	115-IA-2	Air	Tetrachloroethene	0.27		0.14	0.20	ug/m3

Hit Summary Sheet SW-846

SDG No.: Q2317

Client: RTP Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q2317-03	115-IA-2	Air	m/p-Xylene	1.22	J	0.91	4.34	ug/m3
Q2317-03	115-IA-2	Air	1,2,4-Trimethylbenzene	0.64	J	0.39	2.46	ug/m3
Q2317-03	115-IA-2	Air	1,4-Dichlorobenzene	133	E	0.30	3.01	ug/m3
Q2317-03	115-IA-2	Air	Hexane	3.38		0.39	1.76	ug/m3
			Total Voc :	270				
			Total Concentration:	270				
Client ID:	115-IA-2DL							
Q2317-03DL	115-IA-2DL	Air	Chloromethane	1.67	JD	1.01	5.16	ug/m3
Q2317-03DL	115-IA-2DL	Air	Tetrahydrofuran	4.13	JD	1.92	7.37	ug/m3
Q2317-03DL	115-IA-2DL	Air	Trichlorofluoromethane	36.0	D	4.50	14.1	ug/m3
Q2317-03DL	115-IA-2DL	Air	Acetone	85.0	D	2.85	5.94	ug/m3
Q2317-03DL	115-IA-2DL	Air	2-Butanone	7.08	JD	1.77	7.37	ug/m3
Q2317-03DL	115-IA-2DL	Air	1,4-Dichlorobenzene	129	D	1.62	15.0	ug/m3
Q2317-03DL	115-IA-2DL	Air	Hexane	2.68	JD	1.94	8.81	ug/m3
			Total Voc :	266				
			Total Concentration:	266				
Client ID:	115-SG1							
Q2317-04	115-SG1	Air	Chloromethane	13.4		0.81	4.13	ug/m3
Q2317-04	115-SG1	Air	Chloroethane	6.07		1.79	5.28	ug/m3
Q2317-04	115-SG1	Air	Tetrahydrofuran	6.19		1.53	5.90	ug/m3
Q2317-04	115-SG1	Air	Trichlorofluoromethane	238		3.60	11.2	ug/m3
Q2317-04	115-SG1	Air	Heptane	3.20	J	1.80	8.20	ug/m3
Q2317-04	115-SG1	Air	Acetone	52.0		2.28	4.75	ug/m3
Q2317-04	115-SG1	Air	Carbon Disulfide	7.16		1.99	6.23	ug/m3
Q2317-04	115-SG1	Air	Methylene Chloride	28.1		2.64	6.95	ug/m3
Q2317-04	115-SG1	Air	2-Butanone	12.4		1.42	5.90	ug/m3
Q2317-04	115-SG1	Air	Chloroform	3.76	J	1.61	9.77	ug/m3
Q2317-04	115-SG1	Air	Benzene	28.4		1.12	6.39	ug/m3
Q2317-04	115-SG1	Air	Trichloroethene	26.3		0.38	0.64	ug/m3
Q2317-04	115-SG1	Air	4-Methyl-2-Pentanone	5.33	J	1.52	8.20	ug/m3
Q2317-04	115-SG1	Air	Toluene	43.3		1.66	7.54	ug/m3
Q2317-04	115-SG1	Air	Tetrachloroethene	2.44		0.41	0.81	ug/m3
Q2317-04	115-SG1	Air	Styrene	2.09	J	1.87	8.52	ug/m3
Q2317-04	115-SG1	Air	1,4-Dichlorobenzene	3.37	J	1.26	12.0	ug/m3
Q2317-04	115-SG1	Air	Hexane	9.87		1.55	7.05	ug/m3
Q2317-04	115-SG1	Air	Methyl Methacrylate	104		1.39	8.19	ug/m3
			Total Voc :	596				
			Total Concentration:	596				

Hit Summary Sheet SW-846

SDG No.: Q2317

Client: RTP Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	115-SG2							
Q2317-05	115-SG2	Air	Chloromethane	18.4		0.81	4.13	ug/m3
Q2317-05	115-SG2	Air	Chloroethane	8.71		1.79	5.28	ug/m3
Q2317-05	115-SG2	Air	Tetrahydrofuran	5.90		1.53	5.90	ug/m3
Q2317-05	115-SG2	Air	Trichlorofluoromethane	79.8		3.60	11.2	ug/m3
Q2317-05	115-SG2	Air	Acetone	125		2.28	4.75	ug/m3
Q2317-05	115-SG2	Air	Methylene Chloride	9.73		2.64	6.95	ug/m3
Q2317-05	115-SG2	Air	2-Butanone	11.5		1.42	5.90	ug/m3
Q2317-05	115-SG2	Air	Chloroform	40.5		1.61	9.77	ug/m3
Q2317-05	115-SG2	Air	Benzene	10.5		1.12	6.39	ug/m3
Q2317-05	115-SG2	Air	Trichloroethene	69.3		0.38	0.64	ug/m3
Q2317-05	115-SG2	Air	Toluene	1.92	J	1.66	7.54	ug/m3
Q2317-05	115-SG2	Air	Tetrachloroethene	83.4		0.41	0.81	ug/m3
Q2317-05	115-SG2	Air	1,4-Dichlorobenzene	30.1		1.26	12.0	ug/m3
Q2317-05	115-SG2	Air	Hexane	3.10	J	1.55	7.05	ug/m3
Q2317-05	115-SG2	Air	Methyl Methacrylate	72.9		1.39	8.19	ug/m3
Total Voc :					571			
Total Concentration:					571			

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : AA-3

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-01

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.48	J	2.37		
Chloromethane	74-87-3	50.49	0.54		1.12		
Vinyl Chloride	75-01-4	62.5	0.01	U	0.03		
Bromomethane	74-83-9	94.94	0.14	U	0.54		
Chloroethane	75-00-3	64.52	0.17	U	0.45		
Tetrahydrofuran	109-99-9	72.11	0.13	U	0.38		
Trichlorofluoromethane	75-69-4	137.4	0.27	J	1.52		
Dichlorotetrafluoroethane	76-14-2	170.9	0.09	U	0.63		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.16	U	1.23		
tert-Butyl alcohol	75-65-0	74.12	0.19	U	0.58		
Heptane	142-82-5	100.2	0.15	J	0.61		
1,1-Dichloroethene	75-35-4	96.94	0.14	U	0.56		
Acetone	67-64-1	58.08	10.2		24.2		
Carbon Disulfide	75-15-0	76.14	0.16	U	0.5		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.12	U	0.43		
Methylene Chloride	75-09-2	84.94	3		10.4		
trans-1,2-Dichloroethene	156-60-5	96.94	0.19	U	0.75		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.65		1.92		
Carbon Tetrachloride	56-23-5	153.8	0.08		0.5		
cis-1,2-Dichloroethene	156-59-2	96.94	0.09	U	0.36		
Chloroform	67-66-3	119.4	0.08	U	0.39		
1,1,1-Trichloroethane	71-55-6	133.4	0.01	U	0.05		
2,2,4-Trimethylpentane	540-84-1	114.2	0.19	J	0.89		
Benzene	71-43-2	78.11	0.32	J	1.02		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.11	U	0.51		
Bromodichloromethane	75-27-4	163.8	0.08	U	0.54		
4-Methyl-2-Pentanone	108-10-1	100.2	0.18	J	0.74		
Toluene	108-88-3	92.14	0.46	J	1.73		
t-1,3-Dichloropropene	10061-02-6	111	0.06	U	0.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.06	U	0.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.07	U	0.38		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : AA-3

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-01

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.07	U	0.6		
1,2-Dibromoethane	106-93-4	187.9	0.07	U	0.54		
Tetrachloroethene	127-18-4	165.8	0.02	U	0.14		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.12	U	0.52		
m/p-Xylene	179601-23-1	106.2	0.48	J	2.08		
o-Xylene	95-47-6	106.2	0.23	J	1		
Styrene	100-42-5	104.1	0.13	J	0.55		
Bromoform	75-25-2	252.8	0.06	U	0.62		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.01	U	0.07		
2-Chlorotoluene	95-49-8	126.6	0.11	U	0.57		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.11	U	0.54		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.17	J	0.84		
1,3-Dichlorobenzene	541-73-1	147	0.08	U	0.48		
1,4-Dichlorobenzene	106-46-7	147	0.05	U	0.3		
1,2-Dichlorobenzene	95-50-1	147	0.08	U	0.48		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.09	U	0.67		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.09	U	0.96		
Naphthalene	91-20-3	128.17	0.08	U	0.42		
1,3-Butadiene	106-99-0	54.09	0.13	U	0.29		
4-Ethyltoluene	622-96-8	120.2	0.12	U	0.59		
Hexane	110-54-3	86.17	4.8		16.9		
Allyl Chloride	107-05-1	76.53	0.15	U	0.47		
1,4-Dioxane	123-91-1	88.12	0.21	U	0.76		
Methyl Methacrylate	80-62-6	100.117	0.09	U	0.37		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-IA-1

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-02

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.47	J	2.32		
Chloromethane	74-87-3	50.49	0.71		1.47		
Vinyl Chloride	75-01-4	62.5	0.01	U	0.03		
Bromomethane	74-83-9	94.94	0.14	U	0.54		
Chloroethane	75-00-3	64.52	0.17	U	0.45		
Tetrahydrofuran	109-99-9	72.11	1.4		4.13		
Trichlorofluoromethane	75-69-4	137.4	8.1		45.5		
Dichlorotetrafluoroethane	76-14-2	170.9	0.09	U	0.63		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.16	U	1.23		
tert-Butyl alcohol	75-65-0	74.12	0.19	U	0.58		
Heptane	142-82-5	100.2	0.2	J	0.82		
1,1-Dichloroethene	75-35-4	96.94	0.14	U	0.56		
Acetone	67-64-1	58.08	38.3	E	91.0		
Carbon Disulfide	75-15-0	76.14	0.16	U	0.5		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.12	U	0.43		
Methylene Chloride	75-09-2	84.94	1.3		4.52		
trans-1,2-Dichloroethene	156-60-5	96.94	0.19	U	0.75		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	2.1		6.19		
Carbon Tetrachloride	56-23-5	153.8	0.09		0.57		
cis-1,2-Dichloroethene	156-59-2	96.94	0.09	U	0.36		
Chloroform	67-66-3	119.4	0.58		2.83		
1,1,1-Trichloroethane	71-55-6	133.4	0.01	U	0.05		
2,2,4-Trimethylpentane	540-84-1	114.2	0.16	J	0.75		
Benzene	71-43-2	78.11	0.33	J	1.05		
1,2-Dichloroethane	107-06-2	98.96	0.17	J	0.69		
Trichloroethene	79-01-6	131.4	0.05		0.27		
1,2-Dichloropropane	78-87-5	113	0.11	U	0.51		
Bromodichloromethane	75-27-4	163.8	0.08	U	0.54		
4-Methyl-2-Pentanone	108-10-1	100.2	0.26	J	1.07		
Toluene	108-88-3	92.14	0.75		2.83		
t-1,3-Dichloropropene	10061-02-6	111	0.06	U	0.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.06	U	0.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.07	U	0.38		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-IA-1

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-02

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.07	U	0.6		
1,2-Dibromoethane	106-93-4	187.9	0.07	U	0.54		
Tetrachloroethene	127-18-4	165.8	0.07		0.47		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.12	J	0.52		
m/p-Xylene	179601-23-1	106.2	0.39	J	1.69		
o-Xylene	95-47-6	106.2	0.17	J	0.74		
Styrene	100-42-5	104.1	0.14	J	0.6		
Bromoform	75-25-2	252.8	0.06	U	0.62		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.01	U	0.07		
2-Chlorotoluene	95-49-8	126.6	0.11	U	0.57		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.11	U	0.54		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.28	J	1.38		
1,3-Dichlorobenzene	541-73-1	147	0.08	U	0.48		
1,4-Dichlorobenzene	106-46-7	147	17.4	E	104		
1,2-Dichlorobenzene	95-50-1	147	0.08	U	0.48		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.09	U	0.67		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.09	U	0.96		
Naphthalene	91-20-3	128.17	0.08	U	0.42		
1,3-Butadiene	106-99-0	54.09	0.13	U	0.29		
4-Ethyltoluene	622-96-8	120.2	0.12	U	0.59		
Hexane	110-54-3	86.17	4.1		14.4		
Allyl Chloride	107-05-1	76.53	0.15	U	0.47		
1,4-Dioxane	123-91-1	88.12	0.21	U	0.76		
Methyl Methacrylate	80-62-6	100.117	0.09	U	0.37		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-IA-1DL

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-02DL

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	1.1	UD	5.44		
Chloromethane	74-87-3	50.49	0.98	JD	2.02		
Vinyl Chloride	75-01-4	62.5	0.07	UD	0.18		
Bromomethane	74-83-9	94.94	0.7	UD	2.72		
Chloroethane	75-00-3	64.52	0.85	UD	2.24		
Tetrahydrofuran	109-99-9	72.11	1.2	JD	3.54		
Trichlorofluoromethane	75-69-4	137.4	8.6	D	48.3		
Dichlorotetrafluoroethane	76-14-2	170.9	0.44	UD	3.08		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.8	UD	6.13		
tert-Butyl alcohol	75-65-0	74.12	0.95	UD	2.88		
Heptane	142-82-5	100.2	0.55	UD	2.25		
1,1-Dichloroethene	75-35-4	96.94	0.7	UD	2.78		
Acetone	67-64-1	58.08	41.2	D	97.9		
Carbon Disulfide	75-15-0	76.14	0.8	UD	2.49		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.6	UD	2.16		
Methylene Chloride	75-09-2	84.94	2.1	JD	7.3		
trans-1,2-Dichloroethene	156-60-5	96.94	0.95	UD	3.77		
1,1-Dichloroethane	75-34-3	98.96	0.65	UD	2.63		
Cyclohexane	110-82-7	84.16	1.1	UD	3.79		
2-Butanone	78-93-3	72.11	2.1	JD	6.19		
Carbon Tetrachloride	56-23-5	153.8	0.06	UD	0.38		
cis-1,2-Dichloroethene	156-59-2	96.94	0.45	UD	1.78		
Chloroform	67-66-3	119.4	0.42	UD	2.05		
1,1,1-Trichloroethane	71-55-6	133.4	0.05	UD	0.27		
2,2,4-Trimethylpentane	540-84-1	114.2	0.5	UD	2.34		
Benzene	71-43-2	78.11	0.44	UD	1.41		
1,2-Dichloroethane	107-06-2	98.96	0.46	UD	1.86		
Trichloroethene	79-01-6	131.4	0.09	UD	0.48		
1,2-Dichloropropane	78-87-5	113	0.55	UD	2.54		
Bromodichloromethane	75-27-4	163.8	0.39	UD	2.61		
4-Methyl-2-Pentanone	108-10-1	100.2	0.47	UD	1.93		
Toluene	108-88-3	92.14	0.66	JD	2.49		
t-1,3-Dichloropropene	10061-02-6	111	0.3	UD	1.36		
cis-1,3-Dichloropropene	10061-01-5	111	0.32	UD	1.45		
1,1,2-Trichloroethane	79-00-5	133.4	0.36	UD	1.96		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-IA-1DL

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-02DL

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.33	UD	2.81		
1,2-Dibromoethane	106-93-4	187.9	0.36	UD	2.77		
Tetrachloroethene	127-18-4	165.8	0.08	UD	0.54		
Chlorobenzene	108-90-7	112.6	0.38	UD	1.75		
Ethyl Benzene	100-41-4	106.2	0.6	UD	2.61		
m/p-Xylene	179601-23-1	106.2	1.1	UD	4.78		
o-Xylene	95-47-6	106.2	0.6	UD	2.61		
Styrene	100-42-5	104.1	0.55	UD	2.34		
Bromoform	75-25-2	252.8	0.28	UD	2.9		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.06	UD	0.41		
2-Chlorotoluene	95-49-8	126.6	0.55	UD	2.85		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.55	UD	2.7		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.39	UD	1.92		
1,3-Dichlorobenzene	541-73-1	147	0.4	UD	2.4		
1,4-Dichlorobenzene	106-46-7	147	16.1	D	96.8		
1,2-Dichlorobenzene	95-50-1	147	0.38	UD	2.28		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.43	UD	3.19		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.43	UD	4.59		
Naphthalene	91-20-3	128.17	0.38	UD	1.99		
1,3-Butadiene	106-99-0	54.09	0.65	UD	1.44		
4-Ethyltoluene	622-96-8	120.2	0.6	UD	2.95		
Hexane	110-54-3	86.17	3.2	D	11.3		
Allyl Chloride	107-05-1	76.53	0.75	UD	2.35		
1,4-Dioxane	123-91-1	88.12	1.1	UD	3.96		
Methyl Methacrylate	80-62-6	100.117	0.43	UD	1.76		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-IA-2

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-03

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.5		2.47		
Chloromethane	74-87-3	50.49	0.7		1.45		
Vinyl Chloride	75-01-4	62.5	0.01	U	0.03		
Bromomethane	74-83-9	94.94	0.14	U	0.54		
Chloroethane	75-00-3	64.52	0.17	U	0.45		
Tetrahydrofuran	109-99-9	72.11	1.5		4.42		
Trichlorofluoromethane	75-69-4	137.4	5.4		30.4		
Dichlorotetrafluoroethane	76-14-2	170.9	0.09	U	0.63		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.16	U	1.23		
tert-Butyl alcohol	75-65-0	74.12	0.37	J	1.12		
Heptane	142-82-5	100.2	0.21	J	0.86		
1,1-Dichloroethene	75-35-4	96.94	0.14	U	0.56		
Acetone	67-64-1	58.08	32.2	E	76.5		
Carbon Disulfide	75-15-0	76.14	0.16	U	0.5		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.12	U	0.43		
Methylene Chloride	75-09-2	84.94	0.59		2.05		
trans-1,2-Dichloroethene	156-60-5	96.94	0.19	U	0.75		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	2.3		6.78		
Carbon Tetrachloride	56-23-5	153.8	0.08		0.5		
cis-1,2-Dichloroethene	156-59-2	96.94	0.09	U	0.36		
Chloroform	67-66-3	119.4	0.3	J	1.47		
1,1,1-Trichloroethane	71-55-6	133.4	0.01	U	0.05		
2,2,4-Trimethylpentane	540-84-1	114.2	0.1	U	0.47		
Benzene	71-43-2	78.11	0.34	J	1.09		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.11	U	0.51		
Bromodichloromethane	75-27-4	163.8	0.08	U	0.54		
4-Methyl-2-Pentanone	108-10-1	100.2	0.09	U	0.37		
Toluene	108-88-3	92.14	0.6		2.26		
t-1,3-Dichloropropene	10061-02-6	111	0.06	U	0.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.06	U	0.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.07	U	0.38		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-IA-2

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-03

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.07	U	0.6		
1,2-Dibromoethane	106-93-4	187.9	0.07	U	0.54		
Tetrachloroethene	127-18-4	165.8	0.04		0.27		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.12	U	0.52		
m/p-Xylene	179601-23-1	106.2	0.28	J	1.22		
o-Xylene	95-47-6	106.2	0.12	U	0.52		
Styrene	100-42-5	104.1	0.11	U	0.47		
Bromoform	75-25-2	252.8	0.06	U	0.62		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.01	U	0.07		
2-Chlorotoluene	95-49-8	126.6	0.11	U	0.57		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.11	U	0.54		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.13	J	0.64		
1,3-Dichlorobenzene	541-73-1	147	0.08	U	0.48		
1,4-Dichlorobenzene	106-46-7	147	22.2	E	133		
1,2-Dichlorobenzene	95-50-1	147	0.08	U	0.48		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.09	U	0.67		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.09	U	0.96		
Naphthalene	91-20-3	128.17	0.08	U	0.42		
1,3-Butadiene	106-99-0	54.09	0.13	U	0.29		
4-Ethyltoluene	622-96-8	120.2	0.12	U	0.59		
Hexane	110-54-3	86.17	0.96		3.38		
Allyl Chloride	107-05-1	76.53	0.15	U	0.47		
1,4-Dioxane	123-91-1	88.12	0.21	U	0.76		
Methyl Methacrylate	80-62-6	100.117	0.09	U	0.37		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-IA-2DL

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-03DL

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	1.1	UD	5.44		
Chloromethane	74-87-3	50.49	0.81	JD	1.67		
Vinyl Chloride	75-01-4	62.5	0.07	UD	0.18		
Bromomethane	74-83-9	94.94	0.7	UD	2.72		
Chloroethane	75-00-3	64.52	0.85	UD	2.24		
Tetrahydrofuran	109-99-9	72.11	1.4	JD	4.13		
Trichlorofluoromethane	75-69-4	137.4	6.4	D	36.0		
Dichlorotetrafluoroethane	76-14-2	170.9	0.44	UD	3.08		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.8	UD	6.13		
tert-Butyl alcohol	75-65-0	74.12	0.95	UD	2.88		
Heptane	142-82-5	100.2	0.55	UD	2.25		
1,1-Dichloroethene	75-35-4	96.94	0.7	UD	2.78		
Acetone	67-64-1	58.08	35.8	D	85.0		
Carbon Disulfide	75-15-0	76.14	0.8	UD	2.49		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.6	UD	2.16		
Methylene Chloride	75-09-2	84.94	0.95	UD	3.3		
trans-1,2-Dichloroethene	156-60-5	96.94	0.95	UD	3.77		
1,1-Dichloroethane	75-34-3	98.96	0.65	UD	2.63		
Cyclohexane	110-82-7	84.16	1.1	UD	3.79		
2-Butanone	78-93-3	72.11	2.4	JD	7.08		
Carbon Tetrachloride	56-23-5	153.8	0.06	UD	0.38		
cis-1,2-Dichloroethene	156-59-2	96.94	0.45	UD	1.78		
Chloroform	67-66-3	119.4	0.42	UD	2.05		
1,1,1-Trichloroethane	71-55-6	133.4	0.05	UD	0.27		
2,2,4-Trimethylpentane	540-84-1	114.2	0.5	UD	2.34		
Benzene	71-43-2	78.11	0.44	UD	1.41		
1,2-Dichloroethane	107-06-2	98.96	0.46	UD	1.86		
Trichloroethene	79-01-6	131.4	0.09	UD	0.48		
1,2-Dichloropropane	78-87-5	113	0.55	UD	2.54		
Bromodichloromethane	75-27-4	163.8	0.39	UD	2.61		
4-Methyl-2-Pentanone	108-10-1	100.2	0.47	UD	1.93		
Toluene	108-88-3	92.14	0.55	UD	2.07		
t-1,3-Dichloropropene	10061-02-6	111	0.3	UD	1.36		
cis-1,3-Dichloropropene	10061-01-5	111	0.32	UD	1.45		
1,1,2-Trichloroethane	79-00-5	133.4	0.36	UD	1.96		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-IA-2DL

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-03DL

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.33	UD	2.81		
1,2-Dibromoethane	106-93-4	187.9	0.36	UD	2.77		
Tetrachloroethene	127-18-4	165.8	0.08	UD	0.54		
Chlorobenzene	108-90-7	112.6	0.38	UD	1.75		
Ethyl Benzene	100-41-4	106.2	0.6	UD	2.61		
m/p-Xylene	179601-23-1	106.2	1.1	UD	4.78		
o-Xylene	95-47-6	106.2	0.6	UD	2.61		
Styrene	100-42-5	104.1	0.55	UD	2.34		
Bromoform	75-25-2	252.8	0.28	UD	2.9		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.06	UD	0.41		
2-Chlorotoluene	95-49-8	126.6	0.55	UD	2.85		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.55	UD	2.7		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.39	UD	1.92		
1,3-Dichlorobenzene	541-73-1	147	0.4	UD	2.4		
1,4-Dichlorobenzene	106-46-7	147	21.5	D	129		
1,2-Dichlorobenzene	95-50-1	147	0.38	UD	2.28		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.43	UD	3.19		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.43	UD	4.59		
Naphthalene	91-20-3	128.17	0.38	UD	1.99		
1,3-Butadiene	106-99-0	54.09	0.65	UD	1.44		
4-Ethyltoluene	622-96-8	120.2	0.6	UD	2.95		
Hexane	110-54-3	86.17	0.76	JD	2.68		
Allyl Chloride	107-05-1	76.53	0.75	UD	2.35		
1,4-Dioxane	123-91-1	88.12	1.1	UD	3.96		
Methyl Methacrylate	80-62-6	100.117	0.43	UD	1.76		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-SG1

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-04

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.84	U	4.15		
Chloromethane	74-87-3	50.49	6.5		13.4		
Vinyl Chloride	75-01-4	62.5	0.06	U	0.15		
Bromomethane	74-83-9	94.94	0.56	U	2.17		
Chloroethane	75-00-3	64.52	2.3		6.07		
Tetrahydrofuran	109-99-9	72.11	2.1		6.19		
Trichlorofluoromethane	75-69-4	137.4	42.3		237		
Dichlorotetrafluoroethane	76-14-2	170.9	0.35	U	2.45		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.64	U	4.91		
tert-Butyl alcohol	75-65-0	74.12	0.76	U	2.3		
Heptane	142-82-5	100.2	0.78	J	3.2		
1,1-Dichloroethene	75-35-4	96.94	0.56	U	2.22		
Acetone	67-64-1	58.08	21.9		52.0		
Carbon Disulfide	75-15-0	76.14	2.3		7.16		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.48	U	1.73		
Methylene Chloride	75-09-2	84.94	8.1		28.1		
trans-1,2-Dichloroethene	156-60-5	96.94	0.76	U	3.01		
1,1-Dichloroethane	75-34-3	98.96	0.52	U	2.1		
Cyclohexane	110-82-7	84.16	0.88	U	3.03		
2-Butanone	78-93-3	72.11	4.2		12.4		
Carbon Tetrachloride	56-23-5	153.8	0.04	U	0.25		
cis-1,2-Dichloroethene	156-59-2	96.94	0.36	U	1.43		
Chloroform	67-66-3	119.4	0.77	J	3.76		
1,1,1-Trichloroethane	71-55-6	133.4	0.04	U	0.22		
2,2,4-Trimethylpentane	540-84-1	114.2	0.4	U	1.87		
Benzene	71-43-2	78.11	8.9		28.4		
1,2-Dichloroethane	107-06-2	98.96	0.36	U	1.46		
Trichloroethene	79-01-6	131.4	4.9		26.3		
1,2-Dichloropropane	78-87-5	113	0.44	U	2.03		
Bromodichloromethane	75-27-4	163.8	0.31	U	2.08		
4-Methyl-2-Pentanone	108-10-1	100.2	1.3	J	5.33		
Toluene	108-88-3	92.14	11.5		43.3		
t-1,3-Dichloropropene	10061-02-6	111	0.24	U	1.09		
cis-1,3-Dichloropropene	10061-01-5	111	0.26	U	1.18		
1,1,2-Trichloroethane	79-00-5	133.4	0.28	U	1.53		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-SG1

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-04

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.26	U	2.22		
1,2-Dibromoethane	106-93-4	187.9	0.29	U	2.23		
Tetrachloroethene	127-18-4	165.8	0.36		2.44		
Chlorobenzene	108-90-7	112.6	0.3	U	1.38		
Ethyl Benzene	100-41-4	106.2	0.48	U	2.08		
m/p-Xylene	179601-23-1	106.2	0.84	U	3.65		
o-Xylene	95-47-6	106.2	0.48	U	2.08		
Styrene	100-42-5	104.1	0.49	J	2.09		
Bromoform	75-25-2	252.8	0.23	U	2.38		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.04	U	0.27		
2-Chlorotoluene	95-49-8	126.6	0.44	U	2.28		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.44	U	2.16		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.31	U	1.52		
1,3-Dichlorobenzene	541-73-1	147	0.32	U	1.92		
1,4-Dichlorobenzene	106-46-7	147	0.56	J	3.37		
1,2-Dichlorobenzene	95-50-1	147	0.3	U	1.8		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.34	U	2.52		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.34	U	3.63		
Naphthalene	91-20-3	128.17	0.3	U	1.57		
1,3-Butadiene	106-99-0	54.09	0.52	U	1.15		
4-Ethyltoluene	622-96-8	120.2	0.48	U	2.36		
Hexane	110-54-3	86.17	2.8		9.87		
Allyl Chloride	107-05-1	76.53	0.6	U	1.88		
1,4-Dioxane	123-91-1	88.12	0.84	U	3.03		
Methyl Methacrylate	80-62-6	100.117	25.5		104		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-SG2

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-05

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.84	U	4.15		
Chloromethane	74-87-3	50.49	8.9		18.4		
Vinyl Chloride	75-01-4	62.5	0.06	U	0.15		
Bromomethane	74-83-9	94.94	0.56	U	2.17		
Chloroethane	75-00-3	64.52	3.3		8.71		
Tetrahydrofuran	109-99-9	72.11	2		5.9		
Trichlorofluoromethane	75-69-4	137.4	14.2		79.8		
Dichlorotetrafluoroethane	76-14-2	170.9	0.35	U	2.45		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.64	U	4.91		
tert-Butyl alcohol	75-65-0	74.12	0.76	U	2.3		
Heptane	142-82-5	100.2	0.44	U	1.8		
1,1-Dichloroethene	75-35-4	96.94	0.56	U	2.22		
Acetone	67-64-1	58.08	52.7		125		
Carbon Disulfide	75-15-0	76.14	0.64	U	1.99		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.48	U	1.73		
Methylene Chloride	75-09-2	84.94	2.8		9.73		
trans-1,2-Dichloroethene	156-60-5	96.94	0.76	U	3.01		
1,1-Dichloroethane	75-34-3	98.96	0.52	U	2.1		
Cyclohexane	110-82-7	84.16	0.88	U	3.03		
2-Butanone	78-93-3	72.11	3.9		11.5		
Carbon Tetrachloride	56-23-5	153.8	0.04	U	0.25		
cis-1,2-Dichloroethene	156-59-2	96.94	0.36	U	1.43		
Chloroform	67-66-3	119.4	8.3		40.5		
1,1,1-Trichloroethane	71-55-6	133.4	0.04	U	0.22		
2,2,4-Trimethylpentane	540-84-1	114.2	0.4	U	1.87		
Benzene	71-43-2	78.11	3.3		10.5		
1,2-Dichloroethane	107-06-2	98.96	0.36	U	1.46		
Trichloroethene	79-01-6	131.4	12.9		69.3		
1,2-Dichloropropane	78-87-5	113	0.44	U	2.03		
Bromodichloromethane	75-27-4	163.8	0.31	U	2.08		
4-Methyl-2-Pentanone	108-10-1	100.2	0.37	U	1.52		
Toluene	108-88-3	92.14	0.51	J	1.92		
t-1,3-Dichloropropene	10061-02-6	111	0.24	U	1.09		
cis-1,3-Dichloropropene	10061-01-5	111	0.26	U	1.18		
1,1,2-Trichloroethane	79-00-5	133.4	0.28	U	1.53		

Project : Yellow Cab - TO-15

Sampling Date : 06/12/25

Field Id Number : 115-SG2

Analysis Date : 06/16/25

Laboratory Id Number : Q2317-05

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.26	U	2.22		
1,2-Dibromoethane	106-93-4	187.9	0.29	U	2.23		
Tetrachloroethene	127-18-4	165.8	12.3		83.4		
Chlorobenzene	108-90-7	112.6	0.3	U	1.38		
Ethyl Benzene	100-41-4	106.2	0.48	U	2.08		
m/p-Xylene	179601-23-1	106.2	0.84	U	3.65		
o-Xylene	95-47-6	106.2	0.48	U	2.08		
Styrene	100-42-5	104.1	0.44	U	1.87		
Bromoform	75-25-2	252.8	0.23	U	2.38		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.04	U	0.27		
2-Chlorotoluene	95-49-8	126.6	0.44	U	2.28		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.44	U	2.16		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.31	U	1.52		
1,3-Dichlorobenzene	541-73-1	147	0.32	U	1.92		
1,4-Dichlorobenzene	106-46-7	147	5		30.1		
1,2-Dichlorobenzene	95-50-1	147	0.3	U	1.8		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.34	U	2.52		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.34	U	3.63		
1,3-Butadiene	106-99-0	54.09	0.52	U	1.15		
Naphthalene	91-20-3	128.17	0.3	U	1.57		
4-Ethyltoluene	622-96-8	120.2	0.48	U	2.36		
Hexane	110-54-3	86.17	0.88	J	3.1		
Allyl Chloride	107-05-1	76.53	0.6	U	1.88		
1,4-Dioxane	123-91-1	88.12	0.84	U	3.03		
Methyl Methacrylate	80-62-6	100.117	17.8		72.9		

Method	Matrix	Parameter	Spike_Amt_ Added	Calculated MDL	MB Data	MB Datafile	New MDL TO Be Used	MDL Unit
TO-15	Air	1,1,1,2-Tetrachloroethane	0.4	0.090683682	0		0.091	ppbv
TO-15	Air	1,1,1-Trichloroethane	0.03	0.00947589	0		0.0095	ppbv
TO-15	Air	1,1,2,2-Tetrachloroethane	0.03	0.010183321	0		0.011	ppbv
TO-15	Air	1,1,2-Trichloroethane	0.4	0.070112514	0		0.071	ppbv
TO-15	Air	1,1,2-Trichlorotrifluoroethane	0.4	0.155560652	0		0.16	ppbv
TO-15	Air	1,1-Dichloroethane	0.4	0.122378743	0		0.13	ppbv
TO-15	Air	1,1-Dichloroethene	0.4	0.139735617	0		0.14	ppbv
TO-15	Air	1,2,4-Trichlorobenzene	0.4	0.084677012	0		0.085	ppbv
TO-15	Air	1,2,4-Trimethylbenzene	0.4	0.077496367	0		0.078	ppbv
TO-15	Air	1,2-Dibromoethane	0.4	0.071318853	0		0.072	ppbv
TO-15	Air	1,2-Dichlorobenzene	0.4	0.074250332	0		0.075	ppbv
TO-15	Air	1,2-Dichloroethane	0.4	0.090890005	0		0.091	ppbv
TO-15	Air	1,2-Dichloropropane	0.4	0.073346868	0.109	VL038919.D	0.11	ppbv
TO-15	Air	1,3,5-Trimethylbenzene	0.4	0.101198372	0		0.11	ppbv
TO-15	Air	1,3-Butadiene	0.4	0.124009143	0		0.13	ppbv
TO-15	Air	1,3-Dichlorobenzene	0.4	0.079013305	0		0.08	ppbv
TO-15	Air	1,4-Dichlorobenzene	0.4	0.052746151	0		0.053	ppbv
TO-15	Air	1,4-Dioxane	0.4	0.202123585	0		0.21	ppbv
TO-15	Air	2,2,4-Trimethylpentane	0.4	0.09988427	0		0.10	ppbv
TO-15	Air	2-Butanone	0.4	0.118687999	0		0.12	ppbv
TO-15	Air	2-Chlorotoluene	0.4	0.104141644	0		0.11	ppbv
TO-15	Air	2-Hexanone	0.4	0.080656229	0		0.081	ppbv
TO-15	Air	4-Ethyltoluene	0.4	0.112037051	0		0.12	ppbv
TO-15	Air	4-Methyl-2-Pentanone	0.4	0.092836933	0		0.093	ppbv
TO-15	Air	Acetone	0.4	0.234416891	0		0.24	ppbv
TO-15	Air	Allyl Chloride	0.4	0.145618654	0		0.15	ppbv
TO-15	Air	Benzene	0.4	0.086407689	0		0.087	ppbv
TO-15	Air	Benzyl Chloride	0.4	0.173895169	0		0.18	ppbv
TO-15	Air	Bromodichloromethane	0.4	0.076780313	0		0.077	ppbv
TO-15	Air	Bromoethene	0.4	0.115560124	0		0.12	ppbv
TO-15	Air	Bromoform	0.4	0.056947551	0		0.057	ppbv
TO-15	Air	Bromomethane	0.4	0.137678239	0		0.14	ppbv
TO-15	Air	Carbon Disulfide	0.4	0.157343925	0		0.16	ppbv
TO-15	Air	Carbon Tetrachloride	0.03	0.010617308	0		0.011	ppbv
TO-15	Air	Chlorobenzene	0.4	0.075202602	0		0.076	ppbv
TO-15	Air	Chlorodifluoromethane	0.4	0.087466815	0		0.088	ppbv
TO-15	Air	Chloroethane	0.4	0.163530318	0		0.17	ppbv
TO-15	Air	Chloroform	0.4	0.082232394	0		0.083	ppbv
TO-15	Air	Chloromethane	0.4	0.097292998	0		0.098	ppbv
TO-15	Air	cis-1,2-Dichloroethene	0.4	0.088332535	0		0.089	ppbv
TO-15	Air	cis-1,3-Dichloropropene	0.4	0.063248763	0		0.064	ppbv
TO-15	Air	Cyclohexane	0.4	0.217999816	0		0.22	ppbv
TO-15	Air	Dibromochloromethane	0.4	0.065703935	0		0.066	ppbv
TO-15	Air	Dichlorodifluoromethane	0.4	0.20957733	0		0.21	ppbv
TO-15	Air	Dichlorotetrafluoroethane	0.4	0.087348046	0		0.088	ppbv
TO-15	Air	Ethyl Acetate	0.4	0.078161123	0		0.079	ppbv
TO-15	Air	Ethyl Benzene	0.4	0.115413562	0		0.12	ppbv
TO-15	Air	Heptane	0.4	0.100800637	0		0.11	ppbv
TO-15	Air	Hexachloro-1,3-Butadiene	0.4	0.084169773	0		0.085	ppbv
TO-15	Air	Hexane	0.4	0.106485082	0		0.11	ppbv
TO-15	Air	Isopropyl Alcohol	0.4	0.068329134	0		0.069	ppbv
TO-15	Air	Isopropylbenzene	0.4	0.086744608	0		0.087	ppbv
TO-15	Air	m/p-Xylene	0.4	0.205831125	0		0.21	ppbv
TO-15	Air	Methyl Methacrylate	0.4	0.085528753	0		0.086	ppbv
TO-15	Air	Methyl tert-Butyl Ether	0.4	0.119546728	0		0.12	ppbv
TO-15	Air	Methylene Chloride	0.4	0.18405265	0.11	VL039022.D	0.19	ppbv
TO-15	Air	Naphthalene	0.4	0.074961199	0		0.075	ppbv
TO-15	Air	Naphthalene,2-methyl-	0.4	0.148496389	0		0.15	ppbv
TO-15	Air	n-Butylbenzene	0.4	0.08201719	0		0.083	ppbv
TO-15	Air	n-propylbenzene	0.4	0.0778963	0		0.078	ppbv
TO-15	Air	o-Xylene	0.4	0.117630235	0		0.12	ppbv
TO-15	Air	p-Isopropyltoluene	0.4	0.085704866	0		0.086	ppbv
TO-15	Air	Propene	0.4	0.143790814	0.14	VL039337.D	0.15	ppbv
TO-15	Air	sec-Butylbenzene	0.4	0.093164076	0		0.094	ppbv
TO-15	Air	Styrene	0.4	0.109444442	0		0.11	ppbv
TO-15	Air	t-1,3-Dichloropropene	0.4	0.059734415	0		0.06	ppbv
TO-15	Air	tert-Butyl alcohol	0.4	0.187711954	0		0.19	ppbv
TO-15	Air	tert-Butylbenzene	0.4	0.092760327	0		0.093	ppbv
TO-15	Air	Tetrachloroethene	0.03	0.014462962	0		0.015	ppbv
TO-15	Air	Tetrahydrofuran	0.4	0.122655048	0		0.13	ppbv
TO-15	Air	Toluene	0.4	0.109284	0		0.11	ppbv
TO-15	Air	trans-1,2-Dichloroethene	0.4	0.185264638	0		0.19	ppbv
TO-15	Air	Trichloroethene	0.03	0.016558175	0		0.017	ppbv
TO-15	Air	Trichlorofluoromethane	0.4	0.157465794	0		0.16	ppbv
TO-15	Air	Vinyl Acetate	0.4	0.152371485	0		0.16	ppbv
TO-15	Air	Vinyl Chloride	0.03	0.013585954	0		0.014	ppbv
TO-15	Air	Ethanol	0.4	0.317031286	0		0.32	ug/L

StddevVal	StudentsT Value	Instrument ID	Datafile1	Concentration1	Adate1	Datafile2	Concentration2
0.03485153	2.602	MSVOA_L	VL038602.D	0.51	03142022T18:52:00	VL038645.D	0.494
0.003641772	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.029
0.003913651	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.035
0.026945624	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.439
0.059785032	2.602	MSVOA_L	VL038602.D	0.47	03142022T18:52:00	VL038645.D	0.466
0.047032569	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.456
0.053703158	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.411
0.032543048	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.336
0.029783385	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.396
0.027409244	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.352
0.028535869	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.407
0.034930824	2.602	MSVOA_L	VL038602.D	0.41	03142022T18:52:00	VL038645.D	0.445
0.02818865	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.406
0.038892534	2.602	MSVOA_L	VL038602.D	0.33	03142022T18:52:00	VL038645.D	0.374
0.047659163	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.419
0.030366374	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.428
0.020271388	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.418
0.077680086	2.602	MSVOA_L	VL038602.D	0.46	03142022T18:52:00	VL038645.D	0.285
0.038387498	2.602	MSVOA_L	VL038602.D	0.39	03142022T18:52:00	VL038645.D	0.383
0.045614143	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.406
0.040023691	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.375
0.030997782	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.283
0.043058052	2.602	MSVOA_L	VL038602.D	0.34	03142022T18:52:00	VL038645.D	0.356
0.035679067	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.282
0.090091042	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.468
0.055964125	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.392
0.033208182	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.407
0.066831349	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.454
0.029508191	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.408
0.044412039	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.433
0.021886069	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.31
0.052912467	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.404
0.060470378	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.357
0.004080441	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.026
0.028901845	2.602	MSVOA_L	VL038602.D	0.48	03142022T18:52:00	VL038645.D	0.467
0.033615225	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.468
0.062847932	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.453
0.031603534	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.464
0.037391621	2.602	MSVOA_L	VL038602.D	0.47	03142022T18:52:00	VL038645.D	0.439
0.033947938	2.602	MSVOA_L	VL038602.D	0.36	03142022T18:52:00	VL038645.D	0.391
0.024307749	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.317
0.083781636	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.414
0.02525132	2.602	MSVOA_L	VL038602.D	0.39	03142022T18:52:00	VL038645.D	0.359
0.080544708	2.602	MSVOA_L	VL038602.D	0.49	03142022T18:52:00	VL038645.D	0.43
0.033569579	2.602	MSVOA_L	VL038602.D	0.5	03142022T18:52:00	VL038645.D	0.472
0.030038864	2.602	MSVOA_L	VL038602.D	0.39	03142022T18:52:00	VL038645.D	0.412
0.044355712	2.602	MSVOA_L	VL038602.D	0.33	03142022T18:52:00	VL038645.D	0.357
0.038739676	2.602	MSVOA_L	VL038602.D	0.35	03142022T18:52:00	VL038645.D	0.349
0.032348107	2.602	MSVOA_L	VL038602.D	0.48	03142022T18:52:00	VL038645.D	0.5
0.04092432	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.411
0.026260236	2.602	MSVOA_L	VL038602.D	0.47	03142022T18:52:00	VL038645.D	0.44
0.033337666	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.387
0.079104967	2.602	MSVOA_L	VL038602.D	0.76	03142022T18:52:00	VL038645.D	0.762
0.032870389	2.602	MSVOA_L	VL038602.D	0.31	03142022T18:52:00	VL038645.D	0.334
0.045944169	2.602	MSVOA_L	VL038602.D	0.41	03142022T18:52:00	VL038645.D	0.418
0.070735069	2.602	MSVOA_L	VL038602.D	0.49	03142022T18:52:00	VL038645.D	0.479
0.02880907	2.602	MSVOA_L	VL038602.D	0.25	03142022T18:52:00	VL038645.D	0.289
0.057070096	2.602	MSVOA_L	VL038602.D	0.14	03142022T18:52:00	VL038645.D	0.074
0.031520826	2.602	MSVOA_L	VL038602.D	0.35	03142022T18:52:00	VL038645.D	0.366
0.029937087	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.365
0.045207623	2.602	MSVOA_L	VL038602.D	0.36	03142022T18:52:00	VL038645.D	0.368
0.032938073	2.602	MSVOA_L	VL038602.D	0.35	03142022T18:52:00	VL038645.D	0.366
0.05526165	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.443
0.035804795	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.392
0.042061661	2.602	MSVOA_L	VL038602.D	0.26	03142022T18:52:00	VL038645.D	0.275
0.022957116	2.602	MSVOA_L	VL038602.D	0.27	03142022T18:52:00	VL038645.D	0.29
0.072141412	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.459
0.035649626	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.405
0.005558402	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.031
0.047138758	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.324
0.042	2.602	MSVOA_L	VL038602.D	0.31	03142022T18:52:00	VL038645.D	0.332
0.07120086	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.396
0.006363634	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.035
0.060517215	2.602	MSVOA_L	VL038602.D	0.46	03142022T18:52:00	VL038645.D	0.472
0.058559372	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.389
0.00522135	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.031
0.105747594	2.998	MSVOA_L	VL040210.D	0.438	02222023T00:52:00	VL040220.D	0.43

Adate2	Datafile3	Concentration3	Adate3	Datafile4	Concentration4	Adate4	Datafile5
03182022T03:39:00	VL039242.D	0.45	06082022T19:44:00	VL039436.D	0.45	07062022T02:26:00	VL039568.D
03212022T20:11:00	VL039329.D	0.036	06222022T20:53:00	VL039439.D	0.033	07062022T09:13:00	VL039570.D
03212022T20:11:00	VL039329.D	0.029	06222022T20:53:00	VL039439.D	0.035	07062022T09:13:00	VL039570.D
03182022T03:39:00	VL039242.D	0.417	06082022T19:44:00	VL039436.D	0.445	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.615	06082022T19:44:00	VL039436.D	0.498	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.493	06082022T19:44:00	VL039436.D	0.545	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.594	06082022T19:44:00	VL039436.D	0.485	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.318	06082022T19:44:00	VL039436.D	0.328	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.389	06082022T19:44:00	VL039436.D	0.387	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.392	06082022T19:44:00	VL039436.D	0.391	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.409	06082022T19:44:00	VL039436.D	0.393	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.494	06082022T19:44:00	VL039436.D	0.435	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.404	06082022T19:44:00	VL039436.D	0.451	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.421	06082022T19:44:00	VL039436.D	0.397	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.583	06082022T19:44:00	VL039436.D	0.486	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.409	06082022T19:44:00	VL039436.D	0.465	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.392	06082022T19:44:00	VL039436.D	0.41	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.454	06082022T19:44:00	VL039436.D	0.535	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.37	06082022T19:44:00	VL039436.D	0.425	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.438	06082022T19:44:00	VL039436.D	0.482	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.415	06082022T19:44:00	VL039436.D	0.401	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.309	06082022T19:44:00	VL039436.D	0.31	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.391	06082022T19:44:00	VL039436.D	0.393	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.371	06082022T19:44:00	VL039436.D	0.372	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.649	06082022T19:44:00	VL039436.D	0.621	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.58	06082022T19:44:00	VL039436.D	0.45	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.391	06082022T19:44:00	VL039436.D	0.429	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.328	06082022T19:44:00	VL039436.D	0.438	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.398	06082022T19:44:00	VL039436.D	0.395	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.526	06082022T19:44:00	VL039436.D	0.475	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.335	06082022T19:44:00	VL039436.D	0.322	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.535	06082022T19:44:00	VL039436.D	0.503	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.494	06082022T19:44:00	VL039436.D	0.389	07062022T02:26:00	VL039568.D
03212022T20:11:00	VL039329.D	0.028	06222022T20:53:00	VL039439.D	0.03	07062022T09:13:00	VL039570.D
03182022T03:39:00	VL039242.D	0.418	06082022T19:44:00	VL039436.D	0.473	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.489	06082022T19:44:00	VL039436.D	0.483	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.595	06082022T19:44:00	VL039436.D	0.52	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.453	06082022T19:44:00	VL039436.D	0.469	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.5	06082022T19:44:00	VL039436.D	0.482	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.403	06082022T19:44:00	VL039436.D	0.434	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.315	06082022T19:44:00	VL039436.D	0.327	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.428	06082022T19:44:00	VL039436.D	0.512	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.359	06082022T19:44:00	VL039436.D	0.37	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.618	06082022T19:44:00	VL039436.D	0.675	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.551	06082022T19:44:00	VL039436.D	0.51	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.376	06082022T19:44:00	VL039436.D	0.441	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.374	06082022T19:44:00	VL039436.D	0.381	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.33	06082022T19:44:00	VL039436.D	0.41	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.444	06082022T19:44:00	VL039436.D	0.472	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.394	06082022T19:44:00	VL039436.D	0.47	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.504	06082022T19:44:00	VL039436.D	0.465	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.425	06082022T19:44:00	VL039436.D	0.416	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.815	06082022T19:44:00	VL039436.D	0.805	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.347	06082022T19:44:00	VL039436.D	0.361	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.492	06082022T19:44:00	VL039436.D	0.452	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.684	06082022T19:44:00	VL039436.D	0.578	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.282	06082022T19:44:00	VL039436.D	0.302	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.155	06082022T19:44:00	VL039436.D	0.143	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.39	06082022T19:44:00	VL039436.D	0.374	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.394	06082022T19:44:00	VL039436.D	0.397	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.421	06082022T19:44:00	VL039436.D	0.403	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.378	06082022T19:44:00	VL039436.D	0.388	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.435	06082022T19:44:00	VL039436.D	0.521	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.411	06082022T19:44:00	VL039436.D	0.406	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.345	06082022T19:44:00	VL039436.D	0.302	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.309	06082022T19:44:00	VL039436.D	0.279	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.655	06082022T19:44:00	VL039436.D	0.505	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.403	06082022T19:44:00	VL039436.D	0.388	07062022T02:26:00	VL039568.D
03212022T20:11:00	VL039329.D	0.036	06222022T20:53:00	VL039439.D	0.034	07062022T09:13:00	VL039570.D
03182022T03:39:00	VL039242.D	0.346	06082022T19:44:00	VL039436.D	0.38	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.341	06082022T19:44:00	VL039436.D	0.37	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.64	06082022T19:44:00	VL039436.D	0.494	07062022T02:26:00	VL039568.D
03212022T20:11:00	VL039329.D	0.038	06222022T20:53:00	VL039439.D	0.034	07062022T09:13:00	VL039570.D
03182022T03:39:00	VL039242.D	0.596	06082022T19:44:00	VL039436.D	0.492	07062022T02:26:00	VL039568.D
03182022T03:39:00	VL039242.D	0.418	06082022T19:44:00	VL039436.D	0.437	07062022T02:26:00	VL039568.D
03212022T20:11:00	VL039329.D	0.043	06222022T20:53:00	VL039439.D	0.036	07062022T09:13:00	VL039570.D
02232023T13:12:00	VL040329.D	0.51	0412023T17:39:00	VL040349.D	0.633	04122023T23:23:00	VL040559.D

Concentration5	Adate5	Datafile6	Concentration6	Adate6	Datafile7	Concentration7	Adate7
0.433	08182022T21:04:00	VL039600.D	0.463	08262022T12:33:00	VL039709.D	0.402	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.031	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.03	08182022T22:24:00	VL039586.D	0.029	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.404	08182022T21:04:00	VL039600.D	0.424	08262022T12:33:00	VL039709.D	0.412	10102022T23:30:00
0.448	08182022T21:04:00	VL039600.D	0.454	08262022T12:33:00	VL039709.D	0.438	10102022T23:30:00
0.426	08182022T21:04:00	VL039600.D	0.451	08262022T12:33:00	VL039709.D	0.445	10102022T23:30:00
0.394	08182022T21:04:00	VL039600.D	0.387	08262022T12:33:00	VL039709.D	0.427	10102022T23:30:00
0.314	08182022T21:04:00	VL039600.D	0.367	08262022T12:33:00	VL039709.D	0.373	10102022T23:30:00
0.342	08182022T21:04:00	VL039600.D	0.389	08262022T12:33:00	VL039709.D	0.389	10102022T23:30:00
0.361	08182022T21:04:00	VL039600.D	0.402	08262022T12:33:00	VL039709.D	0.375	10102022T23:30:00
0.408	08182022T21:04:00	VL039600.D	0.431	08262022T12:33:00	VL039709.D	0.448	10102022T23:30:00
0.411	08182022T21:04:00	VL039600.D	0.455	08262022T12:33:00	VL039709.D	0.418	10102022T23:30:00
0.444	08182022T21:04:00	VL039600.D	0.429	08262022T12:33:00	VL039709.D	0.373	10102022T23:30:00
0.331	08182022T21:04:00	VL039600.D	0.357	08262022T12:33:00	VL039709.D	0.342	10102022T23:30:00
0.405	08182022T21:04:00	VL039600.D	0.449	08262022T12:33:00	VL039709.D	0.422	10102022T23:30:00
0.368	08182022T21:04:00	VL039600.D	0.414	08262022T12:33:00	VL039709.D	0.399	10102022T23:30:00
0.404	08182022T21:04:00	VL039600.D	0.384	08262022T12:33:00	VL039709.D	0.439	10102022T23:30:00
0.373	08182022T21:04:00	VL039600.D	0.341	08262022T12:33:00	VL039709.D	0.418	10102022T23:30:00
0.348	08182022T21:04:00	VL039600.D	0.379	08262022T12:33:00	VL039709.D	0.392	10102022T23:30:00
0.314	08182022T21:04:00	VL039600.D	0.451	08262022T12:33:00	VL039709.D	0.354	10102022T23:30:00
0.312	08182022T21:04:00	VL039600.D	0.359	08262022T12:33:00	VL039709.D	0.348	10102022T23:30:00
0.27	08182022T21:04:00	VL039600.D	0.294	08262022T12:33:00	VL039709.D	0.326	10102022T23:30:00
0.284	08182022T21:04:00	VL039600.D	0.348	08262022T12:33:00	VL039709.D	0.355	10102022T23:30:00
0.322	08182022T21:04:00	VL039600.D	0.335	08262022T12:33:00	VL039709.D	0.32	10102022T23:30:00
0.363	08182022T21:04:00	VL039600.D	0.484	08262022T12:33:00	VL039709.D	0.5	10102022T23:30:00
0.418	08182022T21:04:00	VL039600.D	0.459	08262022T12:33:00	VL039709.D	0.379	10102022T23:30:00
0.345	08182022T21:04:00	VL039600.D	0.376	08262022T12:33:00	VL039709.D	0.398	10102022T23:30:00
0.394	08182022T21:04:00	VL039600.D	0.336	08262022T12:33:00	VL039709.D	0.356	10102022T23:30:00
0.382	08182022T21:04:00	VL039600.D	0.367	08262022T12:33:00	VL039709.D	0.377	10102022T23:30:00
0.408	08182022T21:04:00	VL039600.D	0.441	08262022T12:33:00	VL039709.D	0.388	10102022T23:30:00
0.305	08182022T21:04:00	VL039600.D	0.303	08262022T12:33:00	VL039709.D	0.293	10102022T23:30:00
0.364	08182022T21:04:00	VL039600.D	0.494	08262022T12:33:00	VL039709.D	0.464	10102022T23:30:00
0.399	08182022T21:04:00	VL039600.D	0.268	08262022T12:33:00	VL039709.D	0.332	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.027	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.431	08182022T21:04:00	VL039600.D	0.466	08262022T12:33:00	VL039709.D	0.46	10102022T23:30:00
0.456	08182022T21:04:00	VL039600.D	0.504	08262022T12:33:00	VL039709.D	0.446	10102022T23:30:00
0.462	08182022T21:04:00	VL039600.D	0.504	08262022T12:33:00	VL039709.D	0.531	10102022T23:30:00
0.423	08182022T21:04:00	VL039600.D	0.46	08262022T12:33:00	VL039709.D	0.411	10102022T23:30:00
0.392	08182022T21:04:00	VL039600.D	0.472	08262022T12:33:00	VL039709.D	0.485	10102022T23:30:00
0.36	08182022T21:04:00	VL039600.D	0.407	08262022T12:33:00	VL039709.D	0.438	10102022T23:30:00
0.277	08182022T21:04:00	VL039600.D	0.291	08262022T12:33:00	VL039709.D	0.284	10102022T23:30:00
0.376	08182022T21:04:00	VL039600.D	0.395	08262022T12:33:00	VL039709.D	0.383	10102022T23:30:00
0.339	08182022T21:04:00	VL039600.D	0.335	08262022T12:33:00	VL039709.D	0.301	10102022T23:30:00
0.62	08182022T21:04:00	VL039600.D	0.643	08262022T12:33:00	VL039709.D	0.694	10102022T23:30:00
0.454	08182022T21:04:00	VL039600.D	0.508	08262022T12:33:00	VL039709.D	0.473	10102022T23:30:00
0.341	08182022T21:04:00	VL039600.D	0.432	08262022T12:33:00	VL039709.D	0.395	10102022T23:30:00
0.297	08182022T21:04:00	VL039600.D	0.338	08262022T12:33:00	VL039709.D	0.334	10102022T23:30:00
0.347	08182022T21:04:00	VL039600.D	0.382	08262022T12:33:00	VL039709.D	0.407	10102022T23:30:00
0.447	08182022T21:04:00	VL039600.D	0.502	08262022T12:33:00	VL039709.D	0.468	10102022T23:30:00
0.385	08182022T21:04:00	VL039600.D	0.405	08262022T12:33:00	VL039709.D	0.435	10102022T23:30:00
0.458	08182022T21:04:00	VL039600.D	0.434	08262022T12:33:00	VL039709.D	0.481	10102022T23:30:00
0.361	08182022T21:04:00	VL039600.D	0.388	08262022T12:33:00	VL039709.D	0.405	10102022T23:30:00
0.673	08182022T21:04:00	VL039600.D	0.719	08262022T12:33:00	VL039709.D	0.688	10102022T23:30:00
0.3	08182022T21:04:00	VL039600.D	0.323	08262022T12:33:00	VL039709.D	0.346	10102022T23:30:00
0.337	08182022T21:04:00	VL039600.D	0.355	08262022T12:33:00	VL039709.D	0.396	10102022T23:30:00
0.536	08182022T21:04:00	VL039600.D	0.634	08262022T12:33:00	VL039709.D	0.499	10102022T23:30:00
0.28	08182022T21:04:00	VL039600.D	0.302	08262022T12:33:00	VL039709.D	0.291	10102022T23:30:00
0.158	08182022T21:04:00	VL039600.D	0.094	08262022T12:33:00	VL039709.D	0.149	10102022T23:30:00
0.321	08182022T21:04:00	VL039600.D	0.354	08262022T12:33:00	VL039709.D	0.358	10102022T23:30:00
0.332	08182022T21:04:00	VL039600.D	0.338	08262022T12:33:00	VL039709.D	0.347	10102022T23:30:00
0.315	08182022T21:04:00	VL039600.D	0.347	08262022T12:33:00	VL039709.D	0.354	10102022T23:30:00
0.316	08182022T21:04:00	VL039600.D	0.362	08262022T12:33:00	VL039709.D	0.358	10102022T23:30:00
0.499	08182022T21:04:00	VL039600.D	0.578	08262022T12:33:00	VL039709.D	0.533	10102022T23:30:00
0.359	08182022T21:04:00	VL039600.D	0.382	08262022T12:33:00	VL039709.D	0.371	10102022T23:30:00
0.23	08182022T21:04:00	VL039600.D	0.257	08262022T12:33:00	VL039709.D	0.278	10102022T23:30:00
0.252	08182022T21:04:00	VL039600.D	0.252	08262022T12:33:00	VL039709.D	0.273	10102022T23:30:00
0.384	08182022T21:04:00	VL039600.D	0.382	08262022T12:33:00	VL039709.D	0.454	10102022T23:30:00
0.345	08182022T21:04:00	VL039600.D	0.399	08262022T12:33:00	VL039709.D	0.371	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.038	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.299	08182022T21:04:00	VL039600.D	0.344	08262022T12:33:00	VL039709.D	0.389	10102022T23:30:00
0.288	08182022T21:04:00	VL039600.D	0.296	08262022T12:33:00	VL039709.D	0.33	10102022T23:30:00
0.375	08182022T21:04:00	VL039600.D	0.466	08262022T12:33:00	VL039709.D	0.437	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.031	08252022T20:23:00	VL039711.D	0.04	10112022T00:45:00
0.436	08182022T21:04:00	VL039600.D	0.47	08262022T12:33:00	VL039709.D	0.385	10102022T23:30:00
0.42	08182022T21:04:00	VL039600.D	0.483	08262022T12:33:00	VL039709.D	0.363	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.029	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.636	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.607	10042023T19:33:00

Datafile8	Concentration8	Adate8	Datafile9	Concentration9	Adate9	Datafile10	Concentration10
VL040134.D	0.42	01302023T15:54:00	VL040210.D	0.383	02222023T00:52:00	VL040220.D	0.42
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.038	02232023T14:26:00	VL040223.D	0.033
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.032	02232023T14:26:00	VL040223.D	0.027
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.395	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.476	02222023T00:52:00	VL040220.D	0.47
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.436	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.452	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.32	01302023T15:54:00	VL040210.D	0.299	02222023T00:52:00	VL040220.D	0.32
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.375	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.38	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.378	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.434	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.432	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.344	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.396	02222023T00:52:00	VL040220.D	0.41
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.395	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.402	02222023T00:52:00	VL040220.D	0.4
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.387	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.352	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.416	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.34	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.32	01302023T15:54:00	VL040210.D	0.318	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.321	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.337	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.45	01302023T15:54:00	VL040210.D	0.482	02222023T00:52:00	VL040220.D	0.5
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.411	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.359	02222023T00:52:00	VL040220.D	0.39
VL040134.D	0.29	01302023T15:54:00	VL040210.D	0.311	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.378	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.413	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.32	01302023T15:54:00	VL040210.D	0.32	02222023T00:52:00	VL040220.D	0.33
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.408	02222023T00:52:00	VL040220.D	0.49
VL040134.D	0.31	01302023T15:54:00	VL040210.D	0.391	02222023T00:52:00	VL040220.D	0.37
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.033	02232023T14:26:00	VL040223.D	0.037
VL040134.D	0.44	01302023T15:54:00	VL040210.D	0.414	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.474	02222023T00:52:00	VL040220.D	0.46
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.416	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.47	01302023T15:54:00	VL040210.D	0.445	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.46	01302023T15:54:00	VL040210.D	0.443	02222023T00:52:00	VL040220.D	0.47
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.435	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.33	01302023T15:54:00	VL040210.D	0.284	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.49	01302023T15:54:00	VL040210.D	0.521	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.343	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.48	01302023T15:54:00	VL040210.D	0.497	02222023T00:52:00	VL040220.D	0.49
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.431	02222023T00:52:00	VL040220.D	0.46
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.398	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.309	02222023T00:52:00	VL040220.D	0.33
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.336	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.421	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.4	02222023T00:52:00	VL040220.D	0.41
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.479	02222023T00:52:00	VL040220.D	0.48
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.361	02222023T00:52:00	VL040220.D	0.4
VL040134.D	0.71	01302023T15:54:00	VL040210.D	0.699	02222023T00:52:00	VL040220.D	0.75
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.314	02222023T00:52:00	VL040220.D	0.35
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.421	02222023T00:52:00	VL040220.D	0.41
VL040134.D	0.44	01302023T15:54:00	VL040210.D	0.558	02222023T00:52:00	VL040220.D	0.57
VL040134.D	0.27	01302023T15:54:00	VL040210.D	0.246	02222023T00:52:00	VL040220.D	0.27
VL040134.D	0.19	01302023T15:54:00	VL040210.D	0.116	02222023T00:52:00	VL040220.D	0.21
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.349	02222023T00:52:00	VL040220.D	0.35
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.354	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.331	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.33	01302023T15:54:00	VL040210.D	0.329	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.42	01302023T15:54:00	VL040210.D	0.432	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.352	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.29	01302023T15:54:00	VL040210.D	0.262	02222023T00:52:00	VL040220.D	0.27
VL040134.D	0.3	01302023T15:54:00	VL040210.D	0.292	02222023T00:52:00	VL040220.D	0.3
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.395	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.349	02222023T00:52:00	VL040220.D	0.36
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.029	02232023T14:26:00	VL040223.D	0.031
VL040134.D	0.42	01302023T15:54:00	VL040210.D	0.368	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.297	02222023T00:52:00	VL040220.D	0.33
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.397	02222023T00:52:00	VL040220.D	0.42
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.027	02232023T14:26:00	VL040223.D	0.028
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.451	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.438	02222023T00:52:00	VL040220.D	0.4
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.033	02232023T14:26:00	VL040223.D	0.038
VL040805.D	0.675	10252023T18:12:00					

A
B
C
D
E
F
G
H

Concentration13	Adate13	Datafile14	Concentration14	Adate14	Datafile15	Concentration15	Adate15
0.48	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.488	10042023T19:33:00
0.03	07312023T17:18:00	VL040581.D	0.031	07312023T20:41:00	VL040772.D	0.035	10122023T01:48:00
0.031	07312023T17:18:00	VL040581.D	0.026	07312023T20:41:00	VL040772.D	0.029	10122023T01:48:00
0.459	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.405	10042023T19:33:00
0.573	07272023T03:27:00	VL040579.D	0.5	07312023T19:20:00	VL040747.D	0.557	10042023T19:33:00
0.535	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.527	10042023T19:33:00
0.527	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.481	10042023T19:33:00
0.33	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.305	10042023T19:33:00
0.389	07272023T03:27:00	VL040579.D	0.3	07312023T19:20:00	VL040747.D	0.361	10042023T19:33:00
0.425	07272023T03:27:00	VL040579.D	0.32	07312023T19:20:00	VL040747.D	0.359	10042023T19:33:00
0.442	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.45	10042023T19:33:00
0.519	07272023T03:27:00	VL040579.D	0.41	07312023T19:20:00	VL040747.D	0.426	10042023T19:33:00
0.452	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.42	10042023T19:33:00
0.33	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.321	10042023T19:33:00
0.412	07272023T03:27:00	VL040579.D	0.41	07312023T19:20:00	VL040747.D	0.466	10042023T19:33:00
0.388	07272023T03:27:00	VL040579.D	0.33	07312023T19:20:00	VL040747.D	0.412	10042023T19:33:00
0.438	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.382	10042023T19:33:00
0.373	07272023T03:27:00	VL040579.D	0.29	07312023T19:20:00	VL040747.D	0.257	10042023T19:33:00
0.398	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.335	10042023T19:33:00
0.448	07272023T03:27:00	VL040579.D	0.37	07312023T19:20:00	VL040747.D	0.443	10042023T19:33:00
0.35	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.325	10042023T19:33:00
0.303	07272023T03:27:00	VL040579.D	0.22	07312023T19:20:00	VL040747.D	0.283	10042023T19:33:00
0.32	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.317	10042023T19:33:00
0.348	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.302	10042023T19:33:00
0.674	07272023T03:27:00	VL040579.D	0.48	07312023T19:20:00	VL040747.D	0.65	10042023T19:33:00
0.436	07272023T03:27:00	VL040579.D	0.37	07312023T19:20:00	VL040747.D	0.511	10042023T19:33:00
0.397	07272023T03:27:00	VL040579.D	0.3	07312023T19:20:00	VL040747.D	0.386	10042023T19:33:00
0.425	07272023T03:27:00	VL040579.D	0.34	07312023T19:20:00	VL040747.D	0.307	10042023T19:33:00
0.467	07272023T03:27:00	VL040579.D	0.4	07312023T19:20:00	VL040747.D	0.409	10042023T19:33:00
0.508	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.506	10042023T19:33:00
0.352	07272023T03:27:00	VL040579.D	0.33	07312023T19:20:00	VL040747.D	0.349	10042023T19:33:00
0.543	07272023T03:27:00	VL040579.D	0.46	07312023T19:20:00	VL040747.D	0.496	10042023T19:33:00
0.46	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.477	10042023T19:33:00
0.036	07312023T17:18:00	VL040581.D	0.033	07312023T20:41:00	VL040772.D	0.031	10122023T01:48:00
0.481	07272023T03:27:00	VL040579.D	0.38	07312023T19:20:00	VL040747.D	0.489	10042023T19:33:00
0.526	07272023T03:27:00	VL040579.D	0.45	07312023T19:20:00	VL040747.D	0.55	10042023T19:33:00
0.516	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.576	10042023T19:33:00
0.521	07272023T03:27:00	VL040579.D	0.41	07312023T19:20:00	VL040747.D	0.501	10042023T19:33:00
0.486	07272023T03:27:00	VL040579.D	0.39	07312023T19:20:00	VL040747.D	0.544	10042023T19:33:00
0.405	07272023T03:27:00	VL040579.D	0.32	07312023T19:20:00	VL040747.D	0.431	10042023T19:33:00
0.295	07272023T03:27:00	VL040579.D	0.25	07312023T19:20:00	VL040747.D	0.298	10042023T19:33:00
0.301	07272023T03:27:00	VL040579.D	0.21	07312023T19:20:00	VL040747.D	0.317	10042023T19:33:00
0.403	07272023T03:27:00	VL040579.D	0.35	07312023T19:20:00	VL040747.D	0.381	10042023T19:33:00
0.54	07272023T03:27:00	VL040579.D	0.47	07312023T19:20:00	VL040747.D	0.562	10042023T19:33:00
0.516	07272023T03:27:00	VL040579.D	0.46	07312023T19:20:00	VL040747.D	0.509	10042023T19:33:00
0.424	07272023T03:27:00	VL040579.D	0.34	07312023T19:20:00	VL040747.D	0.441	10042023T19:33:00
0.32	07272023T03:27:00	VL040579.D	0.21	07312023T19:20:00	VL040747.D	0.299	10042023T19:33:00
0.341	07272023T03:27:00	VL040579.D	0.25	07312023T19:20:00	VL040747.D	0.355	10042023T19:33:00
0.466	07272023T03:27:00	VL040579.D	0.37	07312023T19:20:00	VL040747.D	0.479	10042023T19:33:00
0.38	07272023T03:27:00	VL040579.D	0.31	07312023T19:20:00	VL040747.D	0.465	10042023T19:33:00
0.451	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.426	10042023T19:33:00
0.379	07272023T03:27:00	VL040579.D	0.29	07312023T19:20:00	VL040747.D	0.386	10042023T19:33:00
0.735	07272023T03:27:00	VL040579.D	0.54	07312023T19:20:00	VL040747.D	0.682	10042023T19:33:00
0.283	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.324	10042023T19:33:00
0.468	07272023T03:27:00	VL040579.D	0.38	07312023T19:20:00	VL040747.D	0.472	10042023T19:33:00
0.619	07272023T03:27:00	VL040579.D	0.62	07312023T19:20:00	VL040747.D	0.625	10042023T19:33:00
0.274	07272023T03:27:00	VL040579.D	0.19	07312023T19:20:00	VL040747.D	0.258	10042023T19:33:00
0.107	07272023T03:27:00	VL040579.D	0.06	07312023T19:20:00	VL040747.D	0.133	10042023T19:33:00
0.361	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.343	10042023T19:33:00
0.355	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.369	10042023T19:33:00
0.355	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.301	10042023T19:33:00
0.357	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.313	10042023T19:33:00
0.436	07272023T03:27:00	VL040579.D	0.34	07312023T19:20:00	VL040747.D	0.471	10042023T19:33:00
0.393	07272023T03:27:00	VL040579.D	0.29	07312023T19:20:00	VL040747.D	0.354	10042023T19:33:00
0.242	07272023T03:27:00	VL040579.D	0.17	07312023T19:20:00	VL040747.D	0.266	10042023T19:33:00
0.283	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.312	10042023T19:33:00
0.539	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.394	10042023T19:33:00
0.366	07272023T03:27:00	VL040579.D	0.28	07312023T19:20:00	VL040747.D	0.381	10042023T19:33:00
0.025	07312023T17:18:00	VL040581.D	0.028	07312023T20:41:00	VL040772.D	0.044	10122023T01:48:00
0.353	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.351	10042023T19:33:00
0.285	07272023T03:27:00	VL040579.D	0.2	07312023T19:20:00	VL040747.D	0.288	10042023T19:33:00
0.477	07272023T03:27:00	VL040579.D	0.4	07312023T19:20:00	VL040747.D	0.507	10042023T19:33:00
0.027	07312023T17:18:00	VL040581.D	0.027	07312023T20:41:00	VL040772.D	0.039	10122023T01:48:00
0.589	07272023T03:27:00	VL040579.D	0.52	07312023T19:20:00	VL040747.D	0.561	10042023T19:33:00
0.487	07272023T03:27:00	VL040579.D	0.46	07312023T19:20:00	VL040747.D	0.456	10042023T19:33:00
0.03	07312023T17:18:00	VL040581.D	0.033	07312023T20:41:00	VL040772.D	0.041	10122023T01:48:00

Datafile16	Concentration16	Adate16	Department	AVG Concentration	Recovery
VL040805.D	0.428	10252023T18:12:00	VOC	0.45	110
VL040787.D	0.04	10232023T20:32:00	VOC	0.032	110
VL040787.D	0.04	10232023T20:32:00	VOC	0.03	100
VL040805.D	0.471	10252023T18:12:00	VOC	0.42	110
VL040805.D	0.476	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.427	10252023T18:12:00	VOC	0.46	110
VL040805.D	0.47	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.288	10252023T18:12:00	VOC	0.31	79
VL040805.D	0.435	10252023T18:12:00	VOC	0.37	93
VL040805.D	0.433	10252023T18:12:00	VOC	0.38	95
VL040805.D	0.405	10252023T18:12:00	VOC	0.41	100
VL040805.D	0.501	10252023T18:12:00	VOC	0.44	110
VL040805.D	0.456	10252023T18:12:00	VOC	0.42	100
VL040805.D	0.403	10252023T18:12:00	VOC	0.35	87
VL040805.D	0.455	10252023T18:12:00	VOC	0.44	110
VL040805.D	0.412	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.414	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.486	10252023T18:12:00	VOC	0.38	96
VL040805.D	0.434	10252023T18:12:00	VOC	0.37	93
VL040805.D	0.467	10252023T18:12:00	VOC	0.42	100
VL040805.D	0.42	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.353	10252023T18:12:00	VOC	0.3	75
VL040805.D	0.397	10252023T18:12:00	VOC	0.34	84
VL040805.D	0.403	10252023T18:12:00	VOC	0.34	84
VL040805.D	0.576	10252023T18:12:00	VOC	0.53	130
VL040805.D	0.422	10252023T18:12:00	VOC	0.43	110
VL040805.D	0.445	10252023T18:12:00	VOC	0.38	96
VL040805.D	0.205	10252023T18:12:00	VOC	0.36	91
VL040805.D	0.458	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.436	10252023T18:12:00	VOC	0.44	110
VL040805.D	0.34	10252023T18:12:00	VOC	0.33	82
VL040805.D	0.508	10252023T18:12:00	VOC	0.47	120
VL040805.D	0.361	10252023T18:12:00	VOC	0.38	95
VL040787.D	0.04	10232023T20:32:00	VOC	0.031	100
VL040805.D	0.454	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.517	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.435	10252023T18:12:00	VOC	0.47	120
VL040805.D	0.487	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.458	10252023T18:12:00	VOC	0.46	120
VL040805.D	0.435	10252023T18:12:00	VOC	0.4	99
VL040805.D	0.335	10252023T18:12:00	VOC	0.3	75
VL040805.D	0.416	10252023T18:12:00	VOC	0.39	96
VL040805.D	0.384	10252023T18:12:00	VOC	0.36	89
VL040805.D	0.516	10252023T18:12:00	VOC	0.54	140
VL040805.D	0.505	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.404	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.406	10252023T18:12:00	VOC	0.33	82
VL040805.D	0.404	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.446	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.452	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.442	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.435	10252023T18:12:00	VOC	0.38	96
VL040805.D	0.888	10252023T18:12:00	VOC	0.72	180
VL040805.D	0.348	10252023T18:12:00	VOC	0.32	80
VL040805.D	0.353	10252023T18:12:00	VOC	0.41	100
VL040805.D	0.619	10252023T18:12:00	VOC	0.56	140
VL040805.D	0.226	10252023T18:12:00	VOC	0.27	66
VL040805.D	0	10252023T18:12:00	VOC	0.14	35
VL040805.D	0.415	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.389	10252023T18:12:00	VOC	0.35	89
VL040805.D	0.439	10252023T18:12:00	VOC	0.35	87
VL040805.D	0.381	10252023T18:12:00	VOC	0.34	85
VL040805.D	0.467	10252023T18:12:00	VOC	0.46	110
VL040805.D	0.449	10252023T18:12:00	VOC	0.37	93
VL040805.D	0.35	10252023T18:12:00	VOC	0.27	67
VL040805.D	0.291	10252023T18:12:00	VOC	0.28	70
VL040805.D	0.37	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.427	10252023T18:12:00	VOC	0.37	92
VL040787.D	0.04	10232023T20:32:00	VOC	0.032	110
VL040805.D	0.416	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.38	10252023T18:12:00	VOC	0.31	78
VL040805.D	0.435	10252023T18:12:00	VOC	0.43	110
VL040787.D	0.05	10232023T20:32:00	VOC	0.033	110
VL040805.D	0.506	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.341	10252023T18:12:00	VOC	0.43	110
VL040787.D	0.04	10232023T20:32:00	VOC	0.033	110
			VOC	0.544	136

A
B
C
D
E
F
G
H



SAMPLE DATA

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	AA-3	SDG No.:	Q2317
Lab Sample ID:	Q2317-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042626.D	1		06/16/25 13:27	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.48	2.37	J	1.04	2.47	ug/m3
74-87-3	Chloromethane	0.54	1.12		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.13	0.38	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.27	1.52	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.19	0.58	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.15	0.61	J	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	10.2	24.2		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.16	0.50	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	3.00	10.4		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.65	1.92		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.080	0.50		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.080	0.39	U	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.19	0.89	J	0.47	2.34	ug/m3
71-43-2	Benzene	0.32	1.02	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.18	0.74	J	0.37	2.05	ug/m3
108-88-3	Toluene	0.46	1.73	J	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	AA-3	SDG No.:	Q2317
Lab Sample ID:	Q2317-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042626.D	1		06/16/25 13:27	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.020	0.14	U	0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.48	2.08	J	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.23	1.00	J	0.52	2.17	ug/m3
100-42-5	Styrene	0.13	0.55	J	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.17	0.84	J	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.050	0.30	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	4.80	16.9		0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.3			70 (65) - 130 (135)	103%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	109000		2.79			
540-36-3	1,4-Difluorobenzene	292000		3.962			
3114-55-4	Chlorobenzene-d5	255000		8.885			

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	AA-3	SDG No.:	Q2317
Lab Sample ID:	Q2317-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042626.D	1		06/16/25 13:27	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-1	SDG No.:	Q2317
Lab Sample ID:	Q2317-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042628.D	1		06/16/25 14:35	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.47	2.32	J	1.04	2.47	ug/m3
74-87-3	Chloromethane	0.71	1.47		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	1.40	4.13		0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	8.10	45.5		0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.19	0.58	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.20	0.82	J	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	38.3	91.0	E	0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.16	0.50	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	1.30	4.52		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	2.10	6.19		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.090	0.57		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.58	2.83		0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.16	0.75	J	0.47	2.34	ug/m3
71-43-2	Benzene	0.33	1.05	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.17	0.69	J	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.050	0.27		0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.26	1.07	J	0.37	2.05	ug/m3
108-88-3	Toluene	0.75	2.83		0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-1	SDG No.:	Q2317
Lab Sample ID:	Q2317-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042628.D	1		06/16/25 14:35	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.070	0.47		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	J	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.39	1.69	J	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.17	0.74	J	0.52	2.17	ug/m3
100-42-5	Styrene	0.14	0.60	J	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.28	1.38	J	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	17.4	105	E	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	4.10	14.4		0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.4			70 (65) - 130 (135)	104%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	107000		2.787			
540-36-3	1,4-Difluorobenzene	298000		3.959			
3114-55-4	Chlorobenzene-d5	260000		8.882			

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-1	SDG No.:	Q2317
Lab Sample ID:	Q2317-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042628.D	1		06/16/25 14:35	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-1DL	SDG No.:	Q2317
Lab Sample ID:	Q2317-02DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042636.D	5		06/16/25 19:12	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	1.10	5.44	UD	5.44	12.4	ug/m3
74-87-3	Chloromethane	0.98	2.02	JD	1.01	5.16	ug/m3
75-01-4	Vinyl Chloride	0.070	0.18	UD	0.18	0.38	ug/m3
74-83-9	Bromomethane	0.70	2.72	UD	2.72	9.71	ug/m3
75-00-3	Chloroethane	0.85	2.24	UD	2.24	6.60	ug/m3
109-99-9	Tetrahydrofuran	1.20	3.54	JD	1.92	7.37	ug/m3
75-69-4	Trichlorofluoromethane	8.60	48.3	D	4.50	14.1	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.80	6.13	UD	6.13	19.2	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.44	3.08	UD	3.08	17.5	ug/m3
75-65-0	tert-Butyl alcohol	0.95	2.88	UD	2.88	7.58	ug/m3
142-82-5	Heptane	0.55	2.25	UD	2.25	10.3	ug/m3
75-35-4	1,1-Dichloroethene	0.70	2.78	UD	2.78	9.91	ug/m3
67-64-1	Acetone	41.2	97.9	D	2.85	5.94	ug/m3
75-15-0	Carbon Disulfide	0.80	2.49	UD	2.49	7.79	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.60	2.16	UD	2.16	9.01	ug/m3
75-09-2	Methylene Chloride	2.10	7.30	JD	3.30	8.69	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.95	3.77	UD	3.77	9.91	ug/m3
75-34-3	1,1-Dichloroethane	0.65	2.63	UD	2.63	10.1	ug/m3
110-82-7	Cyclohexane	1.10	3.79	UD	3.79	8.61	ug/m3
78-93-3	2-Butanone	2.10	6.19	JD	1.77	7.37	ug/m3
56-23-5	Carbon Tetrachloride	0.060	0.38	UD	0.38	0.94	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.45	1.78	UD	1.78	9.91	ug/m3
67-66-3	Chloroform	0.42	2.05	UD	2.05	12.2	ug/m3
71-55-6	1,1,1-Trichloroethane	0.050	0.27	UD	0.27	0.82	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.50	2.34	UD	2.34	11.7	ug/m3
71-43-2	Benzene	0.44	1.41	UD	1.41	7.99	ug/m3
107-06-2	1,2-Dichloroethane	0.46	1.86	UD	1.86	10.1	ug/m3
79-01-6	Trichloroethene	0.090	0.48	UD	0.48	0.81	ug/m3
78-87-5	1,2-Dichloropropane	0.55	2.54	UD	2.54	11.6	ug/m3
75-27-4	Bromodichloromethane	0.39	2.61	UD	2.61	16.8	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.47	1.93	UD	1.93	10.3	ug/m3
108-88-3	Toluene	0.66	2.49	JD	2.07	9.42	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.30	1.36	UD	1.36	11.3	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.32	1.45	UD	1.45	11.3	ug/m3
79-00-5	1,1,2-Trichloroethane	0.36	1.96	UD	1.96	13.6	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-1DL	SDG No.:	Q2317
Lab Sample ID:	Q2317-02DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042636.D	5		06/16/25 19:12	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.33	2.81	UD	2.81	21.3	ug/m3
106-93-4	1,2-Dibromoethane	0.36	2.77	UD	2.77	3.84	ug/m3
127-18-4	Tetrachloroethene	0.080	0.54	UD	0.54	1.02	ug/m3
108-90-7	Chlorobenzene	0.38	1.75	UD	1.75	11.5	ug/m3
100-41-4	Ethyl Benzene	0.60	2.61	UD	2.61	10.9	ug/m3
179601-23-1	m/p-Xylene	1.10	4.78	UD	4.78	21.7	ug/m3
95-47-6	o-Xylene	0.60	2.61	UD	2.61	10.9	ug/m3
100-42-5	Styrene	0.55	2.34	UD	2.34	10.6	ug/m3
75-25-2	Bromoform	0.28	2.90	UD	2.90	25.9	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.060	0.41	UD	0.41	1.03	ug/m3
95-49-8	2-Chlorotoluene	0.55	2.85	UD	2.85	12.9	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.55	2.70	UD	2.70	12.3	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.39	1.92	UD	1.92	12.3	ug/m3
541-73-1	1,3-Dichlorobenzene	0.40	2.40	UD	2.40	15.0	ug/m3
106-46-7	1,4-Dichlorobenzene	16.1	96.8	D	1.62	15.0	ug/m3
95-50-1	1,2-Dichlorobenzene	0.38	2.28	UD	2.28	15.0	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.43	3.19	UD	3.19	18.6	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.43	4.59	UD	4.59	26.7	ug/m3
106-99-0	1,3-Butadiene	0.65	1.44	UD	1.44	5.53	ug/m3
91-20-3	Naphthalene	0.38	1.99	UD	1.99	2.62	ug/m3
622-96-8	4-Ethyltoluene	0.60	2.95	UD	2.95	12.3	ug/m3
110-54-3	Hexane	3.20	11.3	D	1.94	8.81	ug/m3
107-05-1	Allyl Chloride	0.75	2.35	UD	2.35	7.83	ug/m3
123-91-1	1,4-Dioxane	1.10	3.96	UD	3.96	9.01	ug/m3
80-62-6	Methyl Methacrylate	0.43	1.76	UD	1.76	10.2	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.5			70 (65) - 130 (135)	105%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	99600		2.79			
540-36-3	1,4-Difluorobenzene	259000		3.962			
3114-55-4	Chlorobenzene-d5	229000		8.885			

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-1DL	SDG No.:	Q2317
Lab Sample ID:	Q2317-02DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042636.D	5		06/16/25 19:12	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-2	SDG No.:	Q2317
Lab Sample ID:	Q2317-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042629.D	1		06/16/25 15:12	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	2.47		1.04	2.47	ug/m3
74-87-3	Chloromethane	0.70	1.45		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	1.50	4.42		0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	5.40	30.4		0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.37	1.12	J	0.58	1.52	ug/m3
142-82-5	Heptane	0.21	0.86	J	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	32.2	76.5	E	0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.16	0.50	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	0.59	2.05		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	2.30	6.78		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.080	0.50		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.30	1.47	J	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.10	0.47	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.34	1.09	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.090	0.37	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.60	2.26		0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-2	SDG No.:	Q2317
Lab Sample ID:	Q2317-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042629.D	1		06/16/25 15:12	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.040	0.27		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.28	1.22	J	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.12	0.52	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.11	0.47	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.13	0.64	J	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	22.2	133	E	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.96	3.38		0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.2			70 (65) - 130 (135)	102%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	106000		2.79			
540-36-3	1,4-Difluorobenzene	287000		3.962			
3114-55-4	Chlorobenzene-d5	265000		8.885			

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-2	SDG No.:	Q2317
Lab Sample ID:	Q2317-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042629.D	1		06/16/25 15:12	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-2DL	SDG No.:	Q2317
Lab Sample ID:	Q2317-03DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042637.D	5		06/16/25 19:46	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	1.10	5.44	UD	5.44	12.4	ug/m3
74-87-3	Chloromethane	0.81	1.67	JD	1.01	5.16	ug/m3
75-01-4	Vinyl Chloride	0.070	0.18	UD	0.18	0.38	ug/m3
74-83-9	Bromomethane	0.70	2.72	UD	2.72	9.71	ug/m3
75-00-3	Chloroethane	0.85	2.24	UD	2.24	6.60	ug/m3
109-99-9	Tetrahydrofuran	1.40	4.13	JD	1.92	7.37	ug/m3
75-69-4	Trichlorofluoromethane	6.40	36.0	D	4.50	14.1	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.80	6.13	UD	6.13	19.2	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.44	3.08	UD	3.08	17.5	ug/m3
75-65-0	tert-Butyl alcohol	0.95	2.88	UD	2.88	7.58	ug/m3
142-82-5	Heptane	0.55	2.25	UD	2.25	10.3	ug/m3
75-35-4	1,1-Dichloroethene	0.70	2.78	UD	2.78	9.91	ug/m3
67-64-1	Acetone	35.8	85.0	D	2.85	5.94	ug/m3
75-15-0	Carbon Disulfide	0.80	2.49	UD	2.49	7.79	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.60	2.16	UD	2.16	9.01	ug/m3
75-09-2	Methylene Chloride	0.95	3.30	UD	3.30	8.69	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.95	3.77	UD	3.77	9.91	ug/m3
75-34-3	1,1-Dichloroethane	0.65	2.63	UD	2.63	10.1	ug/m3
110-82-7	Cyclohexane	1.10	3.79	UD	3.79	8.61	ug/m3
78-93-3	2-Butanone	2.40	7.08	JD	1.77	7.37	ug/m3
56-23-5	Carbon Tetrachloride	0.060	0.38	UD	0.38	0.94	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.45	1.78	UD	1.78	9.91	ug/m3
67-66-3	Chloroform	0.42	2.05	UD	2.05	12.2	ug/m3
71-55-6	1,1,1-Trichloroethane	0.050	0.27	UD	0.27	0.82	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.50	2.34	UD	2.34	11.7	ug/m3
71-43-2	Benzene	0.44	1.41	UD	1.41	7.99	ug/m3
107-06-2	1,2-Dichloroethane	0.46	1.86	UD	1.86	10.1	ug/m3
79-01-6	Trichloroethene	0.090	0.48	UD	0.48	0.81	ug/m3
78-87-5	1,2-Dichloropropane	0.55	2.54	UD	2.54	11.6	ug/m3
75-27-4	Bromodichloromethane	0.39	2.61	UD	2.61	16.8	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.47	1.93	UD	1.93	10.3	ug/m3
108-88-3	Toluene	0.55	2.07	UD	2.07	9.42	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.30	1.36	UD	1.36	11.3	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.32	1.45	UD	1.45	11.3	ug/m3
79-00-5	1,1,2-Trichloroethane	0.36	1.96	UD	1.96	13.6	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-2DL	SDG No.:	Q2317
Lab Sample ID:	Q2317-03DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042637.D	5		06/16/25 19:46	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.33	2.81	UD	2.81	21.3	ug/m3
106-93-4	1,2-Dibromoethane	0.36	2.77	UD	2.77	3.84	ug/m3
127-18-4	Tetrachloroethene	0.080	0.54	UD	0.54	1.02	ug/m3
108-90-7	Chlorobenzene	0.38	1.75	UD	1.75	11.5	ug/m3
100-41-4	Ethyl Benzene	0.60	2.61	UD	2.61	10.9	ug/m3
179601-23-1	m/p-Xylene	1.10	4.78	UD	4.78	21.7	ug/m3
95-47-6	o-Xylene	0.60	2.61	UD	2.61	10.9	ug/m3
100-42-5	Styrene	0.55	2.34	UD	2.34	10.6	ug/m3
75-25-2	Bromoform	0.28	2.90	UD	2.90	25.9	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.060	0.41	UD	0.41	1.03	ug/m3
95-49-8	2-Chlorotoluene	0.55	2.85	UD	2.85	12.9	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.55	2.70	UD	2.70	12.3	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.39	1.92	UD	1.92	12.3	ug/m3
541-73-1	1,3-Dichlorobenzene	0.40	2.40	UD	2.40	15.0	ug/m3
106-46-7	1,4-Dichlorobenzene	21.5	129	D	1.62	15.0	ug/m3
95-50-1	1,2-Dichlorobenzene	0.38	2.28	UD	2.28	15.0	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.43	3.19	UD	3.19	18.6	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.43	4.59	UD	4.59	26.7	ug/m3
106-99-0	1,3-Butadiene	0.65	1.44	UD	1.44	5.53	ug/m3
91-20-3	Naphthalene	0.38	1.99	UD	1.99	2.62	ug/m3
622-96-8	4-Ethyltoluene	0.60	2.95	UD	2.95	12.3	ug/m3
110-54-3	Hexane	0.76	2.68	JD	1.94	8.81	ug/m3
107-05-1	Allyl Chloride	0.75	2.35	UD	2.35	7.83	ug/m3
123-91-1	1,4-Dioxane	1.10	3.96	UD	3.96	9.01	ug/m3
80-62-6	Methyl Methacrylate	0.43	1.76	UD	1.76	10.2	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.3			70 (65) - 130 (135)	103%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	102000		2.79			
540-36-3	1,4-Difluorobenzene	256000		3.962			
3114-55-4	Chlorobenzene-d5	226000		8.885			

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-IA-2DL	SDG No.:	Q2317
Lab Sample ID:	Q2317-03DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042637.D	5		06/16/25 19:46	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-SG1	SDG No.:	Q2317
Lab Sample ID:	Q2317-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042630.D	4		06/16/25 15:46	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.84	4.15	U	4.15	9.89	ug/m3
74-87-3	Chloromethane	6.50	13.4		0.81	4.13	ug/m3
75-01-4	Vinyl Chloride	0.060	0.15	U	0.15	0.31	ug/m3
74-83-9	Bromomethane	0.56	2.17	U	2.17	7.77	ug/m3
75-00-3	Chloroethane	2.30	6.07		1.79	5.28	ug/m3
109-99-9	Tetrahydrofuran	2.10	6.19		1.53	5.90	ug/m3
75-69-4	Trichlorofluoromethane	42.3	238		3.60	11.2	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.64	4.91	U	4.91	15.3	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.35	2.45	U	2.45	14.0	ug/m3
75-65-0	tert-Butyl alcohol	0.76	2.30	U	2.30	6.06	ug/m3
142-82-5	Heptane	0.78	3.20	J	1.80	8.20	ug/m3
75-35-4	1,1-Dichloroethene	0.56	2.22	U	2.22	7.93	ug/m3
67-64-1	Acetone	21.9	52.0		2.28	4.75	ug/m3
75-15-0	Carbon Disulfide	2.30	7.16		1.99	6.23	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.48	1.73	U	1.73	7.21	ug/m3
75-09-2	Methylene Chloride	8.10	28.1		2.64	6.95	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.76	3.01	U	3.01	7.93	ug/m3
75-34-3	1,1-Dichloroethane	0.52	2.10	U	2.10	8.09	ug/m3
110-82-7	Cyclohexane	0.88	3.03	U	3.03	6.88	ug/m3
78-93-3	2-Butanone	4.20	12.4		1.42	5.90	ug/m3
56-23-5	Carbon Tetrachloride	0.040	0.25	U	0.25	0.75	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.36	1.43	U	1.43	7.93	ug/m3
67-66-3	Chloroform	0.77	3.76	J	1.61	9.77	ug/m3
71-55-6	1,1,1-Trichloroethane	0.040	0.22	U	0.22	0.65	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.40	1.87	U	1.87	9.34	ug/m3
71-43-2	Benzene	8.90	28.4		1.12	6.39	ug/m3
107-06-2	1,2-Dichloroethane	0.36	1.46	U	1.46	8.09	ug/m3
79-01-6	Trichloroethene	4.90	26.3		0.38	0.64	ug/m3
78-87-5	1,2-Dichloropropane	0.44	2.03	U	2.03	9.24	ug/m3
75-27-4	Bromodichloromethane	0.31	2.08	U	2.08	13.4	ug/m3
108-10-1	4-Methyl-2-Pentanone	1.30	5.33	J	1.52	8.20	ug/m3
108-88-3	Toluene	11.5	43.3		1.66	7.54	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.24	1.09	U	1.09	9.08	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.26	1.18	U	1.18	9.08	ug/m3
79-00-5	1,1,2-Trichloroethane	0.28	1.53	U	1.53	10.9	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-SG1	SDG No.:	Q2317
Lab Sample ID:	Q2317-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042630.D	4		06/16/25 15:46	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.26	2.22	U	2.22	17.0	ug/m3
106-93-4	1,2-Dibromoethane	0.29	2.23	U	2.23	3.07	ug/m3
127-18-4	Tetrachloroethene	0.36	2.44		0.41	0.81	ug/m3
108-90-7	Chlorobenzene	0.30	1.38	U	1.38	9.21	ug/m3
100-41-4	Ethyl Benzene	0.48	2.08	U	2.08	8.69	ug/m3
179601-23-1	m/p-Xylene	0.84	3.65	U	3.65	17.4	ug/m3
95-47-6	o-Xylene	0.48	2.08	U	2.08	8.69	ug/m3
100-42-5	Styrene	0.49	2.09	J	1.87	8.52	ug/m3
75-25-2	Bromoform	0.23	2.38	U	2.38	20.7	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.040	0.27	U	0.27	0.82	ug/m3
95-49-8	2-Chlorotoluene	0.44	2.28	U	2.28	10.4	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.44	2.16	U	2.16	9.83	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.31	1.52	U	1.52	9.83	ug/m3
541-73-1	1,3-Dichlorobenzene	0.32	1.92	U	1.92	12.0	ug/m3
106-46-7	1,4-Dichlorobenzene	0.56	3.37	J	1.26	12.0	ug/m3
95-50-1	1,2-Dichlorobenzene	0.30	1.80	U	1.80	12.0	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.34	2.52	U	2.52	14.8	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.34	3.63	U	3.63	21.3	ug/m3
106-99-0	1,3-Butadiene	0.52	1.15	U	1.15	4.42	ug/m3
91-20-3	Naphthalene	0.30	1.57	U	1.57	2.10	ug/m3
622-96-8	4-Ethyltoluene	0.48	2.36	U	2.36	9.83	ug/m3
110-54-3	Hexane	2.80	9.87		1.55	7.05	ug/m3
107-05-1	Allyl Chloride	0.60	1.88	U	1.88	6.26	ug/m3
123-91-1	1,4-Dioxane	0.84	3.03	U	3.03	7.21	ug/m3
80-62-6	Methyl Methacrylate	25.5	104		1.39	8.19	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.90			70 (65) - 130 (135)	99%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	111000		2.787			
540-36-3	1,4-Difluorobenzene	292000		3.958			
3114-55-4	Chlorobenzene-d5	267000		8.885			

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-SG1	SDG No.:	Q2317
Lab Sample ID:	Q2317-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042630.D	4		06/16/25 15:46	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-SG2	SDG No.:	Q2317
Lab Sample ID:	Q2317-05	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042632.D	4		06/16/25 16:53	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.84	4.15	U	4.15	9.89	ug/m3
74-87-3	Chloromethane	8.90	18.4		0.81	4.13	ug/m3
75-01-4	Vinyl Chloride	0.060	0.15	U	0.15	0.31	ug/m3
74-83-9	Bromomethane	0.56	2.17	U	2.17	7.77	ug/m3
75-00-3	Chloroethane	3.30	8.71		1.79	5.28	ug/m3
109-99-9	Tetrahydrofuran	2.00	5.90		1.53	5.90	ug/m3
75-69-4	Trichlorofluoromethane	14.2	79.8		3.60	11.2	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.64	4.91	U	4.91	15.3	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.35	2.45	U	2.45	14.0	ug/m3
75-65-0	tert-Butyl alcohol	0.76	2.30	U	2.30	6.06	ug/m3
142-82-5	Heptane	0.44	1.80	U	1.80	8.20	ug/m3
75-35-4	1,1-Dichloroethene	0.56	2.22	U	2.22	7.93	ug/m3
67-64-1	Acetone	52.7	125		2.28	4.75	ug/m3
75-15-0	Carbon Disulfide	0.64	1.99	U	1.99	6.23	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.48	1.73	U	1.73	7.21	ug/m3
75-09-2	Methylene Chloride	2.80	9.73		2.64	6.95	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.76	3.01	U	3.01	7.93	ug/m3
75-34-3	1,1-Dichloroethane	0.52	2.10	U	2.10	8.09	ug/m3
110-82-7	Cyclohexane	0.88	3.03	U	3.03	6.88	ug/m3
78-93-3	2-Butanone	3.90	11.5		1.42	5.90	ug/m3
56-23-5	Carbon Tetrachloride	0.040	0.25	U	0.25	0.75	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.36	1.43	U	1.43	7.93	ug/m3
67-66-3	Chloroform	8.30	40.5		1.61	9.77	ug/m3
71-55-6	1,1,1-Trichloroethane	0.040	0.22	U	0.22	0.65	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.40	1.87	U	1.87	9.34	ug/m3
71-43-2	Benzene	3.30	10.5		1.12	6.39	ug/m3
107-06-2	1,2-Dichloroethane	0.36	1.46	U	1.46	8.09	ug/m3
79-01-6	Trichloroethene	12.9	69.3		0.38	0.64	ug/m3
78-87-5	1,2-Dichloropropane	0.44	2.03	U	2.03	9.24	ug/m3
75-27-4	Bromodichloromethane	0.31	2.08	U	2.08	13.4	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.37	1.52	U	1.52	8.20	ug/m3
108-88-3	Toluene	0.51	1.92	J	1.66	7.54	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.24	1.09	U	1.09	9.08	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.26	1.18	U	1.18	9.08	ug/m3
79-00-5	1,1,2-Trichloroethane	0.28	1.53	U	1.53	10.9	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-SG2	SDG No.:	Q2317
Lab Sample ID:	Q2317-05	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042632.D	4		06/16/25 16:53	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.26	2.22	U	2.22	17.0	ug/m3
106-93-4	1,2-Dibromoethane	0.29	2.23	U	2.23	3.07	ug/m3
127-18-4	Tetrachloroethene	12.3	83.4		0.41	0.81	ug/m3
108-90-7	Chlorobenzene	0.30	1.38	U	1.38	9.21	ug/m3
100-41-4	Ethyl Benzene	0.48	2.08	U	2.08	8.69	ug/m3
179601-23-1	m/p-Xylene	0.84	3.65	U	3.65	17.4	ug/m3
95-47-6	o-Xylene	0.48	2.08	U	2.08	8.69	ug/m3
100-42-5	Styrene	0.44	1.87	U	1.87	8.52	ug/m3
75-25-2	Bromoform	0.23	2.38	U	2.38	20.7	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.040	0.27	U	0.27	0.82	ug/m3
95-49-8	2-Chlorotoluene	0.44	2.28	U	2.28	10.4	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.44	2.16	U	2.16	9.83	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.31	1.52	U	1.52	9.83	ug/m3
541-73-1	1,3-Dichlorobenzene	0.32	1.92	U	1.92	12.0	ug/m3
106-46-7	1,4-Dichlorobenzene	5.00	30.1		1.26	12.0	ug/m3
95-50-1	1,2-Dichlorobenzene	0.30	1.80	U	1.80	12.0	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.34	2.52	U	2.52	14.8	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.34	3.63	U	3.63	21.3	ug/m3
106-99-0	1,3-Butadiene	0.52	1.15	U	1.15	4.42	ug/m3
91-20-3	Naphthalene	0.30	1.57	U	1.57	2.10	ug/m3
622-96-8	4-Ethyltoluene	0.48	2.36	U	2.36	9.83	ug/m3
110-54-3	Hexane	0.88	3.10	J	1.55	7.05	ug/m3
107-05-1	Allyl Chloride	0.60	1.88	U	1.88	6.26	ug/m3
123-91-1	1,4-Dioxane	0.84	3.03	U	3.03	7.21	ug/m3
80-62-6	Methyl Methacrylate	17.8	72.9		1.39	8.19	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.4			70 (65) - 130 (135)	104%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	104000		2.79			
540-36-3	1,4-Difluorobenzene	284000		3.965			
3114-55-4	Chlorobenzene-d5	251000		8.888			

Report of Analysis

Client:	RTP Environmental	Date Collected:	06/12/25
Project:	Yellow Cab - TO-15	Date Received:	06/12/25
Client Sample ID:	115-SG2	SDG No.:	Q2317
Lab Sample ID:	Q2317-05	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042632.D	4		06/16/25 16:53	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements



QC SUMMARY

Surrogate Summary

SDG No.: Q2317

Client: RTP Environmental

Analytical Method: SWTO-15

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2317-01	AA-3	1-Bromo-4-Fluorobenzene	10	10.3	103	70 (65)	130 (135)
Q2317-01DUP	AA-3DUP	1-Bromo-4-Fluorobenzene	10	10.3	103	70 (65)	130 (135)
Q2317-02	115-IA-1	1-Bromo-4-Fluorobenzene	10	10.4	104	70 (65)	130 (135)
Q2317-02DL	115-IA-1DL	1-Bromo-4-Fluorobenzene	10	10.5	105	70 (65)	130 (135)
Q2317-03	115-IA-2	1-Bromo-4-Fluorobenzene	10	10.2	102	70 (65)	130 (135)
Q2317-03DL	115-IA-2DL	1-Bromo-4-Fluorobenzene	10	10.3	103	70 (65)	130 (135)
Q2317-04	115-SG1	1-Bromo-4-Fluorobenzene	10	9.86	99	70 (65)	130 (135)
Q2317-05	115-SG2	1-Bromo-4-Fluorobenzene	10	10.4	104	70 (65)	130 (135)
VL0616ABL01	VL0616ABL01	1-Bromo-4-Fluorobenzene	10	10.2	102	70 (65)	130 (135)
VL0616ABS01	VL0616ABS01	1-Bromo-4-Fluorobenzene	10	10.2	102	70 (65)	130 (135)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2317

Client: RTP Environmental

Analytical Method: SWTO-15

Datafile : VL042624.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0616ABS01	Dichlorodifluoromethane	10	10.4	ppbv	104			40 (70)	160 (130)	
	Chloromethane	10	9.80	ppbv	98			40 (70)	160 (130)	
	Vinyl Chloride	10	9.10	ppbv	91			70 (70)	130 (130)	
	Bromomethane	10	9.30	ppbv	93			40 (70)	160 (130)	
	Chloroethane	10	9.70	ppbv	97			40 (70)	160 (130)	
	Tetrahydrofuran	10	10.7	ppbv	107			70 (70)	130 (130)	
	Trichlorofluoromethane	10	11.4	ppbv	114			40 (70)	160 (130)	
	1,1,2-Trichlorotrifluoroethane	10	11.4	ppbv	114			70 (70)	130 (130)	
	Dichlorotetrafluoroethane	10	10.2	ppbv	102			70 (70)	130 (130)	
	tert-Butyl Alcohol	10	10.2	ppbv	102			70 (70)	130 (130)	
	Heptane	10	9.60	ppbv	96			70 (70)	130 (130)	
	1,1-Dichloroethene	10	10.3	ppbv	103			70 (70)	130 (130)	
	Acetone	10	8.90	ppbv	89			40 (70)	160 (130)	
	Carbon disulfide	10	10.8	ppbv	108			40 (70)	160 (130)	
	Methyl tert-butyl Ether	10	10.2	ppbv	102			70 (70)	130 (130)	
	Methylene Chloride	10	9.30	ppbv	93			70 (70)	130 (130)	
	trans-1,2-Dichloroethene	10	10.6	ppbv	106			70 (70)	130 (130)	
	1,1-Dichloroethane	10	10.6	ppbv	106			70 (70)	130 (130)	
	Cyclohexane	10	9.20	ppbv	92			70 (70)	130 (130)	
	2-Butanone	10	11.1	ppbv	111			40 (70)	160 (130)	
	Carbon Tetrachloride	10	11.1	ppbv	111			70 (70)	130 (130)	
	cis-1,2-Dichloroethene	10	10.0	ppbv	100			70 (70)	130 (130)	
	Chloroform	10	10.1	ppbv	101			70 (70)	130 (130)	
	1,1,1-Trichloroethane	10	9.70	ppbv	97			70 (70)	130 (130)	
	2,2,4-Trimethylpentane	10	10.7	ppbv	107			70 (70)	130 (130)	
	Benzene	10	10.8	ppbv	108			70 (70)	130 (130)	
	1,2-Dichloroethane	10	11.7	ppbv	117			70 (70)	130 (130)	
	Trichloroethene	10	10.2	ppbv	102			70 (70)	130 (130)	
	1,2-Dichloropropane	10	10.6	ppbv	106			70 (70)	130 (130)	
	Bromodichloromethane	10	11.7	ppbv	117			70 (70)	130 (130)	
	4-Methyl-2-Pentanone	10	10.9	ppbv	109			40 (70)	160 (130)	
	Toluene	10	10.3	ppbv	103			70 (70)	130 (130)	
	t-1,3-Dichloropropene	10	10.5	ppbv	105			70 (70)	130 (130)	
	cis-1,3-Dichloropropene	10	10.6	ppbv	106			70 (70)	130 (130)	
	1,1,2-Trichloroethane	10	10.8	ppbv	108			70 (70)	130 (130)	
	Dibromochloromethane	10	10.8	ppbv	108			70 (70)	130 (130)	
	1,2-Dibromoethane	10	10.6	ppbv	106			70 (70)	130 (130)	
	Tetrachloroethene	10	9.70	ppbv	97			70 (70)	130 (130)	
	Chlorobenzene	10	10.5	ppbv	105			70 (70)	130 (130)	
	Ethyl Benzene	10	10.6	ppbv	106			70 (70)	130 (130)	
	m/p-Xylene	20	21.6	ppbv	108			70 (70)	130 (130)	
	o-Xylene	10	10.7	ppbv	107			70 (70)	130 (130)	
	Styrene	10	10.4	ppbv	104			70 (70)	130 (130)	
	Bromoform	10	10.8	ppbv	108			70 (70)	130 (130)	
	1,1,2,2-Tetrachloroethane	10	10.7	ppbv	107			70 (70)	130 (130)	
	2-Chlorotoluene	10	10.5	ppbv	105			70 (70)	130 (130)	
	1,3,5-Trimethylbenzene	10	10.6	ppbv	106			70 (70)	130 (130)	
	1,2,4-Trimethylbenzene	10	10.4	ppbv	104			70 (70)	130 (130)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2317

Client: RTP Environmental

Analytical Method: SWTO-15 **Datafile :** VL042624.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0616ABS01	1,3-Dichlorobenzene	10	10.4	ppbv	104			70 (70)	130 (130)	
	1,4-Dichlorobenzene	10	10.5	ppbv	105			70 (70)	130 (130)	
	1,2-Dichlorobenzene	10	10.4	ppbv	104			70 (70)	130 (130)	
	1,2,4-Trichlorobenzene	10	10.1	ppbv	101			70 (70)	130 (130)	
	Hexachloro-1,3-butadiene	10	9.10	ppbv	91			70 (70)	130 (130)	
	Naphthalene	10	10.8	ppbv	108			40 (70)	160 (130)	
	1,3-Butadiene	10	9.30	ppbv	93			70 (70)	130 (130)	
	4-Ethyltoluene	10	10.5	ppbv	105			70 (70)	130 (130)	
	Hexane	10	9.30	ppbv	93			70 (70)	130 (130)	
	Allyl Chloride	10	9.90	ppbv	99			70 (70)	130 (130)	
	1,4-Dioxane	10	10.3	ppbv	103			40 (70)	160 (130)	
	Methyl methacrylate	10	10.5	ppbv	105			70 (70)	130 (130)	

() = LABORATORY INHOUSE LIMIT

Duplicate Sample Summary

Lab Sample Id :	Q2317-01DUP	Q2317-01
Client Id :	AA-3DUP	AA-3
DF :	1	1
Datafile :	VL042627.D	VL042626.D
Anal Date & Time :	06/16/2025 14:01	06/16/2025 13:27

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0.17	0.17	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0.18	0.19	5.4
2-Butanone	0.61	0.65	6.3
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0.18	0.18	0
Acetone	10.6	10.2	3.8
Allyl Chloride	0	0	0
Benzene	0.31	0.32	3.2
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.08	0.08	0

Duplicate Sample Summary

Lab Sample Id :	Q2317-01DUP	Q2317-01
Client Id :	AA-3DUP	AA-3
DF :	1	1
Datafile :	VL042627.D	VL042626.D
Anal Date & Time :	06/16/2025 14:01	06/16/2025 13:27

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.54	0.54	0
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.49	0.48	2.1
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0.11	0	200 *
Heptane	0.15	0.15	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	4.9	4.8	2.1
m/p-Xylene	0.51	0.48	6.1
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	3.3	3	9.5
Naphthalene	0	0	0
o-Xylene	0.24	0.23	4.3
Styrene	0.14	0.13	7.4
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0	0	0
Tetrahydrofuran	0	0	0
Toluene	0.48	0.46	4.3
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.28	0.27	3.6
Vinyl Chloride	0	0	0

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VL0616ABL01

Lab Name: CHEMTECH

Contract: RTPE01

Lab Code: CHEM Case No.: Q2317

SAS No.: Q2317 SDG NO.: Q2317

Lab File ID: VL042623.D

Lab Sample ID: VL0616ABL01

Date Analyzed: 06/16/2025

Time Analyzed: 09:43

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_L

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VL0616ABS01	VL0616ABS01	VL042624.D	06/16/2025
AA-3	Q2317-01	VL042626.D	06/16/2025
AA-3DUP	Q2317-01DUP	VL042627.D	06/16/2025
115-IA-1	Q2317-02	VL042628.D	06/16/2025
115-IA-2	Q2317-03	VL042629.D	06/16/2025
115-SG1	Q2317-04	VL042630.D	06/16/2025
115-SG2	Q2317-05	VL042632.D	06/16/2025
115-IA-1DL	Q2317-02DL	VL042636.D	06/16/2025
115-IA-2DL	Q2317-03DL	VL042637.D	06/16/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RTPE01
Lab Code: CHEM Case No.: Q2317 SAS No.: Q2317 SDG NO.: Q2317
Lab File ID: VL042578.D BFB Injection Date: 05/29/2025
Instrument ID: MSVOA_L BFB Injection Time: 08:05
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	22.4
75	30.0 - 66.0% of mass 95	53.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 120.0% of mass 95	67.2
175	4.0 - 9.0% of mass 174	5.2 (7.7) 1
176	93.0 - 101.0% of mass 174	63.6 (94.6) 1
177	5.0 - 9.0% of mass 176	4 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICCC010	VSTDICCC010	VL042579.D	05/29/2025	11:09
VSTDICCC002	VSTDICCC002	VL042580.D	05/29/2025	11:43
VSTDICCC001	VSTDICCC001	VL042581.D	05/29/2025	12:15
VSTDICCC0.5	VSTDICCC0.5	VL042582.D	05/29/2025	12:47
VSTDICCC0.1	VSTDICCC0.1	VL042583.D	05/29/2025	13:19
VSTDICCC0.03	VSTDICCC0.03	VL042584.D	05/29/2025	13:51
VSTDICCC015	VSTDICCC015	VL042585.D	05/29/2025	14:24

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RTPE01
Lab Code: CHEM Case No.: Q2317 SAS No.: Q2317 SDG NO.: Q2317
Lab File ID: VL042621.D BFB Injection Date: 06/16/2025
Instrument ID: MSVOA_L BFB Injection Time: 08:07
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	24.4
75	30.0 - 66.0% of mass 95	55.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.6 (0.9) 1
174	50.0 - 120.0% of mass 95	62.6
175	4.0 - 9.0% of mass 174	4.7 (7.6) 1
176	93.0 - 101.0% of mass 174	60.6 (96.9) 1
177	5.0 - 9.0% of mass 176	4.1 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC010	VSTDCCC010	VL042622.D	06/16/2025	08:54
VL0616ABL01	VL0616ABL01	VL042623.D	06/16/2025	09:43
VL0616ABS01	VL0616ABS01	VL042624.D	06/16/2025	10:29
AA-3	Q2317-01	VL042626.D	06/16/2025	13:27
AA-3DUP	Q2317-01DUP	VL042627.D	06/16/2025	14:01
115-IA-1	Q2317-02	VL042628.D	06/16/2025	14:35
115-IA-2	Q2317-03	VL042629.D	06/16/2025	15:12
115-SG1	Q2317-04	VL042630.D	06/16/2025	15:46
115-SG2	Q2317-05	VL042632.D	06/16/2025	16:53
115-IA-1DL	Q2317-02DL	VL042636.D	06/16/2025	19:12
115-IA-2DL	Q2317-03DL	VL042637.D	06/16/2025	19:46

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RTPE01
Lab Code: CHEM Case No.: Q2317 SAS No.: Q2317 SDG NO.: Q2317
Lab File ID: VL042622.D Date Analyzed: 06/16/2025
Instrument ID: MSVOA_L Time Analyzed: 08:54
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	107738	2.79	288897	3.96	262507	8.88
UPPER LIMIT	150833	3.12	404456	4.29	367510	9.21
LOWER LIMIT	64642.8	2.46	173338	3.63	157504	8.55
EPA SAMPLE NO.						
AA-3	109400	2.79	291899	3.96	254916	8.89
AA-3DUP	104604	2.78	283556	3.96	252613	8.88
115-IA-1	106724	2.79	298270	3.96	259856	8.88
115-IA-1DL	99585	2.79	258626	3.96	228826	8.89
115-IA-2	106087	2.79	286951	3.96	264656	8.89
115-IA-2DL	101539	2.79	255635	3.96	225996	8.89
115-SG1	111111	2.79	291956	3.96	266931	8.89
115-SG2	104277	2.79	283950	3.97	250604	8.89
VL0616ABL01	108213	2.79	292209	3.96	252864	8.89
VL0616ABS01	107046	2.79	286719	3.96	259461	8.88

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0616ABL01	SDG No.:	Q2317
Lab Sample ID:	VL0616ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042623.D	1		06/16/25 09:43	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.21	1.04	U	1.04	2.47	ug/m3
74-87-3	Chloromethane	0.10	0.21	U	0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.13	0.38	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.16	0.90	U	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.19	0.58	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.11	0.45	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	0.24	0.57	U	0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.16	0.50	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	0.19	0.66	U	0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.12	0.35	U	0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.010	0.060	U	0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.080	0.39	U	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.10	0.47	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.090	0.29	U	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.090	0.37	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.11	0.41	U	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0616ABL01	SDG No.:	Q2317
Lab Sample ID:	VL0616ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042623.D	1		06/16/25 09:43	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.020	0.14	U	0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.21	0.91	U	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.12	0.52	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.11	0.47	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.080	0.39	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.050	0.30	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.11	0.39	U	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.2			70 (65) - 130 (135)	102%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	108000		2.79			
540-36-3	1,4-Difluorobenzene	292000		3.962			
3114-55-4	Chlorobenzene-d5	253000		8.885			

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0616ABL01	SDG No.:	Q2317
Lab Sample ID:	VL0616ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042623.D	1		06/16/25 09:43	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0616ABS01	SDG No.:	Q2317
Lab Sample ID:	VL0616ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042624.D	1		06/16/25 10:29	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	10.4	51.4		1.04	2.47	ug/m3
74-87-3	Chloromethane	9.80	20.2		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	9.10	23.3		0.030	0.080	ug/m3
74-83-9	Bromomethane	9.30	36.1		0.54	1.94	ug/m3
75-00-3	Chloroethane	9.70	25.6		0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	10.7	31.6		0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	11.4	64.1		0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	11.4	87.4		1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	10.2	71.3		0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	10.2	30.9		0.58	1.52	ug/m3
142-82-5	Heptane	9.60	39.3		0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	10.3	40.8		0.56	1.98	ug/m3
67-64-1	Acetone	8.90	21.1		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	10.8	33.6		0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	10.2	36.8		0.43	1.80	ug/m3
75-09-2	Methylene Chloride	9.30	32.3		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	10.6	42.0		0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	10.6	42.9		0.53	2.02	ug/m3
110-82-7	Cyclohexane	9.20	31.7		0.76	1.72	ug/m3
78-93-3	2-Butanone	11.1	32.7		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	11.1	69.8		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	10.0	39.6		0.36	1.98	ug/m3
67-66-3	Chloroform	10.1	49.3		0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	9.70	52.9		0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	10.7	50.0		0.47	2.34	ug/m3
71-43-2	Benzene	10.8	34.5		0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	11.7	47.4		0.36	2.02	ug/m3
79-01-6	Trichloroethene	10.2	54.8		0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	10.6	49.0		0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	11.7	78.4		0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	10.9	44.7		0.37	2.05	ug/m3
108-88-3	Toluene	10.3	38.8		0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	10.5	47.7		0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	10.6	48.1		0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	10.8	58.9		0.38	2.73	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0616ABS01	SDG No.:	Q2317
Lab Sample ID:	VL0616ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042624.D	1		06/16/25 10:29	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	10.8	92.0		0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	10.6	81.5		0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	9.70	65.8		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	10.5	48.4		0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	10.6	46.0		0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	21.6	93.8		0.91	4.34	ug/m3
95-47-6	o-Xylene	10.7	46.5		0.52	2.17	ug/m3
100-42-5	Styrene	10.4	44.3		0.47	2.13	ug/m3
75-25-2	Bromoform	10.8	112		0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	10.7	73.5		0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	10.5	54.4		0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	10.6	52.1		0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	10.4	51.1		0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	10.4	62.5		0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	10.5	63.1		0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	10.4	62.5		0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	10.1	75.0		0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	9.10	97.1		0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	9.30	20.6		0.29	1.11	ug/m3
91-20-3	Naphthalene	10.8	56.6		0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	10.5	51.6		0.59	2.46	ug/m3
110-54-3	Hexane	9.30	32.8		0.39	1.76	ug/m3
107-05-1	Allyl Chloride	9.90	31.0		0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	10.3	37.1		0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	10.5	43.0		0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.2			70 (65) - 130 (135)	102%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	107000		2.787			
540-36-3	1,4-Difluorobenzene	287000		3.955			
3114-55-4	Chlorobenzene-d5	259000		8.882			

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0616ABS01	SDG No.:	Q2317
Lab Sample ID:	VL0616ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042624.D	1		06/16/25 10:29	VL061625

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements



CALIBRATION SUMMARY

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL052925AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Fri May 30 02:01:56 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL042584.D 0.1 =VL042583.D 0.5 =VL042582.D 1 =VL042581.D 2 =VL042580.D 10 =VL042579.D 15 =VL042585.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.257	1.353	1.300	1.135	1.101	1.229	8.78
3) Chlorodifluoro...			1.881	1.861	1.826	1.475	1.420	1.693	13.32
4) Chloromethane			0.502	0.529	0.507	0.461	0.434	0.487	7.85
5) T Vinyl Chloride	0.675	0.465	0.529	0.487	0.468	0.437	0.409	0.496	17.69
6) T Bromomethane			0.310	0.251	0.230	0.216	0.200	0.241	17.69
7) Chloroethane			0.214	0.212	0.198	0.175	0.164	0.193	11.60
8) T Dichlorotetra...			1.099	1.098	1.075	0.948	0.900	1.024	9.11
9) T Propene			0.710	0.728	0.722	0.650	0.599	0.682	8.18
10) T Heptane			2.079	2.082	1.981	1.777	1.688	1.921	9.37
11) T Trichlorofluor...			1.193	1.209	1.202	1.071	1.043	1.144	6.98
12) T 1,1,2-Trichlor...			0.903	0.980	0.951	0.838	0.818	0.898	7.79
13) Ethanol			0.114	0.095	0.092	0.058	0.047	0.081	33.96
14) T Bromoethene			0.359	0.346	0.362	0.317	0.304	0.338	7.71
15) T Acetone			1.404	1.484	1.299	0.928	0.908	1.205	22.39
16) T 1,3-Butadiene			0.649	0.636	0.597	0.491	0.480	0.571	14.07
17) tert-Butyl alc...			1.337	1.334	1.279	1.237	1.088	1.255	8.14
18) T 1,1-Dichloroet...			0.451	0.451	0.437	0.377	0.364	0.416	10.11
19) T Isopropyl Alcohol			0.825	0.782	0.742	0.659	0.593	0.720	13.03
20) T Methylene Chlo...			0.473	0.457	0.418	0.326	0.314	0.398	18.57
21) T Allyl Chloride			0.924	0.921	0.866	0.789	0.749	0.850	9.22
22) T trans-1,2-Dich...			0.496	0.517	0.466	0.427	0.418	0.465	9.21
23) T Vinyl Acetate			1.750	1.663	1.686	1.458	1.584	1.628	6.89
24) T 1,1-Dichloroet...			0.996	1.046	1.004	0.903	0.833	0.957	9.03
25) T Ethyl Acetate			3.496	3.585	3.412	3.054	2.934	3.296	8.67
26) T Hexane			1.690	1.638	1.526	1.350	1.286	1.498	11.76
27) T Carbon Disulfide			1.174	1.267	1.198	1.069	1.040	1.150	8.16
28) T Methyl tert-Bu...			0.664	0.693	0.621	0.550	0.511	0.608	12.59
29) T Chloroform			2.099	2.120	2.034	1.755	1.731	1.948	9.74
30) T Cyclohexane			1.318	1.363	1.288	1.146	1.092	1.242	9.38
31) T cis-1,2-Dichlo...			1.419	1.461	1.465	1.271	1.235	1.370	7.96
32) T 1,1,1-Trichlor...	2.207	2.083	2.004	2.043	2.029	1.794	1.771	1.990	7.86

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.760	0.754	0.746	0.655	0.651	0.713	7.71
35) T Carbon Tetrach...	0.681	0.706	0.665	0.684	0.670	0.603	0.612	0.660	5.82
36) T Benzene			1.051	1.052	1.027	0.900	0.894	0.985	8.20
37) T 1,2-Dichloroet...			0.518	0.509	0.513	0.448	0.460	0.489	6.72
38) T Trichloroethene	0.389	0.432	0.468	0.455	0.438	0.376	0.373	0.419	9.28
39) T 1,2-Dichloropr...			0.415	0.374	0.373	0.329	0.330	0.364	9.85

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\										
Method File : VL052925AIR.M										
40) T	1,4-Dioxane			0.172	0.178	0.168	0.152	0.145	0.163	8.70
41) T	Tetrahydrofuran			0.445	0.446	0.426	0.388	0.382	0.417	7.33
42) T	Bromodichlorom...			0.702	0.737	0.721	0.656	0.660	0.695	5.17
43) T	Methyl Methacr...			0.398	0.384	0.406	0.366	0.357	0.382	5.36
44) T	2,2,4-Trimethy...			1.811	1.830	1.765	1.522	1.495	1.684	9.67
45) T	t-1,3-Dichloro...			0.433	0.455	0.456	0.424	0.418	0.437	4.01
46) T	cis-1,3-Dichlo...			0.567	0.568	0.555	0.527	0.523	0.548	3.99
47) T	1,1,2-Trichlor...			0.406	0.401	0.394	0.343	0.340	0.377	8.69
48) T	Dibromochlorom...			0.600	0.642	0.635	0.570	0.571	0.603	5.64
49) T	Bromoform			0.524	0.519	0.553	0.508	0.501	0.521	3.83
50) T	4-Methyl-2-Pen...			1.036	1.030	1.002	0.958	0.919	0.989	5.05
51) T	2-Hexanone			0.752	0.818	0.824	0.779	0.749	0.784	4.49
52) T	Tetrachloroethene	0.351	0.407	0.391	0.383	0.390	0.333	0.321	0.368	8.93
53) T	Toluene			1.210	1.190	1.197	1.054	1.040	1.138	7.35
54) T	1,2-Dibromoethane		0.575	0.578	0.590	0.576	0.516	0.513	0.558	6.15
-----ISTD-----										
55) I	Chlorobenzene-d5									
56) T	1,1,1,2-Tetrac...			0.520	0.536	0.528	0.467	0.464	0.503	6.95
57) T	Chlorobenzene			1.081	1.028	1.014	0.868	0.865	0.971	10.17
58) T	Ethyl Benzene			1.770	1.786	1.833	1.547	1.545	1.696	8.19
59) T	m/p-Xylene			1.411	1.412	1.431	1.212	1.208	1.335	8.55
60) T	o-Xylene			1.430	1.410	1.441	1.187	1.188	1.331	9.87
61) T	Styrene			0.626	0.658	0.678	0.590	0.596	0.629	6.07
62) T	Isopropylbenzene			2.060	2.129	2.121	1.772	1.751	1.967	9.62
63) T	1,1,2,2-Tetrac...	1.054	0.828	0.879	0.861	0.855	0.723	0.727	0.847	13.16
64) T	n-propylbenzene			0.579	0.541	0.570	0.471	0.475	0.527	9.74
65) T	tert-Butylbenzene			1.945	1.923	1.942	1.577	1.538	1.785	11.68
66) T	Benzyl Chloride			0.199	0.237	0.252	0.241	0.220	0.230	8.94
67) T	sec-Butylbenzene			2.745	2.743	2.699	2.195	2.165	2.509	12.01
68) S	1-Bromo-4-Fluo...	0.752	0.751	0.755	0.751	0.741	0.744	0.754	0.750	0.68
69) T	p-Isopropyltol...			2.241	2.249	2.299	1.868	1.851	2.101	10.57
70) T	n-Butylbenzene			2.110	2.267	2.299	1.842	1.853	2.074	10.56
71) T	2-Chlorotoluene			1.603	1.620	1.596	1.343	1.339	1.500	9.70
72) T	4-Ethyltoluene			1.704	1.797	1.823	1.481	1.494	1.660	9.86
73) T	1,3,5-Trimethy...			1.441	1.482	1.475	1.235	1.236	1.374	9.25
74) T	1,2,4-Trimethy...			1.640	1.675	1.666	1.335	1.306	1.524	12.26
75) T	1,3-Dichlorobe...			1.031	1.034	1.040	0.839	0.843	0.957	11.13
76) T	1,4-Dichlorobe...			0.998	1.040	1.037	0.849	0.851	0.955	10.18
77) T	1,2-Dichlorobe...			1.019	1.008	1.027	0.798	0.793	0.929	13.14
78) T	Hexachloro-1,3...			0.872	0.893	0.880	0.607	0.593	0.769	20.11
79) T	Naphthalene		1.088	1.366	1.634	1.785	1.383	1.366	1.437	16.89
80) T	Naphthalene,2-...			0.444	0.562	0.781	0.716	0.743	0.649	21.85
81) T	1,2,4-Trichlor...			0.716	0.889	0.944	0.735	0.727	0.802	13.26

(#) = Out of Range

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RTPE01

Lab Code: CHEM Case No.: Q2317 SAS No.: Q2317 SDG No.: Q2317

Instrument ID: MSVOA_L Calibration Date/Time: 06/16/2025 08:54

Lab File ID: VL042622.D Init. Calib. Date(s): 05/29/2025 05/29/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:24

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	1.229	1.252		1.87	30
Chloromethane	0.487	0.473		-2.88	30
Vinyl Chloride	0.496	0.445		-10.28	30
Bromomethane	0.241	0.218		-9.54	30
Chloroethane	0.193	0.181		-6.22	30
Tetrahydrofuran	0.417	0.435		4.32	30
Trichlorofluoromethane	1.144	1.287		12.5	30
1,1,2-Trichlorotrifluoroethane	0.898	0.998		11.14	30
Dichlorotetrafluoroethane	1.024	1.030		0.59	30
tert-Butyl alcohol	1.255	1.291		2.87	30
Heptane	1.921	1.806		-5.99	30
1,1-Dichloroethene	0.416	0.440		5.77	30
Acetone	1.205	1.082		-10.21	30
Carbon Disulfide	1.150	1.256		9.22	30
Methyl tert-Butyl Ether	0.608	0.627		3.13	30
Methylene Chloride	0.398	0.371		-6.78	30
trans-1,2-Dichloroethene	0.465	0.491		5.59	30
1,1-Dichloroethane	0.957	1.001		4.6	30
Cyclohexane	1.242	1.136		-8.53	30
2-Butanone	0.713	0.792		11.08	30
Carbon Tetrachloride	0.660	0.734		11.21	30
cis-1,2-Dichloroethene	1.370	1.358		-0.88	30
Chloroform	1.948	1.998		2.57	30
1,1,1-Trichloroethane	1.990	1.954		-1.81	30
2,2,4-Trimethylpentane	1.684	1.761		4.57	30
Benzene	0.985	1.044		5.99	30
1,2-Dichloroethane	0.489	0.570		16.56	30
Trichloroethene	0.419	0.421		0.48	30
1,2-Dichloropropane	0.364	0.378		3.85	30
Bromodichloromethane	0.695	0.820		17.99	30
4-Methyl-2-Pentanone	0.989	1.057		6.88	30
Toluene	1.138	1.168		2.64	30
t-1,3-Dichloropropene	0.437	0.463		5.95	30
cis-1,3-Dichloropropene	0.548	0.584		6.57	30
1,1,2-Trichloroethane	0.377	0.405		7.43	30
Dibromochloromethane	0.603	0.655		8.62	30
1,2-Dibromoethane	0.558	0.601		7.71	30
Tetrachloroethene	0.368	0.357		-2.99	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RTPE01

Lab Code: CHEM Case No.: Q2317 SAS No.: Q2317 SDG No.: Q2317

Instrument ID: MSVOA_L Calibration Date/Time: 06/16/2025 08:54

Lab File ID: VL042622.D Init. Calib. Date(s): 05/29/2025 05/29/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:24

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Chlorobenzene	0.971	0.982		1.13	30
Ethyl Benzene	1.696	1.773		4.54	30
m/p-Xylene	1.335	1.428		6.97	30
o-Xylene	1.331	1.405		5.56	30
Styrene	0.629	0.638		1.43	30
Bromoform	0.521	0.561		7.68	30
1,1,2,2-Tetrachloroethane	0.847	0.895		5.67	30
2-Chlorotoluene	1.500	1.564		4.27	30
1,3,5-Trimethylbenzene	1.374	1.431		4.15	30
1,2,4-Trimethylbenzene	1.524	1.563		2.56	30
1,3-Dichlorobenzene	0.957	0.975		1.88	30
1,4-Dichlorobenzene	0.955	0.977		2.3	30
1,2-Dichlorobenzene	0.929	0.912		-1.83	30
1,2,4-Trichlorobenzene	0.802	0.797		-0.62	30
Hexachloro-1,3-Butadiene	0.769	0.690		-10.27	30
1,3-Butadiene	0.571	0.531		-7.01	30
Naphthalene	1.437	1.542		7.31	30
4-Ethyltoluene	1.660	1.726		3.98	30
1-Bromo-4-Fluorobenzene	0.750	0.778		3.73	30
Hexane	1.498	1.399		-6.61	30
Allyl Chloride	0.850	0.847		-0.35	30
1,4-Dioxane	162.953	166.679		2.29	30
Methyl Methacrylate	0.382	0.401		4.97	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



SHIPPING DOCUMENTS

Client Contact Information				Bottle Order ID : B2505058				Courier :				1 of 5 COCs																			
Client ID : RTPE01				Project ID : Yellow Cab - TO-15				Sampler Name(s) :				Analysis																			
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :				Project Manager : David gabel				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified				<table border="1" style="width:100%; border-collapse: collapse;"> <tr><td colspan="4">Analysis</td></tr> <tr><td></td><td></td><td></td><td></td></tr> <tr><td></td><td></td><td></td><td></td></tr> <tr><td></td><td></td><td></td><td></td></tr> </table>				Analysis															
				Analysis																											
Phone Number : 732-968-9600																															
Fax Number :																															
Site Details: Bayonne, NJ																															
Analysis Turnaround Time				Standard : 10 business days OR				Data Package Type : NJ Full				<table border="1" style="width:100%; border-collapse: collapse;"> <tr><td colspan="4">Matrix</td></tr> <tr><td></td><td></td><td></td><td></td></tr> <tr><td></td><td></td><td></td><td></td></tr> <tr><td></td><td></td><td></td><td></td></tr> </table>				Matrix															
Matrix																															
Rush (Specify): Days				EDD Type : HAZ site w/ conversion table																											

Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID	Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
AA-3	6-11-25 to 6-12-25	10:56	10:56	-30	-5	76	83	-30	-5.9	10713	10588	6 L	4.16	VL042485.D	X	X

Temperature (Fahrenheit)					GC/MS Analyst Signature (TO-15)
	Ambient	Maximum	Minimum		
Start					
Stop					

Pressure (Inches of Hg)					** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.
	Ambient	Maximum	Minimum		
Start					
Stop					

Special Instructions/QC Requirements & Comments :

 Suspected Contamination: High Medium Low PID Readings:

 Sampling site (State):

Quick Connector required : **NO**

Canisters Shipped by: Sam	Date/Time: 6/10/25	Canisters Received by: Chris Wasek	Date/Time: 6/10/25
Samples Relinquished by: clw	Date/Time:	Received by:	Date/Time:
Relinquished by:	Date/Time:	Received by: clw	Date/Time: 6/12/25 1355

B2505058 - 8

Client Contact Information				Bottle Order ID : B2505058				Courier :				2 of 5 COCs					
Client ID : RTPE01				Project ID : Yellow Cab - TO-15				Sampler Name(s) :				Analysis		Matrix			
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :				Project Manager : David gabel				AIR ANALYSIS CHAIN-OF-CUSTODY Individual Certified									
				Phone Number : 732-968-9600													
				Fax Number :													
				Site Details:													
				Analysis Turnaround Time				Data Package Type : SAME									
				Standard : 10 business days OR				EDD Type : AS PAGE 1									
Rush (Specify): Days																	
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor Ambient Air	Soil Gas
115-IA	6-11-25 to 6-12-25	11:08	10:57	-30	-2	72	72	-30	-4.7	10236	10492	6 L	4.16	VL042485.D	X	X	
115-IA 1	Temperature (Fahrenheit)									GC/MS Analyst Signature (TO-15) [Signature]							
	Ambient		Maximum		Minimum												
	Start																
	Stop																
	Pressure (Inches of Hg)									** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
	Ambient		Maximum		Minimum												
	Start																
	Stop																
Special Instructions/QC Requirements & Comments :																	
Suspected Contamination: High Medium Low PID Readings:																	
Sampling site (State):																	
Quick Connector required : NO																	
Canisters Shipped by: Sam				Date/Time: 6/10/25				Canisters Received by: CV				Date/Time: 6/10/25				B2505058 - 7	
Samples Relinquished by: CV				Date/Time:				Received by:				Date/Time:					
Relinquished by:				Date/Time:				Received by: CP				Date/Time: 6/12/25 1355					

Client Contact Information				Bottle Order ID : B2505058				Courier :				3 of 5 COCs					
Client ID : RTPE01				Project ID : Yellow Cab - TO-15				Sampler Name(s) :				Analysis		Matrix			
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :				Project Manager : David gabel				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified									
				Phone Number : 732-968-9600													
				Fax Number :													
				Site Details:													
				Analysis Turnaround Time													
Standard : 10 business days OR				Data Package Type :													
Rush (Specify): Days				EDD Type :													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
115-1111 2	6-11-25 6-12-25	11:13	11:03	-23	-1	72	72	-30	-1.8	10207	10198	6 L	4.16	VL042330.D	X	X	
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) 							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Special Instructions/QC Requirements & Comments :																	
Suspected Contamination: High Medium Low PID Readings:																	
Sampling site (State):																	
Quick Connector required : NO																	
Canisters Shipped by: Sam				Date/Time: 6/19/25				Canisters Received by: CW				Date/Time: 6/10/25				B2505058 - 3	
Samples Relinquished by: CW				Date/Time:				Received by:				Date/Time:					
Relinquished by:				Date/Time:				Received by: CP				Date/Time: 6/12/25 1355					

Client Contact Information				Bottle Order ID : B2505058				Courier :				4 of 5 COCs			
Client ID : RTPE01				Project ID : Yellow Cab - TO-15				Sampler Name(s) :				Analysis			
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :				Project Manager : David gabel				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified				Indoor/Ambient Air Soil Gas			
				Phone Number : 732-968-9600											
				Fax Number :											
				Site Details:											
Analysis Turnaround Time				Data Package Type :				EDD Type :							
Standard : 10 business days OR				Rush (Specify): Days											

Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15				Indoor/Ambient Air	Soil Gas
115-561	6/12/25	11:40	11:50	-30	-7	72	72	-30	-8.0	10202	10601	6 L	100	VL042485.D	X					X

Temperature (Fahrenheit)					GC/MS Analyst Signature (TO-15) SUT
	Ambient	Maximum	Minimum		
Start					
Stop					

Pressure (Inches of Hg)					** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.
	Ambient	Maximum	Minimum		
Start					
Stop					

Special Instructions/QC Requirements & Comments :

Suspected Contamination: High Medium Low PID Readings:

Sampling site (State):

Quick Connector required : **NO**

Canisters Shipped by: Sam	Date/Time: 6/6/09/K	Canisters Received by: CW	Date/Time: 6/10/25
Samples Relinquished by: CW	Date/Time:	Received by:	Date/Time:
Relinquished by:	Date/Time:	Received by: CP	Date/Time: 6/12/25 1355

B2505058 - 9

Client Contact Information						Bottle Order ID : B2505058				Courier :				5 of 5 COCs								
Client ID : RTPE01				Project ID : Yellow Cab - TO-15				Sampler Name(s) :				Analysis		Matrix								
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :						Project Manager : David gabel		AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified														
						Phone Number : 732-968-9600																
						Fax Number :																
						Site Details:																
						Analysis Turnaround Time																
						Standard : 10 business days OR				Data Package Type :												
						Rush (Specify): Days				EDD Type :												
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15		Indoor/Ambient Air	Soil Gas				
6-12-25	11:43	11:53	27-2	72	72	-30	6.7	10241	10315	6 L	100	VL042485.D	X				X					
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) [Signature]												
		Ambient	Maximum	Minimum																		
Start																						
Stop																						
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.												
		Ambient	Maximum	Minimum																		
Start																						
Stop																						
Special Instructions/QC Requirements & Comments :																						
Suspected Contamination: High Medium Low PID Readings:																						
Sampling site (State):																						
Quick Connector required : NO																						
Canisters Shipped by: Sam				Date/Time: 06/09/25				Canisters Received by: CW				Date/Time: 6/10/25				B2505058 - 5						
Samples Relinquished by: CW				Date/Time:				Received by:				Date/Time:										
Relinquished by:				Date/Time:				Received by: CP				Date/Time: 6/12/25/355										

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

Internal Chain of Custody

Instructions: Use 1 form for each 20 samples of aliquot

Laboratory Person Breaking Field Seal on Sample Shuttle & Accepting Responsibility for Sample

Laboratory: Chemtech

Location: 284 Sheffield Street, Mountainside, NJ 7092

~~Q2317~~

Title: Sample Custodian

Field Sample Seal No. Q2317



Date Broken 5/12/2025

Military Time Seal Broken: 13:55:00

Case No.: Yellow Cab - TO-15

Analytical Parameter/Fraction TO-15

Sample No.	Aliquot/Extract No.	Sample No.	Aliquot/Extract No.
Q2317-01	AA-3		
Q2317-02	115-IA-1		
Q2317-03	115-IA-2		
Q2317-04	115-SG1		
Q2317-05	115-SG2		

Date	Time	Relinquished By		Received By		Purpose of Change of Custody
6/13	1100	Signature		Signature		
		Printed Name	George Nelson	Printed Name	Sample Custodian	
		Signature		Signature		
		Printed Name		Printed Name		
		Signature		Signature		
		Printed Name		Printed Name		
		Signature		Signature		
		Printed Name		Printed Name		
		Signature		Signature		
		Printed Name		Printed Name		
		Signature		Signature		
		Printed Name		Printed Name		
		Signature		Signature		
		Printed Name		Printed Name		
		Signature		Signature		
		Printed Name		Printed Name		

Distribution: White - Original (Sent With Report)

Yellow - Contractor Archive

Pink - Sample Custodian - Interim Copy

CHEMTECH

284 Sheffield Street Mountainside, NJ 07092

AIR SAMPLE PRESSURE & DILUTION LOGBOOK

Analyst Signature: *[Signature]*

Supervisor Signature: *[Signature]*

METHOD: TO-15

Pressure Gauge ID: A755971

Date	Sample Number	Canister #	Initial Pressure psia	Initial Pressure Hg	Final Pressure psia	Final Pressure Hg	Dilution Factor	Comment
04/24/15	Q1927-01	10197	10.9	-7.8				sy
	Q1927-02	10267	12.0	-5.5				sy
	Q1927-03	10283	11.4	-6.7				↓
	Q1927-04	10320	11.8	-5.9				↓
	Q1927-05	10445	11.0	-7.5				↓
✓	Q1927-06	10607	13.6	-2.2				↓
04/28/15	Q1908-01	10598	11.6	-6.3				sy
↓	Q1908-02	10595	12.0	-5.5				↓
↓	Q1908-03	10297	12.2	-5.1				↓
05/14/15	Q2011-01	10154	11.2	-7.1				sy
↓	Q2011-02	10599	12.0	-5.5				↓
5/12/15	Q2037-01	10609	12.0	-5.5				sy
u	Q2037-02	10285	12.0	-5.5				sy
05/23/15	Q2124-01	10334	12.0	-5.5				sy
↓	Q2124-02	10439	11.4	-6.7				↓
↓	Q2124-03	10329	12.0	-5.5				↓
06/12/15	Q2317-01	10588	11.8	-5.9				sy
	Q2317-02	10492	12.4	-4.7				↓
	Q2317-03	10198	13.8	-1.8				↓
	Q2317-04	10601	10.8	-8.0				↓
↓	Q2317-05	10315	11.4	-6.7				↓

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

CHEMTECH

Client Sample ID #: AA-3
 Client Name: RTP Environmental
 Project Name: Yellow Cab
 Date: 6/11/25 to 6/12/25 Time: 10:56 to 10:56
 Analysis: TO-15
 Comments: Ambient Air

CHEMTECH'S SAMPLE ID: _____
 Date Received: _____ Expiration: _____

Storage Location: D41
 Sample: Q2317-01
 Cust #: AA-3

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

CHEMTECH

Client Sample ID #: 115-IA1
 Client Name: RTP
 Project Name: Yellow Cab
 Date: 6-11-25 to 6-12-25 Time: 10:08 to 10:58
 Analysis: 10-15
 Comments: Indoor Air

CHEMTECH'S SAMPLE ID: _____
 Date Received: _____ Expiration: _____

Storage Location: D41
 Sample: Q2317-02
 Cust #: 115-IA-1

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

CHEMTECH

Client Sample ID #: 115-IA2
 Client Name: RTP
 Project Name: Yellow Cab
 Date: 6-11-25 to 6-12-25 Time: 11:13 to 11:03
 Analysis: TO-15
 Comments: Indoor air

CHEMTECH'S SAMPLE ID: _____
 Date Received: _____ Expiration: _____

Storage Location: D41
 Sample: Q2317-03
 Cust #: 115-IA-2

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

Client Sample ID # 10950002-115-SG2
Client Name: RIP Yellow Cab
Project Name: Yellow Cab
Date: 6/12/25 Time: 11:43 to 12:43
Analysis: 90-15
Comments: bedroom?
CHEMTECH'S SAMPLE ID: _____
Date Received: _____ Expiration: _____

Storage Location: D41
Sample: Q2317-05
Cust #: 115-SG2

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

Client Sample ID # 115-SG1
Client Name: RIP
Project Name: Yellow Cab
Date: 6/12/25 Time: 11:40 to 12:40
Analysis: 90-15
Comments: bedroom
CHEMTECH'S SAMPLE ID: _____
Date Received: _____ Expiration: _____

Storage Location: D41
Sample: Q2317-04
Cust #: 115-SG1