

Duplicate Sample Summary

Lab Sample Id :	Q2317-01DUP	Q2317-01
Client Id :	AA-3DUP	AA-3
DF :	1	1
Datafile :	VL042627.D	VL042626.D
Anal Date & Time :	06/16/2025 14:01	06/16/2025 13:27

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0.17	0.17	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0.18	0.19	5.4
2-Butanone	0.61	0.65	6.3
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0.18	0.18	0
Acetone	10.6	10.2	3.8
Allyl Chloride	0	0	0
Benzene	0.31	0.32	3.2
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.08	0.08	0

Duplicate Sample Summary

Lab Sample Id :	Q2317-01DUP	Q2317-01
Client Id :	AA-3DUP	AA-3
DF :	1	1
Datafile :	VL042627.D	VL042626.D
Anal Date & Time :	06/16/2025 14:01	06/16/2025 13:27

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.54	0.54	0
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.49	0.48	2.1
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0.11	0	200 *
Heptane	0.15	0.15	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	4.9	4.8	2.1
m/p-Xylene	0.51	0.48	6.1
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	3.3	3	9.5
Naphthalene	0	0	0
o-Xylene	0.24	0.23	4.3
Styrene	0.14	0.13	7.4
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0	0	0
Tetrahydrofuran	0	0	0
Toluene	0.48	0.46	4.3
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.28	0.27	3.6
Vinyl Chloride	0	0	0