

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2318 SAS No.: Q2318 SDG NO.: Q2318

EPA Sample No.: SSTD020226 Date Analyzed: 06/19/2025

Lab File ID: BN037342.D Time Analyzed: 12:19

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	406161	16.969	433116	21.204	472524	24.051
UPPER LIMIT	812322	17.469	866232	21.704	945048	24.551
LOWER LIMIT	203081	16.469	216558	20.704	236262	23.551
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	318050	16.97	304011	21.20	324491	24.05

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.