

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2318

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Datafile : VX046791.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0620WBS02	Dichlorodifluoromethane	20	18.7	ug/L	94			69	116	
	Chloromethane	20	18.1	ug/L	91			65	116	
	Vinyl chloride	20	18.7	ug/L	94			65	117	
	Ethyl Acetate	20	15.8	ug/L	79			75	115	
	Bromomethane	20	21.0	ug/L	105			58	125	
	Chloroethane	20	19.3	ug/L	97			56	128	
	Trichlorofluoromethane	20	18.6	ug/L	93			73	115	
	1,1,2-Trichlorotrifluoroethane	20	18.8	ug/L	94			80	112	
	Tert butyl alcohol	100	59.9	ug/L	60			48	142	
	1,1-Dichloroethene	20	18.7	ug/L	94			74	110	
	Acrolein	100	77.6	ug/L	78			28	144	
	Acrylonitrile	100	83.5	ug/L	84			73	113	
	Acetone	100	72.4	ug/L	72			60	125	
	Carbon disulfide	20	16.7	ug/L	84			64	112	
	Methyl tert-butyl Ether	20	16.8	ug/L	84			78	114	
	Methyl Acetate	20	17.0	ug/L	85			67	125	
	Methylene Chloride	20	18.2	ug/L	91			72	114	
	trans-1,2-Dichloroethene	20	18.8	ug/L	94			75	108	
	Vinyl Acetate	100	87.6	ug/L	88			76	115	
	1,1-Dichloroethane	20	18.6	ug/L	93			78	112	
	Cyclohexane	20	18.6	ug/L	93			75	110	
	2-Butanone	100	79.5	ug/L	80			65	122	
	Carbon Tetrachloride	20	17.9	ug/L	90			77	113	
	2,2-Dichloropropane	20	17.9	ug/L	90			56	164	
	cis-1,2-Dichloroethene	20	18.5	ug/L	93			77	110	
	Bromochloromethane	20	19.8	ug/L	99			70	124	
	Chloroform	20	19.1	ug/L	96			79	113	
	1,1,1-Trichloroethane	20	18.0	ug/L	90			80	108	
	Methylcyclohexane	20	18.0	ug/L	90			72	115	
	1,1-Dichloropropene	20	18.4	ug/L	92			79	110	
	Benzene	20	18.7	ug/L	94			82	109	
	1,2-Dichloroethane	20	18.2	ug/L	91			80	115	
	Trichloroethene	20	18.7	ug/L	94			77	113	
	1,2-Dichloropropane	20	18.9	ug/L	95			83	111	
	Dibromomethane	20	18.5	ug/L	93			82	110	
	Bromodichloromethane	20	18.5	ug/L	93			83	110	
	4-Methyl-2-Pentanone	100	82.4	ug/L	82			74	118	
	Toluene	20	19.1	ug/L	96			82	110	
	t-1,3-Dichloropropene	20	17.6	ug/L	88			79	110	
	cis-1,3-Dichloropropene	20	18.0	ug/L	90			82	110	
	1,1,2-Trichloroethane	20	19.1	ug/L	96			83	112	
	1,3-Dichloropropane	20	18.8	ug/L	94			83	111	
	2-Chloroethyl vinyl ether	100	93.4	ug/L	93			75	124	
	2-Hexanone	100	79.3	ug/L	79			73	117	
	Dibromochloromethane	20	18.3	ug/L	92			82	110	
	1,2-Dibromoethane	20	18.3	ug/L	92			81	110	
	Tetrachloroethene	20	18.3	ug/L	92			67	123	
	Chlorobenzene	20	18.5	ug/L	93			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2318

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Datafile : VX046791.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0620WBS02	1,1,1,2-Tetrachloroethane	20	17.9	ug/L	90			84	111	
	Hexachloroethane	20	16.4	ug/L	82			80	113	
	Ethyl Benzene	20	18.2	ug/L	91			83	109	
	m/p-Xylenes	40	37.1	ug/L	93			82	110	
	o-Xylene	20	19.2	ug/L	96			83	109	
	Styrene	20	18.0	ug/L	90			80	111	
	Bromoform	20	17.1	ug/L	86			79	109	
	Isopropylbenzene	20	17.8	ug/L	89			83	112	
	1,1,2,2-Tetrachloroethane	20	17.2	ug/L	86			76	118	
	1,2,3-Trichloropropane	20	15.2	ug/L	76			75	112	
	Bromobenzene	20	18.2	ug/L	91			77	116	
	N-propylbenzene	20	18.0	ug/L	90			83	112	
	2-Chlorotoluene	20	17.9	ug/L	90			83	110	
	1,3,5-Trimethylbenzene	20	18.1	ug/L	91			85	112	
	4-Chlorotoluene	20	17.8	ug/L	89			82	110	
	tert-Butylbenzene	20	17.6	ug/L	88			83	112	
	1,2,4-Trimethylbenzene	20	18.2	ug/L	91			85	111	
	Sec-butylbenzene	20	18.4	ug/L	92			81	114	
	p-Isopropyltoluene	20	17.9	ug/L	90			78	116	
	1,3-Dichlorobenzene	20	17.7	ug/L	89			82	108	
	1,4-Dichlorobenzene	20	17.3	ug/L	86			82	107	
	n-Butylbenzene	20	17.7	ug/L	89			75	115	
	1,2-Dichlorobenzene	20	17.9	ug/L	90			82	109	
	1,2-Dibromo-3-Chloropropane	20	14.7	ug/L	74			68	112	
	1,2,4-Trichlorobenzene	20	17.1	ug/L	86			75	113	
	Hexachlorobutadiene	20	16.7	ug/L	84			81	121	
	Naphthalene	20	16.0	ug/L	80			78	119	
	1,2,3-Trichlorobenzene	20	17.0	ug/L	85			76	114	