

**Analysis Method:** 9045D  
**Parameter:** Corrosivity  
**Run Number:** LB136178  
**BalanceID:** WC SC-7

**Analyst By :** jignesh  
**Supervisor Review By :** Iwona  
**Slope :** 99.2  
**pH Meter ID :** WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

True Value of CCV3 = 2.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/17/2025	15:15
2	CAL2	1	Water	NA	NA	20.2	7.00	06/17/2025	15:16
3	CAL3	1	Water	NA	NA	20.2	10.02	06/17/2025	15:16
4	ICV	1	Water	NA	NA	20.3	7.00	06/17/2025	15:19
5	CCV1	1	Water	NA	NA	20.3	2.02	06/17/2025	15:20
6	Q2311-04	1	Solid	20.02	20	20.2	9.17	06/17/2025	15:30
7	Q2311-08	1	Solid	20.03	20	20.1	10.23	06/17/2025	15:35
8	Q2312-04	1	Solid	20.02	20	20.6	8.13	06/17/2025	15:40
9	Q2319-04	1	Solid	20.03	20	20.3	5.76	06/17/2025	15:44
10	Q2320-01	1	Solid	20.04	20	20.9	7.80	06/17/2025	15:50
11	Q2325-04	1	Solid	20.02	20	20.2	6.61	06/17/2025	16:00
12	Q2339-04	1	Solid	20.03	20	20.7	6.20	06/17/2025	16:10
13	Q2340-04	1	Solid	20.04	20	20.4	8.92	06/17/2025	16:11
14	Q2341-04	1	Solid	20.03	20	20.6	8.63	06/17/2025	16:20
15	Q2341-08	1	Solid	20.04	20	20.5	7.04	06/17/2025	16:30
16	CCV2	1	Water	NA	NA	20.2	12.02	06/17/2025	16:35
17	Q2347-04	1	Solid	20.02	20	20.2	7.40	06/17/2025	16:45
18	Q2347-04DUP	1	Solid	20.03	20	20.3	7.41	06/17/2025	16:47
19	CCV3	1	Water	NA	NA	20.3	2.02	06/17/2025	16:50

# WORKLIST(Hardcopy Internal Chain)

136178

WorkList Name : corrosivity Q2347      WorkList ID : 190244      Department : Wet-Chemistry      Date : 06-17-2025 14:40:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2311-04	TP03-MH2MH3-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	06/11/2025	9045D
Q2311-08	TP04-MH2MH3-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	06/11/2025	9045D
Q2312-04	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	D51	06/13/2025	9045D
Q2319-04	MH-B	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	06/13/2025	9045D
Q2320-01	WC	Solid	Corrosivity	Cool 4 deg C	FIRS02	D41	06/13/2025	9045D
Q2325-04	TP-8	Solid	Corrosivity	Cool 4 deg C	PSEG03	D51	06/13/2025	9045D
Q2339-04	MH-C	Solid	Corrosivity	Cool 4 deg C	PSEG03	D51	06/13/2025	9045D
Q2340-04	TP05-MH17A-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	06/16/2025	9045D
Q2341-04	TP-9	Solid	Corrosivity	Cool 4 deg C	PSEG03	D52	06/16/2025	9045D
Q2341-08	EP-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	D52	06/16/2025	9045D
Q2347-04	MH-9	Solid	Corrosivity	Cool 4 deg C	PSEG03	D42	06/17/2025	9045D

Date/Time 06/17/25 15:00  
 Raw Sample Received by: SA WNC  
 Raw Sample Relinquished by: CP SM

Date/Time 06/17/25      78.00  
 Raw Sample Received by: CP SM  
 Raw Sample Relinquished by: CP SM